

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112221\
 Data File : BP008076.D
 Acq On : 22 Nov 2021 11:10
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 23 01:20:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 20 00:21:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.548	0.514	6.2	111	0.00
3	Pyridine	1.568	1.587	-1.2	111	0.00
4	n-Nitrosodimethylamine	0.702	0.674	4.0	109	0.00
5 S	2-Fluorophenol	1.170	1.216	-3.9	114	0.00
6	Aniline	2.048	2.052	-0.2	116	0.00
7 S	Phenol-d6	1.622	1.707	-5.2	116	0.00
8	2-Chlorophenol	1.314	1.254	4.6	115	0.00
9	Benzaldehyde	1.171	1.024	12.6	113	0.00
10 C	Phenol	1.737	1.757	-1.2	115	0.00
11	bis(2-Chloroethyl)ether	1.480	1.402	5.3	114	0.00
12	1,3-Dichlorobenzene	1.508	1.410	6.5	112	0.00
13 C	1,4-Dichlorobenzene	1.530	1.427	6.7	114	0.00
14	1,2-Dichlorobenzene	1.469	1.377	6.3	114	0.00
15	Benzyl Alcohol	1.400	1.385	1.1	113	0.00
16	2,2'-oxybis(1-Chloropropane	2.306	2.042	11.4	108	0.00
17	2-Methylphenol	1.145	1.138	0.6	114	0.00
18	Hexachloroethane	0.545	0.521	4.4	111	0.00
19 P	n-Nitroso-di-n-propylamine	1.202	1.110	7.7	112	0.00
20	3+4-Methylphenols	1.587	1.574	0.8	114	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
22	Acetophenone	0.555	0.531	4.3	114	0.00
23 S	Nitrobenzene-d5	0.441	0.436	1.1	113	0.00
24	Nitrobenzene	0.432	0.425	1.6	112	0.00
25	Isophorone	0.776	0.756	2.6	112	0.00
26 C	2-Nitrophenol	0.184	0.187	-1.6	114	0.00
27	2,4-Dimethylphenol	0.277	0.272	1.8	113	0.00
28	bis(2-Chloroethoxy)methane	0.497	0.475	4.4	111	0.00
29 C	2,4-Dichlorophenol	0.334	0.326	2.4	111	0.00
30	1,2,4-Trichlorobenzene	0.380	0.366	3.7	111	0.00
31	Naphthalene	1.016	0.989	2.7	113	-0.01
32	Benzoic acid	0.238	0.242	-1.7	107	0.00
33	4-Chloroaniline	0.431	0.427	0.9	114	-0.01
34 C	Hexachlorobutadiene	0.257	0.239	7.0	109	0.00
35	Caprolactam	0.075	0.073	2.7	108	0.00
36 C	4-Chloro-3-methylphenol	0.386	0.371	3.9	109	0.00
37	2-Methylnaphthalene	0.750	0.684	8.8	111	0.00
38	1-Methylnaphthalene	0.735	0.672	8.6	111	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	0.642	0.631	1.7	109	0.00
41 P	Hexachlorocyclopentadiene	0.292	0.287	1.7	99	0.00
42 S	2,4,6-Tribromophenol	0.253	0.240	5.1	105	0.00
43 C	2,4,6-Trichlorophenol	0.442	0.439	0.7	108	0.00
44	2,4,5-Trichlorophenol	0.477	0.473	0.8	107	0.00
45 S	2-Fluorobiphenyl	1.277	1.290	-1.0	109	0.00
46	1,1'-Biphenyl	1.485	1.404	5.5	109	0.00
47	2-Chloronaphthalene	1.132	1.113	1.7	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112221\
 Data File : BP008076.D
 Acq On : 22 Nov 2021 11:10
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 23 01:20:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 20 00:21:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.406	0.413	-1.7	108	0.00
49	Acenaphthylene	1.758	1.732	1.5	108	0.00
50	Dimethylphthalate	1.514	1.420	6.2	104	0.00
51	2,6-Dinitrotoluene	0.317	0.306	3.5	105	0.00
52 C	Acenaphthene	1.078	1.013	6.0	108	0.00
53	3-Nitroaniline	0.332	0.326	1.8	105	0.00
54 P	2,4-Dinitrophenol	0.198	0.193	2.5	97	0.00
55	Dibenzofuran	1.855	1.710	7.8	107	0.00
56 P	4-Nitrophenol	0.266	0.262	1.5	100	0.00
57	2,4-Dinitrotoluene	0.439	0.423	3.6	102	0.00
58	Fluorene	1.444	1.262	12.6	107	0.00
59	2,3,4,6-Tetrachlorophenol	0.429	0.411	4.2	104	0.00
60	Diethylphthalate	1.554	1.383	11.0	102	0.00
61	4-Chlorophenyl-phenylether	0.800	0.679	15.1	104	0.00
62	4-Nitroaniline	0.341	0.330	3.2	102	0.00
63	Azobenzene	1.620	1.462	9.8	103	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	0.135	0.135	0.0	97	0.00
66 c	n-Nitrosodiphenylamine	0.567	0.555	2.1	103	0.00
67	4-Bromophenyl-phenylether	0.219	0.212	3.2	102	0.00
68	Hexachlorobenzene	0.234	0.226	3.4	101	0.00
69	Atrazine	0.145	0.142	2.1	99	0.00
70 C	Pentachlorophenol	0.151	0.148	2.0	99	0.00
71	Phenanthrene	1.078	1.008	6.5	102	0.00
72	Anthracene	1.067	0.999	6.4	101	0.00
73	Carbazole	1.083	0.963	11.1	100	0.00
74	Di-n-butylphthalate	1.181	1.089	7.8	97	0.00
75 C	Fluoranthene	1.389	1.195	14.0	97	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
77	Benzydine	0.348	0.310	10.9	65	0.00
78	Pyrene	1.358	1.421	-4.6	96	0.00
79 S	Terphenyl-d14	1.017	1.014	0.3	94	0.00
80	Butylbenzylphthalate	0.567	0.571	-0.7	89	0.00
81	Benzo(a)anthracene	1.341	1.291	3.7	87	0.00
82	3,3'-Dichlorobenzidine	0.636	0.542	14.8	82	0.00
83	Chrysene	1.273	1.227	3.6	86	0.00
84	Bis(2-ethylhexyl)phthalate	0.768	0.702	8.6	84	0.00
85 c	Di-n-octyl phthalate	1.407	1.253	10.9	77	0.00
86 I	Perylene-d12	1.000	1.000	0.0	73	0.00
87	Indeno(1,2,3-cd)pyrene	1.388	1.165	16.1	69	0.00
88	Benzo(b)fluoranthene	1.250	1.192	4.6	74	0.00
89	Benzo(k)fluoranthene	1.244	1.232	1.0	78	0.00
90 C	Benzo(a)pyrene	1.048	1.006	4.0	73	0.00
91	Dibenzo(a,h)anthracene	1.155	0.956	17.2	68	0.00
92	Benzo(g,h,i)perylene	1.132	0.951	16.0	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112221\
Data File : BP008076.D
Acq On : 22 Nov 2021 11:10
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 23 01:20:42 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 20 00:21:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0