

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112321\
 Data File : BP008097.D
 Acq On : 23 Nov 2021 15:39
 Operator : CG/JU
 Sample : M4776-01 10X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 RO-72-11948

Quant Time: Nov 23 16:19:55 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 20 00:21:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	141493	20.000	ng	0.00
21) Naphthalene-d8	10.616	136	488331	20.000	ng	-0.01
39) Acenaphthene-d10	14.463	164	296533	20.000	ng	0.00
64) Phenanthrene-d10	17.222	188	645043	20.000	ng	0.00
76) Chrysene-d12	21.316	240	736485	20.000	ng	0.00
86) Perylene-d12	23.715	264	768558	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.411	112	16118	1.947	ng	0.00
7) Phenol-d6	7.005	99	32888	2.866	ng	0.00
23) Nitrobenzene-d5	8.981	82	27594	2.563	ng	-0.01
42) 2,4,6-Tribromophenol	15.963	330	3085	0.822	ng	0.00
45) 2-Fluorobiphenyl	13.081	172	56564	2.987	ng	-0.01
79) Terphenyl-d14	19.786	244	139748	3.731	ng	-0.01
Target Compounds						
60) Diethylphthalate	15.292	149	678146	29.439	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112321\
Data File : BP008097.D
Acq On : 23 Nov 2021 15:39
Operator : CG/JU
Sample : M4776-01 10X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
RO-72-11948

Quant Time: Nov 23 16:19:55 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 20 00:21:11 2021
Response via : Initial Calibration

