

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112521\
 Data File : BP008119.D
 Acq On : 24 Nov 2021 16:40
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: Nov 25 00:25:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 00:18:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	96677	20.000	ng	0.00
21) Naphthalene-d8	10.616	136	365264	20.000	ng	0.00
39) Acenaphthene-d10	14.463	164	258894	20.000	ng	0.00
64) Phenanthrene-d10	17.222	188	567406	20.000	ng	0.00
76) Chrysene-d12	21.316	240	609938	20.000	ng	0.00
86) Perylene-d12	23.716	264	636193	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.405	112	73608	13.014	ng	0.00
7) Phenol-d6	6.999	99	97575	12.445	ng	0.00
23) Nitrobenzene-d5	8.975	82	94031	11.678	ng	0.00
42) 2,4,6-Tribromophenol	15.957	330	38627	11.790	ng	0.00
45) 2-Fluorobiphenyl	13.081	172	216736	13.109	ng	0.00
79) Terphenyl-d14	19.786	244	384172	12.383	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.317	88	17387	6.568	ng	98
3) Pyridine	3.728	79	42608	5.622	ng	95
4) n-Nitrosodimethylamine	3.628	42	18531	5.458	ng	# 98
6) Aniline	7.146	93	55533	5.611	ng	97
8) 2-Chlorophenol	7.387	128	36791	5.794	ng	100
9) Benzaldehyde	6.964	77	32934	5.819	ng	94
10) Phenol	7.022	94	49947	5.947	ng	95
11) bis(2-Chloroethyl)ether	7.246	93	39249	5.488	ng	96
12) 1,3-Dichlorobenzene	7.711	146	43836	6.014	ng	98
13) 1,4-Dichlorobenzene	7.852	146	44215	5.977	ng	94
14) 1,2-Dichlorobenzene	8.169	146	42745	6.021	ng	96
15) Benzyl Alcohol	8.063	79	35883	5.301	ng	96
16) 2,2'-oxybis(1-Chloropr...	8.346	45	60914	5.464	ng	98
17) 2-Methylphenol	8.269	107	32363	5.849	ng	96
18) Hexachloroethane	8.899	117	16150	6.130	ng	96
19) n-Nitroso-di-n-propyla...	8.628	70	32835	5.653	ng	# 98
20) 3+4-Methylphenols	8.605	107	44407	5.789	ng	97
22) Acetophenone	8.646	105	60078	5.928	ng	# 99
24) Nitrobenzene	9.022	77	45952	5.822	ng	97
25) Isophorone	9.552	82	84208	5.941	ng	99
26) 2-Nitrophenol	9.734	139	17532	5.219	ng	96
27) 2,4-Dimethylphenol	9.805	122	29065	5.741	ng	96
28) bis(2-Chloroethoxy)met...	10.034	93	53366	5.881	ng	98
29) 2,4-Dichlorophenol	10.281	162	33081	5.420	ng	98
30) 1,2,4-Trichlorobenzene	10.481	180	41154	5.930	ng	95
31) Naphthalene	10.669	128	112944	6.086	ng	99
32) Benzoic acid	9.910	122	18765	4.324	ng	94
33) 4-Chloroaniline	10.781	127	45794	5.813	ng	98
34) Hexachlorobutadiene	10.957	225	28458	6.060	ng	95
35) Caprolactam	11.569	113	7661	5.592	ng	94
36) 4-Chloro-3-methylphenol	11.922	107	40182	5.704	ng	95
37) 2-Methylnaphthalene	12.281	142	80557	5.881	ng	99
38) 1-Methylnaphthalene	12.499	142	78393	5.843	ng	99
40) 1,2,4,5-Tetrachloroben...	12.657	216	49116	5.913	ng	100
41) Hexachlorocyclopentadiene	12.634	237	10368	2.747	ng	92
43) 2,4,6-Trichlorophenol	12.899	196	31807	5.559	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112521\
 Data File : BP008119.D
 Acq On : 24 Nov 2021 16:40
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Quant Time: Nov 25 00:25:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 00:18:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.981	196	34711	5.623	ng	98
46) 1,1'-Biphenyl	13.293	154	112248	5.839	ng	98
47) 2-Chloronaphthalene	13.334	162	88425	6.035	ng	99
48) 2-Nitroaniline	13.546	65	26920	5.128	ng	93
49) Acenaphthylene	14.181	152	134711	5.918	ng	99
50) Dimethylphthalate	13.922	163	117362	5.987	ng	99
51) 2,6-Dinitrotoluene	14.040	165	22519	5.483	ng	96
52) Acenaphthene	14.528	154	81854	5.868	ng	99
53) 3-Nitroaniline	14.375	138	20811	4.849	ng	97
54) 2,4-Dinitrophenol	14.598	184	9692	3.782	ng	99
55) Dibenzofuran	14.863	168	141647	5.898	ng	98
56) 4-Nitrophenol	14.728	139	14801	4.302	ng	84
57) 2,4-Dinitrotoluene	14.834	165	30374	5.349	ng	# 94
58) Fluorene	15.516	166	112409	6.012	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.098	232	32617	5.874	ng	# 92
60) Diethylphthalate	15.287	149	117764	5.855	ng	99
61) 4-Chlorophenyl-phenyle...	15.510	204	62005	5.990	ng	96
62) 4-Nitroaniline	15.540	138	20965	4.747	ng	92
63) Azobenzene	15.804	77	116495	5.555	ng	98
65) 4,6-Dinitro-2-methylph...	15.610	198	17749	4.632	ng	97
66) n-Nitrosodiphenylamine	15.728	169	100187	6.223	ng	99
67) 4-Bromophenyl-phenylether	16.404	248	37483	6.021	ng	94
68) Hexachlorobenzene	16.528	284	40561	6.097	ng	94
69) Atrazine	16.687	200	25020	6.079	ng	95
70) Pentachlorophenol	16.881	266	19971	4.669	ng	91
71) Phenanthrene	17.263	178	182454	5.968	ng	100
72) Anthracene	17.351	178	180585	5.966	ng	99
73) Carbazole	17.628	167	176133	5.731	ng	99
74) Di-n-butylphthalate	18.186	149	217056	6.480	ng	99
75) Fluoranthene	19.239	202	240929	6.114	ng	99
77) Benzidine	19.428	184	36634	3.449	ng	# 93
78) Pyrene	19.586	202	274700	6.632	ng	99
80) Butylbenzylphthalate	20.469	149	101962	5.892	ng	95
81) Benzo(a)anthracene	21.298	228	239726	5.861	ng	100
82) 3,3'-Dichlorobenzidine	21.239	252	111817	5.768	ng	97
83) Chrysene	21.351	228	230665	5.943	ng	98
84) Bis(2-ethylhexyl)phtha...	21.233	149	156247	6.674	ng	99
85) Di-n-octyl phthalate	22.151	149	267614	6.238	ng	99
87) Indeno(1,2,3-cd)pyrene	26.204	276	275617	6.243	ng	98
88) Benzo(b)fluoranthene	23.027	252	234566	5.897	ng	99
89) Benzo(k)fluoranthene	23.027	252	234566	5.927	ng	99
90) Benzo(a)pyrene	23.604	252	195350	5.860	ng	99
91) Dibenzo(a,h)anthracene	26.227	278	231145	6.290	ng	97
92) Benzo(g,h,i)perylene	26.974	276	229070	6.363	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 25 00:25:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 25 00:18:05 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.816 min Scan# 81
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:152 Resp: 96677
Ion Ratio Lower Upper
152 100
150 157.2 125.7 188.5
115 63.8 50.7 76.1



#2
1,4-Dioxane
Concen: 6.568 ng
RT: 3.317 min Scan# 56
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 88 Resp: 17387
Ion Ratio Lower Upper
88 100
58 80.2 63.5 95.3
43 31.0 23.4 35.2









#3
Pyridine
Concen: 5.622 ng
RT: 3.728 min Scan# 109
Delta R.T. 0.012 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 79 Resp: 42608
Ion Ratio Lower Upper
79 100
52 78.0 60.0 90.0
51 41.8 29.6 44.4



#4
n-Nitrosodimethylamine
Concen: 5.458 ng
RT: 3.628 min Scan# 109
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 42 Resp: 18531
Ion Ratio Lower Upper
42 100
74 135.3 108.0 162.0
44 21.0 5.8 8.6









#5
2-Fluorophenol
Concen: 13.014 ng
RT: 5.405 min Scan# 411
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	112	Resp:	73608
Ion Ratio	Lower	Upper	
112	100		
64	68.4	53.4	80.2
63	35.9	27.8	41.8



#6
Aniline
Concen: 5.611 ng
RT: 7.146 min Scan# 707
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	93	Resp:	55533
Ion Ratio	Lower	Upper	
93	100		
66	44.3	32.9	49.3
65	22.0	17.5	26.3









#7
Phenol-d6
Concen: 12.445 ng
RT: 6.999 min Scan# 682
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	99	Resp:	97575
Ion	Ratio	Lower	Upper
99	100		
42	19.8	14.6	22.0
71	41.4	31.1	46.7



#8
2-Chlorophenol
Concen: 5.794 ng
RT: 7.387 min Scan# 748
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	128	Resp:	36791
Ion	Ratio	Lower	Upper
128	100		
130	33.3	13.5	53.5
64	57.0	37.4	77.4









#9
Benzaldehyde
Concen: 5.819 ng
RT: 6.964 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 77 Resp: 32934
Ion Ratio Lower Upper
77 100
105 98.8 71.8 111.8
106 82.7 67.3 107.3



#10
Phenol
Concen: 5.947 ng
RT: 7.022 min Scan# 686
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 94 Resp: 49947
Ion Ratio Lower Upper
94 100
65 28.9 10.9 50.9
66 40.3 24.3 64.3









#11
bis(2-Chloroethyl)ether
Concen: 5.488 ng
RT: 7.246 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 93 Resp: 39249
Ion Ratio Lower Upper
93 100
63 77.4 61.8 101.8
95 32.6 12.5 52.5



#12
1,3-Dichlorobenzene
Concen: 6.014 ng
RT: 7.711 min Scan# 803
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 146 Resp: 43836
Ion Ratio Lower Upper
146 100
148 64.9 51.9 77.9
75 32.6 29.0 43.6









#13
1,4-Dichlorobenzene
Concen: 5.977 ng
RT: 7.852 min Scan# 81
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	146	Resp:	44215
Ion Ratio	Lower	Upper	
146	100		
148	67.9	49.6	74.4
111	41.8	34.7	52.1



#14
1,2-Dichlorobenzene
Concen: 6.021 ng
RT: 8.169 min Scan# 881
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	146	Resp:	42745
Ion Ratio	Lower	Upper	
146	100		
148	65.6	51.8	77.6
111	48.6	35.0	52.6









#15
Benzyl Alcohol
Concen: 5.301 ng
RT: 8.063 min Scan# 80
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 79 Resp: 35883
Ion Ratio Lower Upper
79 100
108 72.1 55.5 83.3
77 67.4 51.0 76.6



#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.464 ng
RT: 8.346 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 45 Resp: 60914
Ion Ratio Lower Upper
45 100
77 14.5 0.0 33.8
79 9.7 0.0 30.4









#17
2-Methylphenol
Concen: 5.849 ng
RT: 8.269 min Scan# 898
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	107	Resp:	32363
Ion	Ratio	Lower	Upper
107	100		
108	107.2	90.2	135.2
77	50.1	41.3	61.9
79	49.0	41.0	61.6



#18
Hexachloroethane
Concen: 6.130 ng
RT: 8.899 min Scan# 1005
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	117	Resp:	16150
Ion	Ratio	Lower	Upper
117	100		
119	97.2	74.7	112.1
201	114.1	88.6	132.8









#19
n-Nitroso-di-n-propylamine
Concen: 5.653 ng
RT: 8.628 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	70	Resp:	32835
Ion	Ratio	Lower	Upper
70	100		
42	57.3	45.4	68.2
101	7.2	7.6	11.4#
130	19.7	15.1	22.7



#20
3+4-Methylphenols
Concen: 5.789 ng
RT: 8.605 min Scan# 955
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	107	Resp:	44407
Ion	Ratio	Lower	Upper
107	100		
108	84.6	68.2	108.2
77	38.0	16.9	56.9
79	34.9	12.5	52.5









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.616 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	136	Resp:	365264
Ion Ratio	Lower	Upper	
136	100		
137	10.9	9.0	13.6
54	9.5	8.2	12.4
68	6.6	5.8	8.8



#22
Acetophenone
Concen: 5.928 ng
RT: 8.646 min Scan# 962
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	105	Resp:	60078
Ion Ratio	Lower	Upper	
105	100		
71	2.1	4.0	6.0
51	30.0	23.8	35.6
120	20.4	16.0	24.0









#23
Nitrobenzene-d5
Concen: 11.678 ng
RT: 8.975 min Scan# 1018
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 82 Resp: 94031
Ion Ratio Lower Upper
82 100
128 36.7 29.6 44.4
54 53.0 42.7 64.1



#24
Nitrobenzene
Concen: 5.822 ng
RT: 9.022 min Scan# 1026
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 77 Resp: 45952
Ion Ratio Lower Upper
77 100
123 35.4 30.4 45.6
65 13.7 11.3 16.9









#25
Isophorone
Concen: 5.941 ng
RT: 9.552 min Scan# 1116
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 82 Resp: 84208
Ion Ratio Lower Upper
82 100
95 7.5 5.9 8.9
138 15.2 12.7 19.1



#26
2-Nitrophenol
Concen: 5.219 ng
RT: 9.734 min Scan# 1147
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 139 Resp: 17532
Ion Ratio Lower Upper
139 100
109 42.1 28.3 42.5
65 61.7 49.7 74.5









#27
2,4-Dimethylphenol
Concen: 5.741 ng
RT: 9.805 min Scan# 1158
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:122 Resp: 29065
Ion Ratio Lower Upper
122 100
107 127.0 106.6 160.0
121 61.4 48.1 72.1



#28
bis(2-Chloroethoxy)methane
Concen: 5.881 ng
RT: 10.034 min Scan# 1198
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 93 Resp: 53366
Ion Ratio Lower Upper
93 100
95 33.1 25.4 38.2
123 10.0 7.8 11.6









#29
2,4-Dichlorophenol
Concen: 5.420 ng
RT: 10.281 min Scan# 1240
Delta R.T. 0.012 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:162 Resp: 33081
Ion Ratio Lower Upper
162 100
164 63.8 43.9 83.9
98 40.6 17.9 57.9



#30
1,2,4-Trichlorobenzene
Concen: 5.930 ng
RT: 10.481 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:180 Resp: 41154
Ion Ratio Lower Upper
180 100
182 99.6 74.8 112.2
145 29.5 23.2 34.8









#31
Naphthalene
Concen: 6.086 ng
RT: 10.669 min Scan# 1177
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:128 Resp: 112944
Ion Ratio Lower Upper
128 100
129 10.1 8.5 12.7
127 13.9 10.9 16.3



#32
Benzoic acid
Concen: 4.324 ng
RT: 9.910 min Scan# 1177
Delta R.T. -0.053 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:122 Resp: 18765
Ion Ratio Lower Upper
122 100
105 140.4 117.4 157.4
77 120.4 89.2 129.2









#33
4-Chloroaniline
Concen: 5.813 ng
RT: 10.781 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	127	Resp:	45794
Ion	Ratio	Lower	Upper
127	100		
129	30.6	25.4	38.2
65	35.3	29.5	44.3
92	21.6	18.2	27.4



#34
Hexachlorobutadiene
Concen: 6.060 ng
RT: 10.957 min Scan# 1355
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	225	Resp:	28458
Ion	Ratio	Lower	Upper
225	100		
223	63.3	52.5	78.7
227	69.6	51.0	76.6









#35
Caprolactam
Concen: 5.592 ng
RT: 11.569 min Scan# 1460
Delta R.T. -0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:113 Resp: 7661
Ion Ratio Lower Upper
113 100
55 192.1 174.2 214.2
56 133.8 130.3 170.3



#36
4-Chloro-3-methylphenol
Concen: 5.704 ng
RT: 11.922 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:107 Resp: 40182
Ion Ratio Lower Upper
107 100
144 20.9 19.7 29.5
142 71.2 59.6 89.4









#37
2-Methylnaphthalene
Concen: 5.881 ng
RT: 12.281 min Scan# 1580
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	142	Resp:	80557
Ion Ratio	Lower	Upper	
142	100		
141	88.7	70.7	106.1
115	39.6	30.9	46.3



#38
1-Methylnaphthalene
Concen: 5.843 ng
RT: 12.499 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	142	Resp:	78393
Ion Ratio	Lower	Upper	
142	100		
141	91.3	74.2	111.4
116	4.3	3.4	5.2









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.463 min Scan# 1950
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:164 Resp: 258894
Ion Ratio Lower Upper
164 100
162 100.5 81.9 122.9
160 44.1 35.4 53.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.913 ng
RT: 12.657 min Scan# 1644
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:216 Resp: 49116
Ion Ratio Lower Upper
216 100
214 78.0 62.5 93.7
179 20.9 16.4 24.6
108 16.7 13.8 20.8









#41
Hexachlorocyclopentadiene
Concen: 2.747 ng
RT: 12.634 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	237	Resp:	10368
Ion Ratio	Lower	Upper	
237	100		
235	69.5	42.3	82.3
272	13.3	0.0	32.2



#42
2,4,6-Tribromophenol
Concen: 11.790 ng
RT: 15.957 min Scan# 2205
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	330	Resp:	38627
Ion Ratio	Lower	Upper	
330	100		
332	97.2	77.0	115.4
141	35.9	28.4	42.6









#43
2,4,6-Trichlorophenol
Concen: 5.559 ng
RT: 12.899 min Scan# 1684
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:196 Resp: 31807
Ion Ratio Lower Upper
196 100
198 91.3 74.8 112.2
200 31.2 24.3 36.5



#44
2,4,5-Trichlorophenol
Concen: 5.623 ng
RT: 12.981 min Scan# 1699
Delta R.T. 0.012 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:196 Resp: 34711
Ion Ratio Lower Upper
196 100
198 97.0 76.2 114.4
97 51.3 40.8 61.2
132 25.2 19.0 28.4









#45
2-Fluorobiphenyl
Concen: 13.109 ng
RT: 13.081 min Scan# 1716
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:172 Resp: 216736
Ion Ratio Lower Upper
172 100
171 36.2 28.0 42.0
170 24.9 18.9 28.3



#46
1,1'-Biphenyl
Concen: 5.839 ng
RT: 13.293 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:154 Resp: 112248
Ion Ratio Lower Upper
154 100
153 41.0 19.7 59.7
76 14.7 0.0 33.7









#47
2-Chloronaphthalene
Concen: 6.035 ng
RT: 13.334 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:162 Resp: 88425
Ion Ratio Lower Upper
162 100
127 40.0 31.3 46.9
164 32.7 26.9 40.3



#48
2-Nitroaniline
Concen: 5.128 ng
RT: 13.546 min Scan# 1795
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 65 Resp: 26920
Ion Ratio Lower Upper
65 100
92 70.9 50.6 75.8
138 89.6 68.7 103.1









#49
Acenaphthylene
Concen: 5.918 ng
RT: 14.181 min Scan# 1903
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:152 Resp: 134711
Ion Ratio Lower Upper
152 100
151 19.8 16.2 24.4
153 13.4 10.9 16.3



#50
Dimethylphthalate
Concen: 5.987 ng
RT: 13.922 min Scan# 1859
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:163 Resp: 117362
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 9.9 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 5.483 ng
RT: 14.040 min Scan# 1879
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:165 Resp: 22519
Ion Ratio Lower Upper
165 100
63 68.9 50.8 76.2
89 62.1 50.0 75.0



#52
Acenaphthene
Concen: 5.868 ng
RT: 14.528 min Scan# 1962
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:154 Resp: 81854
Ion Ratio Lower Upper
154 100
153 111.3 90.3 135.5
152 53.7 43.3 64.9









#53
3-Nitroaniline
Concen: 4.849 ng
RT: 14.375 min Scan# 1935
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:138 Resp: 20811
Ion Ratio Lower Upper
138 100
108 13.2 9.2 13.8
92 131.8 102.9 154.3



#54
2,4-Dinitrophenol
Concen: 3.782 ng
RT: 14.598 min Scan# 1974
Delta R.T. 0.012 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:184 Resp: 9692
Ion Ratio Lower Upper
184 100
63 66.9 53.8 80.6
154 66.9 54.2 81.2









#55
Dibenzenofuran
Concen: 5.898 ng
RT: 14.863 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:168 Resp: 141647
Ion Ratio Lower Upper
168 100
139 37.2 30.6 46.0
169 12.5 11.0 16.4



#56
4-Nitrophenol
Concen: 4.302 ng
RT: 14.728 min Scan# 1996
Delta R.T. 0.029 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:139 Resp: 14801
Ion Ratio Lower Upper
139 100
109 66.8 65.9 105.9
65 98.8 91.9 131.9









#57
2,4-Dinitrotoluene
Concen: 5.349 ng
RT: 14.834 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:165 Resp: 30374
Ion Ratio Lower Upper
165 100
63 53.2 38.9 58.3
89 79.8 60.5 90.7
182 2.3 2.3 3.5#



#58
Fluorene
Concen: 6.012 ng
RT: 15.516 min Scan# 2130
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:166 Resp: 112409
Ion Ratio Lower Upper
166 100
165 97.3 79.4 119.2
167 12.5 10.9 16.3









#59
2,3,4,6-Tetrachlorophenol
Concen: 5.874 ng
RT: 15.098 min Scan# 2059
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	232	Resp:	32617
Ion Ratio	Lower	Upper	
232	100		
131	38.8	28.1	42.1
130	2.1	1.8	2.8
166	28.7	18.9	28.3



#60
Diethylphthalate
Concen: 5.855 ng
RT: 15.287 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	149	Resp:	117764
Ion Ratio	Lower	Upper	
149	100		
177	23.2	18.2	27.2
150	13.3	10.3	15.5









#61
4-Chlorophenyl-phenylether
Concen: 5.990 ng
RT: 15.510 min Scan# 2128
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:204 Resp: 62005
Ion Ratio Lower Upper
204 100
206 33.9 26.1 39.1
141 55.8 47.4 71.0



#62
4-Nitroaniline
Concen: 4.747 ng
RT: 15.540 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:138 Resp: 20965
Ion Ratio Lower Upper
138 100
92 68.8 40.8 80.8
108 91.4 77.9 117.9









#63
Azobenzene
Concen: 5.555 ng
RT: 15.804 min Scan# 2178
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion: 77	Resp: 116495
Ion Ratio	Lower Upper
77 100	
182 24.3	3.8 43.8
105 21.7	0.5 40.5
51 30.7	10.0 50.0



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.222 min Scan# 2420
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion: 188	Resp: 567406
Ion Ratio	Lower Upper
188 100	
94 8.8	7.2 10.8
80 10.3	8.2 12.4









#65
4,6-Dinitro-2-methylphenol
Concen: 4.632 ng
RT: 15.610 min Scan# 2145
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:198 Resp: 17749
Ion Ratio Lower Upper
198 100
51 42.7 25.9 65.9
105 39.8 20.2 60.2



#66
n-Nitrosodiphenylamine
Concen: 6.223 ng
RT: 15.728 min Scan# 2166
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:169 Resp: 100187
Ion Ratio Lower Upper
169 100
168 65.7 53.5 80.3
167 37.5 29.3 43.9









#67
4-Bromophenyl-phenylether
Concen: 6.021 ng
RT: 16.404 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	248	Resp:	37483
Ion Ratio	Lower	Upper	
248	100		
250	92.5	78.6	117.8
141	76.2	57.0	85.6



#68
Hexachlorobenzene
Concen: 6.097 ng
RT: 16.528 min Scan# 2302
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	284	Resp:	40561
Ion Ratio	Lower	Upper	
284	100		
142	33.4	28.6	42.8
249	29.1	27.1	40.7









#69
 Atrazine
 Concen: 6.079 ng
 RT: 16.687 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion	Ratio	Lower	Upper
200	100		
173	23.8	6.8	46.8
215	45.8	28.4	68.4



#70
 Pentachlorophenol
 Concen: 4.669 ng
 RT: 16.881 min Scan# 2362
 Delta R.T. 0.006 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40

Tgt Ion	Ratio	Lower	Upper
266	100		
268	57.7	50.1	75.1
264	53.2	49.5	74.3









#71
Phenanthrene
Concen: 5.968 ng
RT: 17.263 min Scan# 2427
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:178 Resp: 182454
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.5 12.5 18.7



#72
Anthracene
Concen: 5.966 ng
RT: 17.351 min Scan# 2442
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:178 Resp: 180585
Ion Ratio Lower Upper
178 100
176 17.9 15.1 22.7
179 15.3 12.3 18.5









#73
 Carbazole
 Concen: 5.731 ng
 RT: 17.628 min Scan# 2489
 Delta R.T. 0.000 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:	167	Resp:	176133
Ion	Ratio	Lower	Upper
167	100		
166	21.1	17.4	26.0
139	12.2	10.2	15.2



#74
 Di-n-butylphthalate
 Concen: 6.480 ng
 RT: 18.186 min Scan# 2584
 Delta R.T. 0.006 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40

Tgt Ion:	149	Resp:	217056
Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.4	11.2
104	5.0	4.4	6.6









#75
Fluoranthene
Concen: 6.114 ng
RT: 19.239 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:202 Resp: 240929
Ion Ratio Lower Upper
202 100
101 9.3 0.0 29.8
203 17.2 0.0 37.7



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.316 min Scan# 3116
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:240 Resp: 609938
Ion Ratio Lower Upper
240 100
120 9.2 7.3 10.9
236 26.4 21.4 32.2









#77
Benzidine
Concen: 3.449 ng
RT: 19.428 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:184 Resp: 36634
Ion Ratio Lower Upper
184 100
185 14.3 10.3 15.5
183 15.2 9.1 13.7#



#78
Pyrene
Concen: 6.632 ng
RT: 19.586 min Scan# 2822
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:202 Resp: 274700
Ion Ratio Lower Upper
202 100
200 22.3 17.1 25.7
203 17.9 14.2 21.2









#79
 Terphenyl-d14
 Concen: 12.383 ng
 RT: 19.786 min Scan# 2856
 Delta R.T. 0.000 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:	244	Resp:	384172
Ion Ratio	Lower	Upper	
244	100		
212	7.3	6.0	9.0
122	9.4	6.6	9.8



#80
 Butylbenzylphthalate
 Concen: 5.892 ng
 RT: 20.469 min Scan# 2972
 Delta R.T. 0.006 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40

Tgt Ion:	149	Resp:	101962
Ion Ratio	Lower	Upper	
149	100		
91	68.9	59.8	89.8
206	23.3	18.6	27.8









#81
Benzo(a)anthracene
Concen: 5.861 ng
RT: 21.298 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	228	Resp:	239726
Ion Ratio	Lower	Upper	
228	100		
226	27.7	21.9	32.9
229	20.1	15.9	23.9



#82
3,3'-Dichlorobenzidine
Concen: 5.768 ng
RT: 21.239 min Scan# 3103
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:	252	Resp:	111817
Ion Ratio	Lower	Upper	
252	100		
254	62.5	52.4	78.6
126	9.4	7.6	11.4









#83
Chrysene
Concen: 5.943 ng
RT: 21.351 min Scan# 3122
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:228	Resp: 230665
Ion Ratio	Lower Upper
228 100	
226 31.4	24.0 36.0
229 19.9	15.8 23.8



#84
Bis(2-ethylhexyl)phthalate
Concen: 6.674 ng
RT: 21.233 min Scan# 3102
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:149	Resp: 156247
Ion Ratio	Lower Upper
149 100	
167 27.9	22.5 33.7
279 4.9	3.5 5.3









#85
Di-n-octyl phthalate
Concen: 6.238 ng
RT: 22.151 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 267614
Ion Ratio Lower Upper
149 100
167 1.8 1.3 1.9
43 10.5 7.9 11.9



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.716 min Scan# 3524
Delta R.T. 0.006 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:264 Resp: 636193
Ion Ratio Lower Upper
264 100
260 24.4 19.4 29.2
265 22.3 17.6 26.4









#87
 Indeno(1,2,3-cd)pyrene
 Concen: 6.243 ng
 RT: 26.204 min Scan# 3407
 Delta R.T. -0.006 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion: 276 Resp: 275617
 Ion Ratio Lower Upper
 276 100
 138 20.8 16.6 25.0
 277 26.6 20.2 30.2



#88
 Benzo(b)fluoranthene
 Concen: 5.897 ng
 RT: 23.027 min Scan# 3407
 Delta R.T. 0.047 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40

Tgt Ion: 252 Resp: 234566
 Ion Ratio Lower Upper
 252 100
 253 22.1 17.7 26.5
 125 8.8 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 5.927 ng
RT: 23.027 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 234566
Ion Ratio Lower Upper
252 100
253 22.1 17.3 25.9
125 8.8 6.5 9.7



#90
Benzo(a)pyrene
Concen: 5.860 ng
RT: 23.604 min Scan# 3505
Delta R.T. 0.000 min
Lab File: BP008119.D
Acq: 24 Nov 2021 16:40

Tgt Ion:252 Resp: 195350
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 10.3 7.4 11.0









#91
 Dibenzo(a,h)anthracene
 Concen: 6.290 ng
 RT: 26.227 min Scan# 3951
 Delta R.T. 0.000 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Tgt Ion:278 Resp: 231145
 Ion Ratio Lower Upper
 278 100
 139 15.3 10.8 16.2
 279 22.9 19.1 28.7



#92
 Benzo(g,h,i)perylene
 Concen: 6.363 ng
 RT: 26.974 min Scan# 4078
 Delta R.T. 0.000 min
 Lab File: BP008119.D
 Acq: 24 Nov 2021 16:40

Tgt Ion:276 Resp: 229070
 Ion Ratio Lower Upper
 276 100
 277 26.0 20.0 30.0
 138 16.2 14.1 21.1

