

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112521\
 Data File : BP008126.D
 Acq On : 24 Nov 2021 20:37
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP112521

Quant Time: Nov 25 01:24:51 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 01:13:46 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.580	0.523	9.8	107	0.00
3	Pyridine	1.547	1.442	6.8	109	0.00
4	n-Nitrosodimethylamine	0.645	0.611	5.3	109	0.00
5 S	2-Fluorophenol	1.242	1.164	6.3	107	0.00
6	Aniline	1.966	1.833	6.8	109	0.00
7 S	Phenol-d6	1.677	1.540	8.2	107	0.00
8	2-Chlorophenol	1.272	1.178	7.4	108	0.00
9	Benzaldehyde	1.114	0.928	16.7	108	0.00
10 C	Phenol	1.723	1.585	8.0	107	0.00
11	bis(2-Chloroethyl)ether	1.377	1.274	7.5	108	0.00
12	1,3-Dichlorobenzene	1.468	1.355	7.7	107	0.00
13 C	1,4-Dichlorobenzene	1.488	1.383	7.1	108	0.00
14	1,2-Dichlorobenzene	1.480	1.324	10.5	108	0.00
15	Benzyl Alcohol	1.329	1.246	6.2	110	0.00
16	2,2'-oxybis(1-Chloropropane)	2.078	1.877	9.7	107	0.00
17	2-Methylphenol	1.121	1.029	8.2	108	0.00
18	Hexachloroethane	0.553	0.510	7.8	107	0.00
19 P	n-Nitroso-di-n-propylamine	1.160	1.015	12.5	109	0.00
20	3+4-Methylphenols	1.547	1.414	8.6	108	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
22	Acetophenone	0.547	0.511	6.6	108	0.00
23 S	Nitrobenzene-d5	0.440	0.416	5.5	108	0.00
24	Nitrobenzene	0.427	0.402	5.9	108	0.00
25	Isophorone	0.769	0.707	8.1	109	0.00
26 C	2-Nitrophenol	0.184	0.179	2.7	108	0.00
27	2,4-Dimethylphenol	0.275	0.257	6.5	108	0.00
28	bis(2-Chloroethoxy)methane	0.490	0.454	7.3	109	0.00
29 C	2,4-Dichlorophenol	0.329	0.312	5.2	108	0.00
30	1,2,4-Trichlorobenzene	0.382	0.356	6.8	107	0.00
31	Naphthalene	1.025	0.955	6.8	108	0.00
32	Benzoic acid	0.225	0.223	0.9	111	0.00
33	4-Chloroaniline	0.425	0.397	6.6	111	0.00
34 C	Hexachlorobutadiene	0.261	0.245	6.1	108	0.00
35	Caprolactam	0.073	0.067	8.2	110	0.00
36 C	4-Chloro-3-methylphenol	0.376	0.349	7.2	110	0.00
37	2-Methylnaphthalene	0.725	0.663	8.6	108	0.00
38	1-Methylnaphthalene	0.705	0.643	8.8	109	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	0.650	0.621	4.5	108	0.00
41 P	Hexachlorocyclopentadiene	0.250	0.264	-5.6	112	0.00
42 S	2,4,6-Tribromophenol	0.256	0.240	6.3	109	0.00
43 C	2,4,6-Trichlorophenol	0.440	0.421	4.3	109	0.00
44	2,4,5-Trichlorophenol	0.472	0.452	4.2	109	0.00
45 S	2-Fluorobiphenyl	1.377	1.290	6.3	109	0.00
46	1,1'-Biphenyl	1.451	1.369	5.7	108	0.00
47	2-Chloronaphthalene	1.146	1.085	5.3	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112521\
 Data File : BP008126.D
 Acq On : 24 Nov 2021 20:37
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP112521

Quant Time: Nov 25 01:24:51 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 01:13:46 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.395	0.382	3.3	109	0.00
49	Acenaphthylene	1.768	1.673	5.4	110	0.00
50	Dimethylphthalate	1.502	1.389	7.5	110	0.00
51	2,6-Dinitrotoluene	0.312	0.296	5.1	109	0.00
52 C	Acenaphthene	1.054	0.988	6.3	110	0.00
53	3-Nitroaniline	0.316	0.313	0.9	113	0.00
54 P	2,4-Dinitrophenol	0.184	0.189	-2.7	109	0.00
55	Dibenzofuran	1.800	1.669	7.3	109	0.00
56 P	4-Nitrophenol	0.242	0.246	-1.7	112	0.00
57	2,4-Dinitrotoluene	0.428	0.410	4.2	110	0.00
58	Fluorene	1.463	1.244	15.0	109	0.00
59	2,3,4,6-Tetrachlorophenol	0.419	0.401	4.3	112	0.00
60	Diethylphthalate	1.490	1.378	7.5	110	0.00
61	4-Chlorophenyl-phenylether	0.802	0.687	14.3	109	0.00
62	4-Nitroaniline	0.328	0.320	2.4	110	0.00
63	Azobenzene	1.524	1.429	6.2	108	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	0.131	0.130	0.8	110	0.00
66 c	n-Nitrosodiphenylamine	0.579	0.538	7.1	110	0.00
67	4-Bromophenyl-phenylether	0.222	0.207	6.8	110	0.00
68	Hexachlorobenzene	0.238	0.220	7.6	110	0.00
69	Atrazine	0.150	0.138	8.0	111	0.00
70 C	Pentachlorophenol	0.141	0.140	0.7	111	0.00
71	Phenanthrene	1.056	0.974	7.8	109	0.00
72	Anthracene	1.048	0.976	6.9	110	0.00
73	Carbazole	1.023	0.944	7.7	109	0.00
74	Di-n-butylphthalate	1.281	1.145	10.6	108	0.00
75 C	Fluoranthene	1.423	1.215	14.6	109	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
77	Benzidine	0.274	0.293	-6.9	104	0.00
78	Pyrene	1.446	1.197	17.2	108	0.00
79 S	Terphenyl-d14	0.957	0.910	4.9	110	0.00
80	Butylbenzylphthalate	0.570	0.538	5.6	110	0.00
81	Benzo(a)anthracene	1.326	1.224	7.7	110	0.00
82	3,3'-Dichlorobenzidine	0.625	0.561	10.2	110	0.00
83	Chrysene	1.261	1.184	6.1	112	0.00
84	Bis(2-ethylhexyl)phthalate	0.865	0.725	16.2	109	0.00
85 c	Di-n-octyl phthalate	1.469	1.378	6.2	110	0.00
86 I	Perylene-d12	1.000	1.000	0.0	113	0.00
87	Indeno(1,2,3-cd)pyrene	1.455	1.358	6.7	113	0.00
88	Benzo(b)fluoranthene	1.206	1.111	7.9	113	0.00
89	Benzo(k)fluoranthene	1.180	1.088	7.8	110	0.00
90 C	Benzo(a)pyrene	1.030	0.955	7.3	111	0.00
91	Dibenzo(a,h)anthracene	1.208	1.133	6.2	113	0.00
92	Benzo(g,h,i)perylene	1.219	1.141	6.4	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112521\
Data File : BP008126.D
Acq On : 24 Nov 2021 20:37
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP112521

Quant Time: Nov 25 01:24:51 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 25 01:13:46 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0