

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112521\
 Data File : BP008125.D
 Acq On : 24 Nov 2021 20:03
 Operator : CG/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021
 Supervised By :Sohil Jodhani 11/30/2021

Quant Time: Nov 25 00:27:38 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 00:18:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	139049	20.000	ng	0.00
21) Naphthalene-d8	10.616	136	511846	20.000	ng	0.00
39) Acenaphthene-d10	14.463	164	336424	20.000	ng	0.00
64) Phenanthrene-d10	17.216	188	723649	20.000	ng	0.00
76) Chrysene-d12	21.322	240	722796	20.000	ng	0.00
86) Perylene-d12	23.710	264	819345	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.405	112	1112476	136.750	ng	0.00
7) Phenol-d6	7.005	99	1518609	134.669	ng	0.00
23) Nitrobenzene-d5	8.981	82	1508176	133.670	ng	0.00
42) 2,4,6-Tribromophenol	15.963	330	583147	136.969	ng	0.00
45) 2-Fluorobiphenyl	13.087	172	2963809	137.947	ng	0.00
79) Terphenyl-d14	19.786	244	4590384	124.862	ng	0.00
Target Compounds						Qvalue
3) Pyridine	3.717	79	741579m	68.037	ng	
4) n-Nitrosodimethylamine	3.628	42	303454	62.141	ng	99
6) Aniline	7.152	93	921379	64.725	ng	97
8) 2-Chlorophenol	7.387	128	581353	63.658	ng	98
10) Phenol	7.028	94	796309	65.924	ng	99
11) bis(2-Chloroethyl)ether	7.246	93	633394	61.571	ng	99
12) 1,3-Dichlorobenzene	7.705	146	662920	63.236	ng	99
13) 1,4-Dichlorobenzene	7.852	146	671233	63.084	ng	99
15) Benzyl Alcohol	8.069	79	638266	65.560	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.358	45	946074	59.006	ng	99
17) 2-Methylphenol	8.275	107	521986	65.589	ng	99
18) Hexachloroethane	8.893	117	251938	66.488	ng	98
20) 3+4-Methylphenols	8.605	107	717601	65.047	ng	99
22) Acetophenone	8.646	105	912190	64.235	ng	# 99
24) Nitrobenzene	9.022	77	744873	67.351	ng	97
25) Isophorone	9.557	82	1326210	66.773	ng	99
26) 2-Nitrophenol	9.734	139	338729	71.959	ng	96
27) 2,4-Dimethylphenol	9.805	122	475494	67.025	ng	96
28) bis(2-Chloroethoxy)met...	10.034	93	831556	65.399	ng	99
29) 2,4-Dichlorophenol	10.275	162	583843	68.267	ng	99
30) 1,2,4-Trichlorobenzene	10.481	180	659681	67.833	ng	97
31) Naphthalene	10.669	128	1735813	66.753	ng	100
32) Benzoic acid	10.010	122	458240	75.346	ng	99
33) 4-Chloroaniline	10.781	127	746814	67.654	ng	99
34) Hexachlorobutadiene	10.957	225	448776	68.198	ng	97
35) Caprolactam	11.604	113	134081m	69.840	ng	
36) 4-Chloro-3-methylphenol	11.922	107	658022	66.656	ng	98
37) 2-Methylnaphthalene	12.281	142	1217748	63.437	ng	98
38) 1-Methylnaphthalene	12.498	142	1186000	63.087	ng	100
40) 1,2,4,5-Tetrachloroben...	12.651	216	748790	69.374	ng	100
41) Hexachlorocyclopentadiene	12.628	237	367784	74.997	ng	99
43) 2,4,6-Trichlorophenol	12.898	196	515601	69.344	ng	98
44) 2,4,5-Trichlorophenol	12.975	196	554687	69.151	ng	98
46) 1,1'-Biphenyl	13.293	154	1621250	64.896	ng	98
47) 2-Chloronaphthalene	13.334	162	1292851	67.905	ng	99
48) 2-Nitroaniline	13.545	65	481335	70.554	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112521\
 Data File : BP008125.D
 Acq On : 24 Nov 2021 20:03
 Operator : CG/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021
 Supervised By :Sohil Jodhani 11/30/2021


Quant Time: Nov 25 00:27:38 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 00:18:05 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
49) Acenaphthylene	14.181	152	1995326	67.461	ng	100
50) Dimethylphthalate	13.934	163	1674183	65.719	ng	100
51) 2,6-Dinitrotoluene	14.045	165	364812	68.353	ng	98
52) Acenaphthene	14.528	154	1169908	64.541	ng	99
53) 3-Nitroaniline	14.375	138	388352	69.637	ng	98
54) 2,4-Dinitrophenol	14.587	184	261729	78.590	ng	96
55) Dibenzofuran	14.863	168	1971190	63.161	ng	100
56) 4-Nitrophenol	14.698	139	316138	70.716	ng	94
57) 2,4-Dinitrotoluene	14.840	165	505005	68.433	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.098	232	484187m	67.104	ng	
60) Diethylphthalate	15.292	149	1662800	63.624	ng	100
62) 4-Nitroaniline	15.545	138	404669	70.512	ng	98
63) Azobenzene	15.804	77	1701022	62.418	ng	98
65) 4,6-Dinitro-2-methylph...	15.610	198	364822	74.647	ng	98
66) n-Nitrosodiphenylamine	15.728	169	1360778	66.273	ng	99
67) 4-Bromophenyl-phenylether	16.404	248	535723	67.473	ng	97
68) Hexachlorobenzene	16.528	284	575470	67.825	ng	97
70) Pentachlorophenol	16.875	266	384201	70.425	ng	98
71) Phenanthrene	17.263	178	2445848	62.729	ng	100
72) Anthracene	17.357	178	2451990	63.520	ng	100
73) Carbazole	17.628	167	2396564	61.138	ng	99
77) Benzidine	19.410	184	644922	51.238	ng	98
80) Butylbenzylphthalate	20.463	149	1387500	67.655	ng	100
81) Benzo(a)anthracene	21.304	228	3206919	66.158	ng	100
83) Chrysene	21.357	228	2998897	65.197	ng	99
85) Di-n-octyl phthalate	22.151	149	3517865	69.199	ng	99
87) Indeno(1,2,3-cd)pyrene	26.227	276	3999971	70.355	ng	99
88) Benzo(b)fluoranthene	22.986	252	3270551	63.843	ng	100
89) Benzo(k)fluoranthene	23.033	252	3198711	62.757	ng	98
90) Benzo(a)pyrene	23.610	252	2851773	66.423	ng	99
91) Dibenzo(a,h)anthracene	26.245	278	3275240	69.204	ng	98
92) Benzo(g,h,i)perylene	26.998	276	3404932	73.442	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations


Quant Time: Nov 25 00:27:38 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 25 00:18:05 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.816 min Scan# 81
Delta R.T. 0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:152 Resp: 139049
Ion Ratio Lower Upper
152 100
150 159.4 125.7 188.5
115 62.6 50.7 76.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#3
Pyridine
Concen: 68.037 ng m
RT: 3.717 min Scan# 124
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Tgt Ion: 79 Resp: 741579
Ion Ratio Lower Upper
79 100
52 75.2 60.0 90.0
51 37.0 29.6 44.4





#4
n-Nitrosodimethylamine
Concen: 62.141 ng
RT: 3.628 min Scan# 108
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion: 42 Resp: 303454
Ion Ratio Lower Upper
42 100
74 136.8 108.0 162.0
44 6.7 5.8 8.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#5
2-Fluorophenol
Concen: 136.750 ng
RT: 5.405 min Scan# 411
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:112 Resp: 1112476
Ion Ratio Lower Upper
112 100
64 66.1 53.4 80.2
63 35.2 27.8 41.8









#6
Aniline
Concen: 64.725 ng
RT: 7.152 min Scan# 707
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
Client Sample Id :
SSTDICC080

Tgt Ion: 93 Resp: 921379
Ion Ratio Lower Upper
93 100
66 43.0 32.9 49.3
65 21.1 17.5 26.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#7
Phenol-d6
Concen: 134.669 ng
RT: 7.005 min Scan# 683
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion: 99 Resp: 1518609
Ion Ratio Lower Upper
99 100
42 18.8 14.6 22.0
71 39.8 31.1 46.7









#8
2-Chlorophenol
Concen: 63.658 ng
RT: 7.387 min Scan# 748
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:128 Resp: 58135
Ion Ratio Lower Upper
128 100
130 31.5 13.5 53.5
64 56.6 37.4 77.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#10
Phenol
Concen: 65.924 ng
RT: 7.028 min Scan# 687
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion: 94 Resp: 796309
Ion Ratio Lower Upper
94 100
65 30.7 10.9 50.9
66 43.5 24.3 64.3





#11
bis(2-Chloroethyl)ether
Concen: 61.571 ng
RT: 7.246 min Scan# 71
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion: 93 Resp: 633394
Ion Ratio Lower Upper
93 100
63 82.7 61.8 101.8
95 32.2 12.5 52.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#12
1,3-Dichlorobenzene
Concen: 63.236 ng
RT: 7.705 min Scan# 802
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Tgt Ion:146 Resp: 662920
Ion Ratio Lower Upper
146 100
148 63.9 51.9 77.9
75 36.5 29.0 43.6





#13
1,4-Dichlorobenzene
Concen: 63.084 ng
RT: 7.852 min Scan# 81
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:146 Resp: 67123
Ion Ratio Lower Upper
146 100
148 63.4 49.6 74.4
111 44.0 34.7 52.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#15
Benzyl Alcohol
Concen: 65.560 ng
RT: 8.069 min Scan# 864
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03



Tgt Ion: 79 Resp: 638266
Ion Ratio Lower Upper
79 100
108 68.5 55.5 83.3
77 62.5 51.0 76.6









#16
2,2'-oxybis(1-Chloropropane)
Concen: 59.006 ng
RT: 8.358 min Scan# 912
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 45 Resp: 946074
Ion Ratio Lower Upper
45 100
77 14.2 0.0 33.8
79 10.9 0.0 30.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#17
2-Methylphenol
Concen: 65.589 ng
RT: 8.275 min Scan# 899
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:107 Resp: 521986
Ion Ratio Lower Upper
107 100
108 110.7 90.2 135.2
77 51.4 41.3 61.9
79 50.9 41.0 61.6






#18
Hexachloroethane
Concen: 66.488 ng
RT: 8.893 min Scan# 1004
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :




SSTDICC080




Tgt Ion:117 Resp: 251938
Ion Ratio Lower Upper
117 100
119 93.6 74.7 112.1
201 114.2 88.6 132.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#20
3+4-Methylphenols
Concen: 65.047 ng
RT: 8.605 min Scan# 955
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03



Tgt Ion:107 Resp: 717601
Ion Ratio Lower Upper
107 100
108 89.3 68.2 108.2
77 36.9 16.9 56.9
79 32.1 12.5 52.5






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.616 min Scan# 1296
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :

SSTDICC080






Tgt Ion:136 Resp: 511840
Ion Ratio Lower Upper
136 100
137 11.0 9.0 13.6
54 9.7 8.2 12.4
68 6.8 5.8 8.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#22
Acetophenone
Concen: 64.235 ng
RT: 8.646 min Scan# 962
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Tgt Ion:105 Resp: 912190
Ion Ratio Lower Upper
105 100
71 6.1 4.0 6.0#
51 29.5 23.8 35.6
120 20.5 16.0 24.0





#23
Nitrobenzene-d5
Concen: 133.670 ng
RT: 8.981 min Scan# 1019
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion: 82 Resp: 1508170
Ion Ratio Lower Upper
82 100
128 37.8 29.6 44.4
54 53.1 42.7 64.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#24
Nitrobenzene
Concen: 67.351 ng
RT: 9.022 min Scan# 1026
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03



Tgt Ion: 77 Resp: 744873
Ion Ratio Lower Upper
77 100
123 40.0 30.4 45.6
65 13.8 11.3 16.9









#25
Isophorone
Concen: 66.773 ng
RT: 9.557 min Scan# 1116
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 1326210
Ion Ratio Lower Upper
82 100
95 7.5 5.9 8.9
138 15.6 12.7 19.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#26
2-Nitrophenol
Concen: 71.959 ng
RT: 9.734 min Scan# 1147
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:139 Resp: 338729
Ion Ratio Lower Upper
139 100
109 37.4 28.3 42.5
65 58.6 49.7 74.5





#27
2,4-Dimethylphenol
Concen: 67.025 ng
RT: 9.805 min Scan# 1158
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :





SSTDICC080

Tgt Ion:122 Resp: 475494
Ion Ratio Lower Upper
122 100
107 127.5 106.6 160.0
121 59.5 48.1 72.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#28
bis(2-Chloroethoxy)methane
Concen: 65.399 ng
RT: 10.034 min Scan# 1198
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion: 93 Resp: 831556
Ion Ratio Lower Upper
93 100
95 32.2 25.4 38.2
123 10.1 7.8 11.6









#29
2,4-Dichlorophenol
Concen: 68.267 ng
RT: 10.275 min Scan# 1239
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 58384
Ion Ratio Lower Upper
162 100
164 65.0 43.9 83.9
98 38.5 17.9 57.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#30
1,2,4-Trichlorobenzene
Concen: 67.833 ng
RT: 10.481 min Scan# 1274
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion:180 Resp: 659681
Ion Ratio Lower Upper
180 100
182 96.5 74.8 112.2
145 29.4 23.2 34.8





#31
Naphthalene
Concen: 66.753 ng
RT: 10.669 min Scan# 1194
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:128 Resp: 173581
Ion Ratio Lower Upper
128 100
129 10.6 8.5 12.7
127 13.5 10.9 16.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#32
Benzoic acid
Concen: 75.346 ng
RT: 10.010 min Scan# 1194
Delta R.T. 0.047 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Tgt Ion:122 Resp: 458240
Ion Ratio Lower Upper
122 100
105 136.6 117.4 157.4
77 109.9 89.2 129.2





#33
4-Chloroaniline
Concen: 67.654 ng
RT: 10.781 min Scan# 1325
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080






Tgt Ion:127 Resp: 746814

Ion	Ratio	Lower	Upper
127	100		
129	31.7	25.4	38.2
65	37.6	29.5	44.3
92	23.2	18.2	27.4

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#34
Hexachlorobutadiene
Concen: 68.198 ng
RT: 10.957 min Scan# 1355
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:225 Resp: 448776


Ion	Ratio	Lower	Upper
225	100		
223	61.5	52.5	78.7
227	62.7	51.0	76.6





#35
Caprolactam
Concen: 69.840 ng m
RT: 11.604 min Scan# 1465
Delta R.T. 0.029 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:113 Resp: 13408
Ion Ratio Lower Upper
113 100
55 198.0 174.2 214.2
56 155.6 130.3 170.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#36
4-Chloro-3-methylphenol
Concen: 66.656 ng
RT: 11.922 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03



Tgt Ion:107 Resp: 658022
Ion Ratio Lower Upper
107 100
144 23.0 19.7 29.5
142 73.5 59.6 89.4





#37
2-Methylnaphthalene
Concen: 63.437 ng
RT: 12.281 min Scan# 1580
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :





SSTDICC080

Tgt Ion:142 Resp: 121774
Ion Ratio Lower Upper
142 100
141 89.4 70.7 106.1
115 40.3 30.9 46.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#38
1-Methylnaphthalene
Concen: 63.087 ng
RT: 12.498 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:142 Resp: 1186000
Ion Ratio Lower Upper
142 100
141 92.3 74.2 111.4
116 4.3 3.4 5.2









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.463 min Scan# 1950
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:164 Resp: 336424
Ion Ratio Lower Upper
164 100
162 102.9 81.9 122.9
160 45.0 35.4 53.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#40
1,2,4,5-Tetrachlorobenzene
Concen: 69.374 ng
RT: 12.651 min Scan# 1643
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion:216 Resp: 748790
Ion Ratio Lower Upper
216 100
214 78.0 62.5 93.7
179 20.6 16.4 24.6
108 17.2 13.8 20.8





#41
Hexachlorocyclopentadiene
Concen: 74.997 ng
RT: 12.628 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion: 237 Resp: 367784
Ion Ratio Lower Upper
237 100
235 62.9 42.3 82.3
272 12.3 0.0 32.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#42
2,4,6-Tribromophenol
Concen: 136.969 ng
RT: 15.963 min Scan# 2206
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Tgt Ion: 330 Resp: 583147
Ion Ratio Lower Upper
330 100
332 97.1 77.0 115.4
141 35.1 28.4 42.6





#43
2,4,6-Trichlorophenol
Concen: 69.344 ng
RT: 12.898 min Scan# 1685
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:196 Resp: 51560
Ion Ratio Lower Upper
196 100
198 95.6 74.8 112.2
200 32.1 24.3 36.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#44
2,4,5-Trichlorophenol
Concen: 69.151 ng
RT: 12.975 min Scan# 1698
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03



Tgt Ion:196 Resp: 554687
Ion Ratio Lower Upper
196 100
198 95.8 76.2 114.4
97 47.9 40.8 61.2
132 24.0 19.0 28.4





#45
2-Fluorobiphenyl
Concen: 137.947 ng
RT: 13.087 min Scan# 1716
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :





SSTDICC080

Tgt Ion:172 Resp: 2963809
Ion Ratio Lower Upper
172 100
171 35.6 28.0 42.0
170 23.7 18.9 28.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#46
1,1'-Biphenyl
Concen: 64.896 ng
RT: 13.293 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion:154 Resp: 1621250
Ion Ratio Lower Upper
154 100
153 41.1 19.7 59.7
76 13.2 0.0 33.7





#47
2-Chloronaphthalene
Concen: 67.905 ng
RT: 13.334 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:162 Resp: 129285
Ion Ratio Lower Upper
162 100
127 39.5 31.3 46.9
164 32.9 26.9 40.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#48
2-Nitroaniline
Concen: 70.554 ng
RT: 13.545 min Scan# 1795
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Tgt Ion: 65 Resp: 481335
Ion Ratio Lower Upper
65 100
92 65.9 50.6 75.8
138 90.2 68.7 103.1





#49
Acenaphthylene
Concen: 67.461 ng
RT: 14.181 min Scan# 1903
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:152 Resp: 1995320
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4
153 13.4 10.9 16.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#50
Dimethylphthalate
Concen: 65.719 ng
RT: 13.934 min Scan# 1861
Delta R.T. 0.012 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Tgt Ion:163 Resp: 1674183
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.2
164 10.2 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 68.353 ng
RT: 14.045 min Scan# 1879
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:165 Resp: 364812
Ion Ratio Lower Upper
165 100
63 61.3 50.8 76.2
89 62.2 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#52
Acenaphthene
Concen: 64.541 ng
RT: 14.528 min Scan# 1962
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:154 Resp: 1169908
Ion Ratio Lower Upper
154 100
153 114.4 90.3 135.5
152 53.3 43.3 64.9









#53
3-Nitroaniline
Concen: 69.637 ng
RT: 14.375 min Scan# 1936
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:138 Resp: 388352
Ion Ratio Lower Upper
138 100
108 12.1 9.2 13.8
92 131.4 102.9 154.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#54
2,4-Dinitrophenol
Concen: 78.590 ng
RT: 14.587 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:184 Resp: 261729
Ion Ratio Lower Upper
184 100
63 72.3 53.8 80.6
154 68.6 54.2 81.2









#55
Dibenzofuran
Concen: 63.161 ng
RT: 14.863 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:168 Resp: 1971190
Ion Ratio Lower Upper
168 100
139 38.3 30.6 46.0
169 13.5 11.0 16.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#56
4-Nitrophenol
Concen: 70.716 ng
RT: 14.698 min Scan# 1991
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:139 Resp: 316138
Ion Ratio Lower Upper
139 100
109 91.8 65.9 105.9
65 117.2 91.9 131.9





#57
2,4-Dinitrotoluene
Concen: 68.433 ng
RT: 14.840 min Scan# 2014
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :

BNA_P





ClientSampleId :

SSTDICC080

Tgt Ion:165 Resp: 50500
Ion Ratio Lower Upper
165 100
63 49.7 38.9 58.3
89 75.6 60.5 90.7
182 2.7 2.3 3.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#59
2,3,4,6-Tetrachlorophenol
Concen: 67.104 ng m
RT: 15.098 min Scan# 2059
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:232 Resp: 484187
Ion Ratio Lower Upper
232 100
131 35.9 28.1 42.1
130 2.4 1.8 2.8
166 24.1 18.9 28.3






#60
Diethylphthalate
Concen: 63.624 ng
RT: 15.292 min Scan# 2091
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :

SSTDICC080






Tgt Ion:149 Resp: 1662800
Ion Ratio Lower Upper
149 100
177 23.0 18.2 27.2
150 12.9 10.3 15.5


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#62
4-Nitroaniline
Concen: 70.512 ng
RT: 15.545 min Scan# 2135
Delta R.T. 0.012 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Tgt Ion:138 Resp: 404669
Ion Ratio Lower Upper
138 100
92 58.0 40.8 80.8
108 99.1 77.9 117.9





#63
Azobenzene
Concen: 62.418 ng
RT: 15.804 min Scan# 2179
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion: 77 Resp: 170102
Ion Ratio Lower Upper
77 100
182 25.5 3.8 43.8
105 19.9 0.5 40.5
51 30.8 10.0 50.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.216 min Scan# 2419
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion:188 Resp: 723649
Ion Ratio Lower Upper
188 100
94 9.0 7.2 10.8
80 10.4 8.2 12.4





#65
4,6-Dinitro-2-methylphenol
Concen: 74.647 ng
RT: 15.610 min Scan# 2146
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:198 Resp: 36482
Ion Ratio Lower Upper
198 100
51 46.7 25.9 65.9
105 42.0 20.2 60.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#66
n-Nitrosodiphenylamine
Concen: 66.273 ng
RT: 15.728 min Scan# 2166
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion:169 Resp: 1360778
Ion Ratio Lower Upper
169 100
168 66.7 53.5 80.3
167 36.0 29.3 43.9





#67
4-Bromophenyl-phenylether
Concen: 67.473 ng
RT: 16.404 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:248 Resp: 53572
Ion Ratio Lower Upper
248 100
250 99.0 78.6 117.8
141 75.9 57.0 85.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#68
Hexachlorobenzene
Concen: 67.825 ng
RT: 16.528 min Scan# 2302
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:284 Resp: 575470
Ion Ratio Lower Upper
284 100
142 33.0 28.6 42.8
249 33.6 27.1 40.7






#70
Pentachlorophenol
Concen: 70.425 ng
RT: 16.875 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :

SSTDICC080






Tgt Ion:266 Resp: 38420
Ion Ratio Lower Upper
266 100
268 63.8 50.1 75.1
264 63.1 49.5 74.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#71
Phenanthrene
Concen: 62.729 ng
RT: 17.263 min Scan# 2427
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion:178 Resp: 2445848
Ion Ratio Lower Upper
178 100
176 19.3 15.3 22.9
179 15.4 12.5 18.7





#72
 Anthracene
 Concen: 63.520 ng
 RT: 17.357 min Scan# 2443
 Delta R.T. 0.006 min
 Lab File: BP008125.D
 Acq: 24 Nov 2021 20:03




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080




Tgt Ion:178 Resp: 2451990
 Ion Ratio Lower Upper
 178 100
 176 18.7 15.1 22.7
 179 15.4 12.3 18.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
 Supervised By :Sohil Jodhani 11/30/2021




#73
 Carbazole
 Concen: 61.138 ng
 RT: 17.628 min Scan# 2489
 Delta R.T. -0.000 min
 Lab File: BP008125.D
 Acq: 24 Nov 2021 20:03




Tgt Ion:167 Resp: 2396564
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.4 26.0
 139 13.3 10.2 15.2





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.322 min Scan# 3116
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:240 Resp: 722790
Ion Ratio Lower Upper
240 100
120 8.2 7.3 10.9
236 26.1 21.4 32.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#77
Benzidine
Concen: 51.238 ng
RT: 19.410 min Scan# 2792
Delta R.T. -0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:184 Resp: 644922
Ion Ratio Lower Upper
184 100
185 13.9 10.3 15.5
183 11.8 9.1 13.7






#79
 Terphenyl-d14
 Concen: 124.862 ng
 RT: 19.786 min Scan# 2856
 Delta R.T. -0.000 min
 Lab File: BP008125.D
 Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :

SSTDICC080






Tgt Ion:244 Resp: 4590384
 Ion Ratio Lower Upper
 244 100
 212 8.0 6.0 9.0
 122 7.8 6.6 9.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
 Supervised By :Sohil Jodhani 11/30/2021




#80
 Butylbenzylphthalate
 Concen: 67.655 ng
 RT: 20.463 min Scan# 2971
 Delta R.T. -0.000 min
 Lab File: BP008125.D
 Acq: 24 Nov 2021 20:03




Tgt Ion:149 Resp: 1387500
 Ion Ratio Lower Upper
 149 100
 91 74.5 59.8 89.8
 206 22.9 18.6 27.8





#81
Benzo(a)anthracene
Concen: 66.158 ng
RT: 21.304 min Scan# 3114
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:228 Resp: 3206919
Ion Ratio Lower Upper
228 100
226 27.4 21.9 32.9
229 19.9 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#83
Chrysene
Concen: 65.197 ng
RT: 21.357 min Scan# 3123
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion:228 Resp: 2998897
Ion Ratio Lower Upper
228 100
226 30.3 24.0 36.0
229 19.6 15.8 23.8





#85
Di-n-octyl phthalate
Concen: 69.199 ng
RT: 22.151 min Scan# 31
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:149 Resp: 351786
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 9.4 7.9 11.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.710 min Scan# 3523
Delta R.T. -0.000 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion:264 Resp: 819345
Ion Ratio Lower Upper
264 100
260 23.7 19.4 29.2
265 21.5 17.6 26.4





#87
 Indeno(1,2,3-cd)pyrene
 Concen: 70.355 ng
 RT: 26.227 min Scan# 3948
 Delta R.T. 0.018 min
 Lab File: BP008125.D
 Acq: 24 Nov 2021 20:03

Instrument :

BNA_P

ClientSampleId :





SSTDICC080

Tgt Ion:276 Resp: 399997
 Ion Ratio Lower Upper
 276 100
 138 19.4 16.6 25.0
 277 25.3 20.2 30.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
 Supervised By :Sohil Jodhani 11/30/2021



#88
 Benzo(b)fluoranthene
 Concen: 63.843 ng
 RT: 22.986 min Scan# 3400
 Delta R.T. 0.006 min
 Lab File: BP008125.D
 Acq: 24 Nov 2021 20:03

Tgt Ion:252 Resp: 3270551
 Ion Ratio Lower Upper
 252 100
 253 21.9 17.7 26.5
 125 7.6 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 62.757 ng
RT: 23.033 min Scan# 3407
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 319871
Ion Ratio Lower Upper
252 100
253 22.4 17.3 25.9
125 7.5 6.5 9.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#90
Benzo(a)pyrene
Concen: 66.423 ng
RT: 23.610 min Scan# 3506
Delta R.T. 0.006 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03


Tgt Ion:252 Resp: 2851773
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 8.5 7.4 11.0





#91
Dibenzo(a,h)anthracene
Concen: 69.204 ng
RT: 26.245 min Scan# 3951
Delta R.T. 0.018 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03




Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion: 278 Resp: 3275240
Ion Ratio Lower Upper
278 100
139 11.8 10.8 16.2
279 23.5 19.1 28.7

Manual Integrations
APPROVED

Reviewed By : Jagrut Upadhyay 11/30/2021
Supervised By : Sohil Jodhani 11/30/2021



#92
Benzo(g,h,i)perylene
Concen: 73.442 ng
RT: 26.998 min Scan# 4082
Delta R.T. 0.023 min
Lab File: BP008125.D
Acq: 24 Nov 2021 20:03

Tgt Ion: 276 Resp: 3404932
Ion Ratio Lower Upper
276 100
277 23.6 20.0 30.0
138 17.1 14.1 21.1

