

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112522\
 Data File : BP012753.D
 Acq On : 26 Nov 2022 06:47
 Operator : CG/JU
 Sample : PB149119BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS119

Quant Time: Nov 26 08:43:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 25 15:59:22 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/26/2022
 Supervised By :mohammad ahmed 11/30/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.022	152	584905	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	2614578	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.657	164	1729855	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.416	188	3524842	20.000	ng/u1	0.00
79) Chrysene-d12	21.498	240	2483363	20.000	ng/u1	0.00
88) Perylene-d12	24.009	264	1937598	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.399	96	79375	5.815	ng/uL	0.00
4) Pyridine-d5	3.816	84	1131058	27.690	ng/u1	0.00
7) Phenol-d5	7.169	99	1553504	32.322	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.346	67	937375	31.651	ng/u1	0.00
11) 2-Chlorophenol-d4	7.546	132	1215448	33.188	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	1267647	32.735	ng/u1	0.00
21) Nitrobenzene-d5	9.187	128	590229	36.883	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	660652	38.881	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	1344188	33.904	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	1647835	30.352	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	4009791	33.676	ng/u1	0.00
49) Acenaphthylene-d8	14.351	160	4555213	32.788	ng/u1	0.00
54) 4-Nitrophenol-d4	14.839	143	711010	33.493	ng/u1	0.00
60) Fluorene-d10	15.651	176	3424324	32.839	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.763	200	603373	35.815	ng/u1	0.00
73) Anthracene-d10	17.510	188	5172742	33.338	ng/u1	0.00
81) Pyrene-d10	19.739	212	5394962	37.388	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.851	264	3289818	34.567	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.434	88	153125	10.446	ng/uL	96
5) Pyridine	3.840	79	1072048	26.822	ng/u1	99
6) Benzaldehyde	7.151	77	800516	34.112	ng/u1	99
8) Phenol	7.199	94	1493827	29.641	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.440	93	1203613	29.148	ng/u1	99
12) 2-Chlorophenol	7.581	128	1181543	29.659	ng/u1	99
13) 2-Methylphenol	8.457	108	1156297	29.521	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.551	45	1695129	28.460	ng/u1	99
16) Acetophenone	8.851	105	1833814	29.878	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.840	70	906826	29.992	ng/u1	98
18) 4-Methylphenol	8.793	108	1288041	29.859	ng/u1	95
19) Hexachloroethane	9.110	117	446758	29.243	ng/u1	98
22) Nitrobenzene	9.228	77	1294715	30.848	ng/u1	99
23) Isophorone	9.757	82	2724710	29.448	ng/u1	99
25) 2-Nitrophenol	9.945	139	693642	34.009	ng/u1	96
26) 2,4-Dimethylphenol	9.998	107	1314540	29.005	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.240	93	1675700	29.539	ng/u1	100
29) 2,4-Dichlorophenol	10.475	162	1247590	30.376	ng/u1	99
30) Naphthalene	10.892	128	3929373	28.701	ng/u1	100
32) 4-Chloroaniline	10.992	127	1547139	27.748	ng/u1	99
33) Hexachlorobutadiene	11.175	225	806404	29.478	ng/u1	99
34) Caprolactam	11.763	113	389612m	30.892	ng/u1	
35) 4-Chloro-3-methylphenol	12.104	107	1282639	30.840	ng/u1	98
36) 2-Methylnaphthalene	12.492	142	2780319	29.491	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112522\
 Data File : BP012753.D
 Acq On : 26 Nov 2022 06:47
 Operator : CG/JU
 Sample : PB149119BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS119

Quant Time: Nov 26 08:43:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 25 15:59:22 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/26/2022
 Supervised By :mohammad ahmed 11/30/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.710	142	2780301	29.425	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.851	216	1606328	29.120	ng/u1	99
40) Hexachlorocyclopentadiene	12.839	237	817928	30.773	ng/u1	99
41) 2,4,6-Trichlorophenol	13.086	196	1063495	30.813	ng/u1	99
42) 2,4,5-Trichlorophenol	13.157	196	1148281	30.948	ng/u1	98
43) 1,1'-Biphenyl	13.492	154	3719586	29.517	ng/u1	99
44) 2-Chloronaphthalene	13.533	162	2938182	29.145	ng/u1	99
45) 2-Nitroaniline	13.733	65	719866	36.496	ng/u1	99
47) Dimethylphthalate	14.110	163	3812418	30.545	ng/u1	99
48) 2,6-Dinitrotoluene	14.228	165	728690	37.305	ng/u1	100
50) Acenaphthylene	14.381	152	4493358	29.578	ng/u1	100
51) 3-Nitroaniline	14.557	138	706770	31.243	ng/u1	95
52) Acenaphthene	14.722	153	3108368	29.063	ng/u1	99
53) 2,4-Dinitrophenol	14.757	184	385638	30.724	ng/u1	96
55) 4-Nitrophenol	14.851	109	461091	29.896	ng/u1	98
56) Dibenzofuran	15.057	168	4377474	29.694	ng/u1	100
57) 2,4-Dinitrotoluene	15.010	165	1039226	36.053	ng/u1	100
58) 2,3,4,6-Tetrachlorophenol	15.275	232	1003557	31.381	ng/u1	99
59) Diethylphthalate	15.475	149	3691903	30.558	ng/u1	99
61) Fluorene	15.704	166	3508994	29.682	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.698	204	1841296	30.018	ng/u1	99
63) 4-Nitroaniline	15.722	138	669228	34.710	ng/u1	99
66) 4,6-Dinitro-2-methylph...	15.775	198	589329	32.589	ng/u1	95
67) N-Nitrosodiphenylamine	15.910	169	3083429	30.780	ng/u1	100
68) 4-Bromophenyl-phenylether	16.598	248	1062931	30.639	ng/u1	99
69) Hexachlorobenzene	16.710	284	1376987	31.142	ng/u1	99
70) Atrazine	16.863	200	1087756	31.648	ng/u1	99
71) Pentachlorophenol	17.051	266	814510	29.996	ng/u1	98
72) Phenanthrene	17.457	178	5626788	29.952	ng/u1	99
74) Anthracene	17.545	178	5637997	30.336	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.457	216	1611439	29.571	ng/uL	100
76) Pentachlorobenzene	14.975	250	1556159	27.471	ng/uL	99
77) Carbazole	17.810	167	4924278	31.320	ng/u1	99
78) Di-n-butylphthalate	18.357	149	5867388	31.524	ng/u1	99
80) Fluoranthene	19.410	202	6093797	33.952	ng/u1	100
82) Pyrene	19.768	202	6137904	33.064	ng/u1	97
83) Butylbenzylphthalate	20.633	149	1691965	38.949	ng/u1	95
84) 3,3'-Dichlorobenzidine	21.404	252	1358158	29.781	ng/u1	99
85) Benzo(a)anthracene	21.480	228	4908574	28.578	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.398	149	2359579	35.518	ng/u1	99
87) Chrysene	21.539	228	4589589	27.636	ng/u1	100
89) Di-n-octyl phthalate	22.368	149	3323773	30.039	ng/u1	100
90) Benzo(b)fluoranthene	23.239	252	3962347	27.616	ng/u1	99
91) Benzo(k)fluoranthene	23.292	252	4121184	28.775	ng/u1	99
93) Benzo(a)pyrene	23.904	252	3456801	29.299	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.650	276	4180482	36.726	ng/u1	99
95) Dibenzo(a,h)anthracene	26.674	278	3666352	35.496	ng/u1	98
96) Benzo(g,h,i)perylene	27.462	276	3642405	37.202	ng/u1	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022


Quant Time: Nov 26 08:43:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 25 15:59:22 2022
Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 10.446 ng/uL
RT: 3.434 min Scan# 41
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion: 88 Resp: 153125
Ion Ratio Lower Upper
88 100
43 27.7 22.6 34.0
58 70.6 59.6 89.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#5
Pyridine
Concen: 26.822 ng/uL
RT: 3.840 min Scan# 111
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion: 79 Resp: 1072048
Ion Ratio Lower Upper
79 100
52 74.1 58.9 88.3
51 34.3 27.3 40.9





#6
Benzaldehyde
Concen: 34.112 ng/ul
RT: 7.151 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion: 77 Resp: 800516
Ion Ratio Lower Upper
77 100
105 103.1 81.4 122.2
106 97.1 77.4 116.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#8
Phenol
Concen: 29.641 ng/ul
RT: 7.199 min Scan# 682
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion: 94 Resp: 1493827
Ion Ratio Lower Upper
94 100
65 26.1 20.9 31.3
66 32.0 26.0 39.0





#10
Bis(2-Chloroethyl)ether
Concen: 29.148 ng/ul
RT: 7.440 min Scan# 723
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion: 93 Resp: 1203613
Ion Ratio Lower Upper
93 100
63 73.4 59.7 89.5
95 32.0 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#12
2-Chlorophenol
Concen: 29.659 ng/ul
RT: 7.581 min Scan# 747
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47



Tgt Ion:128 Resp: 1181543
Ion Ratio Lower Upper
128 100
64 43.8 35.6 53.4
130 32.4 26.0 39.0





#13

2-Methylphenol

Concen: 29.521 ng/ul

RT: 8.457 min Scan# 896

Delta R.T. -0.000 min

Lab File: BP012753.D

Acq: 26 Nov 2022 06:47

Instrument :

BNA_P

ClientSampleId :

SLCS119

Tgt Ion:108 Resp: 115629

Ion Ratio Lower Upper





108 100

107 88.7 69.3 103.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 28.460 ng/ul

RT: 8.551 min Scan# 912

Delta R.T. -0.006 min

Lab File: BP012753.D

Acq: 26 Nov 2022 06:47




Tgt Ion: 45 Resp: 1695129


Ion Ratio Lower Upper

45 100

77 13.7 11.3 16.9


79 10.7 8.6 12.8





#16
Acetophenone
Concen: 29.878 ng/ul
RT: 8.851 min Scan# 963
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:105 Resp: 1833814
Ion Ratio Lower Upper
105 100
77 72.8 59.7 89.5
51 24.7 20.9 31.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#17
N-Nitroso-di-n-propylamine
Concen: 29.992 ng/ul
RT: 8.840 min Scan# 961
Delta R.T. 0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47

Tgt Ion: 70 Resp: 906826
Ion Ratio Lower Upper
70 100
42 56.9 46.1 69.1
101 9.9 7.4 11.2
130 22.0 15.8 23.6









#18
4-Methylphenol
Concen: 29.859 ng/ul
RT: 8.793 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47

Instrument :
BNA_P
ClientSampleId :
SLCS119

Tgt Ion:108 Resp: 128804
Ion Ratio Lower Upper
108 100
107 107.8 90.2 135.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#19
Hexachloroethane
Concen: 29.243 ng/ul
RT: 9.110 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:117 Resp: 446758
Ion Ratio Lower Upper
117 100
201 106.0 85.7 128.5
199 68.0 52.9 79.3





#22
Nitrobenzene
Concen: 30.848 ng/ul
RT: 9.228 min Scan# 1027
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion: 77 Resp: 1294715
Ion Ratio Lower Upper
77 100
123 48.2 38.2 57.2
65 12.7 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#23
Isophorone
Concen: 29.448 ng/ul
RT: 9.757 min Scan# 1117
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47

Tgt Ion: 82 Resp: 2724710
Ion Ratio Lower Upper
82 100
95 6.7 5.2 7.8
138 18.6 14.5 21.7









#25
2-Nitrophenol
Concen: 34.009 ng/ul
RT: 9.945 min Scan# 1149
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47

Instrument :
BNA_P
ClientSampleId :
SLCS119

Tgt Ion:139 Resp: 693642
Ion Ratio Lower Upper
139 100
65 46.0 39.5 59.3
109 25.4 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#26
2,4-Dimethylphenol
Concen: 29.005 ng/ul
RT: 9.998 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:107 Resp: 1314540
Ion Ratio Lower Upper
107 100
121 52.2 41.8 62.6
122 90.6 72.3 108.5





#27
Bis(2-Chloroethoxy)methane
Concen: 29.539 ng/ul
RT: 10.240 min Scan# 1199
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion: 93 Resp: 1675700
Ion Ratio Lower Upper
93 100
95 32.5 26.1 39.1
123 11.2 9.0 13.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#29
2,4-Dichlorophenol
Concen: 30.376 ng/ul
RT: 10.475 min Scan# 1239
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion:162 Resp: 1247590
Ion Ratio Lower Upper
162 100
164 65.3 52.9 79.3
98 33.4 26.6 40.0





#30
Naphthalene
Concen: 28.701 ng/ul
RT: 10.892 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:128 Resp: 3929373
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.1 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#32
4-Chloroaniline
Concen: 27.748 ng/ul
RT: 10.992 min Scan# 1327
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:127 Resp: 1547139
Ion Ratio Lower Upper
127 100
129 32.5 25.7 38.5





#33
Hexachlorobutadiene
Concen: 29.478 ng/ul
RT: 11.175 min Scan# 1358
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:225 Resp: 806404
Ion Ratio Lower Upper
225 100
223 61.8 49.6 74.4
227 62.7 51.0 76.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#34
Caprolactam
Concen: 30.892 ng/ul m
RT: 11.763 min Scan# 1458
Delta R.T. 0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:113 Resp: 389612
Ion Ratio Lower Upper
113 100
55 169.9 145.9 218.9
56 126.9 107.1 160.7





#35
4-Chloro-3-methylphenol
Concen: 30.840 ng/ul
RT: 12.104 min Scan# 1516
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:107 Resp: 1282639
Ion Ratio Lower Upper
107 100
144 27.4 21.4 32.2
142 85.4 66.6 100.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#36
2-Methylnaphthalene
Concen: 29.491 ng/ul
RT: 12.492 min Scan# 1582
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:142 Resp: 2780319
Ion Ratio Lower Upper
142 100
141 90.1 70.2 105.2





#37
1-Methylnaphthalene
Concen: 29.425 ng/ul
RT: 12.710 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:142 Resp: 2780301
Ion Ratio Lower Upper
142 100
141 92.2 73.3 109.9
116 3.5 2.9 4.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.120 ng/ul
RT: 12.851 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:216 Resp: 1606328
Ion Ratio Lower Upper
216 100
214 79.1 63.0 94.6
179 19.6 16.2 24.4
108 15.3 12.7 19.1





#40
Hexachlorocyclopentadiene
Concen: 30.773 ng/ul
RT: 12.839 min Scan# 1641
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:237 Resp: 817928
Ion Ratio Lower Upper
237 100
235 62.5 50.2 75.4
272 13.2 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#41
2,4,6-Trichlorophenol
Concen: 30.813 ng/ul
RT: 13.086 min Scan# 1683
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:196 Resp: 1063495
Ion Ratio Lower Upper
196 100
198 96.1 76.5 114.7
200 31.0 25.3 37.9





#42
2,4,5-Trichlorophenol
Concen: 30.948 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:196 Resp: 1148281
Ion Ratio Lower Upper
196 100
198 95.4 77.6 116.4
200 31.0 24.2 36.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#43
1,1'-Biphenyl
Concen: 29.517 ng/ul
RT: 13.492 min Scan# 1752
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:154 Resp: 3719586
Ion Ratio Lower Upper
154 100
153 39.5 32.0 48.0
76 12.3 10.2 15.4





#44
2-Chloronaphthalene
Concen: 29.145 ng/ul
RT: 13.533 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:162 Resp: 2938182
Ion Ratio Lower Upper
162 100
164 32.4 26.5 39.7
127 36.1 28.8 43.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#45
2-Nitroaniline
Concen: 36.496 ng/ul
RT: 13.733 min Scan# 1793
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion: 65 Resp: 719866
Ion Ratio Lower Upper
65 100
92 71.1 57.8 86.8
138 116.8 94.6 142.0





#47
Dimethylphthalate
Concen: 30.545 ng/ul
RT: 14.110 min Scan# 1857
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:163 Resp: 3812418
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.4 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#48
2,6-Dinitrotoluene
Concen: 37.305 ng/ul
RT: 14.228 min Scan# 1877
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:165 Resp: 728690
Ion Ratio Lower Upper
165 100
89 53.3 42.7 64.1
121 20.7 16.9 25.3





#50
Acenaphthylene
Concen: 29.578 ng/ul
RT: 14.381 min Scan# 1903
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:152 Resp: 4493358
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9
153 13.4 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#51
3-Nitroaniline
Concen: 31.243 ng/ul
RT: 14.557 min Scan# 1933
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion:138 Resp: 706770
Ion Ratio Lower Upper
138 100
108 10.5 9.9 14.9
92 106.9 90.2 135.2





#52
Acenaphthene
Concen: 29.063 ng/ul
RT: 14.722 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:153 Resp: 3108368
Ion Ratio Lower Upper
153 100
152 47.5 38.2 57.4
154 88.2 69.8 104.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#53
2,4-Dinitrophenol
Concen: 30.724 ng/ul
RT: 14.757 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion:184 Resp: 385638
Ion Ratio Lower Upper
184 100
63 56.8 48.1 72.1
154 56.2 47.4 71.2





#55
4-Nitrophenol
Concen: 29.896 ng/ul
RT: 14.851 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:109 Resp: 461091
Ion Ratio Lower Upper
109 100
139 152.6 119.2 178.8
65 136.0 110.7 166.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#56
Dibenzofuran
Concen: 29.694 ng/ul
RT: 15.057 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:168 Resp: 4377474
Ion Ratio Lower Upper
168 100
139 34.3 27.7 41.5





#57
2,4-Dinitrotoluene
Concen: 36.053 ng/ul
RT: 15.010 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:165 Resp: 1039226
Ion Ratio Lower Upper
165 100
63 38.9 31.0 46.4
182 2.5 2.0 3.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 31.381 ng/ul
RT: 15.275 min Scan# 2055
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion:232 Resp: 1003557
Ion Ratio Lower Upper
232 100
131 39.4 32.0 48.0
130 2.1 1.5 2.3
166 28.9 23.2 34.8





#59
Diethylphthalate
Concen: 30.558 ng/ul
RT: 15.475 min Scan# 2089
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:149 Resp: 3691903
Ion Ratio Lower Upper
149 100
177 23.3 18.1 27.1
150 12.5 9.8 14.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#61
Fluorene
Concen: 29.682 ng/ul
RT: 15.704 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion:166 Resp: 3508994
Ion Ratio Lower Upper
166 100
165 96.2 78.2 117.2
167 13.7 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 30.018 ng/ul
RT: 15.698 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:204 Resp: 1841296
Ion Ratio Lower Upper
204 100
206 33.3 26.0 39.0
141 50.7 40.7 61.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#63
4-Nitroaniline
Concen: 34.710 ng/ul
RT: 15.722 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:138 Resp: 669228
Ion Ratio Lower Upper
138 100
92 47.9 38.6 57.8
108 68.9 56.4 84.6





#66
4,6-Dinitro-2-methylphenol
Concen: 32.589 ng/ul
RT: 15.775 min Scan# 2140
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:198 Resp: 589329
Ion Ratio Lower Upper
198 100
182 4.2 4.1 6.1
77 20.9 18.9 28.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#67
N-Nitrosodiphenylamine
Concen: 30.780 ng/ul
RT: 15.910 min Scan# 2163
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:169 Resp: 3083429
Ion Ratio Lower Upper
169 100
168 66.6 53.2 79.8
167 36.5 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 30.639 ng/ul
RT: 16.598 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:248 Resp: 1062931
Ion Ratio Lower Upper
248 100
250 98.0 78.6 118.0
141 59.4 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#69
Hexachlorobenzene
Concen: 31.142 ng/ul
RT: 16.710 min Scan# 2299
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:284 Resp: 1376987
Ion Ratio Lower Upper
284 100
142 30.0 25.0 37.6
249 30.5 24.6 36.8





#70
Atrazine
Concen: 31.648 ng/ul
RT: 16.863 min Scan# 2325
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119



Tgt Ion:200 Resp: 1087756
Ion Ratio Lower Upper
200 100
173 26.3 20.3 30.5
215 49.7 40.1 60.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#71
Pentachlorophenol
Concen: 29.996 ng/ul
RT: 17.051 min Scan# 2357
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:266 Resp: 814510
Ion Ratio Lower Upper
266 100
264 62.7 50.7 76.1
268 63.1 51.8 77.8





#72
Phenanthrene
Concen: 29.952 ng/ul
RT: 17.457 min Scan# 2426
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:178 Resp: 5626788
Ion Ratio Lower Upper
178 100
179 15.7 12.3 18.5
176 19.4 15.4 23.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#74
Anthracene
Concen: 30.336 ng/ul
RT: 17.545 min Scan# 2441
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion:178 Resp: 5637997
Ion Ratio Lower Upper
178 100
179 15.9 12.4 18.6
176 18.9 14.8 22.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 29.571 ng/uL
RT: 13.457 min Scan# 1746
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:216 Resp: 1611439
Ion Ratio Lower Upper
216 100
214 78.1 62.3 93.5
218 48.8 39.1 58.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#76
Pentachlorobenzene
Concen: 27.471 ng/uL
RT: 14.975 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion:250 Resp: 1556159
Ion Ratio Lower Upper
250 100
248 64.0 50.6 76.0
252 64.5 50.5 75.7





#77
 Carbazole
 Concen: 31.320 ng/ul
 RT: 17.810 min Scan# 2486
 Delta R.T. -0.006 min
 Lab File: BP012753.D
 Acq: 26 Nov 2022 06:47




Instrument : BNA_P
 ClientSampleId : SLCS119




Tgt Ion:167 Resp: 4924278
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 12.1 9.8 14.8

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
 Supervised By :mohammad ahmed 11/30/2022




#78
 Di-n-butylphthalate
 Concen: 31.524 ng/ul
 RT: 18.357 min Scan# 2579
 Delta R.T. -0.012 min
 Lab File: BP012753.D
 Acq: 26 Nov 2022 06:47




Tgt Ion:149 Resp: 5867388
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.2 10.8
 104 4.6 3.7 5.5





#80
Fluoranthene
Concen: 33.952 ng/ul
RT: 19.410 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:202 Resp: 609379
Ion Ratio Lower Upper
202 100
101 11.1 8.7 13.1
100 8.4 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#82
Pyrene
Concen: 33.064 ng/ul
RT: 19.768 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47



Tgt Ion:202 Resp: 6137904
Ion Ratio Lower Upper
202 100
101 11.8 10.3 15.5
100 9.5 8.6 13.0





#83

Butylbenzylphthalate

Concen: 38.949 ng/ul

RT: 20.633 min Scan# 2966

Delta R.T. -0.006 min

Lab File: BP012753.D

Acq: 26 Nov 2022 06:47

Instrument :

BNA_P

ClientSampleId :

SLCS119

Tgt Ion:149 Resp: 1691965

Ion Ratio Lower Upper

149 100





91 68.3 58.2 87.2

206 21.5 16.5 24.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#84

3,3'-Dichlorobenzidine

Concen: 29.781 ng/ul

RT: 21.404 min Scan# 3097

Delta R.T. -0.012 min

Lab File: BP012753.D

Acq: 26 Nov 2022 06:47




Tgt Ion:252 Resp: 1358158


Ion Ratio Lower Upper

252 100

254 64.6 52.3 78.5


126 10.8 7.6 11.4





#85
Benzo(a)anthracene
Concen: 28.578 ng/ul
RT: 21.480 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument : BNA_P
ClientSampleId : SLCS119



Tgt Ion:228 Resp: 4908574
Ion Ratio Lower Upper
228 100
229 20.2 16.2 24.2
226 27.8 22.2 33.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 35.518 ng/ul
RT: 21.398 min Scan# 3096
Delta R.T. -0.012 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47

Tgt Ion:149 Resp: 2359579
Ion Ratio Lower Upper
149 100
167 28.1 22.7 34.1
279 4.3 3.8 5.6





#87

Chrysene

Concen: 27.636 ng/ul

RT: 21.539 min Scan# 3119

Delta R.T. -0.006 min

Lab File: BP012753.D

Acq: 26 Nov 2022 06:47

Instrument :

BNA_P

ClientSampleId :

SLCS119

Tgt Ion:228 Resp: 4589589

Ion Ratio Lower Upper

228 100





226 30.2 24.3 36.5

229 19.8 15.9 23.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#89

Di-n-octyl phthalate

Concen: 30.039 ng/ul




RT: 22.368 min Scan# 3261


Delta R.T. -0.012 min

Lab File: BP012753.D

Acq: 26 Nov 2022 06:47

Tgt Ion:149 Resp: 3323773





#90
Benzo(b)fluoranthene
Concen: 27.616 ng/ul
RT: 23.239 min Scan# 3409
Delta R.T. -0.012 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47

Instrument :

BNA_P

ClientSampleId :





SLCS119

Tgt Ion:252 Resp: 3962347
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 9.8 7.4 11.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#91
Benzo(k)fluoranthene
Concen: 28.775 ng/ul
RT: 23.292 min Scan# 3418
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47


Tgt Ion:252 Resp: 4121184
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.8
125 9.3 7.8 11.8





#93
Benzo(a)pyrene
Concen: 29.299 ng/ul
RT: 23.904 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:252 Resp: 345680
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 10.1 8.5 12.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 36.726 ng/ul
RT: 26.650 min Scan# 3989
Delta R.T. -0.018 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Tgt Ion:276 Resp: 4180482
Ion Ratio Lower Upper
276 100
138 22.1 18.3 27.5
277 25.2 19.8 29.8





#95
Dibenzo(a,h)anthracene
Concen: 35.496 ng/ul
RT: 26.674 min Scan# 3993
Delta R.T. -0.018 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47




Instrument :
BNA_P
ClientSampleId :
SLCS119




Tgt Ion:278 Resp: 3666352
Ion Ratio Lower Upper
278 100
139 14.7 12.6 18.8
279 23.4 19.4 29.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/26/2022
Supervised By :mohammad ahmed 11/30/2022



#96
Benzo(g,h,i)perylene
Concen: 37.202 ng/ul
RT: 27.462 min Scan# 4127
Delta R.T. -0.012 min
Lab File: BP012753.D
Acq: 26 Nov 2022 06:47



Tgt Ion:276 Resp: 3642405
Ion Ratio Lower Upper
276 100
138 19.7 16.7 25.1
277 24.4 19.1 28.7

