

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120122\
 Data File : BP012921.D
 Acq On : 01 Dec 2022 17:31
 Operator : CG/JU
 Sample : SSTD00505
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005624

Quant Time: Dec 01 22:55:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 22:54:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	415128	20.000	ng/ul	0.00
20) Naphthalene-d8	10.822	136	1955942	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.640	164	1464817	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.398	188	3478268	20.000	ng/ul	0.00
79) Chrysene-d12	21.480	240	3930106	20.000	ng/ul	0.00
88) Perylene-d12	23.992	264	4224580	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	17468	1.803	ng/uL	0.00
4) Pyridine-d5	3.823	84	127553	4.400	ng/ul	0.00
7) Phenol-d5	7.158	99	153353	4.496	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	98287	4.676	ng/ul	0.00
11) 2-Chlorophenol-d4	7.534	132	122121	4.698	ng/ul	0.00
15) 4-Methylphenol-d8	8.711	113	129970	4.729	ng/ul	0.00
21) Nitrobenzene-d5	9.169	128	58433	4.881	ng/ul	0.00
24) 2-Nitrophenol-d4	9.899	143	63875	5.025	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	138327	4.664	ng/ul	0.00
31) 4-Chloroaniline-d4	10.952	131	204102	5.025	ng/ul	0.00
46) Dimethylphthalate-d6	14.046	166	486995	4.830	ng/ul	0.00
49) Acenaphthylene-d8	14.334	160	508397	4.322	ng/ul	0.00
54) 4-Nitrophenol-d4	14.822	143	85383	4.750	ng/ul	0.00
60) Fluorene-d10	15.634	176	433315	4.907	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	64246	3.865	ng/ul	0.00
73) Anthracene-d10	17.492	188	711690	4.648	ng/ul	0.00
81) Pyrene-d10	19.722	212	878950	3.849	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.822	264	906314	4.368	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	19109	1.837	ng/uL#	53
5) Pyridine	3.846	79	127925	4.510	ng/ul	94
6) Benzaldehyde	7.140	77	87005	5.224	ng/ul	95
8) Phenol	7.181	94	165450	4.625	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.422	93	136183	4.647	ng/ul	97
12) 2-Chlorophenol	7.564	128	129768	4.590	ng/ul	99
13) 2-Methylphenol	8.446	108	126225	4.541	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.540	45	194426	4.599	ng/ul	97
16) Acetophenone	8.834	105	215973	4.958	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.811	70	109187	5.088	ng/ul	97
18) 4-Methylphenol	8.769	108	142050	4.640	ng/ul	99
19) Hexachloroethane	9.093	117	49255	4.543	ng/ul	96
22) Nitrobenzene	9.216	77	144926	4.616	ng/ul	98
23) Isophorone	9.740	82	318136	4.596	ng/ul	99
25) 2-Nitrophenol	9.928	139	72861	4.775	ng/ul	97
26) 2,4-Dimethylphenol	9.981	107	156116	4.605	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.222	93	200288	4.720	ng/ul	99
29) 2,4-Dichlorophenol	10.463	162	140939	4.587	ng/ul	100
30) Naphthalene	10.869	128	478948	4.676	ng/ul	98
32) 4-Chloroaniline	10.981	127	213850	5.127	ng/ul	98
33) Hexachlorobutadiene	11.157	225	94595	4.622	ng/ul	99
34) Caprolactam	11.752	113	52159	5.528	ng/ul	95
35) 4-Chloro-3-methylphenol	12.087	107	157931	5.076	ng/ul	98
36) 2-Methylnaphthalene	12.475	142	346269	4.910	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120122\
 Data File : BP012921.D
 Acq On : 01 Dec 2022 17:31
 Operator : CG/JU
 Sample : SSTD00505
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD005624


Quant Time: Dec 01 22:55:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 22:54:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.693	142	348401	4.929	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.840	216	201255	4.309	ng/ul	96
40) Hexachlorocyclopentadiene	12.822	237	87175	3.873	ng/ul	99
41) 2,4,6-Trichlorophenol	13.075	196	134709	4.609	ng/ul	98
42) 2,4,5-Trichlorophenol	13.140	196	140983	4.487	ng/ul	95
43) 1,1'-Biphenyl	13.475	154	489370	4.586	ng/ul	99
44) 2-Chloronaphthalene	13.516	162	383435	4.492	ng/ul	98
45) 2-Nitroaniline	13.722	65	90499	5.418	ng/ul	95
47) Dimethylphthalate	14.093	163	511185	4.837	ng/ul	100
48) 2,6-Dinitrotoluene	14.210	165	75557	4.568	ng/ul#	87
50) Acenaphthylene	14.363	152	580792	4.515	ng/ul	99
51) 3-Nitroaniline	14.540	138	84052	4.388	ng/ul#	95
52) Acenaphthene	14.704	153	422055	4.660	ng/ul	99
53) 2,4-Dinitrophenol	14.740	184	41333	3.889	ng/ul	91
55) 4-Nitrophenol	14.840	109	67580	5.175	ng/ul	90
56) Dibenzofuran	15.040	168	606888	4.862	ng/ul	98
57) 2,4-Dinitrotoluene	14.998	165	115959	4.751	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.263	232	137184	5.066	ng/ul	96
59) Diethylphthalate	15.457	149	511809	5.003	ng/ul	99
61) Fluorene	15.687	166	510505	5.100	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.681	204	259220	4.991	ng/ul	97
63) 4-Nitroaniline	15.704	138	74692	4.575	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.757	198	73718	4.131	ng/ul#	91
67) N-Nitrosodiphenylamine	15.893	169	437081	4.422	ng/ul	99
68) 4-Bromophenyl-phenylether	16.581	248	140044	4.091	ng/ul	99
69) Hexachlorobenzene	16.698	284	192653	4.415	ng/ul	98
70) Atrazine	16.845	200	164543	4.851	ng/ul	100
71) Pentachlorophenol	17.040	266	110981	4.142	ng/ul	99
72) Phenanthrene	17.439	178	866877	4.676	ng/ul	99
74) Anthracene	17.528	178	863116	4.706	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.446	216	207784	3.864	ng/uL	100
76) Pentachlorobenzene	14.957	250	235570	4.214	ng/uL	100
77) Carbazole	17.798	167	802479	5.172	ng/ul	99
78) Di-n-butylphthalate	18.339	149	1249685	6.804	ng/ul	99
80) Fluoranthene	19.398	202	1066714	3.755	ng/ul	99
82) Pyrene	19.751	202	1137764	3.873	ng/ul	100
83) Butylbenzylphthalate	20.616	149	422517	6.146	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.392	252	310057	4.296	ng/ul	99
85) Benzo(a)anthracene	21.463	228	1167997	4.297	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.375	149	626348	5.958	ng/ul	99
87) Chrysene	21.522	228	1106366	4.210	ng/ul	99
89) Di-n-octyl phthalate	22.339	149	1051825	4.360	ng/ul	100
90) Benzo(b)fluoranthene	23.216	252	1230807	3.934	ng/ul	99
91) Benzo(k)fluoranthene	23.269	252	1158964	3.711	ng/ul	99
93) Benzo(a)pyrene	23.874	252	1074368	4.176	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.610	276	1254532	5.055	ng/ul	99
95) Dibenzo(a,h)anthracene	26.633	278	1131506	5.024	ng/ul	99
96) Benzo(g,h,i)perylene	27.421	276	1075328	5.037	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 01 22:55:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 01 22:54:44 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 1.837 ng/uL
 RT: 3.428 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BP012921.D
 Acq: 01 Dec 2022 17:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005624


Tgt Ion: 88 Resp: 19109
 Ion Ratio Lower Upper
 88 100
 43 120.3 30.6 46.0#
 58 71.6 57.7 86.5



#5
 Pyridine
 Concen: 4.510 ng/uL
 RT: 3.846 min Scan# 112
 Delta R.T. 0.006 min
 Lab File: BP012921.D
 Acq: 01 Dec 2022 17:31

Tgt Ion: 79 Resp: 127925
 Ion Ratio Lower Upper
 79 100
 52 78.9 59.2 88.8
 51 38.9 27.8 41.6









#6
Benzaldehyde
Concen: 5.224 ng/ul
RT: 7.140 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion: 77 Resp: 87005
Ion Ratio Lower Upper
77 100
105 99.0 82.6 124.0
106 93.4 79.8 119.6



#8
Phenol
Concen: 4.625 ng/ul
RT: 7.181 min Scan# 679
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion: 94 Resp: 165450
Ion Ratio Lower Upper
94 100
65 26.5 21.8 32.6
66 33.5 26.6 39.8









#10
Bis(2-Chloroethyl)ether
Concen: 4.647 ng/ul
RT: 7.422 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion: 93 Resp: 136183
Ion Ratio Lower Upper
93 100
63 76.7 59.6 89.4
95 34.8 25.7 38.5



#12
2-Chlorophenol
Concen: 4.590 ng/ul
RT: 7.564 min Scan# 744
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion: 128 Resp: 129768
Ion Ratio Lower Upper
128 100
64 44.5 34.8 52.2
130 31.7 25.4 38.2





#13

2-Methylphenol

Concen: 4.541 ng/ul

RT: 8.446 min Scan# 894

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :





SSTD005624

Tgt Ion:108 Resp: 126225

Ion Ratio Lower Upper

108 100

107 88.2 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.599 ng/ul

RT: 8.540 min Scan# 910

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion: 45 Resp: 194426


Ion Ratio Lower Upper

45 100

77 12.7 11.6 17.4

79 11.2 8.9 13.3









#16
Acetophenone
Concen: 4.958 ng/ul
RT: 8.834 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:105 Resp: 215973
Ion Ratio Lower Upper
105 100
77 76.7 60.4 90.6
51 27.3 20.7 31.1



#17
N-Nitroso-di-n-propylamine
Concen: 5.088 ng/ul
RT: 8.811 min Scan# 956
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion: 70 Resp: 109187
Ion Ratio Lower Upper
70 100
42 59.4 46.0 69.0
101 9.5 7.4 11.0
130 19.5 17.6 26.4





#18

4-Methylphenol

Concen: 4.640 ng/ul

RT: 8.769 min Scan# 94

Delta R.T. -0.006 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :





SSTD005624

Tgt Ion:108 Resp: 142050

Ion Ratio Lower Upper

108 100

107 111.1 89.8 134.6



#19

Hexachloroethane

Concen: 4.543 ng/ul

RT: 9.093 min Scan# 1004

Delta R.T. -0.006 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion:117 Resp: 49255


Ion Ratio Lower Upper

117 100

201 105.1 88.2 132.4

199 69.1 57.4 86.0









#22
Nitrobenzene
Concen: 4.616 ng/ul
RT: 9.216 min Scan# 1025
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion: 77 Resp: 144926
Ion Ratio Lower Upper
77 100
123 45.9 37.6 56.4
65 13.3 9.9 14.9



#23
Isophorone
Concen: 4.596 ng/ul
RT: 9.740 min Scan# 1114
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion: 82 Resp: 318136
Ion Ratio Lower Upper
82 100
95 6.8 5.2 7.8
138 18.3 15.0 22.6









#25
2-Nitrophenol
Concen: 4.775 ng/ul
RT: 9.928 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:139 Resp: 72861
Ion Ratio Lower Upper
139 100
65 51.3 39.3 58.9
109 26.3 21.9 32.9



#26
2,4-Dimethylphenol
Concen: 4.605 ng/ul
RT: 9.981 min Scan# 1155
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:107 Resp: 156116
Ion Ratio Lower Upper
107 100
121 49.9 41.6 62.4
122 86.0 71.4 107.2









#27
Bis(2-Chloroethoxy)methane
Concen: 4.720 ng/ul
RT: 10.222 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Instrument :
BNA_P
ClientSampleId :
SSTD005624

Tgt Ion: 93 Resp: 200288




Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.6	38.4
123	11.7	9.1	13.7




#29
2,4-Dichlorophenol
Concen: 4.587 ng/ul
RT: 10.463 min Scan# 1237
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion: 162 Resp: 140939

Ion	Ratio	Lower	Upper
162	100		
164	65.7	52.2	78.4
98	34.1	27.3	40.9









#30
Naphthalene
Concen: 4.676 ng/ul
RT: 10.869 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:128 Resp: 478948
Ion Ratio Lower Upper
128 100
129 11.9 8.6 13.0
127 12.5 10.2 15.2



#32
4-Chloroaniline
Concen: 5.127 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:127 Resp: 213850
Ion Ratio Lower Upper
127 100
129 33.1 25.7 38.5









#33
Hexachlorobutadiene
Concen: 4.622 ng/ul
RT: 11.157 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:225 Resp: 94595
Ion Ratio Lower Upper
225 100
223 62.0 49.0 73.4
227 63.1 50.2 75.4



#34
Caprolactam
Concen: 5.528 ng/ul
RT: 11.752 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:113 Resp: 52159
Ion Ratio Lower Upper
113 100
55 171.7 129.8 194.6
56 122.9 96.7 145.1









#35
4-Chloro-3-methylphenol
Concen: 5.076 ng/ul
RT: 12.087 min Scan# 1514
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:107 Resp: 157931
Ion Ratio Lower Upper
107 100
144 26.6 21.7 32.5
142 86.2 67.1 100.7



#36
2-Methylnaphthalene
Concen: 4.910 ng/ul
RT: 12.475 min Scan# 1579
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:142 Resp: 346269
Ion Ratio Lower Upper
142 100
141 90.2 71.8 107.8









#37
1-Methylnaphthalene
Concen: 4.929 ng/ul
RT: 12.693 min Scan# 1616
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:142 Resp: 348401
Ion Ratio Lower Upper
142 100
141 91.8 74.2 111.4
116 3.5 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.309 ng/ul
RT: 12.840 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:216 Resp: 201255
Ion Ratio Lower Upper
216 100
214 75.2 64.2 96.4
179 19.4 15.4 23.2
108 14.0 11.7 17.5









#40
Hexachlorocyclopentadiene
Concen: 3.873 ng/ul
RT: 12.822 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:237 Resp: 87175
Ion Ratio Lower Upper
237 100
235 63.3 50.1 75.1
272 12.5 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 4.609 ng/ul
RT: 13.075 min Scan# 1681
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:196 Resp: 134709
Ion Ratio Lower Upper
196 100
198 94.6 76.6 114.8
200 29.4 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 4.487 ng/ul

RT: 13.140 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD005624





Tgt Ion:196 Resp: 140983

Ion Ratio Lower Upper

196 100

198 92.7 77.8 116.6

200 29.2 25.3 37.9



#43

1,1'-Biphenyl

Concen: 4.586 ng/ul

RT: 13.475 min Scan# 1749

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion:154 Resp: 489370


Ion Ratio Lower Upper

154 100

153 38.8 31.8 47.6

76 12.2 10.2 15.2





#44

2-Chloronaphthalene

Concen: 4.492 ng/ul

RT: 13.516 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD005624





Tgt Ion:162 Resp: 383435

Ion Ratio Lower Upper

162 100

164 32.7 26.6 39.8

127 37.1 28.5 42.7



#45

2-Nitroaniline

Concen: 5.418 ng/ul

RT: 13.722 min Scan# 1791

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion: 65 Resp: 90499


Ion Ratio Lower Upper

65 100

92 74.0 56.7 85.1

138 108.7 92.1 138.1









#47
Dimethylphthalate
Concen: 4.837 ng/ul
RT: 14.093 min Scan# 1854
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:163 Resp: 511185
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 4.568 ng/ul
RT: 14.210 min Scan# 1874
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:165 Resp: 75557
Ion Ratio Lower Upper
165 100
89 61.5 41.1 61.7
121 24.6 15.8 23.8









#50
Acenaphthylene
Concen: 4.515 ng/ul
RT: 14.363 min Scan# 1900
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:152 Resp: 580792
Ion Ratio Lower Upper
152 100
151 20.4 16.0 24.0
153 12.7 10.6 16.0



#51
3-Nitroaniline
Concen: 4.388 ng/ul
RT: 14.540 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:138 Resp: 84052
Ion Ratio Lower Upper
138 100
108 14.2 8.9 13.3#
92 114.6 88.2 132.2









#52
Acenaphthene
Concen: 4.660 ng/ul
RT: 14.704 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:153 Resp: 422055
Ion Ratio Lower Upper
153 100
152 49.2 38.5 57.7
154 88.8 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 3.889 ng/ul
RT: 14.740 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:184 Resp: 41333
Ion Ratio Lower Upper
184 100
63 66.6 45.4 68.2
154 60.7 45.7 68.5









#55
4-Nitrophenol
Concen: 5.175 ng/ul
RT: 14.840 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:109 Resp: 67580
Ion Ratio Lower Upper
109 100
139 142.3 124.2 186.4
65 127.8 112.9 169.3



#56
Dibenzofuran
Concen: 4.862 ng/ul
RT: 15.040 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:168 Resp: 606888
Ion Ratio Lower Upper
168 100
139 35.2 27.4 41.2









#57
2,4-Dinitrotoluene
Concen: 4.751 ng/ul
RT: 14.998 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:165 Resp: 115959
Ion Ratio Lower Upper
165 100
63 41.6 30.2 45.2
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 5.066 ng/ul
RT: 15.263 min Scan# 2053
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:232 Resp: 137184
Ion Ratio Lower Upper
232 100
131 41.0 31.0 46.6
130 2.2 1.8 2.6
166 31.1 22.5 33.7









#59
Diethylphthalate
Concen: 5.003 ng/ul
RT: 15.457 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:149 Resp: 511809
Ion Ratio Lower Upper
149 100
177 22.4 18.4 27.6
150 12.0 9.8 14.6



#61
Fluorene
Concen: 5.100 ng/ul
RT: 15.687 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:166 Resp: 510505
Ion Ratio Lower Upper
166 100
165 98.2 77.4 116.2
167 13.1 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 4.991 ng/ul
RT: 15.681 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:204 Resp: 259220
Ion Ratio Lower Upper
204 100
206 32.0 26.6 40.0
141 53.6 41.0 61.4



#63
4-Nitroaniline
Concen: 4.575 ng/ul
RT: 15.704 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:138 Resp: 74692
Ion Ratio Lower Upper
138 100
92 51.3 39.0 58.4
108 77.9 55.6 83.4









#66
4,6-Dinitro-2-methylphenol
Concen: 4.131 ng/ul
RT: 15.757 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:198 Resp: 73718
Ion Ratio Lower Upper
198 100
182 8.1 3.8 5.8#
77 23.6 15.6 23.4#



#67
N-Nitrosodiphenylamine
Concen: 4.422 ng/ul
RT: 15.893 min Scan# 2160
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:169 Resp: 437081
Ion Ratio Lower Upper
169 100
168 66.3 53.4 80.0
167 37.3 28.8 43.2









#68
4-Bromophenyl-phenylether
Concen: 4.091 ng/ul
RT: 16.581 min Scan# 2277
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:248 Resp: 140044
Ion Ratio Lower Upper
248 100
250 96.2 77.5 116.3
141 62.3 49.7 74.5



#69
Hexachlorobenzene
Concen: 4.415 ng/ul
RT: 16.698 min Scan# 2297
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:284 Resp: 192653
Ion Ratio Lower Upper
284 100
142 28.4 23.3 34.9
249 28.7 24.0 36.0









#70
Atrazine
Concen: 4.851 ng/ul
RT: 16.845 min Scan# 21
Delta R.T. -0.005 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:200 Resp: 164543
Ion Ratio Lower Upper
200 100
173 26.7 21.0 31.6
215 49.2 39.2 58.8



#71
Pentachlorophenol
Concen: 4.142 ng/ul
RT: 17.040 min Scan# 2355
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:266 Resp: 110981
Ion Ratio Lower Upper
266 100
264 61.0 50.5 75.7
268 63.9 51.2 76.8









#72
Phenanthrene
Concen: 4.676 ng/ul
RT: 17.439 min Scan# 2423
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:178 Resp: 866877
Ion Ratio Lower Upper
178 100
179 15.1 12.6 18.8
176 19.1 15.4 23.0



#74
Anthracene
Concen: 4.706 ng/ul
RT: 17.528 min Scan# 2438
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:178 Resp: 863116
Ion Ratio Lower Upper
178 100
179 15.4 12.6 19.0
176 18.2 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 3.864 ng/uL

RT: 13.446 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD005624





Tgt Ion:216 Resp: 207784

Ion Ratio Lower Upper

216 100

214 77.8 62.2 93.2

218 48.5 39.4 59.2



#76

Pentachlorobenzene

Concen: 4.214 ng/uL

RT: 14.957 min Scan# 2001

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion:250 Resp: 235570


Ion Ratio Lower Upper

250 100

248 62.4 49.4 74.2

252 63.3 50.6 75.8









#77
 Carbazole
 Concen: 5.172 ng/ul
 RT: 17.798 min Scan# 2484
 Delta R.T. 0.000 min
 Lab File: BP012921.D
 Acq: 01 Dec 2022 17:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005624


Tgt Ion:167 Resp: 802479
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.2 25.8
 139 12.4 9.7 14.5



#78
 Di-n-butylphthalate
 Concen: 6.804 ng/ul
 RT: 18.339 min Scan# 2576
 Delta R.T. 0.000 min
 Lab File: BP012921.D
 Acq: 01 Dec 2022 17:31

Tgt Ion:149 Resp: 1249685
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 4.7 3.7 5.5





#80

Fluoranthene

Concen: 3.755 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD005624





Tgt Ion:202 Resp: 1066714

Ion Ratio Lower Upper

202 100

101 10.4 7.9 11.9

100 7.9 6.2 9.4



#82

Pyrene

Concen: 3.873 ng/ul

RT: 19.751 min Scan# 2816

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion:202 Resp: 1137764


Ion Ratio Lower Upper

202 100

101 12.3 9.7 14.5

100 9.6 7.7 11.5





#83

Butylbenzylphthalate

Concen: 6.146 ng/ul

RT: 20.616 min Scan# 2963

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD005624





Tgt Ion:149 Resp: 422517

Ion Ratio Lower Upper

149 100

91 70.5 55.0 82.6

206 21.4 17.8 26.8



#84

3,3'-Dichlorobenzidine

Concen: 4.296 ng/ul

RT: 21.392 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion:252 Resp: 310057


Ion Ratio Lower Upper

252 100

254 64.2 51.9 77.9

126 11.0 8.0 12.0





#85

Benzo(a)anthracene

Concen: 4.297 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD005624





Tgt Ion:228 Resp: 1167997

Ion Ratio Lower Upper

228 100

229 19.9 16.2 24.2

226 26.7 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 5.958 ng/ul

RT: 21.375 min Scan# 3092

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion:149 Resp: 626348


Ion Ratio Lower Upper

149 100

167 27.9 22.6 33.8

279 4.0 3.5 5.3





#87

Chrysene

Concen: 4.210 ng/ul

RT: 21.522 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD005624





Tgt Ion:228 Resp: 1106366

Ion Ratio Lower Upper

228 100

226 29.8 24.3 36.5

229 19.3 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 4.360 ng/ul




RT: 22.339 min Scan# 3256


Delta R.T. 0.000 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Tgt Ion:149 Resp: 1051825





#90

Benzo(b)fluoranthene

Concen: 3.934 ng/ul

RT: 23.216 min Scan# 3406

Delta R.T. -0.006 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD005624





Tgt Ion:252 Resp: 1230807

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.6

125 9.5 7.1 10.7



#91

Benzo(k)fluoranthene

Concen: 3.711 ng/ul

RT: 23.269 min Scan# 3414

Delta R.T. -0.006 min

Lab File: BP012921.D

Acq: 01 Dec 2022 17:31




Tgt Ion:252 Resp: 1158964


Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 9.2 7.1 10.7









#93
Benzo(a)pyrene
Concen: 4.176 ng/ul
RT: 23.874 min Scan# 3517
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD005624


Tgt Ion:252 Resp: 1074368
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 10.8 8.2 12.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 5.055 ng/ul
RT: 26.610 min Scan# 3982
Delta R.T. -0.006 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:276 Resp: 1254532
Ion Ratio Lower Upper
276 100
138 21.7 17.9 26.9
277 24.9 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 5.024 ng/ul
RT: 26.633 min Scan# 3986
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Instrument :
BNA_P
ClientSampleId :
SSTD005624

Tgt Ion:278 Resp: 1131506
Ion Ratio Lower Upper
278 100
139 15.5 11.8 17.8
279 23.4 19.1 28.7



#96
Benzo(g,h,i)perylene
Concen: 5.037 ng/ul
RT: 27.421 min Scan# 4120
Delta R.T. 0.000 min
Lab File: BP012921.D
Acq: 01 Dec 2022 17:31

Tgt Ion:276 Resp: 1075328
Ion Ratio Lower Upper
276 100
138 20.9 15.9 23.9
277 23.1 19.2 28.8

