

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120122\
 Data File : BP012927.D
 Acq On : 01 Dec 2022 20:56
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV630

Quant Time: Dec 01 23:49:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 23:46:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	305738	20.000	ng/ul	0.00
20) Naphthalene-d8	10.822	136	1491478	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.640	164	1133823	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.392	188	2705982	20.000	ng/ul	0.00
79) Chrysene-d12	21.480	240	3139740	20.000	ng/ul	0.00
88) Perylene-d12	23.986	264	3505876	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	49974	7.273	ng/uL	0.00
4) Pyridine-d5	3.817	84	419992	18.942	ng/ul	0.00
7) Phenol-d5	7.158	99	561545	19.889	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	357121	22.341	ng/ul	0.00
11) 2-Chlorophenol-d4	7.534	132	422250	20.404	ng/ul	0.00
15) 4-Methylphenol-d8	8.710	113	476657	20.560	ng/ul	0.00
21) Nitrobenzene-d5	9.169	128	225740	21.283	ng/ul	0.00
24) 2-Nitrophenol-d4	9.899	143	258458	21.293	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	516490	21.255	ng/ul	0.00
31) 4-Chloroaniline-d4	10.952	131	724029	21.089	ng/ul	0.00
46) Dimethylphthalate-d6	14.045	166	1758125	21.499	ng/ul	0.00
49) Acenaphthylene-d8	14.334	160	1979221	21.920	ng/ul	0.00
54) 4-Nitrophenol-d4	14.828	143	365238	21.836	ng/ul	0.00
60) Fluorene-d10	15.634	176	1538194	21.400	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	323434	20.372	ng/ul	0.00
73) Anthracene-d10	17.492	188	2583884	21.603	ng/ul	0.00
81) Pyrene-d10	19.722	212	3186828	20.649	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.821	264	3571485	20.849	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	51920	7.093	ng/uL	94
5) Pyridine	3.834	79	382959	17.730	ng/ul	97
6) Benzaldehyde	7.140	77	358586	25.847	ng/ul	98
8) Phenol	7.181	94	561481	19.154	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.422	93	440177	19.466	ng/ul	100
12) 2-Chlorophenol	7.563	128	432710	19.539	ng/ul	99
13) 2-Methylphenol	8.446	108	444256	19.383	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.540	45	614916	19.407	ng/ul	99
16) Acetophenone	8.834	105	747634	20.821	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.816	70	375883	20.591	ng/ul	98
18) 4-Methylphenol	8.775	108	506248	19.779	ng/ul	96
19) Hexachloroethane	9.093	117	157572	19.267	ng/ul	97
22) Nitrobenzene	9.210	77	513993	20.090	ng/ul	99
23) Isophorone	9.740	82	1090059	19.945	ng/ul	99
25) 2-Nitrophenol	9.928	139	279927	20.500	ng/ul	98
26) 2,4-Dimethylphenol	9.981	107	546744	20.388	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.222	93	664232	20.070	ng/ul	99
29) 2,4-Dichlorophenol	10.463	162	506067	20.480	ng/ul	99
30) Naphthalene	10.869	128	1562619	19.858	ng/ul	99
32) 4-Chloroaniline	10.975	127	736625	20.860	ng/ul	100
33) Hexachlorobutadiene	11.157	225	311236	19.997	ng/ul	99
34) Caprolactam	11.751	113	182754	19.721	ng/ul	98
35) 4-Chloro-3-methylphenol	12.093	107	563770	20.641	ng/ul	99
36) 2-Methylnaphthalene	12.475	142	1176433	20.620	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120122\
 Data File : BP012927.D
 Acq On : 01 Dec 2022 20:56
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SICV630


Quant Time: Dec 01 23:49:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 23:46:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.693	142	1146155	19.949	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.834	216	680407	19.768	ng/ul	98
40) Hexachlorocyclopentadiene	12.816	237	349824	19.959	ng/ul	99
41) 2,4,6-Trichlorophenol	13.075	196	472343	20.023	ng/ul	99
42) 2,4,5-Trichlorophenol	13.140	196	521635	20.558	ng/ul	99
43) 1,1'-Biphenyl	13.475	154	1640217	20.495	ng/ul	100
44) 2-Chloronaphthalene	13.516	162	1284492	20.053	ng/ul	99
45) 2-Nitroaniline	13.722	65	357656	20.745	ng/ul	98
47) Dimethylphthalate	14.093	163	1724839	20.191	ng/ul	99
48) 2,6-Dinitrotoluene	14.210	165	366701	23.009	ng/ul	99
50) Acenaphthylene	14.363	152	2079868	21.022	ng/ul	100
51) 3-Nitroaniline	14.540	138	400214	22.595	ng/ul	100
52) Acenaphthene	14.704	153	1407594	20.077	ng/ul	99
53) 2,4-Dinitrophenol	14.745	184	211448	18.543	ng/ul	97
55) 4-Nitrophenol	14.840	109	250783	21.407	ng/ul	98
56) Dibenzofuran	15.040	168	2062788	20.774	ng/ul	100
57) 2,4-Dinitrotoluene	14.998	165	504771	21.159	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.263	232	494297	20.687	ng/ul	98
59) Diethylphthalate	15.457	149	1715713	20.151	ng/ul	99
61) Fluorene	15.687	166	1703864	20.757	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.681	204	870922	20.540	ng/ul	100
63) 4-Nitroaniline	15.704	138	422420	25.111	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.757	198	326543	19.410	ng/ul	96
67) N-Nitrosodiphenylamine	15.892	169	1507139	20.375	ng/ul	100
68) 4-Bromophenyl-phenylether	16.581	248	559966	20.905	ng/ul	98
69) Hexachlorobenzene	16.692	284	674651	20.162	ng/ul	99
70) Atrazine	16.845	200	554409	20.196	ng/ul	98
71) Pentachlorophenol	17.039	266	434765	20.062	ng/ul	99
72) Phenanthrene	17.439	178	2858385	20.025	ng/ul	99
74) Anthracene	17.528	178	2938399	20.544	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.440	216	667145	18.507	ng/ul	98
76) Pentachlorobenzene	14.957	250	705498	17.580	ng/ul	100
77) Carbazole	17.798	167	2778098	22.330	ng/ul	100
78) Di-n-butylphthalate	18.339	149	3153797	18.623	ng/ul	100
80) Fluoranthene	19.398	202	3589546	19.623	ng/ul	99
82) Pyrene	19.751	202	3804545	19.991	ng/ul	98
83) Butylbenzylphthalate	20.616	149	1486596	19.382	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.392	252	1493661	21.777	ng/ul	99
85) Benzo(a)anthracene	21.463	228	4003641	20.200	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.375	149	2210077	19.759	ng/ul	99
87) Chrysene	21.522	228	3769627	20.271	ng/ul	99
89) Di-n-octyl phthalate	22.333	149	3763647	21.036	ng/ul	100
90) Benzo(b)fluoranthene	23.216	252	4236563	18.993	ng/ul	100
91) Benzo(k)fluoranthene	23.269	252	4338040	20.508	ng/ul	99
93) Benzo(a)pyrene	23.874	252	3879152	19.696	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.610	276	4915624	20.395	ng/ul	99
95) Dibenzo(a,h)anthracene	26.627	278	4307809	19.962	ng/ul	99
96) Benzo(g,h,i)perylene	27.415	276	4111317	19.651	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 01 23:49:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 01 23:46:27 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.093 ng/uL
 RT: 3.428 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BP012927.D
 Acq: 01 Dec 2022 20:56




Instrument :
 BNA_P
 ClientSampleId :
 SICV630


Tgt Ion: 88	Resp: 51920
Ion Ratio	Lower Upper
88	100
43	34.5 30.6 46.0
58	76.8 57.7 86.5



#5
 Pyridine
 Concen: 17.730 ng/uL
 RT: 3.834 min Scan# 110
 Delta R.T. -0.006 min
 Lab File: BP012927.D
 Acq: 01 Dec 2022 20:56

Tgt Ion: 79	Resp: 382959
Ion Ratio	Lower Upper
79	100
52	78.1 59.2 88.8
51	34.6 27.8 41.6









#6
Benzaldehyde
Concen: 25.847 ng/ul
RT: 7.140 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion: 77 Resp: 358586
Ion Ratio Lower Upper
77 100
105 100.9 82.6 124.0
106 97.8 79.8 119.6



#8
Phenol
Concen: 19.154 ng/ul
RT: 7.181 min Scan# 679
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion: 94 Resp: 561481
Ion Ratio Lower Upper
94 100
65 25.7 21.8 32.6
66 33.5 26.6 39.8









#10
Bis(2-Chloroethyl)ether
Concen: 19.466 ng/ul
RT: 7.422 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion: 93 Resp: 440177
Ion Ratio Lower Upper
93 100
63 74.2 59.6 89.4
95 31.9 25.7 38.5



#12
2-Chlorophenol
Concen: 19.539 ng/ul
RT: 7.563 min Scan# 744
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:128 Resp: 432710
Ion Ratio Lower Upper
128 100
64 44.4 34.8 52.2
130 32.5 25.4 38.2





#13

2-Methylphenol

Concen: 19.383 ng/ul

RT: 8.446 min Scan# 894

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :





SICV630

Tgt Ion:108 Resp: 444256

Ion Ratio Lower Upper

108 100

107 88.8 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.407 ng/ul

RT: 8.540 min Scan# 910

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56




Tgt Ion: 45 Resp: 614916


Ion Ratio Lower Upper

45 100

77 13.9 11.6 17.4

79 11.1 8.9 13.3









#16
Acetophenone
Concen: 20.821 ng/ul
RT: 8.834 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:105 Resp: 747634
Ion Ratio Lower Upper
105 100
77 74.7 60.4 90.6
51 25.3 20.7 31.1



#17
N-Nitroso-di-n-propylamine
Concen: 20.591 ng/ul
RT: 8.816 min Scan# 957
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion: 70 Resp: 375883
Ion Ratio Lower Upper
70 100
42 56.2 46.0 69.0
101 10.4 7.4 11.0
130 22.6 17.6 26.4









#18
4-Methylphenol
Concen: 19.779 ng/ul
RT: 8.775 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:108 Resp: 506248
Ion Ratio Lower Upper
108 100
107 107.8 89.8 134.6



#19
Hexachloroethane
Concen: 19.267 ng/ul
RT: 9.093 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:117 Resp: 157572
Ion Ratio Lower Upper
117 100
201 114.3 88.2 132.4
199 70.5 57.4 86.0









#22
Nitrobenzene
Concen: 20.090 ng/ul
RT: 9.210 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion: 77 Resp: 513993
Ion Ratio Lower Upper
77 100
123 48.1 37.6 56.4
65 12.5 9.9 14.9



#23
Isophorone
Concen: 19.945 ng/ul
RT: 9.740 min Scan# 1114
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion: 82 Resp: 1090059
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 18.5 15.0 22.6









#25
2-Nitrophenol
Concen: 20.500 ng/ul
RT: 9.928 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:139 Resp: 279927
Ion Ratio Lower Upper
139 100
65 47.4 39.3 58.9
109 27.7 21.9 32.9



#26
2,4-Dimethylphenol
Concen: 20.388 ng/ul
RT: 9.981 min Scan# 1155
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:107 Resp: 546744
Ion Ratio Lower Upper
107 100
121 51.9 41.6 62.4
122 88.3 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.070 ng/ul
 RT: 10.222 min Scan# 1196
 Delta R.T. 0.000 min
 Lab File: BP012927.D
 Acq: 01 Dec 2022 20:56




Instrument :
 BNA_P
 ClientSampleId :
 SICV630


Tgt Ion: 93 Resp: 664232
 Ion Ratio Lower Upper
 93 100
 95 33.0 25.6 38.4
 123 11.6 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 20.480 ng/ul
 RT: 10.463 min Scan# 1237
 Delta R.T. 0.000 min
 Lab File: BP012927.D
 Acq: 01 Dec 2022 20:56

Tgt Ion:162 Resp: 506067
 Ion Ratio Lower Upper
 162 100
 164 65.1 52.2 78.4
 98 32.4 27.3 40.9





#30

Naphthalene

Concen: 19.858 ng/ul

RT: 10.869 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:128 Resp: 1562619

Ion Ratio Lower Upper

128 100

129 11.1 8.6 13.0

127 13.0 10.2 15.2



#32

4-Chloroaniline

Concen: 20.860 ng/ul

RT: 10.975 min Scan# 1324

Delta R.T. -0.006 min

Lab File: BP012927.D




Acq: 01 Dec 2022 20:56


Tgt Ion:127 Resp: 736625

Ion Ratio Lower Upper

127 100

129 32.3 25.7 38.5









#33
Hexachlorobutadiene
Concen: 19.997 ng/ul
RT: 11.157 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:225 Resp: 311236
Ion Ratio Lower Upper
225 100
223 60.8 49.0 73.4
227 63.7 50.2 75.4



#34
Caprolactam
Concen: 19.721 ng/ul
RT: 11.751 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:113 Resp: 182754
Ion Ratio Lower Upper
113 100
55 165.7 129.8 194.6
56 122.0 96.7 145.1









#35
4-Chloro-3-methylphenol
Concen: 20.641 ng/ul
RT: 12.093 min Scan# 1514
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:107 Resp: 563770
Ion Ratio Lower Upper
107 100
144 28.2 21.7 32.5
142 84.7 67.1 100.7



#36
2-Methylnaphthalene
Concen: 20.620 ng/ul
RT: 12.475 min Scan# 1579
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:142 Resp: 1176433
Ion Ratio Lower Upper
142 100
141 88.7 71.8 107.8





#37

1-Methylnaphthalene

Concen: 19.949 ng/ul

RT: 12.693 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:142 Resp: 1146155

Ion Ratio Lower Upper

142 100

141 90.4 74.2 111.4

116 3.4 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.768 ng/ul

RT: 12.834 min Scan# 1640

Delta R.T. -0.006 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Tgt Ion:216 Resp: 680407




Ion Ratio Lower Upper


216 100

214 78.3 64.2 96.4

179 19.7 15.4 23.2

108 14.8 11.7 17.5









#40
Hexachlorocyclopentadiene
Concen: 19.959 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:237 Resp: 349824
Ion Ratio Lower Upper
237 100
235 63.0 50.1 75.1
272 13.4 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 20.023 ng/ul
RT: 13.075 min Scan# 1681
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:196 Resp: 472343
Ion Ratio Lower Upper
196 100
198 96.1 76.6 114.8
200 30.4 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 20.558 ng/ul
RT: 13.140 min Scan# 1693
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:196 Resp: 521635
Ion Ratio Lower Upper
196 100
198 96.4 77.8 116.6
200 31.4 25.3 37.9



#43
1,1'-Biphenyl
Concen: 20.495 ng/ul
RT: 13.475 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:154 Resp: 1640217
Ion Ratio Lower Upper
154 100
153 39.8 31.8 47.6
76 12.2 10.2 15.2









#44
2-Chloronaphthalene
Concen: 20.053 ng/ul
RT: 13.516 min Scan# 1756
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:162 Resp: 1284492
Ion Ratio Lower Upper
162 100
164 33.3 26.6 39.8
127 36.7 28.5 42.7



#45
2-Nitroaniline
Concen: 20.745 ng/ul
RT: 13.722 min Scan# 1791
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion: 65 Resp: 357656
Ion Ratio Lower Upper
65 100
92 71.9 56.7 85.1
138 117.9 92.1 138.1









#47
Dimethylphthalate
Concen: 20.191 ng/ul
RT: 14.093 min Scan# 1854
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:163 Resp: 1724839
Ion Ratio Lower Upper
163 100
194 4.3 3.3 4.9
164 10.5 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 23.009 ng/ul
RT: 14.210 min Scan# 1874
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:165 Resp: 366701
Ion Ratio Lower Upper
165 100
89 50.7 41.1 61.7
121 19.2 15.8 23.8









#50
Acenaphthylene
Concen: 21.022 ng/ul
RT: 14.363 min Scan# 1900
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:152 Resp: 2079868
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 13.1 10.6 16.0



#51
3-Nitroaniline
Concen: 22.595 ng/ul
RT: 14.540 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:138 Resp: 400214
Ion Ratio Lower Upper
138 100
108 10.4 8.9 13.3
92 110.0 88.2 132.2









#52
Acenaphthene
Concen: 20.077 ng/ul
RT: 14.704 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:153 Resp: 1407594
Ion Ratio Lower Upper
153 100
152 47.4 38.5 57.7
154 88.0 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 18.543 ng/ul
RT: 14.745 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:184 Resp: 211448
Ion Ratio Lower Upper
184 100
63 53.8 45.4 68.2
154 58.0 45.7 68.5









#55
4-Nitrophenol
Concen: 21.407 ng/ul
RT: 14.840 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:109 Resp: 250783
Ion Ratio Lower Upper
109 100
139 154.7 124.2 186.4
65 136.7 112.9 169.3



#56
Dibenzofuran
Concen: 20.774 ng/ul
RT: 15.040 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:168 Resp: 2062788
Ion Ratio Lower Upper
168 100
139 34.5 27.4 41.2









#57
2,4-Dinitrotoluene
Concen: 21.159 ng/ul
RT: 14.998 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:165 Resp: 504771
Ion Ratio Lower Upper
165 100
63 36.9 30.2 45.2
182 2.3 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.687 ng/ul
RT: 15.263 min Scan# 2053
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:232 Resp: 494297
Ion Ratio Lower Upper
232 100
131 36.8 31.0 46.6
130 1.8 1.8 2.6
166 27.5 22.5 33.7









#59
Diethylphthalate
Concen: 20.151 ng/ul
RT: 15.457 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:149 Resp: 1715713
Ion Ratio Lower Upper
149 100
177 23.2 18.4 27.6
150 12.4 9.8 14.6



#61
Fluorene
Concen: 20.757 ng/ul
RT: 15.687 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:166 Resp: 1703864
Ion Ratio Lower Upper
166 100
165 97.1 77.4 116.2
167 13.5 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.540 ng/ul
RT: 15.681 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:204 Resp: 870922
Ion Ratio Lower Upper
204 100
206 32.7 26.6 40.0
141 51.1 41.0 61.4



#63
4-Nitroaniline
Concen: 25.111 ng/ul
RT: 15.704 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:138 Resp: 422420
Ion Ratio Lower Upper
138 100
92 47.8 39.0 58.4
108 68.8 55.6 83.4









#66
4,6-Dinitro-2-methylphenol
Concen: 19.410 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:198 Resp: 326543
Ion Ratio Lower Upper
198 100
182 4.7 3.8 5.8
77 21.9 15.6 23.4



#67
N-Nitrosodiphenylamine
Concen: 20.375 ng/ul
RT: 15.892 min Scan# 2160
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:169 Resp: 1507139
Ion Ratio Lower Upper
169 100
168 66.6 53.4 80.0
167 36.3 28.8 43.2









#68
4-Bromophenyl-phenylether
Concen: 20.905 ng/ul
RT: 16.581 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt	Ion:248	Resp:	559966
Ion	Ratio	Lower	Upper
248	100		
250	99.0	77.5	116.3
141	61.2	49.7	74.5



#69
Hexachlorobenzene
Concen: 20.162 ng/ul
RT: 16.692 min Scan# 2296
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt	Ion:284	Resp:	674651
Ion	Ratio	Lower	Upper
284	100		
142	29.9	23.3	34.9
249	30.7	24.0	36.0









#70
Atrazine
Concen: 20.196 ng/ul
RT: 16.845 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:200 Resp: 554409
Ion Ratio Lower Upper
200 100
173 25.5 21.0 31.6
215 47.7 39.2 58.8



#71
Pentachlorophenol
Concen: 20.062 ng/ul
RT: 17.039 min Scan# 2355
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:266 Resp: 434765
Ion Ratio Lower Upper
266 100
264 63.8 50.5 75.7
268 63.3 51.2 76.8





#72

Phenanthrene

Concen: 20.025 ng/ul

RT: 17.439 min Scan# 2423

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:178 Resp: 2858385

Ion Ratio Lower Upper

178 100

179 15.5 12.6 18.8

176 19.6 15.4 23.0



#74

Anthracene

Concen: 20.544 ng/ul

RT: 17.528 min Scan# 2438

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56




Tgt Ion:178 Resp: 2938399


Ion Ratio Lower Upper

178 100

179 15.4 12.6 19.0

176 18.6 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.507 ng/uL

RT: 13.440 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:216 Resp: 667145

Ion Ratio Lower Upper

216 100

214 76.7 62.2 93.2

218 47.5 39.4 59.2



#76

Pentachlorobenzene

Concen: 17.580 ng/uL

RT: 14.957 min Scan# 2001

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56




Tgt Ion:250 Resp: 705498


Ion Ratio Lower Upper

250 100

248 62.4 49.4 74.2

252 63.2 50.6 75.8









#77
 Carbazole
 Concen: 22.330 ng/ul
 RT: 17.798 min Scan# 2484
 Delta R.T. 0.000 min
 Lab File: BP012927.D
 Acq: 01 Dec 2022 20:56




Instrument :
 BNA_P
 ClientSampleId :
 SICV630


Tgt Ion:167 Resp: 2778098
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.2 25.8
 139 11.9 9.7 14.5



#78
 Di-n-butylphthalate
 Concen: 18.623 ng/ul
 RT: 18.339 min Scan# 2576
 Delta R.T. 0.000 min
 Lab File: BP012927.D
 Acq: 01 Dec 2022 20:56

Tgt Ion:149 Resp: 3153797
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.2
 104 4.5 3.7 5.5









#80
Fluoranthene
Concen: 19.623 ng/uI
RT: 19.398 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:202 Resp: 3589546
Ion Ratio Lower Upper
202 100
101 9.7 7.9 11.9
100 7.4 6.2 9.4



#82
Pyrene
Concen: 19.991 ng/uI
RT: 19.751 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:202 Resp: 3804545
Ion Ratio Lower Upper
202 100
101 11.2 9.7 14.5
100 9.1 7.7 11.5





#83

Butylbenzylphthalate

Concen: 19.382 ng/ul

RT: 20.616 min Scan# 2963

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:149 Resp: 1486596

Ion Ratio Lower Upper

149 100

91 64.9 55.0 82.6

206 23.1 17.8 26.8



#84

3,3'-Dichlorobenzidine

Concen: 21.777 ng/ul

RT: 21.392 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56




Tgt Ion:252 Resp: 1493661


Ion Ratio Lower Upper

252 100

254 65.2 51.9 77.9

126 8.8 8.0 12.0





#85

Benzo(a)anthracene

Concen: 20.200 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:228 Resp: 4003641

Ion Ratio Lower Upper

228 100

229 19.9 16.2 24.2

226 27.5 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.759 ng/ul

RT: 21.375 min Scan# 3092

Delta R.T. 0.000 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56




Tgt Ion:149 Resp: 2210077


Ion Ratio Lower Upper

149 100

167 28.8 22.6 33.8

279 5.0 3.5 5.3









#87
Chrysene
Concen: 20.271 ng/ul
RT: 21.522 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:228 Resp: 3769627
Ion Ratio Lower Upper
228 100
226 30.6 24.3 36.5
229 19.7 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 21.036 ng/ul
RT: 22.333 min Scan# 3255
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:149 Resp: 3763647





#90

Benzo(b)fluoranthene

Concen: 18.993 ng/ul

RT: 23.216 min Scan# 3406

Delta R.T. -0.006 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:252 Resp: 4236563

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.6

125 8.8 7.1 10.7



#91

Benzo(k)fluoranthene

Concen: 20.508 ng/ul

RT: 23.269 min Scan# 3414

Delta R.T. -0.006 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56




Tgt Ion:252 Resp: 4338040


Ion Ratio Lower Upper

252 100

253 22.4 17.6 26.4

125 8.6 7.1 10.7









#93
Benzo(a)pyrene
Concen: 19.696 ng/ul
RT: 23.874 min Scan# 3517
Delta R.T. 0.000 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:252 Resp: 3879152
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 9.8 8.2 12.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.395 ng/ul
RT: 26.610 min Scan# 3982
Delta R.T. -0.006 min
Lab File: BP012927.D
Acq: 01 Dec 2022 20:56

Tgt Ion:276 Resp: 4915624
Ion Ratio Lower Upper
276 100
138 21.3 17.9 26.9
277 25.1 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 19.962 ng/ul

RT: 26.627 min Scan# 3985

Delta R.T. -0.006 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:278 Resp: 4307809

Ion Ratio Lower Upper

278 100

139 14.7 11.8 17.8

279 24.2 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 19.651 ng/ul

RT: 27.415 min Scan# 4119

Delta R.T. -0.006 min

Lab File: BP012927.D

Acq: 01 Dec 2022 20:56

Tgt Ion:276 Resp: 4111317

Ion Ratio Lower Upper

276 100

138 19.0 15.9 23.9

277 24.0 19.2 28.8

