

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120221\
 Data File : BP008191.D
 Acq On : 02 Dec 2021 14:55
 Operator : CG/JU
 Sample : M4872-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DT-CON-PAD

Quant Time: Dec 02 15:36:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 01 12:06:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	67385	20.000	ng	-0.02
21) Naphthalene-d8	10.593	136	234119	20.000	ng	-0.02
39) Acenaphthene-d10	14.445	164	137790	20.000	ng	-0.01
64) Phenanthrene-d10	17.210	188	294505	20.000	ng	0.00
76) Chrysene-d12	21.310	240	329563	20.000	ng	0.00
86) Perylene-d12	23.710	264	395072	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.387	112	329210	78.661	ng	-0.02
7) Phenol-d6	6.981	99	576579	102.055	ng	-0.02
23) Nitrobenzene-d5	8.958	82	457938	88.946	ng	-0.02
42) 2,4,6-Tribromophenol	15.945	330	71030	40.331	ng	-0.01
45) 2-Fluorobiphenyl	13.063	172	863915	91.071	ng	-0.02
79) Terphenyl-d14	19.780	244	1413080	89.610	ng	0.00
Target Compounds						
60) Diethylphthalate	15.275	149	185307	18.047	ng	99
74) Di-n-butylphthalate	18.175	149	63893	3.388	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120221\
Data File : BP008191.D
Acq On : 02 Dec 2021 14:55
Operator : CG/JU
Sample : M4872-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DT-CON-PAD

Quant Time: Dec 02 15:36:47 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 01 12:06:05 2021
Response via : Initial Calibration

