

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120222\
 Data File : BP012940.D
 Acq On : 02 Dec 2022 15:26
 Operator : CG/JU
 Sample : N5785-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 COL72

Quant Time: Dec 03 01:30:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 23:46:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.004	152	375895	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	1930637	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.645	164	1566200	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	3849803	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	4393912	20.000	ng/u1	0.00
88) Perylene-d12	23.992	264	4420904	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	30515	3.612	ng/uL	0.00
4) Pyridine-d5	3.828	84	158227	5.804	ng/u1	0.01
7) Phenol-d5	7.157	99	190898	5.499	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	486126	24.735	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	508567	19.989	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	378467	13.278	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	350507	25.530	ng/u1	0.00
24) 2-Nitrophenol-d4	9.898	143	315488	20.079	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	689026	21.906	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	522681	11.761	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	3109421	27.526	ng/u1	0.00
49) Acenaphthylene-d8	14.339	160	3324219	26.653	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	89266	3.863	ng/u1	0.00
60) Fluorene-d10	15.633	176	2790262	28.103	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	264413	11.706	ng/u1	0.00
73) Anthracene-d10	17.498	188	5104086	29.995	ng/u1	0.00
81) Pyrene-d10	19.727	212	6307413	29.204	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.827	264	6478718	29.993	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120222\
Data File : BP012940.D
Acq On : 02 Dec 2022 15:26
Operator : CG/JU
Sample : N5785-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0L72

Quant Time: Dec 03 01:30:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 01 23:46:27 2022
Response via : Initial Calibration

