

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120722\
 Data File : BP012969.D
 Acq On : 07 Dec 2022 13:54
 Operator : CG/JU
 Sample : SST02013
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020640

Quant Time: Dec 07 23:59:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 07 23:53:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	564627	20.000	ng/ul	0.00
20) Naphthalene-d8	10.793	136	2475738	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.616	164	1645333	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.375	188	3709689	20.000	ng/ul	0.00
79) Chrysene-d12	21.457	240	3596859	20.000	ng/ul	0.00
88) Perylene-d12	23.951	264	3781497	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	104405	8.228	ng/uL	0.00
4) Pyridine-d5	3.787	84	730239	17.833	ng/ul	0.00
7) Phenol-d5	7.134	99	880685	16.890	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	563971	19.104	ng/ul	0.00
11) 2-Chlorophenol-d4	7.510	132	726104	18.999	ng/ul	0.00
15) 4-Methylphenol-d8	8.687	113	740810	17.303	ng/ul	0.00
21) Nitrobenzene-d5	9.146	128	364147	20.683	ng/ul	0.00
24) 2-Nitrophenol-d4	9.875	143	410645	20.381	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	804840	19.954	ng/ul	0.00
31) 4-Chloroaniline-d4	10.928	131	1145754	20.105	ng/ul	0.00
46) Dimethylphthalate-d6	14.022	166	2549282	21.482	ng/ul	0.00
49) Acenaphthylene-d8	14.310	160	2818715	21.513	ng/ul	0.00
54) 4-Nitrophenol-d4	14.810	143	441915	18.206	ng/ul	0.00
60) Fluorene-d10	15.610	176	2149875	20.611	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.728	200	448051	20.586	ng/ul	0.00
73) Anthracene-d10	17.475	188	3409676	20.794	ng/ul	0.00
81) Pyrene-d10	19.698	212	4055907	22.941	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.786	264	3763320	20.368	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	106695	7.893	ng/uL	100
5) Pyridine	3.805	79	730494	18.313	ng/ul	100
6) Benzaldehyde	7.116	77	357241	13.943	ng/ul	100
8) Phenol	7.163	94	900804	16.640	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.399	93	767446	18.378	ng/ul	100
12) 2-Chlorophenol	7.540	128	743885	18.189	ng/ul	100
13) 2-Methylphenol	8.422	108	714774	16.887	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.516	45	1068746	18.265	ng/ul	100
16) Acetophenone	8.810	105	1218134	18.369	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.793	70	608483	18.050	ng/ul	100
18) 4-Methylphenol	8.752	108	800134	16.928	ng/ul	100
19) Hexachloroethane	9.069	117	294258	19.483	ng/ul	100
22) Nitrobenzene	9.187	77	856747	20.174	ng/ul	100
23) Isophorone	9.722	82	1744502	19.230	ng/ul	100
25) 2-Nitrophenol	9.904	139	448851	19.802	ng/ul	100
26) 2,4-Dimethylphenol	9.957	107	876443	19.689	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.199	93	1108420	20.177	ng/ul	100
29) 2,4-Dichlorophenol	10.440	162	782995	19.089	ng/ul	100
30) Naphthalene	10.846	128	2580251	19.755	ng/ul	100
32) 4-Chloroaniline	10.951	127	1147050	19.568	ng/ul	100
33) Hexachlorobutadiene	11.128	225	524758	20.312	ng/ul	100
34) Caprolactam	11.734	113	162469	10.562	ng/ul	100
35) 4-Chloro-3-methylphenol	12.069	107	814247	17.960	ng/ul	100
36) 2-Methylnaphthalene	12.451	142	1774340	18.736	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120722\
 Data File : BP012969.D
 Acq On : 07 Dec 2022 13:54
 Operator : CG/JU
 Sample : SSTD02013
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020640


Quant Time: Dec 07 23:59:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 07 23:53:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.669	142	1841721	19.312	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.816	216	1056740	21.157	ng/ul	100
40) Hexachlorocyclopentadiene	12.793	237	648169	25.484	ng/ul	100
41) 2,4,6-Trichlorophenol	13.051	196	681859	19.919	ng/ul	100
42) 2,4,5-Trichlorophenol	13.122	196	729769	19.820	ng/ul	100
43) 1,1'-Biphenyl	13.451	154	2480727	21.361	ng/ul	100
44) 2-Chloronaphthalene	13.492	162	1936829	20.837	ng/ul	100
45) 2-Nitroaniline	13.698	65	487094	19.469	ng/ul	100
47) Dimethylphthalate	14.069	163	2504085	20.200	ng/ul	100
48) 2,6-Dinitrotoluene	14.192	165	476373	20.598	ng/ul	100
50) Acenaphthylene	14.339	152	3029844	21.104	ng/ul	100
51) 3-Nitroaniline	14.522	138	412165	16.036	ng/ul	100
52) Acenaphthene	14.681	153	2077904	20.424	ng/ul	100
53) 2,4-Dinitrophenol	14.722	184	273611	16.535	ng/ul	100
55) 4-Nitrophenol	14.822	109	314180	18.481	ng/ul	100
56) Dibenzofuran	15.016	168	2958826	20.535	ng/ul	100
57) 2,4-Dinitrotoluene	14.975	165	702121	20.282	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.239	232	674548	19.455	ng/ul	100
59) Diethylphthalate	15.434	149	2447078	19.806	ng/ul	100
61) Fluorene	15.669	166	2421602	20.330	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.657	204	1266414	20.582	ng/ul	100
63) 4-Nitroaniline	15.686	138	365335	14.966	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.739	198	451584	19.580	ng/ul	100
67) N-Nitrosodiphenylamine	15.869	169	2078728	20.499	ng/ul	100
68) 4-Bromophenyl-phenylether	16.557	248	803447	21.879	ng/ul	100
69) Hexachlorobenzene	16.675	284	954180	20.800	ng/ul	100
70) Atrazine	16.828	200	805991	21.417	ng/ul	100
71) Pentachlorophenol	17.016	266	554214	18.655	ng/ul	100
72) Phenanthrene	17.416	178	3912351	19.993	ng/ul	100
74) Anthracene	17.510	178	3985733	20.327	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.416	216	1049029	21.227	ng/ul	100
76) Pentachlorobenzene	14.934	250	1075793	19.554	ng/ul	100
77) Carbazole	17.775	167	3533816	20.719	ng/ul	100
78) Di-n-butylphthalate	18.316	149	4215554	18.158	ng/ul	100
80) Fluoranthene	19.375	202	4695024	22.404	ng/ul	100
82) Pyrene	19.727	202	4910880	22.525	ng/ul	100
83) Butylbenzylphthalate	20.592	149	1882511	21.424	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.369	252	1625844	20.692	ng/ul	100
85) Benzo(a)anthracene	21.439	228	4807360	21.172	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.351	149	2709715	21.147	ng/ul	100
87) Chrysene	21.498	228	4579870	21.498	ng/ul	100
89) Di-n-octyl phthalate	22.304	149	4520965	23.427	ng/ul	100
90) Benzo(b)fluoranthene	23.186	252	4903267	20.380	ng/ul	100
91) Benzo(k)fluoranthene	23.233	252	4676323	20.496	ng/ul	100
93) Benzo(a)pyrene	23.839	252	4203005	19.785	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.556	276	5105778	19.640	ng/ul	100
95) Dibenzo(a,h)anthracene	26.568	278	4266442	18.330	ng/ul	100
96) Benzo(g,h,i)perylene	27.351	276	4026274	17.842	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 07 23:59:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 07 23:53:54 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.893 ng/uL
RT: 3.393 min Scan# 35
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion: 88 Resp: 106695
Ion Ratio Lower Upper
88 100
43 28.0 22.4 33.6
58 72.8 58.2 87.4



#5
Pyridine
Concen: 18.313 ng/uL
RT: 3.805 min Scan# 105
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion: 79 Resp: 730494
Ion Ratio Lower Upper
79 100
52 75.9 60.7 91.1
51 36.1 28.9 43.3









#6
Benzaldehyde
Concen: 13.943 ng/ul
RT: 7.116 min Scan# 60
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion: 77 Resp: 357241
Ion Ratio Lower Upper
77 100
105 101.9 81.5 122.3
106 96.1 76.9 115.3



#8
Phenol
Concen: 16.640 ng/ul
RT: 7.163 min Scan# 676
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion: 94 Resp: 900804
Ion Ratio Lower Upper
94 100
65 26.4 21.1 31.7
66 33.8 27.0 40.6









#10
Bis(2-Chloroethyl)ether
Concen: 18.378 ng/ul
RT: 7.399 min Scan# 716
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion: 93 Resp: 767446
Ion Ratio Lower Upper
93 100
63 74.8 59.8 89.8
95 33.0 26.4 39.6



#12
2-Chlorophenol
Concen: 18.189 ng/ul
RT: 7.540 min Scan# 740
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:128 Resp: 743885
Ion Ratio Lower Upper
128 100
64 43.6 34.9 52.3
130 32.2 25.8 38.6





#13

2-Methylphenol

Concen: 16.887 ng/ul

RT: 8.422 min Scan# 890

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :





SSTD020640

Tgt Ion:108 Resp: 714774

Ion Ratio Lower Upper

108 100

107 86.4 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 18.265 ng/ul

RT: 8.516 min Scan# 906

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion: 45 Resp: 1068746


Ion Ratio Lower Upper

45 100

77 13.8 11.0 16.6

79 11.6 9.3 13.9









#16
Acetophenone
Concen: 18.369 ng/ul
RT: 8.810 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:105 Resp: 1218134
Ion Ratio Lower Upper
105 100
77 74.1 59.3 88.9
51 24.5 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 18.050 ng/ul
RT: 8.793 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion: 70 Resp: 608483
Ion Ratio Lower Upper
70 100
42 56.9 45.5 68.3
101 9.9 7.9 11.9
130 21.8 17.4 26.2









#18
4-Methylphenol
Concen: 16.928 ng/ul
RT: 8.752 min Scan# 946
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:108 Resp: 800134
Ion Ratio Lower Upper
108 100
107 109.3 87.4 131.2



#19
Hexachloroethane
Concen: 19.483 ng/ul
RT: 9.069 min Scan# 1000
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:117 Resp: 294258
Ion Ratio Lower Upper
117 100
201 117.1 93.7 140.5
199 73.3 58.6 88.0









#22
Nitrobenzene
Concen: 20.174 ng/ul
RT: 9.187 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion: 77 Resp: 856747
Ion Ratio Lower Upper
77 100
123 47.9 38.3 57.5
65 12.9 10.3 15.5



#23
Isophorone
Concen: 19.230 ng/ul
RT: 9.722 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion: 82 Resp: 1744502
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 19.1 15.3 22.9









#25
2-Nitrophenol
Concen: 19.802 ng/ul
RT: 9.904 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:139 Resp: 448851
Ion Ratio Lower Upper
139 100
65 45.7 36.6 54.8
109 27.0 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 19.689 ng/ul
RT: 9.957 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:107 Resp: 876443
Ion Ratio Lower Upper
107 100
121 52.2 41.8 62.6
122 90.7 72.6 108.8





#27

Bis(2-Chloroethoxy)methane

Concen: 20.177 ng/ul

RT: 10.199 min Scan# 1192

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion: 93 Resp: 1108420

Ion Ratio Lower Upper

93 100

95 32.5 26.0 39.0

123 11.5 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 19.089 ng/ul

RT: 10.440 min Scan# 1233

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:162 Resp: 782995


Ion Ratio Lower Upper

162 100

164 65.6 52.5 78.7

98 33.2 26.6 39.8









#30
Naphthalene
Concen: 19.755 ng/ul
RT: 10.846 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:128 Resp: 2580251
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 19.568 ng/ul
RT: 10.951 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:127 Resp: 1147050
Ion Ratio Lower Upper
127 100
129 32.0 25.6 38.4









#33
Hexachlorobutadiene
Concen: 20.312 ng/ul
RT: 11.128 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:225 Resp: 524758
Ion Ratio Lower Upper
225 100
223 61.5 49.2 73.8
227 62.7 50.2 75.2



#34
Caprolactam
Concen: 10.562 ng/ul
RT: 11.734 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:113 Resp: 162469
Ion Ratio Lower Upper
113 100
55 177.4 141.9 212.9
56 124.5 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 17.960 ng/ul
RT: 12.069 min Scan# 1510
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:107 Resp: 814247
Ion Ratio Lower Upper
107 100
144 26.8 21.4 32.2
142 82.2 65.8 98.6



#36
2-Methylnaphthalene
Concen: 18.736 ng/ul
RT: 12.451 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:142 Resp: 1774340
Ion Ratio Lower Upper
142 100
141 88.5 70.8 106.2









#37
1-Methylnaphthalene
Concen: 19.312 ng/ul
RT: 12.669 min Scan# 1612
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:142 Resp: 1841721
Ion Ratio Lower Upper
142 100
141 90.9 72.7 109.1
116 3.4 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.157 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:216 Resp: 1056740
Ion Ratio Lower Upper
216 100
214 77.4 61.9 92.9
179 18.8 15.0 22.6
108 13.7 11.0 16.4





#40

Hexachlorocyclopentadiene

Concen: 25.484 ng/ul

RT: 12.793 min Scan# 1633

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:237 Resp: 648169

Ion Ratio Lower Upper

237 100

235 62.1 49.7 74.5

272 13.5 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 19.919 ng/ul

RT: 13.051 min Scan# 1677

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:196 Resp: 681859


Ion Ratio Lower Upper

196 100

198 96.3 77.0 115.6

200 31.6 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 19.820 ng/ul

RT: 13.122 min Scan# 1689

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:196 Resp: 729769

Ion Ratio Lower Upper

196 100

198 96.4 77.1 115.7

200 31.8 25.4 38.2



#43

1,1'-Biphenyl

Concen: 21.361 ng/ul

RT: 13.451 min Scan# 1745

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:154 Resp: 2480727


Ion Ratio Lower Upper

154 100

153 39.4 31.5 47.3

76 12.0 9.6 14.4









#44
2-Chloronaphthalene
Concen: 20.837 ng/ul
RT: 13.492 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:162 Resp: 1936829
Ion Ratio Lower Upper
162 100
164 32.9 26.3 39.5
127 36.6 29.3 43.9



#45
2-Nitroaniline
Concen: 19.469 ng/ul
RT: 13.698 min Scan# 1787
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion: 65 Resp: 487094
Ion Ratio Lower Upper
65 100
92 70.5 56.4 84.6
138 112.9 90.3 135.5









#47
Dimethylphthalate
Concen: 20.200 ng/ul
RT: 14.069 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:163 Resp: 2504085
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.4 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.598 ng/ul
RT: 14.192 min Scan# 1871
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:165 Resp: 476373
Ion Ratio Lower Upper
165 100
89 53.7 43.0 64.4
121 20.1 16.1 24.1









#50
Acenaphthylene
Concen: 21.104 ng/ul
RT: 14.339 min Scan# 1896
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:152 Resp: 3029844
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.3 10.6 16.0



#51
3-Nitroaniline
Concen: 16.036 ng/ul
RT: 14.522 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:138 Resp: 412165
Ion Ratio Lower Upper
138 100
108 10.8 8.6 13.0
92 110.5 88.4 132.6





#52

Acenaphthene

Concen: 20.424 ng/ul

RT: 14.681 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:153 Resp: 2077904

Ion Ratio Lower Upper

153 100

152 47.8 38.2 57.4

154 88.5 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 16.535 ng/ul

RT: 14.722 min Scan# 1961

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:184 Resp: 273611


Ion Ratio Lower Upper

184 100

63 60.5 48.4 72.6

154 56.9 45.5 68.3









#55
4-Nitrophenol
Concen: 18.481 ng/ul
RT: 14.822 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:109 Resp: 314180
Ion Ratio Lower Upper
109 100
139 151.0 120.8 181.2
65 134.9 107.9 161.9



#56
Dibenzofuran
Concen: 20.535 ng/ul
RT: 15.016 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:168 Resp: 2958826
Ion Ratio Lower Upper
168 100
139 34.9 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 20.282 ng/ul
RT: 14.975 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:165 Resp: 702121
Ion Ratio Lower Upper
165 100
63 39.7 31.8 47.6
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.455 ng/ul
RT: 15.239 min Scan# 2049
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:232 Resp: 674548
Ion Ratio Lower Upper
232 100
131 39.7 31.8 47.6
130 2.2 1.8 2.6
166 28.6 22.9 34.3









#59
Diethylphthalate
Concen: 19.806 ng/ul
RT: 15.434 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:149 Resp: 2447078
Ion Ratio Lower Upper
149 100
177 23.0 18.4 27.6
150 12.4 9.9 14.9



#61
Fluorene
Concen: 20.330 ng/ul
RT: 15.669 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:166 Resp: 2421602
Ion Ratio Lower Upper
166 100
165 96.8 77.4 116.2
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.582 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:204 Resp: 1266414
Ion Ratio Lower Upper
204 100
206 32.9 26.3 39.5
141 50.5 40.4 60.6



#63
4-Nitroaniline
Concen: 14.966 ng/ul
RT: 15.686 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:138 Resp: 365335
Ion Ratio Lower Upper
138 100
92 48.7 39.0 58.4
108 68.9 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 19.580 ng/ul
RT: 15.739 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:198 Resp: 451584
Ion Ratio Lower Upper
198 100
182 4.1 3.3 4.9
77 20.8 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 20.499 ng/ul
RT: 15.869 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:169 Resp: 2078728
Ion Ratio Lower Upper
169 100
168 66.3 53.0 79.6
167 36.4 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 21.879 ng/ul
RT: 16.557 min Scan# 2273
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:248 Resp: 803447
Ion Ratio Lower Upper
248 100
250 96.7 77.4 116.0
141 60.4 48.3 72.5



#69
Hexachlorobenzene
Concen: 20.800 ng/ul
RT: 16.675 min Scan# 2293
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:284 Resp: 954180
Ion Ratio Lower Upper
284 100
142 28.3 22.6 34.0
249 30.8 24.6 37.0





#70

Atrazine

Concen: 21.417 ng/ul

RT: 16.828 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:200 Resp: 805991

Ion Ratio Lower Upper

200 100

173 25.6 20.5 30.7

215 49.9 39.9 59.9



#71

Pentachlorophenol

Concen: 18.655 ng/ul

RT: 17.016 min Scan# 2351

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:266 Resp: 554214


Ion Ratio Lower Upper

266 100

264 63.6 50.9 76.3

268 64.3 51.4 77.2





#72

Phenanthrene

Concen: 19.993 ng/ul

RT: 17.416 min Scan# 2419

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:178 Resp: 3912351

Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 19.2 15.4 23.0



#74

Anthracene

Concen: 20.327 ng/ul

RT: 17.510 min Scan# 2435

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:178 Resp: 3985733


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.5 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.227 ng/uL
RT: 13.416 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:216 Resp: 1049029
Ion Ratio Lower Upper
216 100
214 77.7 62.2 93.2
218 48.6 38.9 58.3



#76
Pentachlorobenzene
Concen: 19.554 ng/uL
RT: 14.934 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:250 Resp: 1075793
Ion Ratio Lower Upper
250 100
248 63.3 50.6 76.0
252 63.9 51.1 76.7









#77
 Carbazole
 Concen: 20.719 ng/ul
 RT: 17.775 min Scan# 2480
 Delta R.T. 0.000 min
 Lab File: BP012969.D
 Acq: 07 Dec 2022 13:54




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020640


Tgt Ion:167 Resp: 3533816
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.2 25.8
 139 11.8 9.4 14.2



#78
 Di-n-butylphthalate
 Concen: 18.158 ng/ul
 RT: 18.316 min Scan# 2572
 Delta R.T. 0.000 min
 Lab File: BP012969.D
 Acq: 07 Dec 2022 13:54

Tgt Ion:149 Resp: 4215554
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 4.7 3.8 5.6









#80
Fluoranthene
Concen: 22.404 ng/u1
RT: 19.375 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:202 Resp: 4695024
Ion Ratio Lower Upper
202 100
101 10.2 8.2 12.2
100 7.6 6.1 9.1



#82
Pyrene
Concen: 22.525 ng/u1
RT: 19.727 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:202 Resp: 4910880
Ion Ratio Lower Upper
202 100
101 11.8 9.4 14.2
100 9.8 7.8 11.8





#83

Butylbenzylphthalate

Concen: 21.424 ng/ul

RT: 20.592 min Scan# 2959

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:149 Resp: 1882511

Ion Ratio Lower Upper

149 100

91 66.7 53.4 80.0

206 22.4 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 20.692 ng/ul

RT: 21.369 min Scan# 3091

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:252 Resp: 1625844


Ion Ratio Lower Upper

252 100

254 64.8 51.8 77.8

126 9.3 7.4 11.2





#85

Benzo(a)anthracene

Concen: 21.172 ng/ul

RT: 21.439 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:228 Resp: 4807360

Ion Ratio Lower Upper

228 100

229 20.1 16.1 24.1

226 27.5 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.147 ng/ul

RT: 21.351 min Scan# 3088

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:149 Resp: 2709715


Ion Ratio Lower Upper

149 100

167 28.1 22.5 33.7

279 4.8 3.8 5.8





#87

Chrysene

Concen: 21.498 ng/ul

RT: 21.498 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:228 Resp: 4579870

Ion Ratio Lower Upper

228 100

226 30.6 24.5 36.7

229 20.0 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 23.427 ng/ul




RT: 22.304 min Scan# 3250


Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Tgt Ion:149 Resp: 4520965









#90
Benzo(b)fluoranthene
Concen: 20.380 ng/ul
RT: 23.186 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:252 Resp: 4903267
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 8.5 6.8 10.2



#91
Benzo(k)fluoranthene
Concen: 20.496 ng/ul
RT: 23.233 min Scan# 3408
Delta R.T. 0.000 min
Lab File: BP012969.D
Acq: 07 Dec 2022 13:54

Tgt Ion:252 Resp: 4676323
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 8.5 6.8 10.2





#93

Benzo(a)pyrene

Concen: 19.785 ng/ul

RT: 23.839 min Scan# 3511

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:252 Resp: 4203005

Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 9.4 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.640 ng/ul

RT: 26.556 min Scan# 3973

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54




Tgt Ion:276 Resp: 5105778


Ion Ratio Lower Upper

276 100

138 21.4 17.1 25.7

277 25.0 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 18.330 ng/ul

RT: 26.568 min Scan# 3975

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:278 Resp: 4266442

Ion Ratio Lower Upper

278 100

139 13.9 11.1 16.7

279 23.8 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 17.842 ng/ul

RT: 27.351 min Scan# 4108

Delta R.T. 0.000 min

Lab File: BP012969.D

Acq: 07 Dec 2022 13:54

Tgt Ion:276 Resp: 4026274

Ion Ratio Lower Upper

276 100

138 19.3 15.4 23.2

277 23.9 19.1 28.7

