

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120722\
 Data File : BP012971.D
 Acq On : 07 Dec 2022 15:02
 Operator : CG/JU
 Sample : SST08015
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080642

Quant Time: Dec 08 00:00:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 07 23:53:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	624078	20.000	ng/u1	0.00
20) Naphthalene-d8	10.799	136	2597391	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.622	164	1634765	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.375	188	3552943	20.000	ng/u1	0.00
79) Chrysene-d12	21.463	240	2637748	20.000	ng/u1	0.00
88) Perylene-d12	23.951	264	2192388	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	468445	33.401	ng/uL	0.00
4) Pyridine-d5	3.787	84	3327607	73.522	ng/u1	0.00
7) Phenol-d5	7.140	99	4111477	71.340	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.311	67	2368672	72.594	ng/u1	0.00
11) 2-Chlorophenol-d4	7.511	132	3334043	78.929	ng/u1	0.00
15) 4-Methylphenol-d8	8.699	113	3286202	69.443	ng/u1	0.01
21) Nitrobenzene-d5	9.157	128	1669183	90.368	ng/u1	0.01
24) 2-Nitrophenol-d4	9.875	143	1952897	92.386	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	3508596	82.911	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	4551610	76.127	ng/u1	0.00
46) Dimethylphthalate-d6	14.034	166	9549534	80.992	ng/u1	0.01
49) Acenaphthylene-d8	14.316	160	10758583	82.642	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	1838826	76.247	ng/u1	0.02
60) Fluorene-d10	15.616	176	7899204	76.222	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	1951514	93.619	ng/u1	0.01
73) Anthracene-d10	17.481	188	12014205	76.502	ng/u1	0.00
81) Pyrene-d10	19.704	212	13135589	101.311	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.792	264	9270341	86.541	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	485925	32.523	ng/uL	100
5) Pyridine	3.805	79	3286021	74.529	ng/u1	99
6) Benzaldehyde	7.116	77	1580816	55.823	ng/u1	98
8) Phenol	7.169	94	4129113	69.007	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.405	93	3270007	70.845	ng/u1	99
12) 2-Chlorophenol	7.546	128	3322033	73.488	ng/u1	99
13) 2-Methylphenol	8.428	108	3196730	68.329	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.516	45	4407275	68.144	ng/u1	99
16) Acetophenone	8.816	105	4694779	64.052	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.816	70	2269477	60.907	ng/u1	98
18) 4-Methylphenol	8.769	108	3459304	66.213	ng/u1	99
19) Hexachloroethane	9.069	117	1305105	78.180	ng/u1	97
22) Nitrobenzene	9.199	77	3580425	80.361	ng/u1	97
23) Isophorone	9.734	82	7044848	74.018	ng/u1	99
25) 2-Nitrophenol	9.910	139	2006943	84.395	ng/u1	97
26) 2,4-Dimethylphenol	9.969	107	3577827	76.610	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.204	93	4424194	76.762	ng/u1	99
29) 2,4-Dichlorophenol	10.446	162	3355912	77.984	ng/u1	98
30) Naphthalene	10.852	128	10282816	75.039	ng/u1	99
32) 4-Chloroaniline	10.963	127	4468342	72.658	ng/u1	99
33) Hexachlorobutadiene	11.128	225	2262325	83.467	ng/u1	98
34) Caprolactam	11.775	113	577291	35.771	ng/u1	91
35) 4-Chloro-3-methylphenol	12.081	107	3347878	70.385	ng/u1	95
36) 2-Methylnaphthalene	12.451	142	7094314	71.403	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120722\
 Data File : BP012971.D
 Acq On : 07 Dec 2022 15:02
 Operator : CG/JU
 Sample : SSTD08015
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080642

Quant Time: Dec 08 00:00:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 07 23:53:54 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.669	142	7155549	71.517	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.816	216	4200309	84.637	ng/ul	98
40) Hexachlorocyclopentadiene	12.798	237	2982712	118.029	ng/ul	100
41) 2,4,6-Trichlorophenol	13.057	196	2861479	84.131	ng/ul	99
42) 2,4,5-Trichlorophenol	13.128	196	3110659	85.029	ng/ul	99
43) 1,1'-Biphenyl	13.457	154	9262102	80.269	ng/ul	99
44) 2-Chloronaphthalene	13.504	162	7421767	80.361	ng/ul	98
45) 2-Nitroaniline	13.710	65	2053976	82.628	ng/ul	91
47) Dimethylphthalate	14.081	163	9305803	75.552	ng/ul	100
48) 2,6-Dinitrotoluene	14.204	165	2093446	91.104	ng/ul	92
50) Acenaphthylene	14.345	152	11127809	78.009	ng/ul	98
51) 3-Nitroaniline	14.534	138	1851162	72.487	ng/ul	95
52) Acenaphthene	14.687	153	7654727	75.724	ng/ul	99
53) 2,4-Dinitrophenol	14.734	184	1426494	86.763	ng/ul	93
55) 4-Nitrophenol	14.845	109	1206136	71.406	ng/ul	95
56) Dibenzofuran	15.022	168	10712844	74.829	ng/ul	100
57) 2,4-Dinitrotoluene	14.987	165	2893965	84.137	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.245	232	2703679	78.481	ng/ul	96
59) Diethylphthalate	15.445	149	9042583	73.662	ng/ul	97
61) Fluorene	15.675	166	7927640	66.984	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.663	204	4376455	71.588	ng/ul	98
63) 4-Nitroaniline	15.704	138	1598231	65.896	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.757	198	1867048	84.524	ng/ul#	94
67) N-Nitrosodiphenylamine	15.881	169	7505915	77.282	ng/ul	99
68) 4-Bromophenyl-phenylether	16.557	248	3146426	89.462	ng/ul	100
69) Hexachlorobenzene	16.681	284	3609659	82.159	ng/ul	97
70) Atrazine	16.839	200	3061926	84.951	ng/ul	100
71) Pentachlorophenol	17.022	266	2308021	81.115	ng/ul	99
72) Phenanthrene	17.422	178	13309680	71.015	ng/ul	95
74) Anthracene	17.516	178	12954003	68.978	ng/ul	93
75) 1,2,3,4-Tetrachloroben...	13.422	216	4213269	89.018	ng/ul	100
76) Pentachlorobenzene	14.940	250	4102716	77.861	ng/ul	100
77) Carbazole	17.781	167	12202148	74.699	ng/ul	98
78) Di-n-butylphthalate	18.316	149	12922194	58.116	ng/ul#	94
80) Fluoranthene	19.375	202	13855007	90.154	ng/ul	95
82) Pyrene	19.733	202	13372972	83.641	ng/ul#	91
83) Butylbenzylphthalate	20.598	149	6023725	93.481	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.369	252	4661947	80.906	ng/ul	100
85) Benzo(a)anthracene	21.439	228	13336840	80.095	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.351	149	7730200	82.264	ng/ul	98
87) Chrysene	21.498	228	12327950	78.909	ng/ul	96
89) Di-n-octyl phthalate	22.310	149	11738701	104.917	ng/ul	100
90) Benzo(b)fluoranthene	23.192	252	12047719	86.370	ng/ul	99
91) Benzo(k)fluoranthene	23.239	252	11467157	86.690	ng/ul	99
93) Benzo(a)pyrene	23.845	252	10060924	81.690	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.568	276	12582174	83.479	ng/ul	100
95) Dibenzo(a,h)anthracene	26.586	278	10370464	76.848	ng/ul	100
96) Benzo(g,h,i)perylene	27.368	276	10340171	79.032	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120722\
 Data File : BP012971.D
 Acq On : 07 Dec 2022 15:02
 Operator : CG/JU
 Sample : SSTD08015
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080642

Quant Time: Dec 08 00:00:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 07 23:53:54 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 32.523 ng/uL
RT: 3.393 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion: 88 Resp: 485925
Ion Ratio Lower Upper
88 100
43 28.1 22.4 33.6
58 72.7 58.2 87.4



#5
Pyridine
Concen: 74.529 ng/uL
RT: 3.805 min Scan# 105
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion: 79 Resp: 3286021
Ion Ratio Lower Upper
79 100
52 76.1 60.7 91.1
51 34.4 28.9 43.3









#6
Benzaldehyde
Concen: 55.823 ng/ul
RT: 7.116 min Scan# 60
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion: 77 Resp: 1580816
Ion Ratio Lower Upper
77 100
105 102.8 81.5 122.3
106 99.8 76.9 115.3



#8
Phenol
Concen: 69.007 ng/ul
RT: 7.169 min Scan# 677
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion: 94 Resp: 4129113
Ion Ratio Lower Upper
94 100
65 25.6 21.1 31.7
66 33.0 27.0 40.6









#10
Bis(2-Chloroethyl)ether
Concen: 70.845 ng/ul
RT: 7.405 min Scan# 716
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion: 93 Resp: 3270007
Ion Ratio Lower Upper
93 100
63 73.3 59.8 89.8
95 33.2 26.4 39.6



#12
2-Chlorophenol
Concen: 73.488 ng/ul
RT: 7.546 min Scan# 741
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:128 Resp: 3322033
Ion Ratio Lower Upper
128 100
64 43.0 34.9 52.3
130 33.2 25.8 38.6





#13

2-Methylphenol

Concen: 68.329 ng/ul

RT: 8.428 min Scan# 890

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :





SSTD080642

Tgt Ion:108 Resp: 3196730

Ion Ratio Lower Upper

108 100

107 87.2 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 68.144 ng/ul

RT: 8.516 min Scan# 906

Delta R.T. 0.000 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion: 45 Resp: 4407275


Ion Ratio Lower Upper

45 100

77 14.3 11.0 16.6

79 11.2 9.3 13.9









#16
Acetophenone
Concen: 64.052 ng/ul
RT: 8.816 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:105 Resp: 4694779
Ion Ratio Lower Upper
105 100
77 72.1 59.3 88.9
51 24.1 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 60.907 ng/ul
RT: 8.816 min Scan# 957
Delta R.T. 0.024 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion: 70 Resp: 2269477
Ion Ratio Lower Upper
70 100
42 55.6 45.5 68.3
101 10.4 7.9 11.9
130 23.8 17.4 26.2





#18

4-Methylphenol

Concen: 66.213 ng/ul

RT: 8.769 min Scan# 946

Delta R.T. 0.018 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :





SSTD080642

Tgt Ion:108 Resp: 3459304

Ion Ratio Lower Upper

108 100

107 108.4 87.4 131.2



#19

Hexachloroethane

Concen: 78.180 ng/ul

RT: 9.069 min Scan# 1000

Delta R.T. 0.000 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion:117 Resp: 1305105


Ion Ratio Lower Upper

117 100

201 120.9 93.7 140.5

199 75.2 58.6 88.0









#22
Nitrobenzene
Concen: 80.361 ng/ul
RT: 9.199 min Scan# 1020
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion: 77 Resp: 3580425
Ion Ratio Lower Upper
77 100
123 50.6 38.3 57.5
65 13.3 10.3 15.5



#23
Isophorone
Concen: 74.018 ng/ul
RT: 9.734 min Scan# 1113
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion: 82 Resp: 7044848
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.2
138 19.4 15.3 22.9









#25
2-Nitrophenol
Concen: 84.395 ng/ul
RT: 9.910 min Scan# 1142
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:139 Resp: 2006943
Ion Ratio Lower Upper
139 100
65 43.1 36.6 54.8
109 25.6 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 76.610 ng/ul
RT: 9.969 min Scan# 1153
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:107 Resp: 3577827
Ion Ratio Lower Upper
107 100
121 53.2 41.8 62.6
122 91.2 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 76.762 ng/ul
 RT: 10.204 min Scan# 1193
 Delta R.T. 0.006 min
 Lab File: BP012971.D
 Acq: 07 Dec 2022 15:02




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080642


Tgt Ion: 93 Resp: 4424194
 Ion Ratio Lower Upper
 93 100
 95 33.1 26.0 39.0
 123 11.9 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 77.984 ng/ul
 RT: 10.446 min Scan# 1234
 Delta R.T. 0.006 min
 Lab File: BP012971.D
 Acq: 07 Dec 2022 15:02

Tgt Ion:162 Resp: 3355912
 Ion Ratio Lower Upper
 162 100
 164 66.4 52.5 78.7
 98 31.8 26.6 39.8









#30
Naphthalene
Concen: 75.039 ng/ul
RT: 10.852 min Scan# 1302
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:128 Resp:10282816
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.5 10.4 15.6



#32
4-Chloroaniline
Concen: 72.658 ng/ul
RT: 10.963 min Scan# 1322
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:127 Resp: 4468342
Ion Ratio Lower Upper
127 100
129 32.8 25.6 38.4





#33

Hexachlorobutadiene

Concen: 83.467 ng/ul

RT: 11.128 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:225 Resp: 2262325

Ion Ratio Lower Upper

225 100

223 63.3 49.2 73.8

227 63.5 50.2 75.2



#34

Caprolactam

Concen: 35.771 ng/ul

RT: 11.775 min Scan# 1460

Delta R.T. 0.041 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion:113 Resp: 577291


Ion Ratio Lower Upper

113 100

55 161.0 141.9 212.9

56 119.8 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 70.385 ng/ul
RT: 12.081 min Scan# 1510
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:107 Resp: 3347878
Ion Ratio Lower Upper
107 100
144 28.2 21.4 32.2
142 86.8 65.8 98.6



#36
2-Methylnaphthalene
Concen: 71.403 ng/ul
RT: 12.451 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:142 Resp: 7094314
Ion Ratio Lower Upper
142 100
141 88.7 70.8 106.2









#37
1-Methylnaphthalene
Concen: 71.517 ng/ul
RT: 12.669 min Scan# 1612
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:142 Resp: 7155549
Ion Ratio Lower Upper
142 100
141 90.9 72.7 109.1
116 3.6 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 84.637 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:216 Resp: 4200309
Ion Ratio Lower Upper
216 100
214 79.4 61.9 92.9
179 19.7 15.0 22.6
108 14.1 11.0 16.4









#40
Hexachlorocyclopentadiene
Concen: 118.029 ng/ul
RT: 12.798 min Scan# 1634
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:237 Resp: 2982712
Ion Ratio Lower Upper
237 100
235 62.3 49.7 74.5
272 13.4 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 84.131 ng/ul
RT: 13.057 min Scan# 1678
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:196 Resp: 2861479
Ion Ratio Lower Upper
196 100
198 97.3 77.0 115.6
200 31.7 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 85.029 ng/ul
RT: 13.128 min Scan# 1689
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:196 Resp: 3110659
Ion Ratio Lower Upper
196 100
198 95.0 77.1 115.7
200 31.4 25.4 38.2



#43
1,1'-Biphenyl
Concen: 80.269 ng/ul
RT: 13.457 min Scan# 1746
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:154 Resp: 9262102
Ion Ratio Lower Upper
154 100
153 40.4 31.5 47.3
76 11.8 9.6 14.4





#44

2-Chloronaphthalene

Concen: 80.361 ng/ul

RT: 13.504 min Scan# 1754

Delta R.T. 0.012 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:162 Resp: 7421767

Ion Ratio Lower Upper

162 100

164 33.6 26.3 39.5

127 35.6 29.3 43.9



#45

2-Nitroaniline

Concen: 82.628 ng/ul

RT: 13.710 min Scan# 1789

Delta R.T. 0.012 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion: 65 Resp: 2053976


Ion Ratio Lower Upper

65 100

92 73.6 56.4 84.6

138 126.6 90.3 135.5









#47
Dimethylphthalate
Concen: 75.552 ng/ul
RT: 14.081 min Scan# 1852
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:163 Resp: 9305803
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.0
164 10.5 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 91.104 ng/ul
RT: 14.204 min Scan# 1873
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:165 Resp: 2093446
Ion Ratio Lower Upper
165 100
89 46.8 43.0 64.4
121 18.3 16.1 24.1









#50
Acenaphthylene
Concen: 78.009 ng/ul
RT: 14.345 min Scan# 1897
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:152 Resp:11127809
Ion Ratio Lower Upper
152 100
151 20.7 15.8 23.6
153 13.8 10.6 16.0



#51
3-Nitroaniline
Concen: 72.487 ng/ul
RT: 14.534 min Scan# 1929
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:138 Resp: 1851162
Ion Ratio Lower Upper
138 100
108 9.6 8.6 13.0
92 105.1 88.4 132.6









#52
Acenaphthene
Concen: 75.724 ng/ul
RT: 14.687 min Scan# 1954
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:153 Resp: 7654727
Ion Ratio Lower Upper
153 100
152 47.9 38.2 57.4
154 90.0 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 86.763 ng/ul
RT: 14.734 min Scan# 1963
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:184 Resp: 1426494
Ion Ratio Lower Upper
184 100
63 52.4 48.4 72.6
154 54.3 45.5 68.3









#55
4-Nitrophenol
Concen: 71.406 ng/ul
RT: 14.845 min Scan# 1978
Delta R.T. 0.024 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:109 Resp: 1206136
Ion Ratio Lower Upper
109 100
139 155.4 120.8 181.2
65 126.6 107.9 161.9



#56
Dibenzofuran
Concen: 74.829 ng/ul
RT: 15.022 min Scan# 2012
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:168 Resp:10712844
Ion Ratio Lower Upper
168 100
139 35.1 27.9 41.9





#57

2,4-Dinitrotoluene

Concen: 84.137 ng/ul

RT: 14.987 min Scan# 2006

Delta R.T. 0.012 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:165 Resp: 2893965

Ion Ratio Lower Upper

165 100

63 35.2 31.8 47.6

182 2.9 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 78.481 ng/ul

RT: 15.245 min Scan# 2050

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Tgt Ion:232 Resp: 2703679




Ion Ratio Lower Upper


232 100

131 36.0 31.8 47.6

130 1.9 1.8 2.6

166 27.7 22.9 34.3









#59
Diethylphthalate
Concen: 73.662 ng/ul
RT: 15.445 min Scan# 2084
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:149 Resp: 9042583
Ion Ratio Lower Upper
149 100
177 24.7 18.4 27.6
150 13.0 9.9 14.9



#61
Fluorene
Concen: 66.984 ng/ul
RT: 15.675 min Scan# 2123
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:166 Resp: 7927640
Ion Ratio Lower Upper
166 100
165 96.0 77.4 116.2
167 13.4 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 71.588 ng/ul
RT: 15.663 min Scan# 2121
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:204 Resp: 4376455
Ion Ratio Lower Upper
204 100
206 33.4 26.3 39.5
141 48.5 40.4 60.6



#63
4-Nitroaniline
Concen: 65.896 ng/ul
RT: 15.704 min Scan# 2128
Delta R.T. 0.018 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:138 Resp: 1598231
Ion Ratio Lower Upper
138 100
92 46.6 39.0 58.4
108 68.3 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 84.524 ng/ul
RT: 15.757 min Scan# 2134
Delta R.T. 0.018 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:198 Resp: 1867048
Ion Ratio Lower Upper
198 100
182 3.2 3.3 4.9#
77 17.5 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 77.282 ng/ul
RT: 15.881 min Scan# 2158
Delta R.T. 0.012 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:169 Resp: 7505915
Ion Ratio Lower Upper
169 100
168 66.6 53.0 79.6
167 37.0 29.1 43.7





#68

4-Bromophenyl-phenylether

Concen: 89.462 ng/ul

RT: 16.557 min Scan# 2273

Delta R.T. 0.000 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:248 Resp: 3146426

Ion Ratio Lower Upper

248 100

250 96.8 77.4 116.0

141 59.6 48.3 72.5



#69

Hexachlorobenzene

Concen: 82.159 ng/ul

RT: 16.681 min Scan# 2294

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion:284 Resp: 3609659


Ion Ratio Lower Upper

284 100

142 26.2 22.6 34.0

249 30.1 24.6 37.0





#70

Atrazine

Concen: 84.951 ng/ul

RT: 16.839 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:200 Resp: 3061926

Ion Ratio Lower Upper

200 100

173 25.6 20.5 30.7

215 49.8 39.9 59.9



#71

Pentachlorophenol

Concen: 81.115 ng/ul

RT: 17.022 min Scan# 2352

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion:266 Resp: 2308021


Ion Ratio Lower Upper

266 100

264 62.3 50.9 76.3

268 63.7 51.4 77.2





#72

Phenanthrene

Concen: 71.015 ng/ul

RT: 17.422 min Scan# 2419

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:178 Resp:13309680

Ion Ratio Lower Upper

178 100

179 17.3 12.5 18.7

176 21.5 15.4 23.0



#74

Anthracene

Concen: 68.978 ng/ul

RT: 17.516 min Scan# 2436

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion:178 Resp:12954003


Ion Ratio Lower Upper

178 100

179 18.2 12.5 18.7

176 22.0 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 89.018 ng/uL
RT: 13.422 min Scan# 1739
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:216 Resp: 4213269
Ion Ratio Lower Upper
216 100
214 78.1 62.2 93.2
218 48.8 38.9 58.3



#76
Pentachlorobenzene
Concen: 77.861 ng/uL
RT: 14.940 min Scan# 1998
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:250 Resp: 4102716
Ion Ratio Lower Upper
250 100
248 63.2 50.6 76.0
252 63.9 51.1 76.7









#77
Carbazole
Concen: 74.699 ng/ul
RT: 17.781 min Scan# 2480
Delta R.T. 0.006 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:167 Resp:12202148
Ion Ratio Lower Upper
167 100
166 22.1 17.2 25.8
139 12.5 9.4 14.2



#78
Di-n-butylphthalate
Concen: 58.116 ng/ul
RT: 18.316 min Scan# 2572
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:149 Resp:12922194
Ion Ratio Lower Upper
149 100
150 12.3 7.5 11.3#
104 5.7 3.8 5.6#





#80

Fluoranthene

Concen: 90.154 ng/ul

RT: 19.375 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :





BNA_P

ClientSampleId :

SSTD080642

Tgt Ion:202 Resp:13855007

Ion	Ratio	Lower	Upper
202	100		
101	12.0	8.2	12.2
100	9.0	6.1	9.1



#82

Pyrene

Concen: 83.641 ng/ul

RT: 19.733 min Scan# 2813




Delta R.T. 0.006 min


Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Tgt Ion:202 Resp:13372972

Ion	Ratio	Lower	Upper
202	100		
101	15.6	9.4	14.2#
100	12.6	7.8	11.8#





#83

Butylbenzylphthalate

Concen: 93.481 ng/ul

RT: 20.598 min Scan# 2959

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:149 Resp: 6023725

Ion Ratio Lower Upper

149 100

91 61.2 53.4 80.0

206 25.9 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 80.906 ng/ul

RT: 21.369 min Scan# 3091

Delta R.T. 0.000 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion:252 Resp: 4661947


Ion Ratio Lower Upper

252 100

254 64.9 51.8 77.8

126 9.8 7.4 11.2









#85
Benzo(a)anthracene
Concen: 80.095 ng/ul
RT: 21.439 min Scan# 3103
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02




Instrument :
BNA_P
ClientSampleId :
SSTD080642


Tgt Ion:228 Resp:13336840
Ion Ratio Lower Upper
228 100
229 21.3 16.1 24.1
226 28.9 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 82.264 ng/ul
RT: 21.351 min Scan# 3088
Delta R.T. 0.000 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:149 Resp: 7730200
Ion Ratio Lower Upper
149 100
167 29.0 22.5 33.7
279 5.2 3.8 5.8





#87

Chrysene

Concen: 78.909 ng/ul

RT: 21.498 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:228 Resp:12327950

Ion Ratio Lower Upper

228 100

226 32.7 24.5 36.7

229 21.5 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 104.917 ng/ul




RT: 22.310 min Scan# 3251


Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Tgt Ion:149 Resp:11738701





#90

Benzo(b)fluoranthene

Concen: 86.370 ng/ul

RT: 23.192 min Scan# 3400

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:252 Resp:12047719

Ion Ratio Lower Upper

252 100

253 22.8 17.7 26.5

125 8.7 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 86.690 ng/ul

RT: 23.239 min Scan# 3409

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion:252 Resp:11467157


Ion Ratio Lower Upper

252 100

253 22.8 17.7 26.5

125 8.7 6.8 10.2





#93

Benzo(a)pyrene

Concen: 81.690 ng/ul

RT: 23.845 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02

Instrument :

BNA_P

ClientSampleId :

SSTD080642





Tgt Ion:252 Resp:10060924

Ion Ratio Lower Upper

252 100

253 22.6 17.5 26.3

125 9.6 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 83.479 ng/ul

RT: 26.568 min Scan# 3975

Delta R.T. 0.012 min

Lab File: BP012971.D

Acq: 07 Dec 2022 15:02




Tgt Ion:276 Resp:12582174


Ion Ratio Lower Upper

276 100

138 21.1 17.1 25.7

277 25.1 20.0 30.0









#95
Dibenzo(a,h)anthracene
Concen: 76.848 ng/ul
RT: 26.586 min Scan# 3975
Delta R.T. 0.018 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Instrument :
BNA_P
ClientSampleId :
SSTD080642

Tgt Ion:278 Resp:10370464
Ion Ratio Lower Upper
278 100
139 13.6 11.1 16.7
279 24.0 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 79.032 ng/ul
RT: 27.368 min Scan# 4111
Delta R.T. 0.018 min
Lab File: BP012971.D
Acq: 07 Dec 2022 15:02

Tgt Ion:276 Resp:10340171
Ion Ratio Lower Upper
276 100
138 19.1 15.4 23.2
277 24.5 19.1 28.7

