


Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP120722\  
 Data File : BP012966.D  
 Acq On : 07 Dec 2022 12:11  
 Operator : CG/JU  
 Sample : DFTPP727  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 DFTPP727

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP120722.M  
 Title : SVOA CALIBRATION  
 Last Update : Thu Dec 08 00:35:07 2022



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2478

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 26.6         | 401397     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.6          | 6523       | PASS                |
| 69             | 198             | 0.00            | 100             | 26.4         | 399165     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 2051       | PASS                |
| 127            | 198             | 10              | 80              | 38.5         | 580602     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.3          | 4998       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 1509366    | PASS                |
| 199            | 198             | 5               | 9               | 6.9          | 104303     | PASS                |
| 275            | 198             | 10              | 60              | 28.1         | 424348     | PASS                |
| 365            | 198             | 1               | 100             | 3.6          | 54141      | PASS                |
| 441            | 198             | 0.01            | 100             | 15.3         | 230391     | PASS                |
| 442            | 198             | 50              | 100             | 98.5         | 1487331    | PASS                |
| 443            | 442             | 15              | 24              | 19.6         | 291264     | PASS                |

| m/z    | Abundance |
|--------|-----------|
| 35.90  | 774.0     |
| 37.00  | 159.0     |
| 37.90  | 302.0     |
| 39.00  | 742.0     |
| 40.00  | 3092.0    |
| 41.00  | 1583.0    |
| 42.10  | 627.0     |
| 43.10  | 1677.0    |
| 44.00  | 13908.0   |
| 44.90  | 662.0     |
| 46.00  | 63.0      |
| 46.90  | 240.0     |
| 47.80  | 163.0     |
| 48.90  | 842.0     |
| 49.90  | 200.0     |
| 50.90  | 587.0     |
| 52.10  | 300.0     |
| 53.00  | 267.0     |
| 54.10  | 176.0     |
| 55.00  | 1138.0    |
| 56.00  | 573.0     |
| 57.00  | 598.0     |
| 58.00  | 108.0     |
| 59.10  | 160.0     |
| 59.80  | 208.0     |
| 60.20  | 244.0     |
| 61.10  | 164.0     |
| 61.90  | 67.0      |
| 62.90  | 222.0     |
| 63.80  | 98.0      |
| 65.00  | 225.0     |
| 65.90  | 162.0     |
| 67.00  | 559.0     |
| 67.90  | 341.0     |
| 69.00  | 686.0     |
| 70.00  | 405.0     |
| 71.00  | 636.0     |
| 71.90  | 149.0     |
| 73.00  | 1431.0    |
| 73.80  | 398.0     |
| 75.00  | 388.0     |
| 75.80  | 82.0      |
| 76.00  | 80.0      |
| 76.90  | 464.0     |
| 77.90  | 395.0     |
| 79.00  | 348.0     |
| 80.10  | 163.0     |
| 81.10  | 414.0     |
| 82.00  | 454.0     |
| 83.10  | 598.0     |
| 84.00  | 932.0     |
| 85.00  | 354.0     |
| 85.90  | 461.0     |
| 86.90  | 182.0     |
| 87.60  | 131.0     |
| 88.00  | 85.0      |
| 88.90  | 232.0     |
| 89.90  | 65.0      |
| 91.00  | 551.0     |
| 92.10  | 189.0     |
| 92.90  | 129.0     |
| 93.80  | 107.0     |
| 95.00  | 401.0     |
| 96.00  | 653.0     |
| 97.00  | 263.0     |
| 98.20  | 194.0     |
| 98.90  | 269.0     |
| 99.80  | 192.0     |
| 100.80 | 95.0      |
| 102.00 | 91.0      |
| 103.10 | 233.0     |
| 104.00 | 442.0     |
| 105.10 | 370.0     |
| 105.90 | 169.0     |
| 106.90 | 234.0     |
| 107.90 | 116.0     |
| 109.10 | 227.0     |
| 109.90 | 188.0     |
| 111.20 | 142.0     |
| 111.90 | 174.0     |
| 112.90 | 155.0     |
| 114.00 | 90.0      |
| 115.00 | 330.0     |
| 116.10 | 155.0     |
| 116.90 | 121.0     |
| 117.90 | 92.0      |
| 118.90 | 384.0     |
| 120.10 | 101.0     |
| 120.90 | 184.0     |
| 121.90 | 118.0     |
| 122.90 | 173.0     |
| 124.00 | 105.0     |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |        |
|--------|--------|
| 125.10 | 114.0  |
| 125.90 | 37.0   |
| 127.10 | 91.0   |
| 127.90 | 255.0  |
| 129.00 | 491.0  |
| 130.00 | 172.0  |
| 131.10 | 175.0  |
| 131.80 | 162.0  |
| 133.10 | 775.0  |
| 134.20 | 128.0  |
| 135.00 | 402.0  |
| 136.20 | 72.0   |
| 136.90 | 100.0  |
| 137.80 | 51.0   |
| 138.40 | 20.0   |
| 139.50 | 9.0    |
| 140.00 | 90.0   |
| 141.10 | 195.0  |
| 141.90 | 67.0   |
| 143.10 | 111.0  |
| 144.20 | 89.0   |
| 144.90 | 359.0  |
| 145.60 | 16.0   |
| 146.30 | 110.0  |
| 146.90 | 548.0  |
| 148.10 | 189.0  |
| 148.90 | 270.0  |
| 150.20 | 19.0   |
| 151.10 | 185.0  |
| 152.00 | 176.0  |
| 152.80 | 99.0   |
| 153.90 | 218.0  |
| 154.60 | 49.0   |
| 155.00 | 57.0   |
| 156.00 | 76.0   |
| 156.90 | 128.0  |
| 158.00 | 60.0   |
| 159.00 | 181.0  |
| 160.00 | 40.0   |
| 161.00 | 34.0   |
| 161.70 | 62.0   |
| 162.20 | 58.0   |
| 163.10 | 412.0  |
| 163.80 | 52.0   |
| 165.20 | 223.0  |
| 166.10 | 175.0  |
| 166.90 | 156.0  |
| 167.90 | 146.0  |
| 169.20 | 13.0   |
| 169.90 | 43.0   |
| 171.00 | 13.0   |
| 171.80 | 128.0  |
| 173.10 | 16.0   |
| 173.60 | 11.0   |
| 174.30 | 21.0   |
| 175.00 | 177.0  |
| 175.90 | 56.0   |
| 176.90 | 268.0  |
| 177.90 | 197.0  |
| 179.00 | 192.0  |
| 180.10 | 45.0   |
| 181.10 | 166.0  |
| 182.00 | 23.0   |
| 183.00 | 62.0   |
| 184.00 | 16.0   |
| 185.00 | 108.0  |
| 186.90 | 101.0  |
| 188.20 | 29.0   |
| 188.70 | 73.0   |
| 189.10 | 133.0  |
| 189.90 | 68.0   |
| 191.00 | 917.0  |
| 192.00 | 198.0  |
| 193.00 | 529.0  |
| 193.90 | 149.0  |
| 195.10 | 208.0  |
| 196.80 | 43.0   |
| 197.40 | 2.0    |
| 197.80 | 10.0   |
| 198.50 | 19.0   |
| 199.00 | 81.0   |
| 200.00 | 12.0   |
| 201.20 | 80.0   |
| 202.10 | 32.0   |
| 203.20 | 57.0   |
| 204.30 | 16.0   |
| 205.10 | 113.0  |
| 207.10 | 5898.0 |
| 208.00 | 1769.0 |
| 208.90 | 930.0  |
| 210.00 | 200.0  |
| 211.00 | 161.0  |
| 212.10 | 34.0   |
| 213.10 | 58.0   |
| 213.80 | 55.0   |
| 215.00 | 15.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |        |
|--------|--------|
| 216.90 | 10.0   |
| 217.90 | 15.0   |
| 218.80 | 22.0   |
| 219.80 | 10.0   |
| 220.90 | 423.0  |
| 221.90 | 234.0  |
| 223.10 | 52.0   |
| 223.80 | 12.0   |
| 225.00 | 203.0  |
| 225.80 | 16.0   |
| 226.70 | 27.0   |
| 227.20 | 79.0   |
| 228.10 | 157.0  |
| 229.20 | 73.0   |
| 231.00 | 9.0    |
| 231.30 | 10.0   |
| 232.80 | 23.0   |
| 234.20 | 8.0    |
| 235.20 | 13.0   |
| 235.80 | 8.0    |
| 236.80 | 16.0   |
| 237.40 | 18.0   |
| 239.10 | 68.0   |
| 240.10 | 48.0   |
| 241.00 | 149.0  |
| 242.00 | 14.0   |
| 243.00 | 41.0   |
| 243.80 | 17.0   |
| 245.00 | 25.0   |
| 246.00 | 9.0    |
| 246.70 | 8.0    |
| 247.30 | 13.0   |
| 248.10 | 13.0   |
| 248.80 | 248.0  |
| 250.00 | 73.0   |
| 250.90 | 206.0  |
| 251.80 | 22.0   |
| 253.00 | 73.0   |
| 253.60 | 4.0    |
| 254.60 | 8.0    |
| 254.90 | 7.0    |
| 256.20 | 7.0    |
| 256.90 | 13.0   |
| 257.30 | 11.0   |
| 258.10 | 35.0   |
| 259.60 | 16.0   |
| 260.90 | 9.0    |
| 262.50 | 9.0    |
| 263.80 | 25.0   |
| 264.90 | 289.0  |
| 265.90 | 158.0  |
| 266.80 | 271.0  |
| 267.90 | 176.0  |
| 268.80 | 10.0   |
| 270.10 | 19.0   |
| 271.30 | 28.0   |
| 272.00 | 7.0    |
| 272.50 | 8.0    |
| 273.30 | 14.0   |
| 273.70 | 15.0   |
| 275.00 | 15.0   |
| 276.10 | 15.0   |
| 276.70 | 25.0   |
| 277.10 | 70.0   |
| 278.10 | 17.0   |
| 278.90 | 20.0   |
| 279.80 | 9.0    |
| 281.00 | 1970.0 |
| 281.90 | 620.0  |
| 283.10 | 573.0  |
| 283.90 | 90.0   |
| 284.50 | 10.0   |
| 284.90 | 8.0    |
| 285.80 | 10.0   |
| 286.10 | 9.0    |
| 286.90 | 2.0    |
| 287.70 | 8.0    |
| 289.10 | 13.0   |
| 290.90 | 16.0   |
| 291.70 | 18.0   |
| 292.40 | 11.0   |
| 293.90 | 16.0   |
| 294.90 | 88.0   |
| 296.80 | 32.0   |
| 298.50 | 11.0   |
| 300.30 | 9.0    |
| 302.20 | 20.0   |
| 302.80 | 12.0   |
| 303.40 | 10.0   |
| 304.00 | 4.0    |
| 305.00 | 7.0    |
| 306.60 | 7.0    |
| 308.10 | 9.0    |
| 308.90 | 13.0   |
| 309.70 | 8.0    |
| 311.40 | 22.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |       |
|--------|-------|
| 313.00 | 12.0  |
| 313.60 | 1.0   |
| 314.30 | 16.0  |
| 315.80 | 18.0  |
| 316.80 | 31.0  |
| 317.60 | 16.0  |
| 319.00 | 14.0  |
| 320.80 | 17.0  |
| 322.20 | 24.0  |
| 322.90 | 22.0  |
| 323.90 | 8.0   |
| 324.90 | 49.0  |
| 326.00 | 26.0  |
| 327.00 | 131.0 |
| 327.60 | 1.0   |
| 328.20 | 9.0   |
| 328.70 | 19.0  |
| 330.20 | 9.0   |
| 331.50 | 13.0  |
| 333.30 | 15.0  |
| 334.80 | 9.0   |
| 336.10 | 19.0  |
| 338.10 | 7.0   |
| 339.30 | 16.0  |
| 341.00 | 154.0 |
| 341.90 | 69.0  |
| 342.90 | 77.0  |
| 343.80 | 9.0   |
| 344.30 | 13.0  |
| 345.50 | 13.0  |
| 347.30 | 17.0  |
| 349.00 | 8.0   |
| 350.60 | 7.0   |
| 352.30 | 7.0   |
| 353.30 | 8.0   |
| 353.80 | 9.0   |
| 355.00 | 340.0 |
| 356.00 | 66.0  |
| 356.90 | 7.0   |
| 358.50 | 8.0   |
| 359.90 | 8.0   |
| 361.70 | 10.0  |
| 363.20 | 14.0  |
| 364.70 | 8.0   |
| 366.40 | 7.0   |
| 366.80 | 12.0  |
| 367.70 | 13.0  |
| 368.90 | 7.0   |
| 369.40 | 12.0  |
| 371.10 | 8.0   |
| 372.60 | 10.0  |
| 374.20 | 9.0   |
| 375.80 | 8.0   |
| 377.30 | 7.0   |
| 378.00 | 7.0   |
| 379.20 | 12.0  |
| 380.90 | 14.0  |
| 381.80 | 10.0  |
| 383.30 | 13.0  |
| 384.80 | 3.0   |
| 385.20 | 9.0   |
| 385.90 | 6.0   |
| 386.70 | 19.0  |
| 388.10 | 15.0  |
| 389.60 | 10.0  |
| 390.40 | 9.0   |
| 391.20 | 16.0  |
| 392.80 | 10.0  |
| 393.40 | 9.0   |
| 394.60 | 9.0   |
| 396.20 | 8.0   |
| 397.80 | 9.0   |
| 399.10 | 19.0  |
| 399.60 | 8.0   |
| 400.90 | 18.0  |
| 401.80 | 87.0  |
| 402.50 | 7.0   |
| 404.10 | 8.0   |
| 405.70 | 8.0   |
| 407.30 | 7.0   |
| 408.30 | 14.0  |
| 408.80 | 9.0   |
| 410.30 | 8.0   |
| 411.00 | 10.0  |
| 412.00 | 9.0   |
| 413.50 | 7.0   |
| 414.80 | 141.0 |
| 415.90 | 8.0   |
| 416.60 | 7.0   |
| 417.10 | 15.0  |
| 417.80 | 15.0  |
| 418.20 | 7.0   |
| 419.00 | 3.0   |
| 419.80 | 9.0   |
| 421.40 | 9.0   |
| 421.90 | 5.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |      |
|--------|------|
| 423.00 | 9.0  |
| 424.50 | 8.0  |
| 426.10 | 9.0  |
| 427.70 | 9.0  |
| 429.00 | 19.0 |
| 430.10 | 65.0 |
| 431.00 | 12.0 |
| 432.00 | 18.0 |
| 432.60 | 9.0  |
| 434.00 | 8.0  |
| 435.50 | 7.0  |
| 437.20 | 12.0 |
| 438.80 | 8.0  |
| 440.30 | 9.0  |
| 441.20 | 9.0  |
| 442.00 | 8.0  |
| 443.30 | 22.0 |
| 444.70 | 6.0  |
| 445.20 | 22.0 |
| 446.50 | 8.0  |
| 448.10 | 10.0 |
| 449.60 | 8.0  |
| 450.10 | 12.0 |
| 451.20 | 8.0  |
| 452.70 | 10.0 |
| 454.60 | 13.0 |
| 456.00 | 9.0  |
| 456.70 | 11.0 |
| 457.60 | 8.0  |
| 459.20 | 10.0 |
| 460.70 | 14.0 |
| 461.70 | 3.0  |
| 462.20 | 17.0 |
| 463.10 | 12.0 |
| 463.90 | 8.0  |
| 465.70 | 12.0 |
| 467.10 | 6.0  |
| 468.70 | 8.0  |
| 469.40 | 15.0 |
| 470.20 | 10.0 |
| 471.80 | 9.0  |
| 473.40 | 9.0  |
| 474.90 | 45.0 |
| 475.40 | 10.0 |
| 476.00 | 22.0 |
| 476.70 | 14.0 |
| 477.30 | 26.0 |
| 478.00 | 10.0 |
| 479.60 | 9.0  |
| 481.10 | 9.0  |
| 482.80 | 8.0  |
| 484.40 | 9.0  |
| 485.90 | 8.0  |
| 487.50 | 9.0  |
| 489.10 | 49.0 |
| 490.10 | 10.0 |
| 490.90 | 12.0 |
| 491.70 | 22.0 |
| 492.20 | 8.0  |
| 493.80 | 10.0 |
| 495.40 | 9.0  |
| 497.30 | 17.0 |
| 498.80 | 16.0 |
| 500.10 | 8.0  |
| 501.70 | 6.0  |
| 502.90 | 8.0  |
| 503.30 | 8.0  |
| 504.00 | 9.0  |
| 504.80 | 13.0 |
| 506.70 | 12.0 |
| 508.00 | 8.0  |
| 509.60 | 25.0 |
| 511.00 | 7.0  |
| 512.70 | 8.0  |
| 514.30 | 10.0 |
| 515.90 | 8.0  |
| 516.90 | 9.0  |
| 517.70 | 7.0  |
| 518.90 | 9.0  |
| 519.70 | 4.0  |
| 520.60 | 8.0  |
| 521.20 | 4.0  |
| 522.10 | 8.0  |
| 522.70 | 13.0 |
| 523.70 | 8.0  |
| 524.30 | 8.0  |
| 525.30 | 8.0  |
| 526.80 | 8.0  |
| 528.50 | 9.0  |
| 530.00 | 9.0  |
| 531.60 | 8.0  |
| 533.20 | 7.0  |
| 534.70 | 10.0 |
| 536.30 | 10.0 |
| 537.90 | 9.0  |
| 539.50 | 9.0  |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |      |
|--------|------|
| 541.30 | 20.0 |
| 542.60 | 9.0  |
| 544.20 | 7.0  |
| 545.80 | 9.0  |
| 547.20 | 9.0  |
| 548.30 | 8.0  |
| 548.80 | 11.0 |

**Instrument :**  
BNA\_P  
**ClientSampleId :**  
DFTPP727

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 542.0     |
| 37.00  | 1522.0    |
| 38.10  | 4662.0    |
| 39.00  | 26768.0   |
| 40.00  | 3224.0    |
| 41.00  | 1522.0    |
| 42.10  | 442.0     |
| 43.00  | 1350.0    |
| 44.00  | 10379.0   |
| 45.00  | 1047.0    |
| 45.90  | 169.0     |
| 47.00  | 282.0     |
| 47.90  | 524.0     |
| 49.00  | 3189.0    |
| 50.00  | 100640.0  |
| 51.00  | 402752.0  |
| 52.00  | 20632.0   |
| 53.10  | 803.0     |
| 54.00  | 289.0     |
| 55.00  | 2615.0    |
| 56.00  | 11420.0   |
| 57.00  | 25704.0   |
| 58.00  | 1207.0    |
| 58.90  | 413.0     |
| 59.90  | 477.0     |
| 61.00  | 5129.0    |
| 62.00  | 5445.0    |
| 63.00  | 15305.0   |
| 64.00  | 1999.0    |
| 65.00  | 7950.0    |
| 66.00  | 750.0     |
| 67.00  | 1057.0    |
| 68.00  | 6402.0    |
| 69.00  | 392768.0  |
| 70.00  | 2353.0    |
| 71.10  | 677.0     |
| 72.00  | 192.0     |
| 73.00  | 4554.0    |
| 74.00  | 41144.0   |
| 75.00  | 63640.0   |
| 76.00  | 23432.0   |
| 77.00  | 480576.0  |
| 78.00  | 35512.0   |
| 79.00  | 29936.0   |
| 80.00  | 22880.0   |
| 81.00  | 29848.0   |
| 82.00  | 8006.0    |
| 83.00  | 7779.0    |
| 83.90  | 1689.0    |
| 85.00  | 5370.0    |
| 85.90  | 8989.0    |
| 87.00  | 3984.0    |
| 88.00  | 1731.0    |
| 89.00  | 865.0     |
| 90.00  | 257.0     |
| 91.00  | 7352.0    |
| 92.00  | 8277.0    |
| 93.00  | 51520.0   |
| 94.00  | 3687.0    |
| 94.90  | 1504.0    |
| 96.00  | 3035.0    |
| 97.10  | 1101.0    |
| 98.00  | 39840.0   |
| 99.00  | 37760.0   |
| 100.00 | 3125.0    |
| 101.00 | 18936.0   |
| 102.00 | 997.0     |
| 103.00 | 6370.0    |
| 104.00 | 11159.0   |
| 105.00 | 11808.0   |
| 106.00 | 3358.0    |
| 107.00 | 149376.0  |
| 108.00 | 23744.0   |
| 109.00 | 3940.0    |
| 110.00 | 272192.0  |
| 111.00 | 43696.0   |
| 112.00 | 4953.0    |
| 113.00 | 1951.0    |
| 114.20 | 636.0     |
| 114.90 | 671.0     |
| 116.00 | 7330.0    |
| 117.00 | 122280.0  |
| 118.00 | 10060.0   |
| 119.00 | 1299.0    |
| 120.00 | 1866.0    |
| 121.10 | 443.0     |
| 122.00 | 9166.0    |
| 123.00 | 15157.0   |
| 124.00 | 7522.0    |
| 124.90 | 6216.0    |
| 126.10 | 2313.0    |
| 127.00 | 541248.0  |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727



|        |           |
|--------|-----------|
| 128.00 | 42864.0   |
| 129.00 | 199168.0  |
| 130.00 | 17440.0   |
| 131.00 | 3261.0    |
| 132.00 | 2115.0    |
| 133.00 | 1147.0    |
| 134.00 | 5588.0    |
| 135.00 | 16656.0   |
| 136.00 | 6895.0    |
| 137.00 | 10437.0   |
| 137.90 | 1822.0    |
| 138.80 | 1302.0    |
| 139.90 | 2258.0    |
| 141.00 | 24688.0   |
| 142.00 | 8615.0    |
| 143.00 | 6157.0    |
| 144.00 | 1786.0    |
| 144.90 | 1323.0    |
| 146.00 | 5226.0    |
| 147.00 | 13964.0   |
| 148.00 | 34264.0   |
| 149.00 | 8022.0    |
| 149.90 | 1649.0    |
| 151.00 | 3501.0    |
| 153.00 | 9027.0    |
| 154.00 | 6082.0    |
| 155.00 | 15011.0   |
| 156.00 | 24184.0   |
| 157.00 | 4655.0    |
| 158.00 | 5614.0    |
| 158.90 | 4180.0    |
| 160.00 | 8512.0    |
| 161.00 | 13500.0   |
| 162.00 | 4215.0    |
| 163.00 | 1284.0    |
| 164.00 | 1647.0    |
| 165.00 | 10099.0   |
| 166.00 | 8546.0    |
| 167.00 | 59136.0   |
| 168.00 | 43272.0   |
| 169.00 | 5615.0    |
| 170.00 | 1836.0    |
| 171.00 | 2749.0    |
| 172.00 | 5723.0    |
| 173.00 | 6362.0    |
| 174.00 | 12262.0   |
| 175.00 | 23424.0   |
| 176.00 | 7054.0    |
| 177.00 | 10258.0   |
| 177.90 | 3385.0    |
| 179.00 | 42432.0   |
| 180.00 | 33936.0   |
| 181.00 | 15311.0   |
| 182.00 | 2667.0    |
| 183.00 | 1068.0    |
| 184.00 | 3333.0    |
| 185.00 | 20664.0   |
| 186.00 | 167552.0  |
| 187.00 | 49424.0   |
| 188.00 | 5439.0    |
| 189.00 | 9767.0    |
| 190.00 | 1584.0    |
| 191.00 | 5185.0    |
| 192.00 | 14894.0   |
| 193.00 | 16033.0   |
| 194.10 | 3237.0    |
| 195.00 | 1772.0    |
| 196.00 | 35896.0   |
| 196.90 | 5801.0    |
| 198.00 | 1272832.0 |
| 199.00 | 88960.0   |
| 200.00 | 7357.0    |
| 201.50 | 5079.0    |
| 203.00 | 8573.0    |
| 204.00 | 44528.0   |
| 205.00 | 78856.0   |
| 206.00 | 331456.0  |
| 207.00 | 51752.0   |
| 208.00 | 11513.0   |
| 209.00 | 4551.0    |
| 210.20 | 5663.0    |
| 211.10 | 12725.0   |
| 212.90 | 1034.0    |
| 213.90 | 587.0     |
| 215.00 | 3379.0    |
| 216.00 | 7171.0    |
| 217.00 | 91120.0   |
| 218.00 | 11111.0   |
| 219.10 | 1135.0    |
| 221.00 | 55232.0   |
| 223.00 | 19160.0   |
| 224.00 | 188992.0  |
| 225.00 | 48216.0   |
| 226.00 | 6175.0    |
| 227.00 | 75400.0   |
| 228.00 | 10000.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |          |
|--------|----------|
| 229.00 | 16153.0  |
| 230.00 | 2205.0   |
| 231.00 | 6915.0   |
| 232.00 | 1178.0   |
| 233.00 | 1359.0   |
| 234.00 | 5026.0   |
| 235.00 | 5599.0   |
| 236.00 | 3182.0   |
| 237.00 | 8113.0   |
| 238.00 | 1513.0   |
| 239.00 | 2486.0   |
| 240.00 | 2576.0   |
| 241.00 | 4756.0   |
| 242.00 | 10768.0  |
| 243.00 | 10829.0  |
| 244.00 | 152448.0 |
| 245.00 | 20008.0  |
| 246.00 | 25224.0  |
| 247.00 | 5924.0   |
| 248.00 | 1080.0   |
| 249.00 | 5173.0   |
| 250.00 | 926.0    |
| 251.00 | 1347.0   |
| 252.00 | 1185.0   |
| 253.00 | 3138.0   |
| 254.00 | 4640.0   |
| 255.00 | 734336.0 |
| 256.00 | 112200.0 |
| 257.00 | 10531.0  |
| 258.00 | 40328.0  |
| 259.00 | 7393.0   |
| 260.00 | 887.0    |
| 261.00 | 1134.0   |
| 261.90 | 144.0    |
| 262.80 | 417.0    |
| 263.80 | 1011.0   |
| 265.00 | 17496.0  |
| 266.00 | 2544.0   |
| 266.90 | 681.0    |
| 268.10 | 156.0    |
| 269.00 | 255.0    |
| 269.90 | 864.0    |
| 271.00 | 1817.0   |
| 272.00 | 1769.0   |
| 273.00 | 21320.0  |
| 274.00 | 57352.0  |
| 275.00 | 329920.0 |
| 276.00 | 47192.0  |
| 277.00 | 26672.0  |
| 278.00 | 5291.0   |
| 279.00 | 1213.0   |
| 280.00 | 56.0     |
| 281.00 | 1780.0   |
| 282.00 | 1000.0   |
| 283.00 | 3218.0   |
| 284.00 | 1833.0   |
| 285.00 | 4135.0   |
| 286.10 | 843.0    |
| 286.90 | 83.0     |
| 287.80 | 407.0    |
| 288.90 | 934.0    |
| 290.10 | 1005.0   |
| 291.00 | 736.0    |
| 291.90 | 1205.0   |
| 293.00 | 6718.0   |
| 294.00 | 1598.0   |
| 295.00 | 1814.0   |
| 296.00 | 84624.0  |
| 297.00 | 12152.0  |
| 298.00 | 1010.0   |
| 299.00 | 436.0    |
| 299.70 | 98.0     |
| 300.90 | 1106.0   |
| 302.00 | 1675.0   |
| 303.00 | 8872.0   |
| 304.00 | 3170.0   |
| 305.00 | 576.0    |
| 305.90 | 237.0    |
| 306.90 | 168.0    |
| 307.90 | 1249.0   |
| 308.90 | 1164.0   |
| 309.90 | 1328.0   |
| 310.90 | 284.0    |
| 312.00 | 250.0    |
| 312.90 | 638.0    |
| 314.00 | 4629.0   |
| 315.00 | 9401.0   |
| 316.00 | 5388.0   |
| 317.00 | 872.0    |
| 318.10 | 146.0    |
| 318.90 | 192.0    |
| 319.90 | 385.0    |
| 321.00 | 2855.0   |
| 322.10 | 1574.0   |
| 323.00 | 28008.0  |
| 324.10 | 5246.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |         |
|--------|---------|
| 325.00 | 583.0   |
| 325.90 | 803.0   |
| 327.00 | 5164.0  |
| 328.00 | 3005.0  |
| 328.90 | 716.0   |
| 330.00 | 110.0   |
| 330.80 | 215.0   |
| 331.90 | 2242.0  |
| 333.00 | 2909.0  |
| 334.00 | 18248.0 |
| 335.00 | 4688.0  |
| 336.00 | 612.0   |
| 337.10 | 159.0   |
| 337.90 | 17.0    |
| 338.90 | 463.0   |
| 340.00 | 370.0   |
| 341.00 | 3809.0  |
| 342.00 | 1146.0  |
| 342.90 | 167.0   |
| 343.90 | 49.0    |
| 345.10 | 161.0   |
| 345.90 | 7102.0  |
| 346.90 | 1140.0  |
| 348.10 | 285.0   |
| 350.00 | 318.0   |
| 351.00 | 592.0   |
| 352.00 | 8179.0  |
| 353.00 | 5853.0  |
| 354.00 | 8829.0  |
| 355.00 | 2183.0  |
| 355.90 | 340.0   |
| 357.00 | 298.0   |
| 357.60 | 156.0   |
| 357.90 | 165.0   |
| 359.10 | 600.0   |
| 359.90 | 138.0   |
| 361.10 | 114.0   |
| 362.10 | 69.0    |
| 362.90 | 88.0    |
| 365.00 | 36864.0 |
| 366.00 | 6142.0  |
| 367.00 | 606.0   |
| 367.80 | 5.0     |
| 368.30 | 9.0     |
| 368.90 | 75.0    |
| 369.90 | 764.0   |
| 371.00 | 2487.0  |
| 372.00 | 13653.0 |
| 373.00 | 3834.0  |
| 373.90 | 550.0   |
| 374.80 | 11.0    |
| 375.70 | 15.0    |
| 376.60 | 107.0   |
| 377.10 | 375.0   |
| 378.00 | 81.0    |
| 379.00 | 20.0    |
| 379.40 | 8.0     |
| 380.40 | 2.0     |
| 381.00 | 72.0    |
| 381.90 | 85.0    |
| 383.00 | 3842.0  |
| 383.90 | 1625.0  |
| 385.00 | 643.0   |
| 386.10 | 116.0   |
| 387.40 | 10.0    |
| 388.10 | 19.0    |
| 389.10 | 103.0   |
| 390.00 | 1916.0  |
| 390.90 | 1513.0  |
| 392.00 | 1117.0  |
| 392.90 | 134.0   |
| 393.70 | 7.0     |
| 394.90 | 146.0   |
| 395.60 | 80.0    |
| 396.20 | 72.0    |
| 396.80 | 151.0   |
| 397.10 | 170.0   |
| 397.80 | 53.0    |
| 399.40 | 18.0    |
| 399.80 | 7.0     |
| 401.00 | 763.0   |
| 402.00 | 6225.0  |
| 403.00 | 10650.0 |
| 404.00 | 3561.0  |
| 405.00 | 886.0   |
| 405.90 | 206.0   |
| 406.90 | 5.0     |
| 408.10 | 74.0    |
| 409.10 | 45.0    |
| 410.00 | 425.0   |
| 411.00 | 38.0    |
| 412.50 | 9.0     |
| 413.00 | 6.0     |
| 414.00 | 28.0    |
| 414.90 | 574.0   |
| 415.80 | 308.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |          |
|--------|----------|
| 416.70 | 53.0     |
| 417.60 | 21.0     |
| 418.10 | 20.0     |
| 418.90 | 57.0     |
| 421.00 | 7810.0   |
| 422.00 | 7550.0   |
| 423.00 | 54392.0  |
| 424.00 | 20104.0  |
| 425.00 | 3208.0   |
| 426.00 | 208.0    |
| 426.80 | 60.0     |
| 428.30 | 8.0      |
| 429.20 | 17.0     |
| 430.20 | 23.0     |
| 430.90 | 21.0     |
| 433.10 | 47.0     |
| 433.70 | 9.0      |
| 434.50 | 11.0     |
| 435.80 | 5.0      |
| 436.30 | 9.0      |
| 436.80 | 4.0      |
| 437.70 | 250.0    |
| 439.00 | 286.0    |
| 439.60 | 131.0    |
| 441.00 | 137856.0 |
| 442.00 | 902592.0 |
| 443.00 | 184384.0 |
| 444.00 | 18128.0  |
| 445.10 | 1094.0   |
| 445.90 | 15.0     |
| 446.90 | 8.0      |
| 448.90 | 17.0     |
| 450.30 | 9.0      |
| 452.00 | 8.0      |
| 452.70 | 7.0      |
| 453.40 | 8.0      |
| 455.00 | 9.0      |
| 456.40 | 9.0      |
| 457.30 | 6.0      |
| 458.20 | 8.0      |
| 459.00 | 20.0     |
| 459.60 | 7.0      |
| 460.90 | 114.0    |
| 461.60 | 12.0     |
| 462.70 | 7.0      |
| 464.50 | 13.0     |
| 466.00 | 7.0      |
| 467.60 | 7.0      |
| 469.20 | 8.0      |
| 470.70 | 8.0      |
| 471.50 | 14.0     |
| 472.30 | 8.0      |
| 473.40 | 3.0      |
| 473.90 | 10.0     |
| 475.10 | 75.0     |
| 475.80 | 47.0     |
| 476.90 | 16.0     |
| 478.70 | 9.0      |
| 479.30 | 13.0     |
| 480.20 | 10.0     |
| 481.60 | 8.0      |
| 483.30 | 8.0      |
| 484.90 | 7.0      |
| 485.30 | 15.0     |
| 486.40 | 8.0      |
| 488.00 | 10.0     |
| 489.20 | 13.0     |
| 489.70 | 9.0      |
| 490.80 | 17.0     |
| 492.90 | 27.0     |
| 494.20 | 6.0      |
| 495.70 | 11.0     |
| 497.50 | 10.0     |
| 499.00 | 9.0      |
| 500.70 | 8.0      |
| 501.60 | 0.0      |
| 502.40 | 10.0     |
| 503.70 | 9.0      |
| 504.80 | 18.0     |
| 505.40 | 7.0      |
| 506.90 | 8.0      |
| 508.50 | 8.0      |
| 510.10 | 8.0      |
| 511.70 | 9.0      |
| 513.20 | 8.0      |
| 514.80 | 9.0      |
| 516.40 | 9.0      |
| 516.60 | 11.0     |
| 517.70 | 22.0     |
| 518.30 | 18.0     |
| 519.20 | 7.0      |
| 519.60 | 7.0      |
| 521.30 | 10.0     |
| 522.60 | 8.0      |
| 524.20 | 7.0      |
| 525.80 | 8.0      |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |      |
|--------|------|
| 526.70 | 4.0  |
| 527.50 | 18.0 |
| 529.00 | 21.0 |
| 530.20 | 20.0 |
| 532.10 | 15.0 |
| 533.70 | 7.0  |
| 534.70 | 12.0 |
| 535.20 | 9.0  |
| 535.70 | 9.0  |
| 536.90 | 11.0 |
| 538.40 | 7.0  |
| 540.20 | 15.0 |
| 541.40 | 14.0 |
| 543.10 | 8.0  |
| 544.60 | 8.0  |
| 546.20 | 8.0  |
| 548.00 | 18.0 |
| 548.70 | 0.0  |
| 549.30 | 59.0 |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 662.0     |
| 37.00  | 2114.0    |
| 38.00  | 5372.0    |
| 39.00  | 30616.0   |
| 40.00  | 3434.0    |
| 41.00  | 1586.0    |
| 42.00  | 628.0     |
| 43.10  | 1060.0    |
| 44.00  | 8544.0    |
| 45.00  | 993.0     |
| 46.00  | 170.0     |
| 46.80  | 335.0     |
| 48.00  | 292.0     |
| 49.00  | 3793.0    |
| 50.00  | 116824.0  |
| 51.00  | 455360.0  |
| 52.00  | 22672.0   |
| 53.00  | 757.0     |
| 54.00  | 344.0     |
| 55.00  | 2654.0    |
| 56.00  | 13682.0   |
| 57.00  | 29464.0   |
| 57.90  | 1255.0    |
| 58.90  | 440.0     |
| 60.00  | 537.0     |
| 61.00  | 5663.0    |
| 62.00  | 6418.0    |
| 63.00  | 19592.0   |
| 64.00  | 2564.0    |
| 65.00  | 9109.0    |
| 65.90  | 526.0     |
| 66.90  | 1066.0    |
| 68.00  | 7815.0    |
| 69.00  | 456192.0  |
| 70.00  | 2960.0    |
| 71.00  | 644.0     |
| 72.10  | 355.0     |
| 73.00  | 4519.0    |
| 74.00  | 49712.0   |
| 75.00  | 75704.0   |
| 76.00  | 27080.0   |
| 77.00  | 566912.0  |
| 78.00  | 39328.0   |
| 79.00  | 33728.0   |
| 80.00  | 26104.0   |
| 81.00  | 39272.0   |
| 82.00  | 9636.0    |
| 83.00  | 9388.0    |
| 83.90  | 1837.0    |
| 85.00  | 6621.0    |
| 86.00  | 11106.0   |
| 87.00  | 5272.0    |
| 87.90  | 2138.0    |
| 89.00  | 1097.0    |
| 89.90  | 478.0     |
| 90.90  | 8172.0    |
| 92.00  | 10052.0   |
| 93.00  | 63576.0   |
| 93.90  | 4467.0    |
| 95.00  | 1339.0    |
| 96.00  | 3604.0    |
| 97.00  | 1417.0    |
| 98.00  | 49496.0   |
| 99.00  | 43424.0   |
| 100.00 | 3609.0    |
| 101.00 | 22816.0   |
| 102.00 | 1299.0    |
| 103.00 | 7092.0    |
| 104.00 | 14269.0   |
| 105.00 | 13081.0   |
| 106.00 | 5036.0    |
| 107.00 | 183104.0  |
| 108.00 | 27392.0   |
| 109.10 | 5060.0    |
| 110.00 | 338176.0  |
| 111.00 | 49792.0   |
| 112.00 | 6825.0    |
| 113.00 | 2051.0    |
| 113.80 | 410.0     |
| 114.20 | 461.0     |
| 115.00 | 901.0     |
| 116.00 | 9392.0    |
| 117.00 | 150656.0  |
| 118.00 | 12457.0   |
| 119.00 | 1722.0    |
| 120.00 | 2490.0    |
| 121.00 | 982.0     |
| 122.00 | 10907.0   |
| 123.00 | 18048.0   |
| 124.00 | 9286.0    |
| 125.00 | 8011.0    |
| 126.10 | 3125.0    |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |           |
|--------|-----------|
| 127.00 | 674432.0  |
| 128.00 | 49888.0   |
| 129.00 | 246656.0  |
| 130.00 | 21216.0   |
| 131.00 | 4714.0    |
| 132.00 | 2573.0    |
| 132.90 | 1570.0    |
| 134.00 | 7088.0    |
| 135.00 | 20376.0   |
| 136.00 | 8489.0    |
| 137.00 | 12392.0   |
| 138.00 | 2246.0    |
| 139.00 | 1531.0    |
| 140.00 | 3573.0    |
| 141.00 | 34128.0   |
| 142.00 | 12090.0   |
| 143.00 | 7424.0    |
| 144.00 | 2054.0    |
| 145.00 | 2276.0    |
| 146.00 | 5130.0    |
| 147.00 | 17968.0   |
| 148.00 | 43424.0   |
| 149.00 | 8821.0    |
| 150.10 | 2557.0    |
| 151.10 | 4425.0    |
| 151.70 | 3418.0    |
| 153.00 | 11761.0   |
| 154.00 | 7840.0    |
| 155.00 | 19808.0   |
| 156.00 | 30872.0   |
| 157.10 | 6533.0    |
| 158.00 | 7339.0    |
| 159.00 | 4426.0    |
| 160.00 | 11087.0   |
| 161.00 | 17224.0   |
| 162.00 | 4557.0    |
| 162.90 | 1232.0    |
| 164.00 | 1725.0    |
| 165.00 | 14102.0   |
| 166.00 | 10664.0   |
| 167.00 | 77952.0   |
| 168.00 | 50232.0   |
| 169.00 | 6804.0    |
| 170.00 | 2486.0    |
| 170.90 | 3382.0    |
| 171.90 | 6466.0    |
| 173.00 | 10248.0   |
| 174.00 | 15765.0   |
| 175.00 | 32256.0   |
| 176.10 | 9419.0    |
| 177.00 | 14126.0   |
| 178.00 | 4328.0    |
| 179.00 | 55744.0   |
| 180.00 | 44976.0   |
| 181.00 | 21008.0   |
| 182.00 | 3867.0    |
| 183.00 | 1995.0    |
| 184.00 | 4685.0    |
| 185.00 | 28904.0   |
| 186.00 | 232832.0  |
| 187.00 | 67600.0   |
| 188.00 | 7427.0    |
| 189.00 | 12790.0   |
| 190.00 | 2750.0    |
| 190.90 | 6893.0    |
| 192.00 | 20296.0   |
| 193.00 | 22936.0   |
| 194.00 | 4809.0    |
| 195.00 | 2551.0    |
| 196.00 | 49608.0   |
| 197.00 | 9322.0    |
| 198.00 | 1808896.0 |
| 199.00 | 126448.0  |
| 200.00 | 8912.0    |
| 201.50 | 8339.0    |
| 203.00 | 13096.0   |
| 204.00 | 64856.0   |
| 205.00 | 109560.0  |
| 206.00 | 474432.0  |
| 207.00 | 67792.0   |
| 208.00 | 15493.0   |
| 209.00 | 5438.0    |
| 210.20 | 7105.0    |
| 211.00 | 17928.0   |
| 212.90 | 1236.0    |
| 213.80 | 651.0     |
| 215.00 | 5087.0    |
| 216.00 | 9560.0    |
| 217.00 | 127384.0  |
| 218.00 | 16384.0   |
| 219.10 | 1905.0    |
| 221.00 | 77856.0   |
| 221.80 | 17720.0   |
| 223.00 | 29824.0   |
| 224.00 | 272064.0  |
| 225.00 | 71264.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |           |
|--------|-----------|
| 226.00 | 7704.0    |
| 227.00 | 105688.0  |
| 228.00 | 15261.0   |
| 229.00 | 22448.0   |
| 230.10 | 2660.0    |
| 231.00 | 9551.0    |
| 232.00 | 1615.0    |
| 233.10 | 2045.0    |
| 234.00 | 6697.0    |
| 235.00 | 7810.0    |
| 236.00 | 5713.0    |
| 237.00 | 11149.0   |
| 238.00 | 1628.0    |
| 239.00 | 5065.0    |
| 240.00 | 3759.0    |
| 241.00 | 6866.0    |
| 242.00 | 14111.0   |
| 243.10 | 16314.0   |
| 244.00 | 222144.0  |
| 245.10 | 31024.0   |
| 246.00 | 40248.0   |
| 247.00 | 8872.0    |
| 248.00 | 2766.0    |
| 249.00 | 7774.0    |
| 250.00 | 1361.0    |
| 251.00 | 1590.0    |
| 252.10 | 2089.0    |
| 253.00 | 4586.0    |
| 254.10 | 6762.0    |
| 255.00 | 1106944.0 |
| 256.00 | 164672.0  |
| 257.00 | 13004.0   |
| 258.00 | 61768.0   |
| 259.00 | 10696.0   |
| 260.00 | 1803.0    |
| 261.00 | 1698.0    |
| 261.90 | 363.0     |
| 263.00 | 577.0     |
| 264.00 | 1705.0    |
| 265.00 | 22808.0   |
| 266.00 | 3332.0    |
| 267.00 | 804.0     |
| 268.00 | 482.0     |
| 269.00 | 303.0     |
| 270.00 | 1232.0    |
| 271.00 | 3069.0    |
| 272.00 | 3577.0    |
| 273.00 | 33432.0   |
| 274.00 | 91696.0   |
| 275.00 | 504896.0  |
| 276.00 | 70224.0   |
| 277.00 | 41688.0   |
| 278.00 | 8175.0    |
| 279.00 | 1601.0    |
| 280.00 | 355.0     |
| 281.10 | 2299.0    |
| 282.00 | 1291.0    |
| 283.00 | 4923.0    |
| 284.00 | 3260.0    |
| 285.00 | 6535.0    |
| 286.00 | 1357.0    |
| 287.10 | 294.0     |
| 288.00 | 347.0     |
| 288.90 | 1801.0    |
| 289.90 | 1649.0    |
| 291.00 | 1220.0    |
| 292.00 | 2254.0    |
| 293.00 | 9642.0    |
| 294.00 | 2476.0    |
| 295.00 | 2500.0    |
| 296.00 | 137536.0  |
| 297.00 | 19664.0   |
| 298.00 | 1640.0    |
| 298.90 | 509.0     |
| 299.90 | 154.0     |
| 301.00 | 1781.0    |
| 302.00 | 2656.0    |
| 303.00 | 17008.0   |
| 304.00 | 4946.0    |
| 305.00 | 747.0     |
| 305.90 | 229.0     |
| 306.90 | 269.0     |
| 308.00 | 2195.0    |
| 309.00 | 1293.0    |
| 310.00 | 2388.0    |
| 311.00 | 336.0     |
| 312.10 | 293.0     |
| 313.10 | 1173.0    |
| 314.00 | 6490.0    |
| 315.00 | 14944.0   |
| 316.00 | 8901.0    |
| 317.00 | 1616.0    |
| 317.90 | 100.0     |
| 319.00 | 213.0     |
| 319.90 | 502.0     |
| 321.10 | 4398.0    |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727



|        |         |
|--------|---------|
| 322.00 | 2087.0  |
| 323.00 | 46728.0 |
| 324.00 | 8678.0  |
| 325.10 | 867.0   |
| 325.90 | 1081.0  |
| 327.00 | 9542.0  |
| 327.90 | 4635.0  |
| 329.00 | 885.0   |
| 330.10 | 160.0   |
| 331.10 | 277.0   |
| 332.00 | 3567.0  |
| 333.00 | 4837.0  |
| 334.00 | 32024.0 |
| 335.00 | 7180.0  |
| 335.90 | 1025.0  |
| 336.90 | 170.0   |
| 337.80 | 4.0     |
| 339.00 | 658.0   |
| 340.10 | 537.0   |
| 341.00 | 5529.0  |
| 342.10 | 1589.0  |
| 343.00 | 543.0   |
| 344.00 | 180.0   |
| 345.00 | 121.0   |
| 346.00 | 11489.0 |
| 347.00 | 2175.0  |
| 347.90 | 435.0   |
| 349.00 | 42.0    |
| 349.90 | 448.0   |
| 351.00 | 897.0   |
| 352.00 | 15670.0 |
| 353.00 | 10442.0 |
| 354.00 | 15648.0 |
| 355.00 | 3070.0  |
| 356.00 | 308.0   |
| 357.10 | 276.0   |
| 358.10 | 329.0   |
| 359.00 | 969.0   |
| 359.90 | 354.0   |
| 360.90 | 253.0   |
| 361.90 | 87.0    |
| 362.90 | 146.0   |
| 365.00 | 64208.0 |
| 366.00 | 9829.0  |
| 367.00 | 851.0   |
| 368.00 | 54.0    |
| 368.90 | 87.0    |
| 370.00 | 2125.0  |
| 371.00 | 3979.0  |
| 372.00 | 26208.0 |
| 373.10 | 6847.0  |
| 374.00 | 648.0   |
| 374.70 | 111.0   |
| 375.50 | 18.0    |
| 377.10 | 737.0   |
| 378.10 | 154.0   |
| 378.90 | 46.0    |
| 379.80 | 24.0    |
| 381.00 | 54.0    |
| 381.90 | 162.0   |
| 383.00 | 7473.0  |
| 384.00 | 2398.0  |
| 385.00 | 800.0   |
| 386.00 | 292.0   |
| 386.80 | 5.0     |
| 387.60 | 19.0    |
| 390.10 | 3812.0  |
| 391.00 | 2854.0  |
| 392.10 | 1422.0  |
| 392.80 | 313.0   |
| 394.00 | 8.0     |
| 394.90 | 648.0   |
| 395.80 | 294.0   |
| 396.90 | 301.0   |
| 397.70 | 145.0   |
| 398.90 | 13.0    |
| 399.70 | 0.0     |
| 401.00 | 1635.0  |
| 402.00 | 11515.0 |
| 403.00 | 16800.0 |
| 404.00 | 6247.0  |
| 405.00 | 1541.0  |
| 406.00 | 215.0   |
| 407.90 | 105.0   |
| 408.80 | 75.0    |
| 410.00 | 621.0   |
| 411.00 | 99.0    |
| 411.80 | 22.0    |
| 413.00 | 14.0    |
| 414.90 | 771.0   |
| 415.90 | 183.0   |
| 417.00 | 61.0    |
| 417.90 | 18.0    |
| 418.80 | 141.0   |
| 419.90 | 311.0   |
| 421.00 | 14389.0 |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |           |
|--------|-----------|
| 422.00 | 14007.0   |
| 423.00 | 101528.0  |
| 424.00 | 29864.0   |
| 425.00 | 4167.0    |
| 425.90 | 245.0     |
| 427.00 | 65.0      |
| 427.80 | 7.0       |
| 428.20 | 5.0       |
| 429.10 | 72.0      |
| 430.80 | 16.0      |
| 431.70 | 18.0      |
| 433.10 | 86.0      |
| 433.70 | 7.0       |
| 435.50 | 21.0      |
| 437.00 | 22.0      |
| 438.00 | 329.0     |
| 438.90 | 289.0     |
| 439.90 | 559.0     |
| 441.00 | 271552.0  |
| 442.00 | 1763328.0 |
| 443.00 | 341376.0  |
| 444.00 | 30584.0   |
| 445.00 | 2178.0    |
| 446.00 | 21.0      |
| 447.60 | 9.0       |
| 449.40 | 7.0       |
| 451.10 | 16.0      |
| 452.50 | 8.0       |
| 453.90 | 9.0       |
| 455.60 | 9.0       |
| 456.20 | 15.0      |
| 457.00 | 8.0       |
| 457.50 | 13.0      |
| 458.00 | 15.0      |
| 459.00 | 19.0      |
| 460.30 | 9.0       |
| 460.90 | 50.0      |
| 461.80 | 11.0      |
| 462.90 | 5.0       |
| 463.70 | 21.0      |
| 465.10 | 9.0       |
| 466.60 | 10.0      |
| 468.10 | 7.0       |
| 469.80 | 8.0       |
| 471.30 | 12.0      |
| 473.20 | 25.0      |
| 475.00 | 125.0     |
| 475.80 | 56.0      |
| 477.60 | 9.0       |
| 479.20 | 13.0      |
| 480.80 | 9.0       |
| 482.40 | 10.0      |
| 483.40 | 2.0       |
| 484.00 | 9.0       |
| 485.30 | 13.0      |
| 487.10 | 9.0       |
| 489.30 | 36.0      |
| 490.10 | 17.0      |
| 491.00 | 22.0      |
| 492.00 | 11.0      |
| 492.90 | 18.0      |
| 494.90 | 9.0       |
| 496.20 | 10.0      |
| 497.70 | 7.0       |
| 498.10 | 8.0       |
| 499.60 | 5.0       |
| 500.60 | 11.0      |
| 501.20 | 7.0       |
| 501.90 | 6.0       |
| 502.80 | 10.0      |
| 504.50 | 7.0       |
| 506.00 | 8.0       |
| 507.40 | 6.0       |
| 507.60 | 7.0       |
| 509.20 | 12.0      |
| 510.70 | 7.0       |
| 512.30 | 8.0       |
| 513.90 | 9.0       |
| 515.40 | 9.0       |
| 516.90 | 10.0      |
| 519.00 | 20.0      |
| 520.10 | 10.0      |
| 521.70 | 8.0       |
| 523.20 | 11.0      |
| 524.90 | 6.0       |
| 526.40 | 10.0      |
| 528.00 | 9.0       |
| 529.60 | 10.0      |
| 530.00 | 6.0       |
| 530.90 | 15.0      |
| 532.70 | 10.0      |
| 534.00 | 15.0      |
| 534.40 | 11.0      |
| 536.00 | 14.0      |
| 537.20 | 20.0      |
| 537.60 | 20.0      |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |      |
|--------|------|
| 539.10 | 22.0 |
| 540.60 | 10.0 |
| 542.10 | 8.0  |
| 543.70 | 8.0  |
| 545.20 | 10.0 |
| 546.90 | 8.0  |
| 548.70 | 24.0 |
| 549.30 | 45.0 |

**Instrument :**  
BNA\_P  
**ClientSampleId :**  
DFTPP727

| m/z    | Abundance |
|--------|-----------|
| 36.10  | 878.0     |
| 37.00  | 1584.0    |
| 38.10  | 3894.0    |
| 39.10  | 24240.0   |
| 40.00  | 3220.0    |
| 41.00  | 1726.0    |
| 41.90  | 563.0     |
| 43.00  | 966.0     |
| 44.00  | 11145.0   |
| 45.00  | 1123.0    |
| 45.90  | 116.0     |
| 46.90  | 266.0     |
| 47.90  | 388.0     |
| 49.00  | 3209.0    |
| 50.00  | 86608.0   |
| 51.00  | 347840.0  |
| 52.00  | 17072.0   |
| 53.00  | 886.0     |
| 54.10  | 125.0     |
| 55.00  | 2100.0    |
| 56.00  | 9660.0    |
| 57.00  | 23608.0   |
| 58.00  | 1181.0    |
| 59.10  | 445.0     |
| 60.10  | 440.0     |
| 61.00  | 4460.0    |
| 62.00  | 5167.0    |
| 63.00  | 12782.0   |
| 64.00  | 2397.0    |
| 65.00  | 7390.0    |
| 66.00  | 706.0     |
| 67.00  | 826.0     |
| 68.00  | 6375.0    |
| 69.00  | 350592.0  |
| 70.00  | 2056.0    |
| 71.10  | 493.0     |
| 72.10  | 287.0     |
| 73.10  | 4540.0    |
| 74.00  | 36192.0   |
| 75.00  | 58520.0   |
| 76.00  | 20784.0   |
| 77.00  | 439040.0  |
| 78.00  | 30032.0   |
| 79.00  | 25208.0   |
| 80.00  | 19712.0   |
| 81.00  | 29080.0   |
| 82.00  | 6899.0    |
| 83.00  | 6663.0    |
| 83.90  | 1269.0    |
| 85.00  | 4707.0    |
| 86.00  | 7646.0    |
| 87.00  | 3499.0    |
| 87.90  | 1601.0    |
| 89.00  | 1071.0    |
| 90.20  | 380.0     |
| 91.00  | 6866.0    |
| 92.00  | 7130.0    |
| 93.00  | 46720.0   |
| 93.90  | 3390.0    |
| 95.00  | 919.0     |
| 96.00  | 2919.0    |
| 97.00  | 809.0     |
| 98.00  | 36800.0   |
| 99.00  | 32184.0   |
| 100.00 | 2796.0    |
| 101.00 | 18224.0   |
| 102.00 | 1134.0    |
| 103.00 | 5400.0    |
| 104.00 | 9922.0    |
| 105.00 | 10898.0   |
| 106.00 | 3452.0    |
| 107.00 | 135744.0  |
| 108.00 | 20832.0   |
| 109.00 | 3815.0    |
| 110.00 | 264512.0  |
| 111.00 | 38080.0   |
| 112.00 | 4467.0    |
| 113.10 | 1928.0    |
| 113.90 | 486.0     |
| 115.00 | 829.0     |
| 116.00 | 7834.0    |
| 117.00 | 116704.0  |
| 118.00 | 9167.0    |
| 119.00 | 1206.0    |
| 120.00 | 2063.0    |
| 121.00 | 816.0     |
| 122.00 | 9424.0    |
| 123.00 | 15300.0   |
| 124.00 | 6747.0    |
| 125.00 | 6575.0    |
| 126.00 | 2789.0    |
| 127.00 | 526400.0  |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |           |
|--------|-----------|
| 128.00 | 37656.0   |
| 129.00 | 192320.0  |
| 130.00 | 15402.0   |
| 131.00 | 3099.0    |
| 132.00 | 1676.0    |
| 133.00 | 1258.0    |
| 134.00 | 5091.0    |
| 135.00 | 15432.0   |
| 136.00 | 6583.0    |
| 137.00 | 9925.0    |
| 138.00 | 1763.0    |
| 139.00 | 1559.0    |
| 140.00 | 2293.0    |
| 141.00 | 24528.0   |
| 142.00 | 8630.0    |
| 143.00 | 6362.0    |
| 144.00 | 1956.0    |
| 145.00 | 1874.0    |
| 146.00 | 4674.0    |
| 147.00 | 13801.0   |
| 148.00 | 34040.0   |
| 149.00 | 7418.0    |
| 149.90 | 2174.0    |
| 151.10 | 3824.0    |
| 151.70 | 2291.0    |
| 153.00 | 8950.0    |
| 154.00 | 5912.0    |
| 155.00 | 16217.0   |
| 156.00 | 25008.0   |
| 157.00 | 4747.0    |
| 157.90 | 5006.0    |
| 159.00 | 3849.0    |
| 160.00 | 9381.0    |
| 161.00 | 13270.0   |
| 161.90 | 3812.0    |
| 162.90 | 1270.0    |
| 164.10 | 1611.0    |
| 165.00 | 11188.0   |
| 166.00 | 9142.0    |
| 167.00 | 63664.0   |
| 168.00 | 37952.0   |
| 169.00 | 5916.0    |
| 170.00 | 2271.0    |
| 171.00 | 2323.0    |
| 172.00 | 4622.0    |
| 173.00 | 7561.0    |
| 174.00 | 12532.0   |
| 175.00 | 24576.0   |
| 176.10 | 7601.0    |
| 177.00 | 10996.0   |
| 178.00 | 3711.0    |
| 179.00 | 47312.0   |
| 180.00 | 32728.0   |
| 181.00 | 16680.0   |
| 182.00 | 2488.0    |
| 183.10 | 1364.0    |
| 184.00 | 3864.0    |
| 185.00 | 23712.0   |
| 186.00 | 182848.0  |
| 187.00 | 52720.0   |
| 188.00 | 5688.0    |
| 189.00 | 10082.0   |
| 190.00 | 1944.0    |
| 191.00 | 5759.0    |
| 192.00 | 16249.0   |
| 193.00 | 18672.0   |
| 194.00 | 3838.0    |
| 195.00 | 2409.0    |
| 196.00 | 40224.0   |
| 198.00 | 1446400.0 |
| 199.00 | 97744.0   |
| 200.00 | 7723.0    |
| 201.40 | 5310.0    |
| 203.00 | 9849.0    |
| 204.00 | 51568.0   |
| 205.00 | 89072.0   |
| 206.00 | 383232.0  |
| 207.00 | 56616.0   |
| 208.00 | 12872.0   |
| 209.00 | 4324.0    |
| 210.10 | 5659.0    |
| 211.00 | 14398.0   |
| 213.00 | 828.0     |
| 213.80 | 422.0     |
| 215.00 | 4027.0    |
| 216.00 | 8190.0    |
| 217.00 | 99112.0   |
| 218.00 | 12436.0   |
| 219.00 | 1299.0    |
| 221.10 | 68440.0   |
| 221.90 | 12992.0   |
| 223.00 | 24336.0   |
| 224.00 | 217408.0  |
| 225.00 | 54712.0   |
| 226.10 | 6328.0    |
| 227.00 | 87832.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |          |
|--------|----------|
| 228.00 | 12038.0  |
| 229.00 | 19784.0  |
| 230.00 | 2313.0   |
| 231.00 | 8592.0   |
| 232.00 | 1118.0   |
| 233.00 | 1456.0   |
| 234.00 | 5358.0   |
| 235.00 | 6903.0   |
| 236.00 | 4753.0   |
| 237.00 | 7798.0   |
| 238.10 | 1346.0   |
| 239.00 | 4068.0   |
| 239.90 | 2959.0   |
| 240.90 | 5002.0   |
| 242.00 | 11577.0  |
| 243.00 | 12970.0  |
| 244.10 | 187840.0 |
| 245.00 | 24544.0  |
| 246.00 | 33208.0  |
| 247.00 | 6763.0   |
| 248.00 | 1598.0   |
| 249.00 | 6716.0   |
| 249.90 | 1075.0   |
| 251.00 | 1387.0   |
| 252.00 | 1818.0   |
| 253.00 | 4065.0   |
| 254.00 | 5947.0   |
| 255.00 | 932352.0 |
| 256.00 | 137536.0 |
| 257.00 | 10367.0  |
| 258.00 | 49680.0  |
| 259.00 | 7677.0   |
| 259.90 | 1583.0   |
| 261.00 | 1770.0   |
| 262.10 | 272.0    |
| 262.90 | 671.0    |
| 264.10 | 1256.0   |
| 265.00 | 21648.0  |
| 266.00 | 3120.0   |
| 266.90 | 634.0    |
| 267.90 | 432.0    |
| 268.80 | 165.0    |
| 270.00 | 929.0    |
| 271.00 | 2524.0   |
| 272.10 | 2796.0   |
| 273.00 | 28168.0  |
| 274.00 | 76296.0  |
| 275.00 | 438272.0 |
| 276.00 | 56192.0  |
| 277.00 | 34752.0  |
| 278.00 | 5486.0   |
| 279.00 | 1553.0   |
| 279.90 | 135.0    |
| 281.00 | 1953.0   |
| 282.00 | 1333.0   |
| 283.10 | 4263.0   |
| 284.00 | 2568.0   |
| 285.00 | 6673.0   |
| 286.10 | 1144.0   |
| 287.10 | 149.0    |
| 287.90 | 440.0    |
| 288.90 | 1486.0   |
| 290.00 | 1372.0   |
| 291.00 | 850.0    |
| 292.00 | 1606.0   |
| 293.00 | 7356.0   |
| 293.90 | 2733.0   |
| 295.00 | 2463.0   |
| 296.00 | 124192.0 |
| 297.00 | 16312.0  |
| 297.90 | 720.0    |
| 299.00 | 438.0    |
| 300.00 | 175.0    |
| 301.00 | 1501.0   |
| 302.00 | 2015.0   |
| 303.00 | 14344.0  |
| 304.00 | 4202.0   |
| 305.00 | 704.0    |
| 306.00 | 85.0     |
| 307.00 | 221.0    |
| 307.90 | 1575.0   |
| 309.00 | 1283.0   |
| 310.00 | 1227.0   |
| 310.90 | 267.0    |
| 311.90 | 580.0    |
| 313.00 | 1284.0   |
| 314.00 | 6067.0   |
| 315.00 | 13429.0  |
| 316.00 | 7925.0   |
| 317.00 | 1526.0   |
| 317.90 | 72.0     |
| 319.00 | 425.0    |
| 320.00 | 407.0    |
| 321.00 | 4411.0   |
| 322.00 | 2288.0   |
| 323.00 | 42120.0  |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |         |
|--------|---------|
| 324.00 | 7834.0  |
| 325.10 | 794.0   |
| 326.00 | 1029.0  |
| 327.00 | 7997.0  |
| 328.00 | 4558.0  |
| 329.00 | 520.0   |
| 330.10 | 291.0   |
| 330.90 | 236.0   |
| 332.10 | 3523.0  |
| 333.00 | 3405.0  |
| 334.00 | 27680.0 |
| 335.00 | 7248.0  |
| 336.00 | 971.0   |
| 337.10 | 122.0   |
| 338.00 | 13.0    |
| 338.90 | 692.0   |
| 339.90 | 567.0   |
| 341.00 | 5280.0  |
| 341.90 | 1296.0  |
| 342.90 | 240.0   |
| 344.90 | 125.0   |
| 345.90 | 10638.0 |
| 347.00 | 2217.0  |
| 347.90 | 174.0   |
| 348.90 | 33.0    |
| 349.60 | 174.0   |
| 351.00 | 750.0   |
| 352.00 | 14014.0 |
| 353.00 | 9814.0  |
| 354.00 | 14946.0 |
| 355.00 | 3018.0  |
| 356.00 | 551.0   |
| 357.00 | 323.0   |
| 358.00 | 338.0   |
| 359.00 | 918.0   |
| 360.10 | 261.0   |
| 361.00 | 285.0   |
| 361.90 | 94.0    |
| 362.90 | 77.0    |
| 365.00 | 61352.0 |
| 366.00 | 9940.0  |
| 367.00 | 556.0   |
| 368.10 | 68.0    |
| 368.70 | 87.0    |
| 369.90 | 1217.0  |
| 370.90 | 3452.0  |
| 372.00 | 26352.0 |
| 373.00 | 6501.0  |
| 374.10 | 778.0   |
| 374.70 | 109.0   |
| 376.10 | 13.0    |
| 376.90 | 614.0   |
| 378.20 | 153.0   |
| 378.90 | 165.0   |
| 380.80 | 35.0    |
| 381.90 | 151.0   |
| 383.00 | 6922.0  |
| 384.00 | 2153.0  |
| 385.00 | 625.0   |
| 386.00 | 153.0   |
| 387.10 | 18.0    |
| 388.30 | 16.0    |
| 389.00 | 185.0   |
| 390.00 | 2874.0  |
| 391.10 | 2636.0  |
| 392.00 | 1906.0  |
| 393.00 | 288.0   |
| 393.80 | 39.0    |
| 394.60 | 201.0   |
| 394.90 | 231.0   |
| 396.10 | 108.0   |
| 396.90 | 320.0   |
| 397.70 | 83.0    |
| 399.10 | 97.0    |
| 401.00 | 1414.0  |
| 402.00 | 11550.0 |
| 403.00 | 16904.0 |
| 404.00 | 6046.0  |
| 404.90 | 1211.0  |
| 405.80 | 382.0   |
| 406.70 | 21.0    |
| 407.20 | 6.0     |
| 407.80 | 231.0   |
| 409.00 | 63.0    |
| 410.10 | 561.0   |
| 411.10 | 75.0    |
| 412.20 | 8.0     |
| 412.80 | 12.0    |
| 413.90 | 12.0    |
| 415.00 | 949.0   |
| 415.90 | 235.0   |
| 417.00 | 142.0   |
| 418.00 | 13.0    |
| 418.80 | 60.0    |
| 420.10 | 305.0   |
| 421.00 | 15460.0 |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727

|        |           |
|--------|-----------|
| 422.00 | 13797.0   |
| 423.00 | 101760.0  |
| 424.00 | 25544.0   |
| 425.00 | 3420.0    |
| 426.00 | 289.0     |
| 427.00 | 68.0      |
| 427.60 | 16.0      |
| 429.10 | 47.0      |
| 430.10 | 8.0       |
| 430.20 | 8.0       |
| 431.10 | 10.0      |
| 432.80 | 12.0      |
| 434.20 | 11.0      |
| 435.30 | 11.0      |
| 436.00 | 6.0       |
| 436.70 | 12.0      |
| 437.10 | 0.0       |
| 437.90 | 409.0     |
| 439.10 | 442.0     |
| 441.00 | 281792.0  |
| 442.00 | 1796096.0 |
| 443.00 | 348032.0  |
| 444.00 | 30848.0   |
| 445.00 | 1694.0    |
| 445.90 | 35.0      |
| 446.80 | 10.0      |
| 448.30 | 7.0       |
| 449.80 | 9.0       |
| 451.40 | 8.0       |
| 453.00 | 9.0       |
| 454.70 | 8.0       |
| 456.20 | 9.0       |
| 457.90 | 18.0      |
| 459.50 | 13.0      |
| 460.30 | 0.0       |
| 461.00 | 126.0     |
| 461.60 | 8.0       |
| 462.40 | 8.0       |
| 463.00 | 11.0      |
| 464.00 | 7.0       |
| 465.70 | 9.0       |
| 467.20 | 8.0       |
| 468.80 | 9.0       |
| 470.30 | 8.0       |
| 471.20 | 11.0      |
| 471.90 | 10.0      |
| 473.60 | 10.0      |
| 474.50 | 1.0       |
| 475.00 | 103.0     |
| 476.00 | 19.0      |
| 476.90 | 18.0      |
| 478.20 | 9.0       |
| 479.90 | 11.0      |
| 481.30 | 9.0       |
| 481.90 | 6.0       |
| 482.90 | 11.0      |
| 484.50 | 11.0      |
| 486.10 | 9.0       |
| 487.60 | 9.0       |
| 488.80 | 7.0       |
| 489.30 | 38.0      |
| 490.80 | 8.0       |
| 491.70 | 10.0      |
| 492.30 | 10.0      |
| 493.40 | 0.0       |
| 493.90 | 7.0       |
| 494.70 | 5.0       |
| 495.60 | 15.0      |
| 496.40 | 9.0       |
| 497.10 | 9.0       |
| 498.60 | 8.0       |
| 500.30 | 9.0       |
| 501.70 | 10.0      |
| 503.70 | 16.0      |
| 504.80 | 8.0       |
| 506.50 | 11.0      |
| 508.10 | 10.0      |
| 509.70 | 22.0      |
| 511.20 | 9.0       |
| 512.80 | 10.0      |
| 514.40 | 10.0      |
| 516.00 | 9.0       |
| 517.00 | 7.0       |
| 517.60 | 9.0       |
| 519.20 | 9.0       |
| 520.40 | 21.0      |
| 521.80 | 12.0      |
| 524.00 | 15.0      |
| 525.50 | 8.0       |
| 527.00 | 10.0      |
| 528.60 | 10.0      |
| 530.10 | 8.0       |
| 531.60 | 24.0      |
| 532.70 | 7.0       |
| 533.50 | 16.0      |
| 534.30 | 0.0       |

Instrument :  
BNA\_P  
ClientSampleId :  
DFTPP727



|        |      |
|--------|------|
| 534.90 | 42.0 |
| 535.60 | 11.0 |
| 536.30 | 13.0 |
| 536.90 | 18.0 |
| 538.00 | 8.0  |
| 538.90 | 3.0  |
| 539.70 | 14.0 |
| 541.20 | 9.0  |
| 542.80 | 10.0 |
| 544.30 | 9.0  |
| 544.80 | 7.0  |
| 545.70 | 8.0  |
| 546.20 | 10.0 |
| 547.40 | 21.0 |
| 549.10 | 9.0  |

**Instrument :**  
BNA\_P  
**ClientSampleId :**  
DFTPP727