


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120821\
 Data File : BP008273.D
 Acq On : 08 Dec 2021 10:10
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Dec 11 03:53:01 2021




AutoFind: Scans 2492, 2493, 2494; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.6	38763	PASS
68	69	0.00	2	1.7	673	PASS
69	198	0.00	100	33.8	39013	PASS
70	69	0.00	2	0.6	245	PASS
127	198	10	80	39.6	45645	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	115257	PASS
199	198	5	9	6.8	7827	PASS
275	198	10	60	28.9	33309	PASS
365	198	1	100	3.7	4249	PASS
441	198	0.01	100	13.9	16043	PASS
442	442	100	100	100.0	105848	PASS
443	442	15	24	20.1	21246	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120821\
 Data File : BP008273.D
 Acq On : 08 Dec 2021 10:10
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Dec 09 10:17:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 09 01:43:36 2021
 Response via : Initial Calibration



(70) Pentachlorophenol (C)

16.863min (+ 0.012) 265004.78 ng




response 103019

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	64.05
264.00	61.90	62.76
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120821\
 Data File : BP008273.D
 Acq On : 08 Dec 2021 10:10
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Dec 09 10:17:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 09 01:43:36 2021
 Response via : Initial Calibration



TIC: BP008273.D\data.ms

(77) Benzidine

19.404min (+ 0.006) 52521.42 ng

response 781686

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	12.90	14.36
183.00	11.40	12.22
0.00	0.00	0.00

DDT Breakdown

Instrument :
BNA_P
ClientSampleId :
DFTPP

Date	Instrument Name	DFTPP Data File
12/8/2021	BNA_P	BP008273.D
Compound Name	Response	Retention Time
DDT	414479	20.61
DDD	9488	20.186
DDE	1228	19.686
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10716	425195	2.52