

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP121019\
 Data File : BP001294.D
 Acq On : 10 Dec 2019 10:43
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 11 08:04:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP112719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 05 12:34:47 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	60	0.00
2	1,4-Dioxane	0.563	0.498	11.5	54	0.00
3	Pyridine	1.562	1.385	11.3	54	0.00
4	n-Nitrosodimethylamine	0.676	0.849	-25.6#	79	0.01
5 S	2-Fluorophenol	1.205	1.209	-0.3	60	0.00
6	Aniline	2.129	2.086	2.0	61	0.00
7 S	Phenol-d6	1.606	1.637	-1.9	63	0.00
8	2-Chlorophenol	1.247	1.252	-0.4	61	0.00
9	Benzaldehyde	1.044	0.966	7.5	61	0.00
10 C	Phenol	1.827	1.844	-0.9	62	0.00
11	bis(2-Chloroethyl)ether	1.423	1.349	5.2	59	0.00
12	1,3-Dichlorobenzene	1.478	1.541	-4.3	64	0.00
13 C	1,4-Dichlorobenzene	1.489	1.560	-4.8	64	0.00
14	1,2-Dichlorobenzene	1.402	1.491	-6.3	65	0.00
15	Benzyl Alcohol	1.318	1.554	-17.9	73	0.00
16	2,2'-oxybis(1-Chloropropane	2.287	2.165	5.3	59	0.00
17	2-Methylphenol	1.144	1.126	1.6	61	0.00
18	Hexachloroethane	0.616	0.676	-9.7	67	0.00
19 P	n-Nitroso-di-n-propylamine	1.159	1.192	-2.8	65	0.00
20	3+4-Methylphenols	1.442	1.494	-3.6	64	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	58	0.00
22	Acetophenone	0.588	0.637	-8.3	65	0.00
23 S	Nitrobenzene-d5	0.461	0.521	-13.0	67	0.00
24	Nitrobenzene	0.458	0.506	-10.5	65	0.00
25	Isophorone	0.827	0.850	-2.8	63	0.00
26 C	2-Nitrophenol	0.168	0.181	-7.7	62	0.00
27	2,4-Dimethylphenol	0.266	0.268	-0.8	60	0.00
28	bis(2-Chloroethoxy)methane	0.519	0.507	2.3	59	0.00
29 C	2,4-Dichlorophenol	0.303	0.332	-9.6	65	0.00
30	1,2,4-Trichlorobenzene	0.354	0.397	-12.1	66	0.00
31	Naphthalene	1.021	1.075	-5.3	62	0.00
32	Benzoic acid	0.192	0.181	5.7	54	0.03
33	4-Chloroaniline	0.443	0.446	-0.7	60	0.00
34 C	Hexachlorobutadiene	0.222	0.277	-24.8#	73	0.00
35	Caprolactam	0.104	0.101	2.9	60	0.02
36 C	4-Chloro-3-methylphenol	0.359	0.388	-8.1	65	0.00
37	2-Methylnaphthalene	0.697	0.739	-6.0	63	0.00
38	1-Methylnaphthalene	0.658	0.691	-5.0	63	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	61	0.00
40	1,2,4,5-Tetrachlorobenzene	0.630	0.722	-14.6	71	0.00
41 P	Hexachlorocyclopentadiene	0.291	0.345	-18.6	70	0.00
42 S	2,4,6-Tribromophenol	0.192	0.246	-28.1#	79	0.00
43 C	2,4,6-Trichlorophenol	0.412	0.453	-10.0	68	0.00
44	2,4,5-Trichlorophenol	0.426	0.467	-9.6	68	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP121019\
 Data File : BP001294.D
 Acq On : 10 Dec 2019 10:43
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 11 08:04:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP112719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 05 12:34:47 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.367	1.578	-15.4	70	0.00
46	1,1'-Biphenyl	1.546	1.657	-7.2	66	0.00
47	2-Chloronaphthalene	1.235	1.324	-7.2	66	0.00
48	2-Nitroaniline	0.484	0.543	-12.2	71	0.00
49	Acenaphthylene	1.851	1.944	-5.0	65	0.00
50	Dimethylphthalate	1.496	1.639	-9.6	68	0.00
51	2,6-Dinitrotoluene	0.320	0.335	-4.7	65	0.00
52 C	Acenaphthene	1.113	1.169	-5.0	65	0.00
53	3-Nitroaniline	0.360	0.364	-1.1	64	0.00
54 P	2,4-Dinitrophenol	0.112	0.160	-42.9#	91	0.00
55	Dibenzofuran	1.673	1.830	-9.4	67	0.00
56 P	4-Nitrophenol	0.255	0.261	-2.4	64	0.00
57	2,4-Dinitrotoluene	0.401	0.465	-16.0	70	0.00
58	Fluorene	1.340	1.483	-10.7	68	0.00
59	2,3,4,6-Tetrachlorophenol	0.351	0.396	-12.8	70	0.00
60	Diethylphthalate	1.549	1.701	-9.8	69	0.00
61	4-Chlorophenyl-phenylether	0.683	0.774	-13.3	70	0.00
62	4-Nitroaniline	0.417	0.421	-1.0	64	0.00
63	Azobenzene	1.674	1.778	-6.2	66	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	64	0.00
65	4,6-Dinitro-2-methylphenol	0.080	0.088	-10.0	73	0.00
66 c	n-Nitrosodiphenylamine	0.608	0.631	-3.8	69	0.00
67	4-Bromophenyl-phenylether	0.217	0.234	-7.8	70	0.00
68	Hexachlorobenzene	0.239	0.262	-9.6	73	0.00
69	Atrazine	0.186	0.194	-4.3	68	0.00
70 C	Pentachlorophenol	0.124	0.145	-16.9	75	0.00
71	Phenanthrene	1.051	1.117	-6.3	70	0.00
72	Anthracene	1.036	1.114	-7.5	70	0.00
73	Carbazole	0.943	0.998	-5.8	69	0.00
74	Di-n-butylphthalate	1.268	1.399	-10.3	70	0.00
75 C	Fluoranthene	1.113	1.250	-12.3	71	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
77	Benzidine	0.516	0.475	7.9	51	0.00
78	Pyrene	1.575	1.587	-0.8	72	0.00
79 S	Terphenyl-d14	1.081	1.203	-11.3	78	0.00
80	Butylbenzylphthalate	0.728	0.770	-5.8	73	0.00
81	Benzo(a)anthracene	1.387	1.435	-3.5	72	0.00
82	3,3'-Dichlorobenzidine	0.458	0.487	-6.3	70	0.00
83	Chrysene	1.242	1.340	-7.9	74	0.00
84	Bis(2-ethylhexyl)phthalate	1.000	1.099	-9.9	73	0.00
85 c	Di-n-octyl phthalate	1.777	1.792	-0.8	67	0.00
86	Indeno(1,2,3-cd)pyrene	1.463	1.366	6.6	64	0.00
87 I	Perylene-d12	1.000	1.000	0.0	61	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP121019\
Data File : BP001294.D
Acq On : 10 Dec 2019 10:43
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 11 08:04:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP112719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 05 12:34:47 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.222	1.275	-4.3	65	0.00
89	Benzo(k)fluoranthene	1.177	1.300	-10.5	69	0.00
90 C	Benzo(a)pyrene	1.125	1.188	-5.6	66	0.00
91	Dibenzo(a,h)anthracene	1.101	1.131	-2.7	64	-0.01
92	Benzo(q,h,i)perylene	1.087	1.076	1.0	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1