

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP121020\
 Data File : BP004213.D
 Acq On : 10 Dec 2020 09:51
 Operator : CG/JU
 Sample : SSTD00581
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581

Quant Time: Dec 10 12:23:58 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP121020MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 10 11:56:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	346267	20.00	ng/ul	0.00
18) Naphthalene-d8	10.62	136	1357666	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.47	164	800912	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.23	188	1712688	20.00	ng/ul	0.00
78) Chrysene-d12	21.32	240	1861249	20.00	ng/ul	0.00
86) Perylene-d12	23.67	264	1867187	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.33	96	21491	2.29	ng/uL	0.00
5) Phenol-d5	6.99	99	127175	4.01	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.16	67	96585	5.04	ng/ul	0.00
9) 2-Chlorophenol-d4	7.36	132	96594	4.08	ng/ul	0.00
13) 4-Methylphenol-d8	8.53	113	91990	3.79	ng/ul	0.00
19) Nitrobenzene-d5	8.98	128	52846	4.58	ng/ul	0.00
22) 2-Nitrophenol-d4	9.70	143	34303	3.62	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.24	165	83428	3.82	ng/ul	0.00
29) 4-Chloroaniline-d4	10.76	131	106781	4.06	ng/ul	0.00
44) Dimethylphthalate-d6	13.88	166	299652	4.64	ng/ul	0.00
47) Acenaphthylene-d8	14.16	160	391455	4.54	ng/ul	0.00
52) 4-Nitrophenol-d4	14.65	143	30889	2.88	ng/ul	0.00
58) Fluorene-d10	15.47	176	284663	5.02	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.58	200	24368	2.94	ng/ul	0.00
71) Anthracene-d10	17.33	188	424359	4.83	ng/ul	0.00
79) Pyrene-d10	19.57	212	474088	4.58	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.52	264	451929	4.38	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.36	88	21981	2.163	ng/uL 94
4) Benzaldehyde	6.98	77	90060	3.955	ng/ul 100
6) Phenol	7.02	94	143366	4.229	ng/ul 99
8) Bis(2-Chloroethyl)ether	7.26	93	131712	4.768	ng/ul 99
10) 2-Chlorophenol	7.39	128	108270	4.318	ng/ul 99
11) 2-Methylphenol	8.26	108	92248	3.705	ng/ul 99
12) 2,2'-oxybis(1-Chloropropan	8.36	45	160905	4.832	ng/ul 99
14) Acetophenone	8.65	105	185485	4.673	ng/ul 99
15) N-Nitroso-di-n-propylamine	8.63	70	80998	4.123	ng/ul 97
16) 4-Methylphenol	8.59	108	105483	3.923	ng/ul 97
17) Hexachloroethane	8.91	117	54821	4.937	ng/ul 98
20) Nitrobenzene	9.02	77	137201	4.621	ng/ul 95
21) Isophorone	9.55	82	208538	3.946	ng/ul 99
23) 2-Nitrophenol	9.73	139	41196	3.685	ng/ul 96
24) 2,4-Dimethylphenol	9.79	107	110180	4.274	ng/ul 100
25) Bis(2-Chloroethoxy)methane	10.04	93	159210	4.608	ng/ul 98
27) 2,4-Dichlorophenol	10.27	162	91744	4.125	ng/ul 98
28) Naphthalene	10.67	128	409826	4.962	ng/ul 100
30) 4-Chloroaniline	10.78	127	107558	4.055	ng/ul 99
31) Hexachlorobutadiene	10.97	225	81036	4.896	ng/ul 97
32) Caprolactam	11.52	113	16786	2.537	ng/ul 98
33) 4-Chloro-3-methylphenol	11.90	107	77052	3.548	ng/ul 99
34) 2-Methylnaphthalene	12.29	142	268319	4.726	ng/ul 100

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP121020\
 Data File : BP004213.D
 Acq On : 10 Dec 2020 09:51
 Operator : CG/JU
 Sample : SSTD00581
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581

Quant Time: Dec 10 12:23:58 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP121020MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 10 11:56:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 1-Methylnaphthalene	12.51	142	318676	5.941	ng/ul	97
37) 1,2,4,5-Tetrachlorobenzene	12.66	216	149118	4.899	ng/ul	98
38) Hexachlorocyclopentadiene	12.65	237	73360	4.357	ng/ul	99
39) 2,4,6-Trichlorophenol	12.90	196	52294	3.432	ng/ul	98
40) 2,4,5-Trichlorophenol	12.96	196	62287	3.671	ng/ul	94
41) 1,1'-Biphenyl	13.30	154	353954	4.939	ng/ul	99
42) 2-Chloronaphthalene	13.35	162	280233	4.916	ng/ul	99
43) 2-Nitroaniline	13.54	65	44364	3.437	ng/ul	98
45) Dimethylphthalate	13.93	163	315313	4.671	ng/ul	100
46) 2,6-Dinitrotoluene	14.04	165	46247	3.645	ng/ul#	92
48) Acenaphthylene	14.19	152	397276	4.752	ng/ul	100
49) 3-Nitroaniline	14.37	138	35200	2.882	ng/ul	95
50) Acenaphthene	14.54	153	288346	4.886	ng/ul	98
51) 2,4-Dinitrophenol	14.57	184	14268	2.804	ng/ul	96
53) 4-Nitrophenol	14.67	109	26442	3.234	ng/ul	89
54) Dibenzofuran	14.87	168	414316	4.982	ng/ul	98
55) 2,4-Dinitrotoluene	14.83	165	64064	3.559	ng/ul	92
56) 2,3,4,6-Tetrachlorophenol	15.10	232	47333	3.428	ng/ul#	96
57) Diethylphthalate	15.30	149	277480	4.376	ng/ul	100
59) Fluorene	15.53	166	329951	4.948	ng/ul	98
60) 4-Chlorophenyl-phenylether	15.52	204	174176	4.937	ng/ul	96
61) 4-Nitroaniline	15.53	138	39903	2.874	ng/ul	94
64) 4,6-Dinitro-2-methylphenol	15.59	198	26736	2.880	ng/ul#	91
65) N-Nitrosodiphenylamine	15.74	169	251329	4.497	ng/ul	99
66) 4-Bromophenyl-phenylether	16.42	248	98598	4.572	ng/ul	99
67) Hexachlorobenzene	16.53	284	126316	4.982	ng/ul	98
68) Atrazine	16.69	200	65715	3.393	ng/ul	99
69) Pentachlorophenol	16.88	266	32498	2.749	ng/ul	91
70) Phenanthrene	17.27	178	540563	4.981	ng/ul	100
72) Anthracene	17.36	178	521558	4.800	ng/ul	99
73) 1,2,3,4-Tetrachlorobenzene	13.27	216	146887	4.807	ng/uL	99
74) Pentachlorobenzene	14.79	250	146039	4.913	ng/uL	99
75) Carbazole	17.63	167	392762	4.395	ng/ul	100
76) Di-n-butylphthalate	18.20	149	349887	3.659	ng/ul	99
77) Fluoranthene	19.25	202	575745	4.922	ng/ul	100
80) Pvrene	19.60	202	652337	4.745	ng/ul	100
81) Butylbenzylphthalate	20.48	149	111246	3.004	ng/ul	97
82) 3,3'-Dichlorobenzidine	21.24	252	88549	2.394	ng/ul#	98
83) Benzo(a)anthracene	21.31	228	619732	4.687	ng/ul	99
84) Bis(2-ethylhexyl)phthalate	21.25	149	201131	3.416	ng/ul	99
85) Chrysene	21.36	228	635416	4.893	ng/ul	99
87) Di-n-octyl phthalate	22.16	149	278949	2.870	ng/ul	100
88) Benzo(b)fluoranthene	22.96	252	570974	4.423	ng/ul	99
89) Benzo(k)fluoranthene	23.00	252	632511	4.866	ng/ul	98
91) Benzo(a)pyrene	23.57	252	537523	4.505	ng/ul	99
92) Indeno(1,2,3-cd)pyrene	26.09	276	655631	4.494	ng/ul	99
93) Dibenzo(a,h)anthracene	26.10	278	558974	4.553	ng/ul	99
94) Benzo(a,h,i)perylene	26.82	276	557099	4.550	ng/ul	99


Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP121020\
Data File : BP004213.D
Acq On : 10 Dec 2020 09:51
Operator : CG/JU
Sample : SSTD00581
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD00581

Quant Time: Dec 10 12:23:58 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP121020MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 10 11:56:16 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						



#2

1,4-Dioxane

Concen: 2.163 ng/uL

RT: 3.36 min Scan# 30

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581


Tot Ion: 88 Resp: 21981

Ion Ratio Lower Upper

88 100

43 34.3 23.2 34.8




58 64.7 55.0 82.6



Abundance Ion 88.00 (87.70 to 88.70): BP0

Ion 43.00 (42.70 to 43.70): BP0

Ion 58.00 (57.70 to 58.70): BP0



#4

Benzaldehyde

Concen: 3.955 ng/uL

RT: 6.98 min Scan# 644

Delta R.T. 0.01 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51


Tgt Ion: 77 Resp: 90060

Ion Ratio Lower Upper

77 100

105 96.6 77.1 115.7



106 94.3 75.6 113.4




Abundance Ion 77.00 (76.70 to 77.70): BP0

Ion 105.00 (104.70 to 105.70): E

Ion 106.00 (105.70 to 106.70): E





#6

Phenol

Concen: 4.229 ng/ul

RT: 7.02 min Scan# 651

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581





Tot Ion: 94 Resp: 143366

Ion Ratio Lower Upper

94 100

65 25.5 20.8 31.2

66 32.3 26.5 39.7



#8

Bis(2-Chloroethyl)ether

Concen: 4.768 ng/ul

RT: 7.26 min Scan# 692

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51




Tgt Ion: 93 Resp: 131712


Ion Ratio Lower Upper

93 100

63 72.7 58.7 88.1

95 31.6 26.2 39.2





#10

2-Chlorophenol

Concen: 4.318 ng/ul

RT: 7.39 min Scan# 715

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581


Tot Ion:128 Resp: 108270

Ion Ratio Lower Upper

128 100

64 45.6 36.9 55.3




130 32.2 25.9 38.9



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 64.00 (63.70 to 64.70): BP0

Ion 130.00 (129.70 to 130.70): E



#11

2-Methylphenol

Concen: 3.705 ng/ul

RT: 8.26 min Scan# 863

Delta R.T. -0.00 min

Lab File: BP004213.D


Acq: 10 Dec 2020 09:51

Tgt Ion:108 Resp: 92248

Ion Ratio Lower Upper



108 100


107 88.4 71.8 107.8



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E









#12
2,2'-oxvbis(1-Chloropropane)
Concen: 4.832 ng/ul
RT: 8.36 min Scan# 880
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion: 45 Resp: 160905
Ion Ratio Lower Upper
45 100
77 16.7 13.0 19.4
79 13.2 10.4 15.6



#14
Acetophenone
Concen: 4.673 ng/ul
RT: 8.65 min Scan# 928
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion: 105 Resp: 185485
Ion Ratio Lower Upper
105 100
77 77.2 61.1 91.7
51 24.4 18.5 27.7









#15
N-Nitroso-di-n-propylamine
Concen: 4.123 ng/ul
RT: 8.63 min Scan# 926
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion:	70	Resp:	80998
Ion	Ratio	Lower	Upper
70	100		
42	51.3	39.0	58.6
101	9.5	7.6	11.4
130	22.7	16.6	25.0



#16
4-Methylphenol
Concen: 3.923 ng/ul
RT: 8.59 min Scan# 918
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion:	108	Resp:	105483
Ion	Ratio	Lower	Upper
108	100		
107	112.0	87.5	131.3









#17
Hexachloroethane
Concen: 4.937 ng/ul
RT: 8.91 min Scan# 973
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion:117 Resp: 54821
Ion Ratio Lower Upper
117 100
201 104.2 85.4 128.2
199 62.3 51.2 76.8



#20
Nitrobenzene
Concen: 4.621 ng/ul
RT: 9.02 min Scan# 992
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion: 77 Resp: 137201
Ion Ratio Lower Upper
77 100
123 41.7 36.8 55.2
65 12.1 10.1 15.1





#21

Isophorone

Concen: 3.946 ng/ul

RT: 9.55 min Scan# 1081

Delta R.T. -0.01 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581


Tot Ion: 82 Resp: 208538

Ion Ratio Lower Upper

82 100

95 7.2 5.4 8.0




138 17.8 14.4 21.6



Abundance Ion 82.00 (81.70 to 82.70): BP0

Ion 95.00 (94.70 to 95.70): BP0

Ion 138.00 (137.70 to 138.70): BP0



#23

2-Nitrophenol

Concen: 3.685 ng/ul

RT: 9.73 min Scan# 1113

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51


Tgt Ion: 139 Resp: 41196

Ion Ratio Lower Upper

139 100

65 50.4 38.2 57.4



109 26.5 22.9 34.3




Abundance Ion 139.00 (138.70 to 139.70): BP0

Ion 65.00 (64.70 to 65.70): BP0

Ion 109.00 (108.70 to 109.70): BP0





#24

2,4-Dimethylphenol

Concen: 4.274 ng/ul

RT: 9.79 min Scan# 1123

Delta R.T. -0.01 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581





Tot Ion:107 Resp: 110180

Ion Ratio Lower Upper

107 100

121 51.4 41.2 61.8

122 86.2 69.3 103.9



#25

Bis(2-Chloroethoxy)methane

Concen: 4.608 ng/ul

RT: 10.04 min Scan# 1165

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51




Tgt Ion: 93 Resp: 159210


Ion Ratio Lower Upper

93 100

95 33.6 25.7 38.5

123 11.5 9.0 13.6









#27
2,4-Dichlorophenol
Concen: 4.125 ng/ul
RT: 10.27 min Scan# 1204
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion:162 Resp: 91744
Ion Ratio Lower Upper
162 100
164 66.7 52.2 78.2
98 32.8 27.8 41.6



#28
Naphthalene
Concen: 4.962 ng/ul
RT: 10.67 min Scan# 1273
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion:128 Resp: 409826
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.1 10.6 16.0









#30
4-Chloroaniline
Concen: 4.055 ng/ul
RT: 10.78 min Scan# 1291
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion:127 Resp: 107558
Ion Ratio Lower Upper
127 100
129 32.1 26.2 39.4



#31
Hexachlorobutadiene
Concen: 4.896 ng/ul
RT: 10.97 min Scan# 1323
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion:225 Resp: 81036
Ion Ratio Lower Upper
225 100
223 66.5 50.1 75.1
227 64.8 51.2 76.8









#32
 Caprolactam
 Concen: 2.537 ng/ul
 RT: 11.52 min Scan# 1417
 Delta R.T. -0.01 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SST00581


Tot Ion:113 Resp: 16786
 Ion Ratio Lower Upper
 113 100
 55 162.5 129.3 193.9
 56 111.9 92.2 138.4



#33
 4-Chloro-3-methylphenol
 Concen: 3.548 ng/ul
 RT: 11.90 min Scan# 1482
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion:107 Resp: 77052
 Ion Ratio Lower Upper
 107 100
 144 26.7 22.0 33.0
 142 82.3 66.2 99.2









#34
 2-Methylnaphthalene
 Concen: 4.726 ng/ul
 RT: 12.29 min Scan# 1547
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581


Tot Ion:142 Resp: 268319
 Ion Ratio Lower Upper
 142 100
 141 87.5 69.7 104.5



#35
 1-Methylnaphthalene
 Concen: 5.941 ng/ul
 RT: 12.51 min Scan# 1585
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion:142 Resp: 318676
 Ion Ratio Lower Upper
 142 100
 141 89.7 73.8 110.8
 116 3.6 2.8 4.2









#37
 1,2,4,5-Tetrachlorobenzene
 Concen: 4.899 ng/ul
 RT: 12.66 min Scan# 1610
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581


Tot Ion:	216	Resp:	149118
Ion	Ratio	Lower	Upper
216	100		
214	80.1	62.0	93.0
179	20.3	15.8	23.6
108	16.0	12.6	18.8



#38
 Hexachlorocyclopentadiene
 Concen: 4.357 ng/ul
 RT: 12.65 min Scan# 1608
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion:	237	Resp:	73360
Ion	Ratio	Lower	Upper
237	100		
235	61.1	49.5	74.3
272	13.7	10.3	15.5









#39
 2,4,6-Trichlorophenol
 Concen: 3.432 ng/ul
 RT: 12.90 min Scan# 1651
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581


Tot Ion:196 Resp: 52294
 Ion Ratio Lower Upper
 196 100
 198 96.7 75.9 113.9
 200 30.7 25.3 37.9



#40
 2,4,5-Trichlorophenol
 Concen: 3.671 ng/ul
 RT: 12.96 min Scan# 1662
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion:196 Resp: 62287
 Ion Ratio Lower Upper
 196 100
 198 100.8 75.4 113.0
 200 29.4 24.3 36.5









#41
1,1'-Biphenyl
Concen: 4.939 ng/ul
RT: 13.30 min Scan# 1720
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion	Ratio	Lower	Upper
154	100		
153	41.0	32.2	48.2
76	14.0	11.0	16.6



#42
2-Chloronaphthalene
Concen: 4.916 ng/ul
RT: 13.35 min Scan# 1727
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion	Ratio	Lower	Upper
162	100		
164	32.2	26.4	39.6
127	36.1	29.0	43.4









#43
2-Nitroaniline
Concen: 3.437 ng/ul
RT: 13.54 min Scan# 1760
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion: 65 Resp: 44364
Ion Ratio Lower Upper
65 100
92 71.7 59.8 89.6
138 112.5 91.0 136.4



#45
Dimethylphthalate
Concen: 4.671 ng/ul
RT: 13.93 min Scan# 1826
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion: 163 Resp: 315313
Ion Ratio Lower Upper
163 100
194 4.5 3.4 5.2
164 10.5 8.4 12.6









#46
 2,6-Dinitrotoluene
 Concen: 3.645 ng/ul
 RT: 14.04 min Scan# 1845
 Delta R.T. -0.01 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581


Tot Ion:165 Resp: 46247
 Ion Ratio Lower Upper
 165 100
 89 62.2 45.6 68.4
 121 26.9 17.1 25.7#



#48
 Acenaphthylene
 Concen: 4.752 ng/ul
 RT: 14.19 min Scan# 1871
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion:152 Resp: 397276
 Ion Ratio Lower Upper
 152 100
 151 20.0 15.8 23.8
 153 13.4 10.6 16.0









#49
3-Nitroaniline
Concen: 2.882 ng/ul
RT: 14.37 min Scan# 1901
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion:138 Resp: 35200
Ion Ratio Lower Upper
138 100
108 13.3 9.0 13.4
92 124.7 95.4 143.2



#50
Acenaphthene
Concen: 4.886 ng/ul
RT: 14.54 min Scan# 1930
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion:153 Resp: 288346
Ion Ratio Lower Upper
153 100
152 49.6 37.8 56.6
154 88.8 69.8 104.8









#51
2,4-Dinitrophenol
Concen: 2.804 ng/ul
RT: 14.57 min Scan# 1935
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion:184 Resp: 14268
Ion Ratio Lower Upper
184 100
63 55.6 44.4 66.6
154 62.4 45.8 68.8



#53
4-Nitrophenol
Concen: 3.234 ng/ul
RT: 14.67 min Scan# 1952
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion:109 Resp: 26442
Ion Ratio Lower Upper
109 100
139 154.9 113.5 170.3
65 144.9 105.1 157.7





#54

Dibenzofuran

Concen: 4.982 ng/ul

RT: 14.87 min Scan# 1987

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :


SSTD00581

Tot Ion:168 Resp: 414316

Ion Ratio Lower Upper




168 100

139 35.3 29.0 43.6



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E



#55

2,4-Dinitrotoluene

Concen: 3.559 ng/ul

RT: 14.83 min Scan# 1979

Delta R.T. -0.01 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51


Tgt Ion:165 Resp: 64064

Ion Ratio Lower Upper

165 100

63 41.9 29.4 44.2



182 2.7 2.6 3.8




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BP0

Ion 182.00 (181.70 to 182.70): E









#56
 2,3,4,6-Tetrachlorophenol
 Concen: 3.428 ng/ul
 RT: 15.10 min Scan# 2025
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SST00581


Tot Ion:	232	Resp:	47333
Ion	Ratio	Lower	Upper
232	100		
131	39.4	33.4	50.0
130	1.8	1.8	2.8
166	27.4	24.1	36.1



#57
 Diethylphthalate
 Concen: 4.376 ng/ul
 RT: 15.30 min Scan# 2060
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion:	149	Resp:	277480
Ion	Ratio	Lower	Upper
149	100		
177	22.9	18.2	27.2
150	12.1	9.8	14.8





#59

Fluorene

Concen: 4.948 ng/ul

RT: 15.53 min Scan# 2098

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581


Tot Ion:166 Resp: 329951

Ion Ratio Lower Upper

166 100

165 95.3 78.2 117.2




167 13.2 10.8 16.2



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#60

4-Chlorophenyl-phenylether

Concen: 4.937 ng/ul

RT: 15.52 min Scan# 2097

Delta R.T. -0.01 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51


Tgt Ion:204 Resp: 174176

Ion Ratio Lower Upper

204 100

206 31.9 26.4 39.6



141 54.9 41.0 61.6




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 2.874 ng/ul
RT: 15.53 min Scan# 2099
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion:138 Resp: 39903
Ion Ratio Lower Upper
138 100
92 53.6 39.0 58.4
108 75.4 56.9 85.3



#64
4,6-Dinitro-2-methylphenol
Concen: 2.880 ng/ul
RT: 15.59 min Scan# 2109
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion:198 Resp: 26736
Ion Ratio Lower Upper
198 100
182 6.6 3.5 5.3#
77 26.3 17.3 25.9#









#65
N-Nitrosodiphenylamine
Concen: 4.497 ng/ul
RT: 15.74 min Scan# 2134
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion	Ion	Ratio	Lower	Upper
169	100			
168	67.9	53.0	79.6	
167	36.4	29.1	43.7	



#66
4-Bromophenyl-phenylether
Concen: 4.572 ng/ul
RT: 16.42 min Scan# 2250
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion	Ion	Ratio	Lower	Upper
248	100			
250	98.0	79.1	118.7	
141	67.2	53.7	80.5	









#67
Hexachlorobenzene
Concen: 4.982 ng/ul
RT: 16.53 min Scan# 2269
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion	Ratio	Lower	Upper
284	100		
142	33.8	26.1	39.1
249	30.6	25.0	37.4



#68
Atrazine
Concen: 3.393 ng/ul
RT: 16.69 min Scan# 2296
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion	Ratio	Lower	Upper
200	100		
173	25.1	20.6	30.8
215	50.2	39.8	59.8





#69

Pentachlorophenol

Concen: 2.749 ng/ul

RT: 16.88 min Scan# 2328

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581


Tot Ion:266 Resp: 32498

Ion Ratio Lower Upper

266 100

264 56.2 49.8 74.6




268 56.7 52.0 78.0



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E



#70

Phenanthrene

Concen: 4.981 ng/ul

RT: 17.27 min Scan# 2395

Delta R.T. -0.00 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51


Tgt Ion:178 Resp: 540563

Ion Ratio Lower Upper

178 100

179 15.4 12.4 18.6



176 19.1 15.4 23.0




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E









#72
 Anthracene
 Concen: 4.800 ng/ul
 RT: 17.36 min Scan# 2410
 Delta R.T. -0.01 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581


Tot Ion	Ratio	Lower	Upper
178	100		
179	15.1	12.3	18.5
176	18.3	14.9	22.3



#73
 1,2,3,4-Tetrachlorobenzene
 Concen: 4.807 ng/uL
 RT: 13.27 min Scan# 1714
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.0	62.0	93.0
218	48.5	38.5	57.7






#74
Pentachlorobenzene
Concen: 4.913 ng/uL
RT: 14.79 min Scan# 1973
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Instrument :
BNA_P
ClientSampleId :
SSTD00581

Tot Ion	Ratio	Lower	Upper
250	100		
248	64.8	50.5	75.7
252	64.0	51.0	76.4






Abundance

Ion 250.00 (249.70 to 250.70): E


Ion 248.00 (247.70 to 248.70): E

Ion 252.00 (251.70 to 252.70): E



#75
Carbazole
Concen: 4.395 ng/uL
RT: 17.63 min Scan# 2456
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.1	25.7
139	12.0	9.8	14.6






Abundance

Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E









#76
Di-n-butylphthalate
Concen: 3.659 ng/ul
RT: 18.20 min Scan# 2552
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion:149 Resp: 349887
Ion Ratio Lower Upper
149 100
150 9.5 7.3 10.9
104 5.0 3.7 5.5



#77
Fluoranthene
Concen: 4.922 ng/ul
RT: 19.25 min Scan# 2730
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion:202 Resp: 575745
Ion Ratio Lower Upper
202 100
101 11.4 9.1 13.7
100 8.8 7.1 10.7









#80
Pvrene
Concen: 4.745 ng/ul
RT: 19.60 min Scan# 2790
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51




Instrument :
BNA_P
ClientSampleId :
SSTD00581


Tot Ion	Ratio	Lower	Upper
202	100		
101	13.2	10.6	15.8
100	10.8	8.6	12.8



#81
Butylbenzylphthalate
Concen: 3.004 ng/ul
RT: 20.48 min Scan# 2940
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion	Ratio	Lower	Upper
149	100		
91	73.4	56.5	84.7
206	23.3	17.7	26.5









#82
 3,3'-Dichlorobenzidine
 Concen: 2.394 ng/ul
 RT: 21.24 min Scan# 3069
 Delta R.T. -0.01 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581


Tot Ion:252 Resp: 88549
 Ion Ratio Lower Upper
 252 100
 254 65.5 51.4 77.0
 126 12.0 7.6 11.4#



#83
 Benzo(a)anthracene
 Concen: 4.687 ng/ul
 RT: 21.31 min Scan# 3081
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion:228 Resp: 619732
 Ion Ratio Lower Upper
 228 100
 229 18.9 15.8 23.8
 226 27.0 21.7 32.5









#84
 Bis(2-ethylhexyl)phthalate
 Concen: 3.416 ng/ul
 RT: 21.25 min Scan# 3070
 Delta R.T. -0.00 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTD00581


Tot Ion:149 Resp: 201131
 Ion Ratio Lower Upper
 149 100
 167 27.4 22.2 33.4
 279 4.4 3.6 5.4



#85
 Chrysene
 Concen: 4.893 ng/ul
 RT: 21.36 min Scan# 3089
 Delta R.T. -0.01 min
 Lab File: BP004213.D
 Acq: 10 Dec 2020 09:51

Tgt Ion:228 Resp: 635416
 Ion Ratio Lower Upper
 228 100
 226 29.5 24.2 36.2
 229 19.6 15.8 23.8









#87
Di-n-octyl phthalate
Concen: 2.870 ng/ul
RT: 22.16 min Scan# 3225
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Instrument :
BNA_P
ClientSampleId :
SSTD00581

Tgt Ion:149 Resp: 278949




Abundance Ion 149.00 (148.70 to 149.70): E



#88
Benzo(b)fluoranthene
Concen: 4.423 ng/ul
RT: 22.96 min Scan# 3361
Delta R.T. -0.01 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51



Tgt Ion:252 Resp: 570974
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 9.6 7.7 11.5




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(k)fluoranthene

Concen: 4.866 ng/ul

RT: 23.00 min Scan# 3369

Delta R.T. -0.01 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581


Tot Ion:252 Resp: 632511

Ion Ratio Lower Upper

252 100

253 20.9 17.4 26.2




125 9.4 7.3 10.9



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Benzo(a)pyrene

Concen: 4.505 ng/ul

RT: 23.57 min Scan# 3465

Delta R.T. -0.01 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51


Tgt Ion:252 Resp: 537523

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9



125 10.6 7.8 11.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#92

Indeno(1,2,3-cd)pyrene
Concen: 4.494 ng/ul
RT: 26.09 min Scan# 3893
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Instrument :
BNA_P
ClientSampleId :
SSTD00581




Tot Ion:276 Resp: 655631
Ion Ratio Lower Upper
276 100
138 21.9 17.8 26.6
277 24.5 19.8 29.6




#93

Dibenzo(a,h)anthracene
Concen: 4.553 ng/ul
RT: 26.10 min Scan# 3896
Delta R.T. -0.00 min
Lab File: BP004213.D
Acq: 10 Dec 2020 09:51

Tgt Ion:278 Resp: 558974
Ion Ratio Lower Upper
278 100
139 15.9 11.9 17.9
279 23.4 18.6 28.0





#94

Benzo(a,h,i)perylene

Concen: 4.550 ng/ul

RT: 26.82 min Scan# 4018

Delta R.T. -0.01 min

Lab File: BP004213.D

Acq: 10 Dec 2020 09:51

Instrument :

BNA_P

ClientSampleId :

SSTD00581


Tot Ion:276 Resp: 557099

Ion Ratio Lower Upper

276 100

138 20.7 16.6 25.0

277 23.8 18.5 27.7



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

