

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121022\
 Data File : BPCCC075.D
 Acq On : 11 Dec 2022 12:20
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Dec 12 03:37:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 12 01:32:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	657971	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	2779604	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.610	164	1798903	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.363	188	4067039	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	3715184	20.000	ng/u1	0.00
88) Perylene-d12	23.933	264	3524514	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	113657	7.386	ng/uL	0.00
4) Pyridine-d5	3.775	84	835498	19.113	ng/u1	0.00
7) Phenol-d5	7.122	99	1045885	19.515	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	644937	19.949	ng/u1	0.00
11) 2-Chlorophenol-d4	7.493	132	847251	20.082	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	835210	19.020	ng/u1	0.00
21) Nitrobenzene-d5	9.134	128	414038	20.274	ng/u1	0.00
24) 2-Nitrophenol-d4	9.863	143	471959	20.527	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.399	165	903718	20.236	ng/u1	0.00
31) 4-Chloroaniline-d4	10.922	131	1233371	19.563	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	2659731	19.238	ng/u1	0.00
49) Acenaphthylene-d8	14.304	160	3066468	20.185	ng/u1	0.00
54) 4-Nitrophenol-d4	14.804	143	481768	19.372	ng/u1	0.00
60) Fluorene-d10	15.598	176	2315600	19.798	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.722	200	465573	17.789	ng/u1	0.00
73) Anthracene-d10	17.463	188	3667919	19.991	ng/u1	0.00
81) Pyrene-d10	19.692	212	4337857	20.342	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.774	264	3489831	19.865	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.387	88	118728	7.391	ng/uL	99
5) Pyridine	3.799	79	843259	19.409	ng/u1	98
6) Benzaldehyde	7.105	77	404613	19.281	ng/u1	98
8) Phenol	7.152	94	1079410	19.920	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.387	93	869118	19.542	ng/u1	99
12) 2-Chlorophenol	7.528	128	858402	19.841	ng/u1	98
13) 2-Methylphenol	8.411	108	820456	19.421	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.493	45	1239424	20.309	ng/u1	98
16) Acetophenone	8.799	105	1352538	20.088	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.781	70	672119	19.851	ng/u1	99
18) 4-Methylphenol	8.740	108	913466	19.522	ng/u1	98
19) Hexachloroethane	9.052	117	340679	19.705	ng/u1	94
22) Nitrobenzene	9.181	77	997312	20.811	ng/u1	98
23) Isophorone	9.705	82	1882905	19.431	ng/u1	98
25) 2-Nitrophenol	9.893	139	515260	20.683	ng/u1	97
26) 2,4-Dimethylphenol	9.946	107	972110	19.999	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.187	93	1206221	19.468	ng/u1	100
29) 2,4-Dichlorophenol	10.428	162	890892	20.365	ng/u1	100
30) Naphthalene	10.834	128	2872650	19.761	ng/u1	100
32) 4-Chloroaniline	10.940	127	1245325	19.961	ng/u1	99
33) Hexachlorobutadiene	11.116	225	594656	19.798	ng/u1	98
34) Caprolactam	11.722	113	190977	12.803	ng/u1	97
35) 4-Chloro-3-methylphenol	12.057	107	899824	20.156	ng/u1	99
36) 2-Methylnaphthalene	12.434	142	1952673	19.580	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121022\
 Data File : BPCCC075.D
 Acq On : 11 Dec 2022 12:20
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Dec 12 03:37:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 12 01:32:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.657	142	2002600	19.526	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.804	216	1166910	20.145	ng/ul	99
40) Hexachlorocyclopentadiene	12.781	237	529648	14.051	ng/ul	98
41) 2,4,6-Trichlorophenol	13.040	196	748528	20.107	ng/ul	100
42) 2,4,5-Trichlorophenol	13.110	196	822260	20.563	ng/ul	99
43) 1,1'-Biphenyl	13.440	154	2673763	19.790	ng/ul	99
44) 2-Chloronaphthalene	13.487	162	2125802	19.992	ng/ul	99
45) 2-Nitroaniline	13.693	65	568096	22.036	ng/ul	100
47) Dimethylphthalate	14.063	163	2662222	19.428	ng/ul	99
48) 2,6-Dinitrotoluene	14.181	165	527477	20.713	ng/ul	98
50) Acenaphthylene	14.328	152	3292267	20.158	ng/ul	100
51) 3-Nitroaniline	14.510	138	450788	18.747	ng/ul	96
52) Acenaphthene	14.669	153	2250596	19.885	ng/ul	100
53) 2,4-Dinitrophenol	14.716	184	281526	16.173	ng/ul	100
55) 4-Nitrophenol	14.816	109	350853	20.676	ng/ul	96
56) Dibenzofuran	15.004	168	3204204	19.912	ng/ul	99
57) 2,4-Dinitrotoluene	14.969	165	770324	20.825	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.234	232	726105	19.898	ng/ul	97
59) Diethylphthalate	15.422	149	2580921	19.375	ng/ul	99
61) Fluorene	15.657	166	2600728	20.152	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.645	204	1355479	19.861	ng/ul	99
63) 4-Nitroaniline	15.675	138	424461	20.060	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.734	198	468240	18.308	ng/ul	96
67) N-Nitrosodiphenylamine	15.863	169	2236927	19.839	ng/ul	99
68) 4-Bromophenyl-phenylether	16.545	248	850347	19.253	ng/ul	98
69) Hexachlorobenzene	16.663	284	1010389	19.266	ng/ul	99
70) Atrazine	16.816	200	849811	19.340	ng/ul	99
71) Pentachlorophenol	17.010	266	576381	18.148	ng/ul	98
72) Phenanthrene	17.404	178	4228922	19.963	ng/ul	100
74) Anthracene	17.498	178	4275254	20.212	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.410	216	1150658	19.830	ng/ul	99
76) Pentachlorobenzene	14.928	250	1152527	19.482	ng/ul	99
77) Carbazole	17.769	167	3824878	21.472	ng/ul	99
78) Di-n-butylphthalate	18.310	149	4398459	20.280	ng/ul	100
80) Fluoranthene	19.369	202	5032366	20.792	ng/ul	99
82) Pyrene	19.722	202	5247758	21.019	ng/ul	99
83) Butylbenzylphthalate	20.586	149	1958602	20.306	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.363	252	1449293	17.268	ng/ul	100
85) Benzo(a)anthracene	21.433	228	4989045	20.229	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.339	149	2833956	20.829	ng/ul	99
87) Chrysene	21.486	228	4648813	19.932	ng/ul	100
89) Di-n-octyl phthalate	22.298	149	4617214	23.384	ng/ul	100
90) Benzo(b)fluoranthene	23.174	252	4605044	20.490	ng/ul	100
91) Benzo(k)fluoranthene	23.221	252	4493993	20.165	ng/ul	100
93) Benzo(a)pyrene	23.827	252	3917356	20.018	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.539	276	4784751	19.629	ng/ul	99
95) Dibenzo(a,h)anthracene	26.557	278	3977277	19.707	ng/ul	99
96) Benzo(g,h,i)perylene	27.339	276	3657809	18.691	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Dec 12 03:37:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 12 01:32:54 2022
Response via : Initial Calibration








#2
 1,4-Dioxane
 Concen: 7.391 ng/uL
 RT: 3.387 min Scan# 34
 Delta R.T. 0.000 min
 Lab File: BPCCC075.D
 Acq: 11 Dec 2022 12:20

Instrument :
 BNA_P
 ClientSampleId :



Tgt Ion: 88 Resp: 118728




Ion	Ratio	Lower	Upper
88	100		
43	28.5	22.4	33.6
58	73.9	58.2	87.4




#5
 Pyridine
 Concen: 19.409 ng/uL
 RT: 3.799 min Scan# 104
 Delta R.T. 0.000 min
 Lab File: BPCCC075.D
 Acq: 11 Dec 2022 12:20

Tgt Ion: 79 Resp: 843259

Ion	Ratio	Lower	Upper
79	100		
52	78.3	60.7	91.1
51	36.9	28.9	43.3









#6
Benzaldehyde
Concen: 19.281 ng/ul
RT: 7.105 min Scan# 666
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 404613
Ion Ratio Lower Upper
77 100
105 99.1 81.5 122.3
106 96.6 76.9 115.3



#8
Phenol
Concen: 19.920 ng/ul
RT: 7.152 min Scan# 674
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion: 94 Resp: 1079410
Ion Ratio Lower Upper
94 100
65 27.2 21.1 31.7
66 34.4 27.0 40.6









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.542 ng/ul
 RT: 7.387 min Scan# 714
 Delta R.T. 0.000 min
 Lab File: BPCCC075.D
 Acq: 11 Dec 2022 12:20




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 93 Resp: 869118
 Ion Ratio Lower Upper
 93 100
 63 76.0 59.8 89.8
 95 32.8 26.4 39.6



#12
 2-Chlorophenol
 Concen: 19.841 ng/ul
 RT: 7.528 min Scan# 738
 Delta R.T. -0.006 min
 Lab File: BPCCC075.D
 Acq: 11 Dec 2022 12:20

Tgt Ion:128 Resp: 858402
 Ion Ratio Lower Upper
 128 100
 64 45.5 34.9 52.3
 130 32.5 25.8 38.6





#13

2-Methylphenol

Concen: 19.421 ng/ul

RT: 8.411 min Scan# 889

Delta R.T. 0.000 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 820456

Ion Ratio Lower Upper

108 100

107 88.2 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.309 ng/ul

RT: 8.493 min Scan# 902

Delta R.T. -0.012 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20




Tgt Ion: 45 Resp: 1239424


Ion Ratio Lower Upper

45 100

77 13.5 11.0 16.6

79 10.3 9.3 13.9









#16
Acetophenone
Concen: 20.088 ng/ul
RT: 8.799 min Scan# 91
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 1352538
Ion Ratio Lower Upper
105 100
77 75.0 59.3 88.9
51 26.6 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 19.851 ng/ul
RT: 8.781 min Scan# 951
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion: 70 Resp: 672119
Ion Ratio Lower Upper
70 100
42 58.1 45.5 68.3
101 9.3 7.9 11.9
130 22.1 17.4 26.2









#18
4-Methylphenol
Concen: 19.522 ng/ul
RT: 8.740 min Scan# 945
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 913466
Ion Ratio Lower Upper
108 100
107 111.2 87.4 131.2



#19
Hexachloroethane
Concen: 19.705 ng/ul
RT: 9.052 min Scan# 997
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:117 Resp: 340679
Ion Ratio Lower Upper
117 100
201 109.8 93.7 140.5
199 69.9 58.6 88.0









#22
Nitrobenzene
Concen: 20.811 ng/ul
RT: 9.181 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 997312
Ion Ratio Lower Upper
77 100
123 46.2 38.3 57.5
65 13.3 10.3 15.5



#23
Isophorone
Concen: 19.431 ng/ul
RT: 9.705 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion: 82 Resp: 1882905
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.2
138 18.1 15.3 22.9









#25
2-Nitrophenol
Concen: 20.683 ng/ul
RT: 9.893 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 515260
Ion Ratio Lower Upper
139 100
65 48.3 36.6 54.8
109 26.3 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 19.999 ng/ul
RT: 9.946 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20


Tgt Ion:107 Resp: 972110
Ion Ratio Lower Upper
107 100
121 51.4 41.8 62.6
122 89.2 72.6 108.8








#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.468 ng/ul
 RT: 10.187 min Scan# 1190
 Delta R.T. -0.006 min
 Lab File: BPCCC075.D
 Acq: 11 Dec 2022 12:20

Instrument :
 BNA_P
 ClientSampleId :



Tgt Ion: 93 Resp: 1206221




Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.0	39.0
123	11.2	9.2	13.8




#29
 2,4-Dichlorophenol
 Concen: 20.365 ng/ul
 RT: 10.428 min Scan# 1231
 Delta R.T. 0.000 min
 Lab File: BPCCC075.D
 Acq: 11 Dec 2022 12:20

Tgt Ion:162 Resp: 890892

Ion	Ratio	Lower	Upper
162	100		
164	65.6	52.5	78.7
98	33.0	26.6	39.8









#30
Naphthalene
Concen: 19.761 ng/ul
RT: 10.834 min Scan# 11
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 2872650
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.2 10.4 15.6



#32
4-Chloroaniline
Concen: 19.961 ng/ul
RT: 10.940 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:127 Resp: 1245325
Ion Ratio Lower Upper
127 100
129 32.7 25.6 38.4









#33
Hexachlorobutadiene
Concen: 19.798 ng/ul
RT: 11.116 min Scan# 11
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:225 Resp: 594656
Ion Ratio Lower Upper
225 100
223 63.0 49.2 73.8
227 63.9 50.2 75.2



#34
Caprolactam
Concen: 12.803 ng/ul
RT: 11.722 min Scan# 1451
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:113 Resp: 190977
Ion Ratio Lower Upper
113 100
55 172.7 141.9 212.9
56 126.9 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 20.156 ng/ul
RT: 12.057 min Scan# 1508
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 899824
Ion Ratio Lower Upper
107 100
144 26.9 21.4 32.2
142 83.5 65.8 98.6



#36
2-Methylnaphthalene
Concen: 19.580 ng/ul
RT: 12.434 min Scan# 1572
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:142 Resp: 1952673
Ion Ratio Lower Upper
142 100
141 88.2 70.8 106.2









#37
1-Methylnaphthalene
Concen: 19.526 ng/ul
RT: 12.657 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 2002600
Ion Ratio Lower Upper
142 100
141 92.5 72.7 109.1
116 3.5 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.145 ng/ul
RT: 12.804 min Scan# 1635
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:216 Resp: 1166910
Ion Ratio Lower Upper
216 100
214 78.8 61.9 92.9
179 19.3 15.0 22.6
108 14.3 11.0 16.4









#40
Hexachlorocyclopentadiene
Concen: 14.051 ng/ul
RT: 12.781 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 529648
Ion Ratio Lower Upper
237 100
235 63.5 49.7 74.5
272 13.3 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 20.107 ng/ul
RT: 13.040 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:196 Resp: 748528
Ion Ratio Lower Upper
196 100
198 96.7 77.0 115.6
200 31.6 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 20.563 ng/ul
RT: 13.110 min Scan# 1
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:196 Resp: 822260
Ion Ratio Lower Upper
196 100
198 95.9 77.1 115.7
200 31.0 25.4 38.2



#43
1,1'-Biphenyl
Concen: 19.790 ng/ul
RT: 13.440 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:154 Resp: 2673763
Ion Ratio Lower Upper
154 100
153 40.0 31.5 47.3
76 12.6 9.6 14.4





#44

2-Chloronaphthalene

Concen: 19.992 ng/ul

RT: 13.487 min Scan# 1751

Delta R.T. 0.000 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:162 Resp: 2125802

Ion Ratio Lower Upper

162 100

164 32.5 26.3 39.5

127 36.3 29.3 43.9



#45

2-Nitroaniline

Concen: 22.036 ng/ul

RT: 13.693 min Scan# 1786

Delta R.T. 0.000 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20




Tgt Ion: 65 Resp: 568096


Ion Ratio Lower Upper

65 100

92 70.4 56.4 84.6

138 112.4 90.3 135.5









#47
Dimethylphthalate
Concen: 19.428 ng/ul
RT: 14.063 min Scan# 1849
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 2662222
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.1 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.713 ng/ul
RT: 14.181 min Scan# 1869
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:165 Resp: 527477
Ion Ratio Lower Upper
165 100
89 55.5 43.0 64.4
121 21.1 16.1 24.1









#50
Acenaphthylene
Concen: 20.158 ng/ul
RT: 14.328 min Scan# 1894
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 3292267
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.6
153 13.5 10.6 16.0



#51
3-Nitroaniline
Concen: 18.747 ng/ul
RT: 14.510 min Scan# 1925
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:138 Resp: 450788
Ion Ratio Lower Upper
138 100
108 10.5 8.6 13.0
92 115.3 88.4 132.6





#52

Acenaphthene

Concen: 19.885 ng/ul

RT: 14.669 min Scan# 1952

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:153 Resp: 2250596

Ion Ratio Lower Upper

153 100

152 47.5 38.2 57.4

154 88.7 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 16.173 ng/ul

RT: 14.716 min Scan# 1960

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20




Tgt Ion:184 Resp: 281526


Ion Ratio Lower Upper

184 100

63 60.9 48.4 72.6

154 57.1 45.5 68.3









#55
4-Nitrophenol
Concen: 20.676 ng/ul
RT: 14.816 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 350853
Ion Ratio Lower Upper
109 100
139 143.8 120.8 181.2
65 138.3 107.9 161.9



#56
Dibenzofuran
Concen: 19.912 ng/ul
RT: 15.004 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:168 Resp: 3204204
Ion Ratio Lower Upper
168 100
139 35.3 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 20.825 ng/ul
RT: 14.969 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 770324
Ion Ratio Lower Upper
165 100
63 40.1 31.8 47.6
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.898 ng/ul
RT: 15.234 min Scan# 2048
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:232 Resp: 726105
Ion Ratio Lower Upper
232 100
131 37.6 31.8 47.6
130 2.0 1.8 2.6
166 27.6 22.9 34.3









#59
Diethylphthalate
Concen: 19.375 ng/ul
RT: 15.422 min Scan# 2081
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 2580921
Ion Ratio Lower Upper
149 100
177 23.2 18.4 27.6
150 12.8 9.9 14.9



#61
Fluorene
Concen: 20.152 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:166 Resp: 2600728
Ion Ratio Lower Upper
166 100
165 97.8 77.4 116.2
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 19.861 ng/ul
RT: 15.645 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 1355479
Ion Ratio Lower Upper
204 100
206 33.1 26.3 39.5
141 52.0 40.4 60.6



#63
4-Nitroaniline
Concen: 20.060 ng/ul
RT: 15.675 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:138 Resp: 424461
Ion Ratio Lower Upper
138 100
92 51.2 39.0 58.4
108 74.1 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.308 ng/ul
RT: 15.734 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 468240
Ion Ratio Lower Upper
198 100
182 4.5 3.3 4.9
77 22.7 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 19.839 ng/ul
RT: 15.863 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:169 Resp: 2236927
Ion Ratio Lower Upper
169 100
168 67.0 53.0 79.6
167 36.1 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 19.253 ng/ul
RT: 16.545 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 850347
Ion Ratio Lower Upper
248 100
250 97.8 77.4 116.0
141 62.7 48.3 72.5



#69
Hexachlorobenzene
Concen: 19.266 ng/ul
RT: 16.663 min Scan# 2291
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:284 Resp: 1010389
Ion Ratio Lower Upper
284 100
142 29.7 22.6 34.0
249 30.7 24.6 37.0





#70

Atrazine

Concen: 19.340 ng/ul

RT: 16.816 min Scan# 21

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 849811

Ion Ratio Lower Upper

200 100

173 25.2 20.5 30.7

215 49.5 39.9 59.9



#71

Pentachlorophenol

Concen: 18.148 ng/ul

RT: 17.010 min Scan# 2350

Delta R.T. 0.000 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20




Tgt Ion:266 Resp: 576381


Ion Ratio Lower Upper

266 100

264 61.9 50.9 76.3

268 63.0 51.4 77.2





#72

Phenanthrene

Concen: 19.963 ng/ul

RT: 17.404 min Scan# 2419

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 4228922

Ion Ratio Lower Upper

178 100

179 15.5 12.5 18.7

176 19.3 15.4 23.0



#74

Anthracene

Concen: 20.212 ng/ul

RT: 17.498 min Scan# 2433

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20




Tgt Ion:178 Resp: 4275254


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.9 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.830 ng/uL
RT: 13.410 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 1150658
Ion Ratio Lower Upper
216 100
214 78.7 62.2 93.2
218 48.4 38.9 58.3



#76
Pentachlorobenzene
Concen: 19.482 ng/uL
RT: 14.928 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:250 Resp: 1152527
Ion Ratio Lower Upper
250 100
248 63.9 50.6 76.0
252 64.5 51.1 76.7









#77
 Carbazole
 Concen: 21.472 ng/ul
 RT: 17.769 min Scan# 2480
 Delta R.T. 0.000 min
 Lab File: BPCCC075.D
 Acq: 11 Dec 2022 12:20




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:167 Resp: 3824878
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.2 25.8
 139 12.1 9.4 14.2



#78
 Di-n-butylphthalate
 Concen: 20.280 ng/ul
 RT: 18.310 min Scan# 2571
 Delta R.T. 0.000 min
 Lab File: BPCCC075.D
 Acq: 11 Dec 2022 12:20

Tgt Ion:149 Resp: 4398459
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.5 11.3
 104 4.6 3.8 5.6





#80

Fluoranthene

Concen: 20.792 ng/ul

RT: 19.369 min Scan# 21

Delta R.T. 0.000 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:202 Resp: 5032366

Ion Ratio Lower Upper

202 100

101 9.6 8.2 12.2

100 7.3 6.1 9.1



#82

Pyrene

Concen: 21.019 ng/ul

RT: 19.722 min Scan# 2811

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20




Tgt Ion:202 Resp: 5247758


Ion Ratio Lower Upper

202 100

101 11.6 9.4 14.2

100 9.5 7.8 11.8









#83
Butylbenzylphthalate
Concen: 20.306 ng/ul
RT: 20.586 min Scan# 2958
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 1958602
Ion Ratio Lower Upper
149 100
91 67.0 53.4 80.0
206 23.4 17.9 26.9



#84
3,3'-Dichlorobenzidine
Concen: 17.268 ng/ul
RT: 21.363 min Scan# 3090
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:252 Resp: 1449293
Ion Ratio Lower Upper
252 100
254 64.7 51.8 77.8
126 9.5 7.4 11.2









#85
Benzo(a)anthracene
Concen: 20.229 ng/ul
RT: 21.433 min Scan# 3103
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 4989045
Ion Ratio Lower Upper
228 100
229 19.9 16.1 24.1
226 27.2 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.829 ng/ul
RT: 21.339 min Scan# 3086
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:149 Resp: 2833956
Ion Ratio Lower Upper
149 100
167 27.6 22.5 33.7
279 4.3 3.8 5.8





#87

Chrysene

Concen: 19.932 ng/ul

RT: 21.486 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 4648813

Ion Ratio Lower Upper

228 100

226 30.6 24.5 36.7

229 20.0 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 23.384 ng/ul




RT: 22.298 min Scan# 3249


Delta R.T. 0.000 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Tgt Ion:149 Resp: 4617214





#90

Benzo(b)fluoranthene

Concen: 20.490 ng/ul

RT: 23.174 min Scan# 31

Delta R.T. 0.000 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 4605044

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 8.7 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.165 ng/ul

RT: 23.221 min Scan# 3406

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20




Tgt Ion:252 Resp: 4493993


Ion Ratio Lower Upper

252 100

253 22.2 17.7 26.5

125 8.9 6.8 10.2





#93

Benzo(a)pyrene

Concen: 20.018 ng/ul

RT: 23.827 min Scan# 3

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 3917356

Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3

125 9.5 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.629 ng/ul

RT: 26.539 min Scan# 3970

Delta R.T. -0.006 min

Lab File: BPCCC075.D

Acq: 11 Dec 2022 12:20




Tgt Ion:276 Resp: 4784751


Ion Ratio Lower Upper

276 100

138 21.3 17.1 25.7

277 25.6 20.0 30.0









#95
Dibenzo(a,h)anthracene
Concen: 19.707 ng/ul
RT: 26.557 min Scan# 3975
Delta R.T. 0.000 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Instrument :
BNA_P
ClientSampleId :

Tgt Ion:278 Resp: 3977277
Ion Ratio Lower Upper
278 100
139 13.9 11.1 16.7
279 24.2 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 18.691 ng/ul
RT: 27.339 min Scan# 4106
Delta R.T. -0.006 min
Lab File: BPCCC075.D
Acq: 11 Dec 2022 12:20

Tgt Ion:276 Resp: 3657809
Ion Ratio Lower Upper
276 100
138 18.1 15.4 23.2
277 23.9 19.1 28.7

