

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121022\
 Data File : BP013061.D
 Acq On : 11 Dec 2022 03:49
 Operator : CG/JU
 Sample : N5927-22
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EXZA8

Quant Time: Dec 12 03:31:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 12 01:32:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	744412	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	3276941	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.610	164	2190188	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.363	188	5028226	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	4448176	20.000	ng/u1	0.00
88) Perylene-d12	23.939	264	4294308	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	78663	4.518	ng/uL	0.00
4) Pyridine-d5	3.781	84	393643	7.959	ng/u1	0.00
7) Phenol-d5	7.122	99	395573	6.524	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	1470610	40.206	ng/u1	0.00
11) 2-Chlorophenol-d4	7.499	132	1382619	28.966	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	837481	16.857	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	988307	41.049	ng/u1	0.00
24) 2-Nitrophenol-d4	9.863	143	1084881	40.024	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.399	165	1794762	34.089	ng/u1	0.00
31) 4-Chloroaniline-d4	10.922	131	2027189	27.274	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	6652069	39.519	ng/u1	0.00
49) Acenaphthylene-d8	14.304	160	7480538	40.444	ng/u1	0.00
54) 4-Nitrophenol-d4	14.798	143	152785	5.046	ng/u1	0.00
60) Fluorene-d10	15.604	176	5741888	40.321	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.722	200	1109270	34.281	ng/u1	0.00
73) Anthracene-d10	17.463	188	9095266	40.094	ng/u1	0.00
81) Pyrene-d10	19.692	212	10408001	40.764	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.780	264	8825053	41.230	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121022\
Data File : BP013061.D
Acq On : 11 Dec 2022 03:49
Operator : CG/JU
Sample : N5927-22
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EXZA8

Quant Time: Dec 12 03:31:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 12 01:32:54 2022
Response via : Initial Calibration

