

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121024\
 Data File : BP023356.D
 Acq On : 10 Dec 2024 21:57
 Operator : RC/JU
 Sample : PB165469BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK469

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/12/2024
 Supervised By :mohammad ahmed 12/13/2024

Quant Time: Dec 11 03:18:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 00:30:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.540	152	96110	20.000	ng/u1	0.00
20) Naphthalene-d8	10.281	136	349851	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.140	164	252662	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.945	188	637341	20.000	ng/u1	-0.01
79) Chrysene-d12	21.386	240	710874	20.000	ng/u1	0.00
88) Perylene-d12	24.557	264	798130	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.105	96	14021	5.371	ng/uL	0.00
4) Pyridine-d5	3.523	84	164879	24.562	ng/u1	0.00
7) Phenol-d5	6.752	99	195788	25.963	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.893	67	125885	26.459	ng/u1	0.00
11) 2-Chlorophenol-d4	7.093	132	146004	26.233	ng/u1	0.00
15) 4-Methylphenol-d8	8.258	113	151625	24.913	ng/u1	0.00
21) Nitrobenzene-d5	8.687	128	67929	26.002	ng/u1	0.00
24) 2-Nitrophenol-d4	9.399	143	77366	25.695	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.946	165	157129m	24.726	ng/u1	0.01
31) 4-Chloroaniline-d4	10.446	131	167876	23.582	ng/u1	0.00
46) Dimethylphthalate-d6	13.557	166	538898	27.475	ng/u1	0.00
49) Acenaphthylene-d8	13.834	160	531099	25.671	ng/u1	0.00
54) 4-Nitrophenol-d4	14.434	143	64654m	26.899	ng/u1	0.02
60) Fluorene-d10	15.151	176	443862	25.728	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.322	200	85181	22.114	ng/u1	0.01
73) Anthracene-d10	17.057	188	752941	25.885	ng/u1	0.00
81) Pyrene-d10	19.433	212	988841	26.625	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.339	264	1111337	26.011	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121024\
Data File : BP023356.D
Acq On : 10 Dec 2024 21:57
Operator : RC/JU
Sample : PB165469BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK469

Quant Time: Dec 11 03:18:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112624.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 00:30:30 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/12/2024
Supervised By :mohammad ahmed 12/13/2024

