


Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP121119\  
 Data File : BP001306.D  
 Acq On : 11 Dec 2019 09:46  
 Operator : JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_P\METHODS\8270-BP112719.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Dec 05 12:34:47 2019



AutoFind: Scans 2380, 2381, 2382; Background Corrected with Scan 2374

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 44.3      | 27152   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 485     | PASS             |
| 69          | 198          | 0.00         | 100          | 44.8      | 27455   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 159     | PASS             |
| 127         | 198          | 10           | 80           | 46.4      | 28477   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 61338   | PASS             |
| 199         | 198          | 5            | 9            | 6.6       | 4028    | PASS             |
| 275         | 198          | 10           | 60           | 26.4      | 16181   | PASS             |
| 365         | 198          | 1            | 100          | 3.5       | 2170    | PASS             |
| 441         | 198          | 0.01         | 100          | 13.0      | 7951    | PASS             |
| 442         | 198          | 50           | 100          | 86.8      | 53232   | PASS             |
| 443         | 442          | 15           | 24           | 18.2      | 9708    | PASS             |