

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121224\
 Data File : BP023394.D
 Acq On : 12 Dec 2024 13:49
 Operator : RC/JU
 Sample : P5153-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E28N9

Quant Time: Dec 13 01:32:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 00:30:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.534	152	113156	20.000	ng/u1	0.00
20) Naphthalene-d8	10.281	136	419550	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.151	164	319255	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.963	188	740904	20.000	ng/u1	0.00
79) Chrysene-d12	21.392	240	793790	20.000	ng/u1	0.00
88) Perylene-d12	24.557	264	886683	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.105	96	9531	3.101	ng/uL	0.00
4) Pyridine-d5	3.517	84	121989	15.435	ng/u1	0.00
7) Phenol-d5	6.752	99	168662m	18.997	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.887	67	109296	19.512	ng/u1	0.00
11) 2-Chlorophenol-d4	7.087	132	126513	19.307	ng/u1	0.00
15) 4-Methylphenol-d8	8.258	113	137302	19.161	ng/u1	0.00
21) Nitrobenzene-d5	8.681	128	60202	19.216	ng/u1	0.00
24) 2-Nitrophenol-d4	9.393	143	70761	19.597	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.940	165	146873	19.273	ng/u1	0.00
31) 4-Chloroaniline-d4	10.440	131	159170m	18.644	ng/u1	0.00
46) Dimethylphthalate-d6	13.557	166	497046	20.055	ng/u1	0.00
49) Acenaphthylene-d8	13.840	160	514662	19.688	ng/u1	0.00
54) 4-Nitrophenol-d4	14.463	143	64548m	21.253	ng/u1	0.05
60) Fluorene-d10	15.175	176	430393	19.743	ng/u1	0.02
65) 4,6-Dinitro-2-methylph...	15.340	200	35760m	7.986	ng/u1	0.03
73) Anthracene-d10	17.057	188	711234	21.034	ng/u1	0.00
81) Pyrene-d10	19.451	212	840579	20.269	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.351	264	858348	18.083	ng/u1	0.00


Target Compounds					Qvalue
8) Phenol	6.781	94	9314m	1.008	ng/u1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121224\
Data File : BP023394.D
Acq On : 12 Dec 2024 13:49
Operator : RC/JU
Sample : P5153-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E28N9


Quant Time: Dec 13 01:32:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112624.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 00:30:30 2024
Response via : Initial Calibration





#8
Phenol
Concen: 1.008 ng/ul m
RT: 6.781 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP023394.D
Acq: 12 Dec 2024 13:49

Instrument :
BNA_P
ClientSampleId :
E28N9



Tgt Ion: 94 Resp: 9314
Ion Ratio Lower Upper
94 100
65 36.9 28.0 42.0
66 76.4 42.0 63.0#

