

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121322\
 Data File : BP013096.D
 Acq On : 13 Dec 2022 09:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 14 04:41:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 13 03:29:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	40.508	-1.3	93	0.00
3	Pyridine	40.000	39.308	1.7	90	0.00
4	n-Nitrosodimethylamine	40.000	43.364	-8.4	98	0.00
5 S	2-Fluorophenol	80.000	83.705	-4.6	94	0.00
6	Aniline	40.000	41.366	-3.4	92	0.00
7 S	Phenol-d6	80.000	84.343	-5.4	93	0.00
8	2-Chlorophenol	40.000	40.887	-2.2	92	0.00
9	Benzaldehyde	40.000	39.697	0.8	94	0.00
10 C	Phenol	40.000	41.646	-4.1	92	0.00
11	bis(2-Chloroethyl)ether	40.000	40.518	-1.3	91	0.00
12	1,3-Dichlorobenzene	40.000	40.599	-1.5	92	0.00
13 C	1,4-Dichlorobenzene	40.000	40.365	-0.9	92	0.00
14	1,2-Dichlorobenzene	40.000	40.465	-1.2	92	0.00
15	Benzyl Alcohol	40.000	42.034	-5.1	91	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.097	-2.7	92	0.00
17	2-Methylphenol	40.000	41.454	-3.6	91	0.00
18	Hexachloroethane	40.000	40.951	-2.4	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.204	-5.5	91	0.00
20	3+4-Methylphenols	40.000	41.916	-4.8	91	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	89	0.00
22	Acetophenone	40.000	41.777	-4.4	91	0.00
23 S	Nitrobenzene-d5	80.000	85.717	-7.1	92	0.00
24	Nitrobenzene	40.000	42.077	-5.2	92	0.00
25	Isophorone	40.000	42.428	-6.1	90	0.00
26 C	2-Nitrophenol	40.000	38.793	3.0	88	0.00
27	2,4-Dimethylphenol	40.000	41.926	-4.8	91	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.397	-3.5	90	0.00
29 C	2,4-Dichlorophenol	40.000	42.152	-5.4	90	0.00
30	1,2,4-Trichlorobenzene	40.000	40.644	-1.6	90	0.00
31	Naphthalene	40.000	40.962	-2.4	90	0.00
32	Benzoic acid	40.000	37.581	6.0	87	-0.01
33	4-Chloroaniline	40.000	42.452	-6.1	91	0.00
34 C	Hexachlorobutadiene	40.000	40.534	-1.3	90	0.00
35	Caprolactam	40.000	40.718	-1.8	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.674	-6.7	91	0.00
37	2-Methylnaphthalene	40.000	41.406	-3.5	90	0.00
38	1-Methylnaphthalene	40.000	41.366	-3.4	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.527	-3.8	90	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.376	-0.9	85	0.00
42 S	2,4,6-Tribromophenol	80.000	85.692	-7.1	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.949	-7.4	90	0.00
44	2,4,5-Trichlorophenol	40.000	43.255	-8.1	92	0.00
45 S	2-Fluorobiphenyl	80.000	84.754	-5.9	92	0.00
46	1,1'-Biphenyl	40.000	41.719	-4.3	91	0.00
47	2-Chloronaphthalene	40.000	41.838	-4.6	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121322\
 Data File : BP013096.D
 Acq On : 13 Dec 2022 09:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 14 04:41:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 13 03:29:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.830	-4.6	95	0.00
49	Acenaphthylene	40.000	42.699	-6.7	92	0.00
50	Dimethylphthalate	40.000	41.939	-4.8	94	0.00
51	2,6-Dinitrotoluene	40.000	44.070	-10.2	94	0.00
52 C	Acenaphthene	40.000	42.270	-5.7	93	0.00
53	3-Nitroaniline	40.000	41.248	-3.1	95	0.00
54 P	2,4-Dinitrophenol	40.000	39.833	0.4	95	0.00
55	Dibenzofuran	40.000	41.866	-4.7	94	0.00
56 P	4-Nitrophenol	40.000	42.496	-6.2	97	0.00
57	2,4-Dinitrotoluene	40.000	41.315	-3.3	96	0.00
58	Fluorene	40.000	42.697	-6.7	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.708	-6.8	94	0.00
60	Diethylphthalate	40.000	42.265	-5.7	97	0.00
61	4-Chlorophenyl-phenylether	40.000	42.377	-5.9	95	0.00
62	4-Nitroaniline	40.000	41.754	-4.4	99	0.00
63	Azobenzene	40.000	44.245	-10.6	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.602	-1.5	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.927	-4.8	96	0.00
67	4-Bromophenyl-phenylether	40.000	41.260	-3.1	95	0.00
68	Hexachlorobenzene	40.000	40.710	-1.8	96	0.00
69	Atrazine	40.000	43.295	-8.2	99	0.00
70 C	Pentachlorophenol	40.000	41.116	-2.8	97	0.00
71	Phenanthrene	40.000	41.002	-2.5	98	0.00
72	Anthracene	40.000	41.553	-3.9	98	0.00
73	Carbazole	40.000	41.659	-4.1	99	0.00
74	Di-n-butylphthalate	40.000	41.832	-4.6	97	0.00
75 C	Fluoranthene	40.000	40.891	-2.2	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzidine	40.000	42.528	-6.3	92	0.00
78	Pyrene	40.000	43.067	-7.7	99	0.00
79 S	Terphenyl-d14	80.000	88.398	-10.5	100	0.00
80	Butylbenzylphthalate	40.000	40.360	-0.9	95	0.00
81	Benzo(a)anthracene	40.000	43.274	-8.2	98	0.00
82	3,3'-Dichlorobenzidine	40.000	45.840	-14.6	97	0.00
83	Chrysene	40.000	42.926	-7.3	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.110	-2.8	94	0.00
85 c	Di-n-octyl phthalate	40.000	41.862	-4.7	93	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.144	-5.4	98	0.00
88	Benzo(b)fluoranthene	40.000	41.064	-2.7	99	0.00
89	Benzo(k)fluoranthene	40.000	40.535	-1.3	95	0.00
90 C	Benzo(a)pyrene	40.000	40.986	-2.5	96	0.00
91	Dibenzo(a,h)anthracene	40.000	42.106	-5.3	98	0.00
92	Benzo(g,h,i)perylene	40.000	42.102	-5.3	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121322\
Data File : BP013096.D
Acq On : 13 Dec 2022 09:47
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 14 04:41:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 13 03:29:24 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0