


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121322\
 Data File : BP013095.D
 Acq On : 13 Dec 2022 09:14
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 13 03:29:24 2022






AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.6	312791	PASS
68	69	0.00	2	1.7	5052	PASS
69	198	0.00	100	29.1	297524	PASS
70	69	0.00	2	0.5	1509	PASS
127	198	10	80	39.7	405393	PASS
197	198	0.00	2	0.5	4881	PASS
198	198	100	100	100.0	1021664	PASS
199	198	5	9	6.6	67859	PASS
275	198	10	60	27.9	285141	PASS
365	198	1	100	3.8	38918	PASS
441	198	0.01	100	15.3	156664	PASS
442	442	100	100	100.0	1009981	PASS
443	442	15	24	19.3	194926	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121322\
 Data File : BP013095.D
 Acq On : 13 Dec 2022 09:14
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Dec 14 06:37:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 13 03:29:24 2022
 Response via : Initial Calibration



TIC: BP013095.D\data.ms

(70) Pentachlorophenol (C)

17.016min (+ 0.012) 164487.34 ng




response 1404195

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.60	63.91
264.00	63.30	62.69
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121322\
Data File : BP013095.D
Acq On : 13 Dec 2022 09:14
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Dec 14 06:37:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 13 03:29:24 2022
Response via : Initial Calibration



TIC: BP013095.D\data.ms

(77) Benzidine

19.557min (+ 0.012) 77450.74 ng

response 6233265

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.93
183.00	11.70	11.79
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/13/2022	BNA_P	<u>BP013095.D</u>
Compound Name	Response	Retention Time
DDT	3156927	20.757
DDD	30802	20.374
DDE	816	19.827
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
31618	3188545	0.99

Instrument :
BNA_P
ClientSampleId :
DFTPP