

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP121421\  
 Data File : BP008394.D  
 Acq On : 14 Dec 2021 15:09  
 Operator : CG/JU  
 Sample : SSTDICCC040  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICCC040

Manual Integrations  
 APPROVED

Reviewed By : Christian Giraldo 12/15/2021  
 Supervised By : Jagrut Upadhyay 12/15/2021

Quant Time: Dec 14 16:54:28 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP121421.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Dec 14 16:50:46 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| -----                         |        |      |          |         |       |          |
| Internal Standards            |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.746  | 152  | 71215    | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8            | 10.540 | 136  | 272264   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10          | 14.398 | 164  | 170501   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10          | 17.163 | 188  | 319869   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12              | 21.269 | 240  | 268715   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12              | 23.639 | 264  | 289679   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds   |        |      |          |         |       |          |
| 5) 2-Fluorophenol             | 5.346  | 112  | 362406   | 81.936  | ng    | 0.00     |
| 7) Phenol-d6                  | 6.940  | 99   | 509667   | 85.360  | ng    | 0.00     |
| 23) Nitrobenzene-d5           | 8.910  | 82   | 529622   | 88.457  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol      | 15.904 | 330  | 206198   | 94.617  | ng    | 0.00     |
| 45) 2-Fluorobiphenyl          | 13.016 | 172  | 1104772  | 94.118  | ng    | 0.00     |
| 79) Terphenyl-d14             | 19.733 | 244  | 1291252  | 100.427 | ng    | 0.00     |
| Target Compounds              |        |      |          |         |       |          |
|                               |        |      |          |         |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.246  | 88   | 74950    | 36.322  | ng    | 93       |
| 3) Pyridine                   | 3.652  | 79   | 221065m  | 40.139  | ng    |          |
| 4) n-Nitrosodimethylamine     | 3.564  | 42   | 86446    | 37.631  | ng    | # 96     |
| 6) Aniline                    | 7.081  | 93   | 295620   | 42.237  | ng    | 95       |
| 8) 2-Chlorophenol             | 7.316  | 128  | 190771   | 42.134  | ng    | 99       |
| 9) Benzaldehyde               | 6.893  | 77   | 140161   | 35.342  | ng    | 97       |
| 10) Phenol                    | 6.969  | 94   | 257137   | 41.908  | ng    | 98       |
| 11) bis(2-Chloroethyl)ether   | 7.181  | 93   | 202879   | 41.368  | ng    | 100      |
| 12) 1,3-Dichlorobenzene       | 7.634  | 146  | 215920   | 41.314  | ng    | 98       |
| 13) 1,4-Dichlorobenzene       | 7.781  | 146  | 220368   | 41.589  | ng    | 97       |
| 14) 1,2-Dichlorobenzene       | 8.099  | 146  | 214352   | 40.676  | ng    | 98       |
| 15) Benzyl Alcohol            | 7.999  | 79   | 213060   | 45.038  | ng    | 96       |
| 16) 2,2'-oxybis(1-Chloropr... | 8.281  | 45   | 289096   | 39.068  | ng    | 98       |
| 17) 2-Methylphenol            | 8.210  | 107  | 172710   | 43.267  | ng    | 97       |
| 18) Hexachloroethane          | 8.816  | 117  | 81391    | 41.311  | ng    | 90       |
| 19) n-Nitroso-di-n-propyla... | 8.563  | 70   | 181391   | 43.904  | ng    | 97       |
| 20) 3+4-Methylphenols         | 8.546  | 107  | 238132   | 43.225  | ng    | 97       |
| 22) Acetophenone              | 8.575  | 105  | 327054   | 43.917  | ng    | 99       |
| 24) Nitrobenzene              | 8.952  | 77   | 248386   | 42.778  | ng    | 97       |
| 25) Isophorone                | 9.487  | 82   | 446293   | 42.606  | ng    | 99       |
| 26) 2-Nitrophenol             | 9.663  | 139  | 110694   | 44.185  | ng    | # 95     |
| 27) 2,4-Dimethylphenol        | 9.740  | 122  | 159558   | 42.587  | ng    | 99       |
| 28) bis(2-Chloroethoxy)met... | 9.963  | 93   | 279537   | 41.872  | ng    | # 98     |
| 29) 2,4-Dichlorophenol        | 10.210 | 162  | 197084   | 44.044  | ng    | 98       |
| 30) 1,2,4-Trichlorobenzene    | 10.410 | 180  | 225550   | 43.428  | ng    | 96       |
| 31) Naphthalene               | 10.593 | 128  | 577741   | 41.423  | ng    | 100      |
| 32) Benzoic acid              | 9.963  | 122  | 128694   | 41.991  | ng    | 99       |
| 33) 4-Chloroaniline           | 10.716 | 127  | 237802   | 41.098  | ng    | 97       |
| 34) Hexachlorobutadiene       | 10.875 | 225  | 165735   | 46.629  | ng    | 98       |
| 35) Caprolactam               | 11.534 | 113  | 37384    | 37.610  | ng    | 96       |
| 36) 4-Chloro-3-methylphenol   | 11.863 | 107  | 219674   | 42.944  | ng    | 97       |
| 37) 2-Methylnaphthalene       | 12.210 | 142  | 406453   | 41.204  | ng    | 98       |
| 38) 1-Methylnaphthalene       | 12.428 | 142  | 398177   | 41.469  | ng    | 98       |
| 40) 1,2,4,5-Tetrachloroben... | 12.587 | 216  | 265965   | 48.021  | ng    | 99       |
| 41) Hexachlorocyclopentadiene | 12.557 | 237  | 62524    | 29.363  | ng    | 99       |
| 43) 2,4,6-Trichlorophenol     | 12.840 | 196  | 173841   | 46.328  | ng    | 97       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP121421\  
 Data File : BP008394.D  
 Acq On : 14 Dec 2021 15:09  
 Operator : CG/JU  
 Sample : SSTDICCC040  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICCC040

Manual Integrations  
 APPROVED


Reviewed By : Christian Giraldo 12/15/2021  
 Supervised By : Jagrut Upadhyay 12/15/2021

Quant Time: Dec 14 16:54:28 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP121421.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Dec 14 16:50:46 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol     | 12.928 | 196  | 178746   | 44.454 | ng    | 96       |
| 46) 1,1'-Biphenyl             | 13.228 | 154  | 544834   | 44.043 | ng    | 97       |
| 47) 2-Chloronaphthalene       | 13.269 | 162  | 431231   | 44.141 | ng    | 98       |
| 48) 2-Nitroaniline            | 13.493 | 65   | 148296   | 44.050 | ng    | 94       |
| 49) Acenaphthylene            | 14.122 | 152  | 636790   | 42.252 | ng    | 99       |
| 50) Dimethylphthalate         | 13.869 | 163  | 529182   | 41.332 | ng    | 100      |
| 51) 2,6-Dinitrotoluene        | 13.992 | 165  | 110190   | 41.469 | ng    | 97       |
| 52) Acenaphthene              | 14.463 | 154  | 380782   | 42.395 | ng    | 98       |
| 53) 3-Nitroaniline            | 14.328 | 138  | 105713   | 39.257 | ng    | 99       |
| 54) 2,4-Dinitrophenol         | 14.557 | 184  | 60413    | 38.453 | ng    | 94       |
| 55) Dibenzofuran              | 14.798 | 168  | 637396   | 41.528 | ng    | 98       |
| 56) 4-Nitrophenol             | 14.681 | 139  | 64039    | 31.026 | ng    | 86       |
| 57) 2,4-Dinitrotoluene        | 14.792 | 165  | 146848   | 40.270 | ng    | # 83     |
| 58) Fluorene                  | 15.457 | 166  | 491061   | 39.379 | ng    | 98       |
| 59) 2,3,4,6-Tetrachlorophenol | 15.039 | 232  | 149573   | 41.832 | ng    | 93       |
| 60) Diethylphthalate          | 15.234 | 149  | 493110   | 38.811 | ng    | 99       |
| 61) 4-Chlorophenyl-phenyle... | 15.445 | 204  | 281842   | 41.205 | ng    | 99       |
| 62) 4-Nitroaniline            | 15.498 | 138  | 94097    | 33.675 | ng    | 89       |
| 63) Azobenzene                | 15.745 | 77   | 512494   | 39.434 | ng    | 96       |
| 65) 4,6-Dinitro-2-methylph... | 15.563 | 198  | 92562    | 44.242 | ng    | 97       |
| 66) n-Nitrosodiphenylamine    | 15.669 | 169  | 421297   | 45.469 | ng    | 100      |
| 67) 4-Bromophenyl-phenylether | 16.345 | 248  | 173942   | 48.903 | ng    | 99       |
| 68) Hexachlorobenzene         | 16.469 | 284  | 189055   | 49.724 | ng    | 96       |
| 69) Atrazine                  | 16.634 | 200  | 94962    | 39.490 | ng    | 96       |
| 70) Pentachlorophenol         | 16.828 | 266  | 89980    | 39.799 | ng    | 97       |
| 71) Phenanthrene              | 17.204 | 178  | 700135   | 41.456 | ng    | 99       |
| 72) Anthracene                | 17.298 | 178  | 697850   | 41.637 | ng    | 100      |
| 73) Carbazole                 | 17.575 | 167  | 625896   | 38.255 | ng    | 100      |
| 74) Di-n-butylphthalate       | 18.128 | 149  | 717621   | 35.038 | ng    | 100      |
| 75) Fluoranthene              | 19.186 | 202  | 760220   | 33.415 | ng    | 97       |
| 77) Benzidine                 | 19.375 | 184  | 120858   | 32.788 | ng    | 96       |
| 78) Pyrene                    | 19.539 | 202  | 769103   | 39.599 | ng    | 99       |
| 80) Butylbenzylphthalate      | 20.416 | 149  | 283332   | 37.028 | ng    | 98       |
| 81) Benzo(a)anthracene        | 21.251 | 228  | 734543   | 41.245 | ng    | 99       |
| 82) 3,3'-Dichlorobenzidine    | 21.186 | 252  | 334022   | 39.808 | ng    | 99       |
| 83) Chrysene                  | 21.304 | 228  | 690702   | 40.757 | ng    | 99       |
| 84) Bis(2-ethylhexyl)phtha... | 21.180 | 149  | 400105   | 34.439 | ng    | # 99     |
| 85) Di-n-octyl phthalate      | 22.086 | 149  | 687941   | 34.857 | ng    | 99       |
| 87) Indeno(1,2,3-cd)pyrene    | 26.109 | 276  | 943159   | 44.754 | ng    | 96       |
| 88) Benzo(b)fluoranthene      | 22.916 | 252  | 707311   | 40.503 | ng    | 99       |
| 89) Benzo(k)fluoranthene      | 22.963 | 252  | 715469   | 41.848 | ng    | 99       |
| 90) Benzo(a)pyrene            | 23.533 | 252  | 615095   | 41.211 | ng    | # 98     |
| 91) Dibenzo(a,h)anthracene    | 26.121 | 278  | 788727   | 45.064 | ng    | # 97     |
| 92) Benzo(g,h,i)perylene      | 26.862 | 276  | 782646   | 44.325 | ng    | 96       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed









#1  
1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.746 min Scan# 809  
Delta R.T. -0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:152 Resp: 71215  
Ion Ratio Lower Upper  
152 100  
150 156.9 125.7 188.5  
115 63.5 50.7 76.1




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#2  
1,4-Dioxane  
Concen: 36.322 ng  
RT: 3.246 min Scan# 44  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion: 88 Resp: 74950  
Ion Ratio Lower Upper  
88 100  
58 72.9 63.5 95.3  
43 26.8 23.4 35.2





#3  
Pyridine  
Concen: 40.139 ng m  
RT: 3.652 min Scan# 113  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 79 Resp: 221065  
Ion Ratio Lower Upper  
79 100  
52 68.5 60.0 90.0  
51 34.7 29.6 44.4

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#4  
n-Nitrosodimethylamine  
Concen: 37.631 ng  
RT: 3.564 min Scan# 98  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion: 42 Resp: 86446  
Ion Ratio Lower Upper  
42 100  
74 139.9 108.0 162.0  
44 10.3 5.8 8.6





#5  
2-Fluorophenol  
Concen: 81.936 ng  
RT: 5.346 min Scan# 401  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 112 Resp: 362400  
Ion Ratio Lower Upper  
112 100  
64 66.0 53.4 80.2  
63 34.3 27.8 41.8

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#6  
Aniline  
Concen: 42.237 ng  
RT: 7.081 min Scan# 696  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion: 93 Resp: 295620  
Ion Ratio Lower Upper  
93 100  
66 45.3 32.9 49.3  
65 21.2 17.5 26.3





#7  
Phenol-d6  
Concen: 85.360 ng  
RT: 6.940 min Scan# 61  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 99 Resp: 50966  
Ion Ratio Lower Upper  
99 100  
42 18.7 14.6 22.0  
71 41.5 31.1 46.7

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#8  
2-Chlorophenol  
Concen: 42.134 ng  
RT: 7.316 min Scan# 736  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:128 Resp: 190771  
Ion Ratio Lower Upper  
128 100  
130 32.3 13.5 53.5  
64 56.6 37.4 77.4





#9  
Benzaldehyde  
Concen: 35.342 ng  
RT: 6.893 min Scan# 664  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 77 Resp: 14016  
Ion Ratio Lower Upper  
77 100  
105 89.6 71.8 111.8  
106 84.7 67.3 107.3

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#10  
Phenol  
Concen: 41.908 ng  
RT: 6.969 min Scan# 677  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion: 94 Resp: 257137  
Ion Ratio Lower Upper  
94 100  
65 30.7 10.9 50.9  
66 46.4 24.3 64.3





#11  
bis(2-Chloroethyl)ether  
Concen: 41.368 ng  
RT: 7.181 min Scan# 713  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 93 Resp: 202879  
Ion Ratio Lower Upper  
93 100  
63 82.0 61.8 101.8  
95 33.1 12.5 52.5

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#12  
1,3-Dichlorobenzene  
Concen: 41.314 ng  
RT: 7.634 min Scan# 790  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09



Tgt Ion:146 Resp: 215920  
Ion Ratio Lower Upper  
146 100  
148 63.5 51.9 77.9  
75 35.1 29.0 43.6









#13  
1,4-Dichlorobenzene  
Concen: 41.589 ng  
RT: 7.781 min Scan# 815  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:146 Resp: 220368  
Ion Ratio Lower Upper  
146 100  
148 65.1 49.6 74.4  
111 43.7 34.7 52.1




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#14  
1,2-Dichlorobenzene  
Concen: 40.676 ng  
RT: 8.099 min Scan# 869  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:146 Resp: 214352  
Ion Ratio Lower Upper  
146 100  
148 65.0 51.8 77.6  
111 46.5 35.0 52.6





#15  
Benzyl Alcohol  
Concen: 45.038 ng  
RT: 7.999 min Scan# 852  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 79 Resp: 213060  
Ion Ratio Lower Upper  
79 100  
108 65.3 55.5 83.3  
77 61.8 51.0 76.6

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 39.068 ng  
RT: 8.281 min Scan# 900  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion: 45 Resp: 289096  
Ion Ratio Lower Upper  
45 100  
77 14.7 0.0 33.8  
79 11.1 0.0 30.4





#17  
2-Methylphenol  
Concen: 43.267 ng  
RT: 8.210 min Scan# 888  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:107 Resp: 172710  
Ion Ratio Lower Upper  
107 100  
108 109.6 90.2 135.2  
77 53.3 41.3 61.9  
79 53.7 41.0 61.6

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#18  
Hexachloroethane  
Concen: 41.311 ng  
RT: 8.816 min Scan# 991  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:117 Resp: 81391  
Ion Ratio Lower Upper  
117 100  
119 100.3 74.7 112.1  
201 123.8 88.6 132.8





#19  
n-Nitroso-di-n-propylamine  
Concen: 43.904 ng  
RT: 8.563 min Scan# 948  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 70 Resp: 18139  
Ion Ratio Lower Upper  
70 100  
42 53.9 45.4 68.2  
101 9.8 7.6 11.4  
130 18.5 15.1 22.7

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#20  
3+4-Methylphenols  
Concen: 43.225 ng  
RT: 8.546 min Scan# 945  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:107 Resp: 238132  
Ion Ratio Lower Upper  
107 100  
108 87.6 68.2 108.2  
77 41.4 16.9 56.9  
79 35.9 12.5 52.5





#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.540 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 136   | Resp: | 272264 |
| Ion Ratio | Lower | Upper |        |
| 136       | 100   |       |        |
| 137       | 10.9  | 9.0   | 13.6   |
| 54        | 10.3  | 8.2   | 12.4   |
| 68        | 6.6   | 5.8   | 8.8    |




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#22  
Acetophenone  
Concen: 43.917 ng  
RT: 8.575 min Scan# 950  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 105   | Resp: | 327054 |
| Ion Ratio | Lower | Upper |        |
| 105       | 100   |       |        |
| 71        | 4.6   | 4.0   | 6.0    |
| 51        | 30.3  | 23.8  | 35.6   |
| 120       | 20.1  | 16.0  | 24.0   |









#23  
Nitrobenzene-d5  
Concen: 88.457 ng  
RT: 8.910 min Scan# 1007  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion: 82 Resp: 52962  
Ion Ratio Lower Upper  
82 100  
128 37.1 29.6 44.4  
54 52.1 42.7 64.1




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#24  
Nitrobenzene  
Concen: 42.778 ng  
RT: 8.952 min Scan# 1014  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion: 77 Resp: 248386  
Ion Ratio Lower Upper  
77 100  
123 39.9 30.4 45.6  
65 15.3 11.3 16.9





#25  
Isophorone  
Concen: 42.606 ng  
RT: 9.487 min Scan# 1105  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 82 Resp: 44629  
Ion Ratio Lower Upper  
82 100  
95 8.0 5.9 8.9  
138 15.6 12.7 19.1




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#26  
2-Nitrophenol  
Concen: 44.185 ng  
RT: 9.663 min Scan# 1135  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:139 Resp: 110694  
Ion Ratio Lower Upper  
139 100  
109 43.1 28.3 42.5#  
65 62.8 49.7 74.5





#27  
2,4-Dimethylphenol  
Concen: 42.587 ng  
RT: 9.740 min Scan# 1148  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:122 Resp: 159558  
Ion Ratio Lower Upper  
122 100  
107 131.5 106.6 160.0  
121 60.4 48.1 72.1




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#28  
bis(2-Chloroethoxy)methane  
Concen: 41.872 ng  
RT: 9.963 min Scan# 1186  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion: 93 Resp: 279537  
Ion Ratio Lower Upper  
93 100  
95 32.3 25.4 38.2  
123 11.8 7.8 11.6#





#29  
2,4-Dichlorophenol  
Concen: 44.044 ng  
RT: 10.210 min Scan# 1228  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:162 Resp: 197084  
Ion Ratio Lower Upper  
162 100  
164 62.4 43.9 83.9  
98 38.9 17.9 57.9

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#30  
1,2,4-Trichlorobenzene  
Concen: 43.428 ng  
RT: 10.410 min Scan# 1262  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:180 Resp: 225550  
Ion Ratio Lower Upper  
180 100  
182 97.3 74.8 112.2  
145 30.2 23.2 34.8





#31  
Naphthalene  
Concen: 41.423 ng  
RT: 10.593 min Scan# 1293  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:128 Resp: 57774  
Ion Ratio Lower Upper  
128 100  
129 10.6 8.5 12.7  
127 13.6 10.9 16.3




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#32  
Benzoic acid  
Concen: 41.991 ng  
RT: 9.963 min Scan# 1186  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:122 Resp: 128694  
Ion Ratio Lower Upper  
122 100  
105 137.2 117.4 157.4  
77 111.1 89.2 129.2





#33  
4-Chloroaniline  
Concen: 41.098 ng  
RT: 10.716 min Scan# 1314  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040






Tgt Ion:127 Resp: 23780


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 127 | 100   |       |       |
| 129 | 33.0  | 25.4  | 38.2  |
| 65  | 38.8  | 29.5  | 44.3  |
| 92  | 24.7  | 18.2  | 27.4  |

Manual Integrations  
APPROVED

Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021






#34  
Hexachlorobutadiene  
Concen: 46.629 ng  
RT: 10.875 min Scan# 1341  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09



Tgt Ion:225 Resp: 165735


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 225 | 100   |       |       |
| 223 | 62.4  | 52.5  | 78.7  |
| 227 | 63.7  | 51.0  | 76.6  |





#35  
Caprolactam  
Concen: 37.610 ng  
RT: 11.534 min Scan# 1453  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:113 Resp: 37384  
Ion Ratio Lower Upper  
113 100  
55 185.9 174.2 214.2  
56 152.6 130.3 170.3

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#36  
4-Chloro-3-methylphenol  
Concen: 42.944 ng  
RT: 11.863 min Scan# 1509  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:107 Resp: 219674  
Ion Ratio Lower Upper  
107 100  
144 23.0 19.7 29.5  
142 72.2 59.6 89.4





#37  
2-Methylnaphthalene  
Concen: 41.204 ng  
RT: 12.210 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:142 Resp: 40645  
Ion Ratio Lower Upper  
142 100  
141 89.8 70.7 106.1  
115 40.5 30.9 46.3

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#38  
1-Methylnaphthalene  
Concen: 41.469 ng  
RT: 12.428 min Scan# 1605  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:142 Resp: 398177  
Ion Ratio Lower Upper  
142 100  
141 94.4 74.2 111.4  
116 4.7 3.4 5.2





#39  
Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.398 min Scan# 1940  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:164 Resp: 17050  
Ion Ratio Lower Upper  
164 100  
162 104.5 81.9 122.9  
160 46.2 35.4 53.0

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#40  
1,2,4,5-Tetrachlorobenzene  
Concen: 48.021 ng  
RT: 12.587 min Scan# 1632  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:216 Resp: 265965  
Ion Ratio Lower Upper  
216 100  
214 78.0 62.5 93.7  
179 20.3 16.4 24.6  
108 15.9 13.8 20.8





#41  
Hexachlorocyclopentadiene  
Concen: 29.363 ng  
RT: 12.557 min Scan# 1627  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 237 Resp: 62524  
Ion Ratio Lower Upper  
237 100  
235 62.9 42.3 82.3  
272 11.6 0.0 32.2

Manual Integrations  
APPROVED



Reviewed By : Christian Giraldo 12/15/2021  
Supervised By : Jagrut Upadhyay 12/15/2021




#42  
2,4,6-Tribromophenol  
Concen: 94.617 ng  
RT: 15.904 min Scan# 2196  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion: 330 Resp: 206198  
Ion Ratio Lower Upper  
330 100  
332 96.5 77.0 115.4  
141 31.6 28.4 42.6





#43  
2,4,6-Trichlorophenol  
Concen: 46.328 ng  
RT: 12.840 min Scan# 1675  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:196 Resp: 17384  
Ion Ratio Lower Upper  
196 100  
198 96.5 74.8 112.2  
200 31.4 24.3 36.5

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#44  
2,4,5-Trichlorophenol  
Concen: 44.454 ng  
RT: 12.928 min Scan# 1690  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:196 Resp: 178746  
Ion Ratio Lower Upper  
196 100  
198 99.8 76.2 114.4  
97 49.1 40.8 61.2  
132 24.3 19.0 28.4





#45  
2-Fluorobiphenyl  
Concen: 94.118 ng  
RT: 13.016 min Scan# 1705  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:172 Resp: 1104772  
Ion Ratio Lower Upper  
172 100  
171 35.4 28.0 42.0  
170 24.1 18.9 28.3

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#46  
1,1'-Biphenyl  
Concen: 44.043 ng  
RT: 13.228 min Scan# 1741  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:154 Resp: 544834  
Ion Ratio Lower Upper  
154 100  
153 41.6 19.7 59.7  
76 12.7 0.0 33.7





#47  
2-Chloronaphthalene  
Concen: 44.141 ng  
RT: 13.269 min Scan# 1748  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:162 Resp: 43123  
Ion Ratio Lower Upper  
162 100  
127 40.5 31.3 46.9  
164 32.5 26.9 40.3

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#48  
2-Nitroaniline  
Concen: 44.050 ng  
RT: 13.493 min Scan# 1786  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion: 65 Resp: 148296  
Ion Ratio Lower Upper  
65 100  
92 66.4 50.6 75.8  
138 92.0 68.7 103.1





#49  
Acenaphthylene  
Concen: 42.252 ng  
RT: 14.122 min Scan# 1893  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:152 Resp: 636790  
Ion Ratio Lower Upper  
152 100  
151 20.2 16.2 24.4  
153 13.1 10.9 16.3

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#50  
Dimethylphthalate  
Concen: 41.332 ng  
RT: 13.869 min Scan# 1850  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:163 Resp: 529182  
Ion Ratio Lower Upper  
163 100  
194 4.4 3.4 5.2  
164 10.1 8.2 12.2





#51  
2,6-Dinitrotoluene  
Concen: 41.469 ng  
RT: 13.992 min Scan# 1871  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:165 Resp: 110190  
Ion Ratio Lower Upper  
165 100  
63 61.8 50.8 76.2  
89 59.8 50.0 75.0

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#52  
Acenaphthene  
Concen: 42.395 ng  
RT: 14.463 min Scan# 1951  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09



Tgt Ion:154 Resp: 380782  
Ion Ratio Lower Upper  
154 100  
153 110.7 90.3 135.5  
152 53.5 43.3 64.9









#53  
3-Nitroaniline  
Concen: 39.257 ng  
RT: 14.328 min Scan# 1928  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:138 Resp: 10571  
Ion Ratio Lower Upper  
138 100  
108 13.5 9.2 13.8  
92 128.1 102.9 154.3




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#54  
2,4-Dinitrophenol  
Concen: 38.453 ng  
RT: 14.557 min Scan# 1967  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:184 Resp: 60413  
Ion Ratio Lower Upper  
184 100  
63 71.0 53.8 80.6  
154 74.2 54.2 81.2





#55  
Dibenzenofuran  
Concen: 41.528 ng  
RT: 14.798 min Scan# 2008  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:168 Resp: 637390  
Ion Ratio Lower Upper  
168 100  
139 40.1 30.6 46.0  
169 13.8 11.0 16.4

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#56  
4-Nitrophenol  
Concen: 31.026 ng  
RT: 14.681 min Scan# 1988  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:139 Resp: 64039  
Ion Ratio Lower Upper  
139 100  
109 102.0 65.9 105.9  
65 124.5 91.9 131.9





#57  
2,4-Dinitrotoluene  
Concen: 40.270 ng  
RT: 14.792 min Scan# 2007  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
Client Sample Id :  
SSTDICCC040




Tgt Ion:165 Resp: 14684  
Ion Ratio Lower Upper  
165 100  
63 66.1 38.9 58.3  
89 85.6 60.5 90.7  
182 2.8 2.3 3.5

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#58  
Fluorene  
Concen: 39.379 ng  
RT: 15.457 min Scan# 2120  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09



Tgt Ion:166 Resp: 491061  
Ion Ratio Lower Upper  
166 100  
165 97.5 79.4 119.2  
167 13.0 10.9 16.3









#59  
2,3,4,6-Tetrachlorophenol  
Concen: 41.832 ng  
RT: 15.039 min Scan# 2049  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 232   | Resp: | 14957 |
| Ion Ratio | Lower | Upper |       |
| 232       | 100   |       |       |
| 131       | 38.6  | 28.1  | 42.1  |
| 130       | 2.4   | 1.8   | 2.8   |
| 166       | 28.1  | 18.9  | 28.3  |




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#60  
Diethylphthalate  
Concen: 38.811 ng  
RT: 15.234 min Scan# 2082  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 149   | Resp: | 493110 |
| Ion Ratio | Lower | Upper |        |
| 149       | 100   |       |        |
| 177       | 23.0  | 18.2  | 27.2   |
| 150       | 12.7  | 10.3  | 15.5   |





#61  
4-Chlorophenyl-phenylether  
Concen: 41.205 ng  
RT: 15.445 min Scan# 2118  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:204 Resp: 28184  
Ion Ratio Lower Upper  
204 100  
206 33.1 26.1 39.1  
141 59.0 47.4 71.0

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#62  
4-Nitroaniline  
Concen: 33.675 ng  
RT: 15.498 min Scan# 2127  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:138 Resp: 94097  
Ion Ratio Lower Upper  
138 100  
92 61.9 40.8 80.8  
108 115.1 77.9 117.9





#63  
Azobenzene  
Concen: 39.434 ng  
RT: 15.745 min Scan# 2169  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 77 Resp: 512494  
Ion Ratio Lower Upper  
77 100  
182 26.3 3.8 43.8  
105 19.6 0.5 40.5  
51 32.3 10.0 50.0




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.163 min Scan# 2410  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Tgt Ion:188 Resp: 319869  
Ion Ratio Lower Upper  
188 100  
94 7.5 7.2 10.8  
80 8.9 8.2 12.4









#65  
4,6-Dinitro-2-methylphenol  
Concen: 44.242 ng  
RT: 15.563 min Scan# 2138  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:198 Resp: 92562  
Ion Ratio Lower Upper  
198 100  
51 49.0 25.9 65.9  
105 40.8 20.2 60.2




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#66  
n-Nitrosodiphenylamine  
Concen: 45.469 ng  
RT: 15.669 min Scan# 2156  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:169 Resp: 421297  
Ion Ratio Lower Upper  
169 100  
168 66.9 53.5 80.3  
167 36.9 29.3 43.9





#67  
4-Bromophenyl-phenylether  
Concen: 48.903 ng  
RT: 16.345 min Scan# 2271  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:248 Resp: 17394  
Ion Ratio Lower Upper  
248 100  
250 97.2 78.6 117.8  
141 73.1 57.0 85.6

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#68  
Hexachlorobenzene  
Concen: 49.724 ng  
RT: 16.469 min Scan# 2292  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:284 Resp: 189055  
Ion Ratio Lower Upper  
284 100  
142 31.0 28.6 42.8  
249 33.8 27.1 40.7





#69  
Atrazine  
Concen: 39.490 ng  
RT: 16.634 min Scan# 21  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:200 Resp: 94962  
Ion Ratio Lower Upper  
200 100  
173 23.2 6.8 46.8  
215 46.6 28.4 68.4

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#70  
Pentachlorophenol  
Concen: 39.799 ng  
RT: 16.828 min Scan# 2353  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09



Tgt Ion:266 Resp: 89980  
Ion Ratio Lower Upper  
266 100  
268 60.3 50.1 75.1  
264 64.1 49.5 74.3









#71  
Phenanthrene  
Concen: 41.456 ng  
RT: 17.204 min Scan# 2417  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:178 Resp: 700135  
Ion Ratio Lower Upper  
178 100  
176 19.3 15.3 22.9  
179 15.3 12.5 18.7




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#72  
Anthracene  
Concen: 41.637 ng  
RT: 17.298 min Scan# 2433  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:178 Resp: 697850  
Ion Ratio Lower Upper  
178 100  
176 18.9 15.1 22.7  
179 15.7 12.3 18.5





#73  
 Carbazole  
 Concen: 38.255 ng  
 RT: 17.575 min Scan# 2480  
 Delta R.T. 0.000 min  
 Lab File: BP008394.D  
 Acq: 14 Dec 2021 15:09




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICC040




Tgt Ion:167 Resp: 625890  
 Ion Ratio Lower Upper  
 167 100  
 166 21.8 17.4 26.0  
 139 13.0 10.2 15.2

Manual Integrations  
 APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
 Supervised By :Jagrut Upadhyay 12/15/2021




#74  
 Di-n-butylphthalate  
 Concen: 35.038 ng  
 RT: 18.128 min Scan# 2574  
 Delta R.T. 0.000 min  
 Lab File: BP008394.D  
 Acq: 14 Dec 2021 15:09




Tgt Ion:149 Resp: 717621  
 Ion Ratio Lower Upper  
 149 100  
 150 9.2 7.4 11.2  
 104 5.7 4.4 6.6





#75  
Fluoranthene  
Concen: 33.415 ng  
RT: 19.186 min Scan# 2754  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:202 Resp: 760220  
Ion Ratio Lower Upper  
202 100  
101 7.9 0.0 29.8  
203 17.2 0.0 37.7




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#76  
Chrysene-d12  
Concen: 20.000 ng  
RT: 21.269 min Scan# 3108  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:240 Resp: 268715  
Ion Ratio Lower Upper  
240 100  
120 7.8 7.3 10.9  
236 26.3 21.4 32.2





#77  
Benzidine  
Concen: 32.788 ng  
RT: 19.375 min Scan# 21  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:184 Resp: 120858  
Ion Ratio Lower Upper  
184 100  
185 15.2 10.3 15.5  
183 11.7 9.1 13.7

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#78  
Pyrene  
Concen: 39.599 ng  
RT: 19.539 min Scan# 2814  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09



Tgt Ion:202 Resp: 769103  
Ion Ratio Lower Upper  
202 100  
200 21.9 17.1 25.7  
203 17.8 14.2 21.2





#79  
 Terphenyl-d14  
 Concen: 100.427 ng  
 RT: 19.733 min Scan# 2847  
 Delta R.T. 0.000 min  
 Lab File: BP008394.D  
 Acq: 14 Dec 2021 15:09

Instrument :

BNA\_P

ClientSampleId :





SSTDICCC040

Tgt Ion:244 Resp: 129125  
 Ion Ratio Lower Upper  
 244 100  
 212 7.4 6.0 9.0  
 122 7.9 6.6 9.8

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
 Supervised By :Jagrut Upadhyay 12/15/2021



#80  
 Butylbenzylphthalate  
 Concen: 37.028 ng  
 RT: 20.416 min Scan# 2963  
 Delta R.T. 0.000 min  
 Lab File: BP008394.D  
 Acq: 14 Dec 2021 15:09


Tgt Ion:149 Resp: 283332  
 Ion Ratio Lower Upper  
 149 100  
 91 73.2 59.8 89.8  
 206 25.0 18.6 27.8





#81  
Benzo(a)anthracene  
Concen: 41.245 ng  
RT: 21.251 min Scan# 3105  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:228 Resp: 73454  
Ion Ratio Lower Upper  
228 100  
226 27.2 21.9 32.9  
229 19.5 15.9 23.9




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#82  
3,3'-Dichlorobenzidine  
Concen: 39.808 ng  
RT: 21.186 min Scan# 3094  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:252 Resp: 334022  
Ion Ratio Lower Upper  
252 100  
254 64.6 52.4 78.6  
126 8.4 7.6 11.4





#83  
Chrysene  
Concen: 40.757 ng  
RT: 21.304 min Scan# 3114  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:228 Resp: 690702  
Ion Ratio Lower Upper  
228 100  
226 29.6 24.0 36.0  
229 20.1 15.8 23.8

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#84  
Bis(2-ethylhexyl)phthalate  
Concen: 34.439 ng  
RT: 21.180 min Scan# 3093  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:149 Resp: 400105  
Ion Ratio Lower Upper  
149 100  
167 28.5 22.5 33.7  
279 5.8 3.5 5.3





#85  
Di-n-octyl phthalate  
Concen: 34.857 ng  
RT: 22.086 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:149 Resp: 68794  
Ion Ratio Lower Upper  
149 100  
167 1.6 1.3 1.9  
43 9.4 7.9 11.9

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#86  
Perylene-d12  
Concen: 20.000 ng  
RT: 23.639 min Scan# 3511  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:264 Resp: 289679  
Ion Ratio Lower Upper  
264 100  
260 25.0 19.4 29.2  
265 23.6 17.6 26.4





#87  
Indeno(1,2,3-cd)pyrene  
Concen: 44.754 ng  
RT: 26.109 min Scan# 3931  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:276 Resp: 943159  
Ion Ratio Lower Upper  
276 100  
138 16.8 16.6 25.0  
277 25.7 20.2 30.2




Manual Integrations  
APPROVED


Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021



#88  
Benzo(b)fluoranthene  
Concen: 40.503 ng  
RT: 22.916 min Scan# 3388  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09


Tgt Ion:252 Resp: 707311  
Ion Ratio Lower Upper  
252 100  
253 22.0 17.7 26.5  
125 6.4 6.2 9.4





#89  
Benzo(k)fluoranthene  
Concen: 41.848 ng  
RT: 22.963 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:252 Resp: 715469  
Ion Ratio Lower Upper  
252 100  
253 22.0 17.3 25.9  
125 7.0 6.5 9.7

Manual Integrations  
APPROVED



Reviewed By :Christian Giraldo 12/15/2021  
Supervised By :Jagrut Upadhyay 12/15/2021




#90  
Benzo(a)pyrene  
Concen: 41.211 ng  
RT: 23.533 min Scan# 3493  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Tgt Ion:252 Resp: 615095  
Ion Ratio Lower Upper  
252 100  
253 21.9 17.6 26.4  
125 7.3 7.4 11.0#





#91  
Dibenzo(a,h)anthracene  
Concen: 45.064 ng  
RT: 26.121 min Scan# 3933  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 278 Resp: 78872  
Ion Ratio Lower Upper  
278 100  
139 10.4 10.8 16.2  
279 23.7 19.1 28.7

Manual Integrations  
APPROVED

Reviewed By : Christian Giraldo 12/15/2021  
Supervised By : Jagrut Upadhyay 12/15/2021



#92  
Benzo(g,h,i)perylene  
Concen: 44.325 ng  
RT: 26.862 min Scan# 4059  
Delta R.T. 0.000 min  
Lab File: BP008394.D  
Acq: 14 Dec 2021 15:09



Tgt Ion: 276 Resp: 782646  
Ion Ratio Lower Upper  
276 100  
277 23.2 20.0 30.0  
138 15.1 14.1 21.1

