

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121823\
 Data File : BP018898.D
 Acq On : 18 Dec 2023 10:06
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732

Quant Time: Dec 18 21:37:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 13 23:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	225923	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.622	136	806604	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.457	164	460140	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.210	188	1043006	20.000	ng/ul	0.00
79) Chrysene-d12	21.310	240	1215548	20.000	ng/ul	0.00
88) Perylene-d12	23.674	264	1432883	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	62189	9.380	ng/uL	0.00
4) Pyridine-d5	3.675	84	390464	21.988	ng/ul	0.00
7) Phenol-d5	6.993	99	438877	19.358	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.163	67	273603	20.361	ng/ul	0.00
11) 2-Chlorophenol-d4	7.357	132	344802	19.490	ng/ul	0.00
15) 4-Methylphenol-d8	8.540	113	321195	17.505	ng/ul	0.00
21) Nitrobenzene-d5	8.987	128	145535	20.272	ng/ul	0.00
24) 2-Nitrophenol-d4	9.704	143	141426	19.001	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.245	165	315278	20.716	ng/ul	0.00
31) 4-Chloroaniline-d4	10.763	131	399543	20.088	ng/ul	0.00
46) Dimethylphthalate-d6	13.875	166	813593	19.811	ng/ul	0.00
49) Acenaphthylene-d8	14.151	160	950669	21.023	ng/ul	0.00
54) 4-Nitrophenol-d4	14.663	143	135694	20.129	ng/ul	0.00
60) Fluorene-d10	15.451	176	739346	21.448	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	108973	17.425	ng/ul	0.00
73) Anthracene-d10	17.310	188	1202663	21.933	ng/ul	0.00
81) Pyrene-d10	19.551	212	1579414	16.692	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.521	264	1755962	21.018	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	63896	8.789	ng/uL	97
5) Pyridine	3.693	79	402961	22.470	ng/ul	98
6) Benzaldehyde	6.969	77	248683	21.245	ng/ul	100
8) Phenol	7.022	94	447284	20.117	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.257	93	368062	20.157	ng/ul	99
12) 2-Chlorophenol	7.387	128	351812	19.649	ng/ul	98
13) 2-Methylphenol	8.269	108	313444	18.414	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.351	45	448357	19.352	ng/ul	99
16) Acetophenone	8.651	105	501409	18.337	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.640	70	227670	15.135	ng/ul	98
18) 4-Methylphenol	8.598	108	331591	17.840	ng/ul	98
19) Hexachloroethane	8.898	117	149656	20.389	ng/ul	95
22) Nitrobenzene	9.028	77	376681	21.581	ng/ul	98
23) Isophorone	9.557	82	607207	17.062	ng/ul	99
25) 2-Nitrophenol	9.740	139	159728	20.043	ng/ul	99
26) 2,4-Dimethylphenol	9.804	107	312860	20.073	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.040	93	433927	19.313	ng/ul	99
29) 2,4-Dichlorophenol	10.269	162	308715	21.059	ng/ul	99
30) Naphthalene	10.669	128	1059974	21.662	ng/ul	100
32) 4-Chloroaniline	10.787	127	383895	20.251	ng/ul	99
33) Hexachlorobutadiene	10.951	225	234743	23.910	ng/ul	100
34) Caprolactam	11.563	113	101359	22.549	ng/ul	99
35) 4-Chloro-3-methylphenol	11.916	107	291370	17.742	ng/ul	99
36) 2-Methylnaphthalene	12.281	142	680820	19.640	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121823\
 Data File : BP018898.D
 Acq On : 18 Dec 2023 10:06
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732

Quant Time: Dec 18 21:37:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 13 23:34:47 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.498	142	681156	19.457	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.651	216	405100	24.265	ng/ul	96
40) Hexachlorocyclopentadiene	12.628	237	186673	18.865	ng/ul	96
41) 2,4,6-Trichlorophenol	12.892	196	224039	20.861	ng/ul	98
42) 2,4,5-Trichlorophenol	12.963	196	245419	21.079	ng/ul	100
43) 1,1'-Biphenyl	13.292	154	887523	22.302	ng/ul	99
44) 2-Chloronaphthalene	13.333	162	711809	22.647	ng/ul	99
45) 2-Nitroaniline	13.545	65	158888	18.647	ng/ul	95
47) Dimethylphthalate	13.922	163	811310	20.069	ng/ul	99
48) 2,6-Dinitrotoluene	14.045	165	141322	18.503	ng/ul	97
50) Acenaphthylene	14.180	152	1084316	21.327	ng/ul	100
51) 3-Nitroaniline	14.375	138	147489	19.345	ng/ul	98
52) Acenaphthene	14.522	153	724603	21.555	ng/ul	99
53) 2,4-Dinitrophenol	14.575	184	59726	14.755	ng/ul	96
55) 4-Nitrophenol	14.680	109	110089	21.616	ng/ul	96
56) Dibenzofuran	14.857	168	1025653	21.952	ng/ul	98
57) 2,4-Dinitrotoluene	14.828	165	213465	19.869	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.086	232	206063	21.099	ng/ul	94
59) Diethylphthalate	15.286	149	783678	19.811	ng/ul	100
61) Fluorene	15.510	166	829153	22.428	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.504	204	434151	22.391	ng/ul	97
63) 4-Nitroaniline	15.533	138	153736	21.166	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.592	198	120878	18.079	ng/ul	93
67) N-Nitrosodiphenylamine	15.722	169	687597	19.895	ng/ul	100
68) 4-Bromophenyl-phenylether	16.404	248	265271	20.146	ng/ul	97
69) Hexachlorobenzene	16.510	284	316915	21.557	ng/ul	98
70) Atrazine	16.680	200	195530	19.029	ng/ul	99
71) Pentachlorophenol	16.857	266	182817	20.909	ng/ul	98
72) Phenanthrene	17.251	178	1376331	21.849	ng/ul	99
74) Anthracene	17.345	178	1406694	22.278	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.257	216	390636	21.660	ng/ul	99
76) Pentachlorobenzene	14.775	250	378359	21.358	ng/ul	99
77) Carbazole	17.621	167	1269149	25.220	ng/ul	99
78) Di-n-butylphthalate	18.174	149	1321847	19.726	ng/ul	99
80) Fluoranthene	19.227	202	1763783	15.770	ng/ul	98
82) Pyrene	19.580	202	1901982	16.857	ng/ul	95
83) Butylbenzylphthalate	20.457	149	621309	15.041	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.227	252	582808	19.730	ng/ul	98
85) Benzo(a)anthracene	21.292	228	2053409	20.812	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.221	149	927081	16.329	ng/ul	100
87) Chrysene	21.345	228	1958600	21.557	ng/ul	100
89) Di-n-octyl phthalate	22.133	149	1639922	15.675	ng/ul	100
90) Benzo(b)fluoranthene	22.951	252	2162564	20.470	ng/ul	99
91) Benzo(k)fluoranthene	22.998	252	2120065	20.740	ng/ul	99
93) Benzo(a)pyrene	23.568	252	2045243	21.187	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.133	276	2517139	21.824	ng/ul	97
95) Dibenzo(a,h)anthracene	26.150	278	2060555	21.476	ng/ul	97
96) Benzo(g,h,i)perylene	26.886	276	2043393	22.014	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121823\
 Data File : BP018898.D
 Acq On : 18 Dec 2023 10:06
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732

Quant Time: Dec 18 21:37:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 13 23:34:47 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.789 ng/uL
RT: 3.293 min Scan# 31
Delta R.T. -0.000 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 88 Resp: 63896
Ion Ratio Lower Upper
88 100
43 23.3 21.8 32.6
58 72.1 56.5 84.7



#5
Pyridine
Concen: 22.470 ng/uL
RT: 3.693 min Scan# 103
Delta R.T. -0.000 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion: 79 Resp: 402961
Ion Ratio Lower Upper
79 100
52 69.0 57.4 86.0
51 33.3 26.2 39.4









#6
Benzaldehyde
Concen: 21.245 ng/ul
RT: 6.969 min Scan# 609
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 77 Resp: 248683
Ion Ratio Lower Upper
77 100
105 97.6 77.9 116.9
106 93.7 74.6 111.8



#8
Phenol
Concen: 20.117 ng/ul
RT: 7.022 min Scan# 669
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion: 94 Resp: 447284
Ion Ratio Lower Upper
94 100
65 26.9 21.0 31.6
66 35.5 28.6 43.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.157 ng/ul
 RT: 7.257 min Scan# 709
 Delta R.T. -0.006 min
 Lab File: BP018898.D
 Acq: 18 Dec 2023 10:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732


Tgt Ion: 93 Resp: 368062
 Ion Ratio Lower Upper
 93 100
 63 74.4 58.3 87.5
 95 32.9 25.9 38.9



#12
 2-Chlorophenol
 Concen: 19.649 ng/ul
 RT: 7.387 min Scan# 731
 Delta R.T. -0.006 min
 Lab File: BP018898.D
 Acq: 18 Dec 2023 10:06

Tgt Ion: 128 Resp: 351812
 Ion Ratio Lower Upper
 128 100
 64 47.5 37.0 55.4
 130 32.2 26.8 40.2





#13

2-Methylphenol

Concen: 18.414 ng/ul

RT: 8.269 min Scan# 881

Delta R.T. -0.012 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :





SSTD020732

Tgt Ion:108 Resp: 313444

Ion Ratio Lower Upper

108 100

107 88.8 70.6 105.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.352 ng/ul

RT: 8.351 min Scan# 895

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion: 45 Resp: 448357


Ion Ratio Lower Upper

45 100

77 16.0 12.7 19.1

79 12.1 10.5 15.7









#16
Acetophenone
Concen: 18.337 ng/ul
RT: 8.651 min Scan# 946
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:105 Resp: 501409
Ion Ratio Lower Upper
105 100
77 76.9 59.4 89.0
51 25.0 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 15.135 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion: 70 Resp: 227670
Ion Ratio Lower Upper
70 100
42 53.5 41.0 61.4
101 9.7 7.8 11.8
130 20.5 16.9 25.3









#18
4-Methylphenol
Concen: 17.840 ng/ul
RT: 8.598 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:108 Resp: 331591
Ion Ratio Lower Upper
108 100
107 115.0 90.6 136.0



#19
Hexachloroethane
Concen: 20.389 ng/ul
RT: 8.898 min Scan# 988
Delta R.T. -0.012 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:117 Resp: 149656
Ion Ratio Lower Upper
117 100
201 110.7 83.7 125.5
199 67.9 52.3 78.5









#22
Nitrobenzene
Concen: 21.581 ng/ul
RT: 9.028 min Scan# 1010
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 77 Resp: 376681
Ion Ratio Lower Upper
77 100
123 43.3 35.8 53.8
65 12.7 10.5 15.7



#23
Isophorone
Concen: 17.062 ng/ul
RT: 9.557 min Scan# 1100
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion: 82 Resp: 607207
Ion Ratio Lower Upper
82 100
95 7.2 5.5 8.3
138 17.3 13.5 20.3









#25
2-Nitrophenol
Concen: 20.043 ng/ul
RT: 9.740 min Scan# 1131
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:139 Resp: 159728
Ion Ratio Lower Upper
139 100
65 49.7 39.1 58.7
109 30.0 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 20.073 ng/ul
RT: 9.804 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:107 Resp: 312860
Ion Ratio Lower Upper
107 100
121 49.9 40.0 60.0
122 85.5 69.1 103.7





#27

Bis(2-Chloroethoxy)methane

Concen: 19.313 ng/ul

RT: 10.040 min Scan# 1182

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion: 93 Resp: 433927

Ion Ratio Lower Upper

93 100

95 31.8 26.1 39.1

123 10.8 8.9 13.3



#29

2,4-Dichlorophenol

Concen: 21.059 ng/ul

RT: 10.269 min Scan# 1221

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:162 Resp: 308715


Ion Ratio Lower Upper

162 100

164 65.7 51.7 77.5

98 35.9 29.2 43.8









#30
Naphthalene
Concen: 21.662 ng/ul
RT: 10.669 min Scan# 1289
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:128 Resp: 1059974
Ion Ratio Lower Upper
128 100
129 11.3 8.9 13.3
127 12.9 10.5 15.7



#32
4-Chloroaniline
Concen: 20.251 ng/ul
RT: 10.787 min Scan# 1309
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:127 Resp: 383895
Ion Ratio Lower Upper
127 100
129 32.7 25.6 38.4









#33
Hexachlorobutadiene
Concen: 23.910 ng/ul
RT: 10.951 min Scan# 1337
Delta R.T. -0.012 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:225 Resp: 234743
Ion Ratio Lower Upper
225 100
223 61.7 49.5 74.3
227 63.5 50.9 76.3



#34
Caprolactam
Concen: 22.549 ng/ul
RT: 11.563 min Scan# 1441
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:113 Resp: 101359
Ion Ratio Lower Upper
113 100
55 164.7 133.7 200.5
56 122.4 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 17.742 ng/ul
RT: 11.916 min Scan# 1501
Delta R.T. -0.000 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:107 Resp: 291370
Ion Ratio Lower Upper
107 100
144 26.2 21.0 31.6
142 80.5 63.6 95.4



#36
2-Methylnaphthalene
Concen: 19.640 ng/ul
RT: 12.281 min Scan# 1563
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:142 Resp: 680820
Ion Ratio Lower Upper
142 100
141 88.2 71.4 107.0





#37

1-Methylnaphthalene

Concen: 19.457 ng/ul

RT: 12.498 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:142 Resp: 681156

Ion Ratio Lower Upper

142 100

141 92.3 73.5 110.3

116 3.8 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 24.265 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Tgt Ion:216 Resp: 405100




Ion Ratio Lower Upper


216 100

214 76.2 63.9 95.9

179 19.2 16.3 24.5

108 14.8 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 18.865 ng/ul

RT: 12.628 min Scan# 1622

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:237 Resp: 186673

Ion Ratio Lower Upper

237 100

235 64.7 49.4 74.0

272 13.3 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 20.861 ng/ul

RT: 12.892 min Scan# 1667

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:196 Resp: 224039


Ion Ratio Lower Upper

196 100

198 97.9 76.9 115.3

200 31.0 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 21.079 ng/ul
RT: 12.963 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:196 Resp: 245419
Ion Ratio Lower Upper
196 100
198 97.3 77.9 116.9
200 30.9 25.2 37.8



#43
1,1'-Biphenyl
Concen: 22.302 ng/ul
RT: 13.292 min Scan# 1735
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:154 Resp: 887523
Ion Ratio Lower Upper
154 100
153 40.0 32.8 49.2
76 13.4 10.7 16.1









#44
2-Chloronaphthalene
Concen: 22.647 ng/ul
RT: 13.333 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:162 Resp: 711809
Ion Ratio Lower Upper
162 100
164 32.7 25.7 38.5
127 37.0 30.3 45.5



#45
2-Nitroaniline
Concen: 18.647 ng/ul
RT: 13.545 min Scan# 1778
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion: 65 Resp: 158888
Ion Ratio Lower Upper
65 100
92 69.3 56.9 85.3
138 102.9 88.2 132.2









#47
Dimethylphthalate
Concen: 20.069 ng/ul
RT: 13.922 min Scan# 1842
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:163 Resp: 811310
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.7 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 18.503 ng/ul
RT: 14.045 min Scan# 1863
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:165 Resp: 141322
Ion Ratio Lower Upper
165 100
89 55.7 46.6 70.0
121 20.6 17.3 25.9









#50
Acenaphthylene
Concen: 21.327 ng/ul
RT: 14.180 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:152 Resp: 1084316
Ion Ratio Lower Upper
152 100
151 19.7 15.9 23.9
153 13.4 10.6 16.0



#51
3-Nitroaniline
Concen: 19.345 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:138 Resp: 147489
Ion Ratio Lower Upper
138 100
108 13.7 9.5 14.3
92 118.5 93.1 139.7





#52

Acenaphthene

Concen: 21.555 ng/ul

RT: 14.522 min Scan# 1944

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:153 Resp: 724603

Ion Ratio Lower Upper

153 100

152 47.4 38.1 57.1

154 88.3 71.7 107.5



#53

2,4-Dinitrophenol

Concen: 14.755 ng/ul

RT: 14.575 min Scan# 1953

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:184 Resp: 59726


Ion Ratio Lower Upper

184 100

63 58.2 48.2 72.4

154 57.1 49.1 73.7









#55
4-Nitrophenol
Concen: 21.616 ng/ul
RT: 14.680 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:109 Resp: 110089
Ion Ratio Lower Upper
109 100
139 131.5 109.0 163.4
65 127.7 106.2 159.4



#56
Dibenzofuran
Concen: 21.952 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:168 Resp: 1025653
Ion Ratio Lower Upper
168 100
139 35.1 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 19.869 ng/ul
RT: 14.828 min Scan# 1996
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:165 Resp: 213465
Ion Ratio Lower Upper
165 100
63 42.1 32.7 49.1
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.099 ng/ul
RT: 15.086 min Scan# 2040
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:232 Resp: 206063
Ion Ratio Lower Upper
232 100
131 37.6 33.8 50.8
130 1.9 1.8 2.8
166 29.0 24.7 37.1









#59
Diethylphthalate
Concen: 19.811 ng/ul
RT: 15.286 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:149 Resp: 783678
Ion Ratio Lower Upper
149 100
177 22.5 18.2 27.4
150 12.4 9.9 14.9



#61
Fluorene
Concen: 22.428 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:166 Resp: 829153
Ion Ratio Lower Upper
166 100
165 98.2 78.0 117.0
167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 22.391 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:204 Resp: 434151
Ion Ratio Lower Upper
204 100
206 33.7 27.1 40.7
141 54.1 45.8 68.8



#63
4-Nitroaniline
Concen: 21.166 ng/ul
RT: 15.533 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:138 Resp: 153736
Ion Ratio Lower Upper
138 100
92 54.5 42.3 63.5
108 76.9 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 18.079 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:198 Resp: 120878
Ion Ratio Lower Upper
198 100
182 4.8 3.4 5.2
77 21.8 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 19.895 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:169 Resp: 687597
Ion Ratio Lower Upper
169 100
168 67.0 53.1 79.7
167 36.4 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 20.146 ng/ul
RT: 16.404 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:248 Resp: 265271
Ion Ratio Lower Upper
248 100
250 98.6 79.3 118.9
141 66.4 57.2 85.8



#69
Hexachlorobenzene
Concen: 21.557 ng/ul
RT: 16.510 min Scan# 2282
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:284 Resp: 316915
Ion Ratio Lower Upper
284 100
142 31.2 25.8 38.6
249 32.0 24.9 37.3





#70

Atrazine

Concen: 19.029 ng/ul

RT: 16.680 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:200 Resp: 195530

Ion Ratio Lower Upper

200 100

173 25.5 21.0 31.4

215 47.8 38.2 57.2



#71

Pentachlorophenol

Concen: 20.909 ng/ul

RT: 16.857 min Scan# 2341

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:266 Resp: 182817


Ion Ratio Lower Upper

266 100

264 61.6 50.2 75.2

268 61.8 50.5 75.7





#72

Phenanthrene

Concen: 21.849 ng/ul

RT: 17.251 min Scan# 2408

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:178 Resp: 1376331

Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 18.8 15.4 23.2



#74

Anthracene

Concen: 22.278 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. -0.000 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:178 Resp: 1406694


Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.3 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.660 ng/uL
RT: 13.257 min Scan# 1729
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:216 Resp: 390636
Ion Ratio Lower Upper
216 100
214 77.5 61.4 92.2
218 49.1 39.2 58.8



#76
Pentachlorobenzene
Concen: 21.358 ng/uL
RT: 14.775 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP018898.D
Acq: 18 Dec 2023 10:06

Tgt Ion:250 Resp: 378359
Ion Ratio Lower Upper
250 100
248 63.3 50.6 75.8
252 62.1 51.3 76.9





#77

Carbazole

Concen: 25.220 ng/ul

RT: 17.621 min Scan# 2471

Delta R.T. -0.000 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:167 Resp: 1269149

Ion Ratio Lower Upper

167 100

166 21.5 17.1 25.7

139 12.2 10.2 15.4



#78

Di-n-butylphthalate

Concen: 19.726 ng/ul

RT: 18.174 min Scan# 2565

Delta R.T. -0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:149 Resp: 1321847


Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.2

104 4.8 3.9 5.9





#80

Fluoranthene

Concen: 15.770 ng/ul

RT: 19.227 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:202 Resp: 1763783

Ion Ratio Lower Upper

202 100

101 10.5 9.1 13.7

100 8.2 7.1 10.7



#82

Pyrene

Concen: 16.857 ng/ul

RT: 19.580 min Scan# 2804

Delta R.T. -0.000 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:202 Resp: 1901982


Ion Ratio Lower Upper

202 100

101 11.8 11.0 16.4

100 9.5 9.1 13.7





#83

Butylbenzylphthalate

Concen: 15.041 ng/ul

RT: 20.457 min Scan# 2953

Delta R.T. -0.000 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:149 Resp: 621309

Ion Ratio Lower Upper

149 100

91 72.7 56.6 85.0

206 21.7 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 19.730 ng/ul

RT: 21.227 min Scan# 3084

Delta R.T. -0.000 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:252 Resp: 582808


Ion Ratio Lower Upper

252 100

254 65.1 51.0 76.4

126 10.1 7.8 11.6





#85

Benzo(a)anthracene

Concen: 20.812 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. 0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:228 Resp: 2053409

Ion Ratio Lower Upper

228 100

229 19.8 15.7 23.5

226 27.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.329 ng/ul

RT: 21.221 min Scan# 3083

Delta R.T. 0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:149 Resp: 927081


Ion Ratio Lower Upper

149 100

167 27.9 22.2 33.2

279 4.5 3.3 4.9





#87

Chrysene

Concen: 21.557 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. 0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:228 Resp: 1958600

Ion Ratio Lower Upper

228 100

226 30.0 23.9 35.9

229 19.3 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 15.675 ng/ul




RT: 22.133 min Scan# 3238


Delta R.T. 0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Tgt Ion:149 Resp: 1639922





#90

Benzo(b)fluoranthene

Concen: 20.470 ng/ul

RT: 22.951 min Scan# 3377

Delta R.T. 0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:252 Resp: 2162564

Ion Ratio Lower Upper

252 100

253 22.1 17.3 25.9

125 8.9 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 20.740 ng/ul

RT: 22.998 min Scan# 3385

Delta R.T. 0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:252 Resp: 2120065


Ion Ratio Lower Upper

252 100

253 21.6 17.7 26.5

125 9.2 8.0 12.0





#93

Benzo(a)pyrene

Concen: 21.187 ng/ul

RT: 23.568 min Scan# 3482

Delta R.T. 0.006 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:252 Resp: 2045243

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8

125 10.0 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.824 ng/ul

RT: 26.133 min Scan# 3918

Delta R.T. 0.012 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06




Tgt Ion:276 Resp: 2517139


Ion Ratio Lower Upper

276 100

138 20.9 19.4 29.2

277 24.6 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 21.476 ng/ul

RT: 26.150 min Scan# 3921

Delta R.T. 0.012 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:278 Resp: 2060555

Ion Ratio Lower Upper

278 100

139 14.2 13.0 19.6

279 24.4 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 22.014 ng/ul

RT: 26.886 min Scan# 4046

Delta R.T. 0.012 min

Lab File: BP018898.D

Acq: 18 Dec 2023 10:06

Tgt Ion:276 Resp: 2043393

Ion Ratio Lower Upper

276 100

138 19.5 18.1 27.1

277 23.9 19.4 29.0

