

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121823\
 Data File : BP018910.D
 Acq On : 18 Dec 2023 17:01
 Operator : MA/JU
 Sample : SSTDCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020733

Quant Time: Dec 18 21:41:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 13 23:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	262823	20.000	ng/u1	0.00
20) Naphthalene-d8	10.622	136	960458	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.463	164	539460	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	1123545	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	1129244	20.000	ng/u1	0.00
88) Perylene-d12	23.674	264	1369734	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	62807	8.143	ng/uL	0.00
4) Pyridine-d5	3.675	84	407385	19.720	ng/u1	0.00
7) Phenol-d5	7.005	99	506237	19.194	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	293538	18.778	ng/u1	0.00
11) 2-Chlorophenol-d4	7.363	132	414916	20.161	ng/u1	0.00
15) 4-Methylphenol-d8	8.546	113	388704	18.210	ng/u1	0.00
21) Nitrobenzene-d5	8.993	128	187217	21.901	ng/u1	0.00
24) 2-Nitrophenol-d4	9.710	143	203746	22.989	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.251	165	400998	22.127	ng/u1	0.00
31) 4-Chloroaniline-d4	10.769	131	473161	19.978	ng/u1	0.00
46) Dimethylphthalate-d6	13.881	166	1023766	21.263	ng/u1	0.00
49) Acenaphthylene-d8	14.157	160	1169173	22.053	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	156637	19.819	ng/u1	0.00
60) Fluorene-d10	15.457	176	858646	21.246	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	109304	16.225	ng/u1	0.00
73) Anthracene-d10	17.310	188	1293968	21.907	ng/u1	0.00
81) Pyrene-d10	19.551	212	1550689	17.641	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.521	264	1689298	21.152	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	67205	7.946	ng/uL	97
5) Pyridine	3.699	79	411290	19.715	ng/u1	94
6) Benzaldehyde	6.975	77	282838	20.770	ng/u1	94
8) Phenol	7.028	94	499884	19.326	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.263	93	413237	19.454	ng/u1	97
12) 2-Chlorophenol	7.393	128	413998	19.876	ng/u1	96
13) 2-Methylphenol	8.281	108	371727	18.772	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.357	45	463864	17.210	ng/u1	99
16) Acetophenone	8.657	105	588832	18.510	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.646	70	278207	15.898	ng/u1	97
18) 4-Methylphenol	8.610	108	391158	18.090	ng/u1	99
19) Hexachloroethane	8.904	117	169932	19.900	ng/u1	84
22) Nitrobenzene	9.034	77	429759	20.678	ng/u1	96
23) Isophorone	9.563	82	790755	18.660	ng/u1	98
25) 2-Nitrophenol	9.746	139	215486	22.708	ng/u1	96
26) 2,4-Dimethylphenol	9.810	107	379144	20.429	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.046	93	513587	19.197	ng/u1	99
29) 2,4-Dichlorophenol	10.275	162	378584	21.688	ng/u1	98
30) Naphthalene	10.675	128	1242243	21.321	ng/u1	99
32) 4-Chloroaniline	10.793	127	451935	20.021	ng/u1	98
33) Hexachlorobutadiene	10.957	225	304496	26.046	ng/u1	99
34) Caprolactam	11.575	113	127432	23.808	ng/u1	89
35) 4-Chloro-3-methylphenol	11.922	107	350256	17.912	ng/u1	99
36) 2-Methylnaphthalene	12.287	142	814668	19.736	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121823\
 Data File : BP018910.D
 Acq On : 18 Dec 2023 17:01
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020733


Quant Time: Dec 18 21:41:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 13 23:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.504	142	812874	19.500	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.651	216	510083	26.061	ng/ul	98
40) Hexachlorocyclopentadiene	12.628	237	162678	14.023	ng/ul	99
41) 2,4,6-Trichlorophenol	12.898	196	298981	23.746	ng/ul	98
42) 2,4,5-Trichlorophenol	12.969	196	312040	22.860	ng/ul	99
43) 1,1'-Biphenyl	13.298	154	1060084	22.721	ng/ul	98
44) 2-Chloronaphthalene	13.340	162	851689	23.113	ng/ul	99
45) 2-Nitroaniline	13.551	65	191362	19.156	ng/ul	94
47) Dimethylphthalate	13.928	163	996096	21.017	ng/ul	100
48) 2,6-Dinitrotoluene	14.045	165	191028	21.334	ng/ul	96
50) Acenaphthylene	14.181	152	1310874	21.992	ng/ul	100
51) 3-Nitroaniline	14.375	138	193661	21.666	ng/ul	98
52) Acenaphthene	14.528	153	866391	21.983	ng/ul	99
53) 2,4-Dinitrophenol	14.587	184	59998	12.643	ng/ul	91
55) 4-Nitrophenol	14.687	109	120619	20.201	ng/ul	95
56) Dibenzofuran	14.863	168	1193066	21.781	ng/ul	97
57) 2,4-Dinitrotoluene	14.834	165	262922	20.874	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.092	232	259456	22.660	ng/ul#	89
59) Diethylphthalate	15.286	149	954381	20.579	ng/ul	99
61) Fluorene	15.510	166	955633	22.048	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.510	204	515748	22.688	ng/ul	94
63) 4-Nitroaniline	15.534	138	199094	23.381	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.592	198	118281	16.422	ng/ul	95
67) N-Nitrosodiphenylamine	15.722	169	780855	20.973	ng/ul	99
68) 4-Bromophenyl-phenylether	16.404	248	327338	23.077	ng/ul	95
69) Hexachlorobenzene	16.510	284	373404	23.578	ng/ul	96
70) Atrazine	16.681	200	240160	21.697	ng/ul	98
71) Pentachlorophenol	16.863	266	218900	23.241	ng/ul	99
72) Phenanthrene	17.257	178	1606315	23.672	ng/ul	99
74) Anthracene	17.345	178	1542411	22.677	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.257	216	491952	25.323	ng/ul	98
76) Pentachlorobenzene	14.781	250	457590	23.979	ng/ul	99
77) Carbazole	17.622	167	1320094	24.352	ng/ul	99
78) Di-n-butylphthalate	18.175	149	1601582	22.188	ng/ul	100
80) Fluoranthene	19.227	202	1979107	19.047	ng/ul	95
82) Pyrene	19.580	202	2003816	19.117	ng/ul#	93
83) Butylbenzylphthalate	20.457	149	732747	19.095	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.227	252	603757	22.001	ng/ul	99
85) Benzo(a)anthracene	21.292	228	1983336	21.638	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.216	149	1053797	19.980	ng/ul	100
87) Chrysene	21.345	228	1856445	21.994	ng/ul	99
89) Di-n-octyl phthalate	22.133	149	1887734	18.876	ng/ul	100
90) Benzo(b)fluoranthene	22.951	252	2046765	20.267	ng/ul	98
91) Benzo(k)fluoranthene	22.998	252	1971375	20.175	ng/ul	98
93) Benzo(a)pyrene	23.568	252	1905584	20.650	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.133	276	2413917	21.894	ng/ul	95
95) Dibenzo(a,h)anthracene	26.151	278	2015135	21.971	ng/ul	96
96) Benzo(g,h,i)perylene	26.892	276	1928834	21.738	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 18 21:41:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 13 23:34:47 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.946 ng/uL
RT: 3.293 min Scan# 31
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion: 88	Resp: 67205
Ion Ratio	Lower Upper
88	100
43	24.0 21.8 32.6
58	72.3 56.5 84.7



#5
Pyridine
Concen: 19.715 ng/uL
RT: 3.699 min Scan# 104
Delta R.T. 0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion: 79	Resp: 411290
Ion Ratio	Lower Upper
79	100
52	65.6 57.4 86.0
51	31.2 26.2 39.4









#6
Benzaldehyde
Concen: 20.770 ng/ul
RT: 6.975 min Scan# 60
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion: 77 Resp: 282838
Ion Ratio Lower Upper
77 100
105 103.2 77.9 116.9
106 99.1 74.6 111.8



#8
Phenol
Concen: 19.326 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion: 94 Resp: 499884
Ion Ratio Lower Upper
94 100
65 26.3 21.0 31.6
66 36.1 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 19.454 ng/ul
RT: 7.263 min Scan# 710
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion: 93 Resp: 413237
Ion Ratio Lower Upper
93 100
63 69.7 58.3 87.5
95 32.5 25.9 38.9



#12
2-Chlorophenol
Concen: 19.876 ng/ul
RT: 7.393 min Scan# 732
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:128 Resp: 413998
Ion Ratio Lower Upper
128 100
64 42.9 37.0 55.4
130 31.9 26.8 40.2





#13

2-Methylphenol

Concen: 18.772 ng/ul

RT: 8.281 min Scan# 881

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :





SSTD020733

Tgt Ion:108 Resp: 371727

Ion Ratio Lower Upper

108 100

107 89.0 70.6 105.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 17.210 ng/ul

RT: 8.357 min Scan# 896

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion: 45 Resp: 463864


Ion Ratio Lower Upper

45 100

77 17.0 12.7 19.1

79 13.1 10.5 15.7









#16
Acetophenone
Concen: 18.510 ng/ul
RT: 8.657 min Scan# 946
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:105 Resp: 588832
Ion Ratio Lower Upper
105 100
77 74.0 59.4 89.0
51 23.2 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 15.898 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion: 70 Resp: 278207
Ion Ratio Lower Upper
70 100
42 48.5 41.0 61.4
101 10.9 7.8 11.8
130 22.0 16.9 25.3









#18
4-Methylphenol
Concen: 18.090 ng/ul
RT: 8.610 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:108 Resp: 391158
Ion Ratio Lower Upper
108 100
107 112.3 90.6 136.0



#19
Hexachloroethane
Concen: 19.900 ng/ul
RT: 8.904 min Scan# 989
Delta R.T. -0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:117 Resp: 169932
Ion Ratio Lower Upper
117 100
201 122.0 83.7 125.5
199 76.1 52.3 78.5









#22
Nitrobenzene
Concen: 20.678 ng/ul
RT: 9.034 min Scan# 1101
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion: 77 Resp: 429759
Ion Ratio Lower Upper
77 100
123 47.9 35.8 53.8
65 12.9 10.5 15.7



#23
Isophorone
Concen: 18.660 ng/ul
RT: 9.563 min Scan# 1101
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion: 82 Resp: 790755
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 17.9 13.5 20.3









#25
2-Nitrophenol
Concen: 22.708 ng/ul
RT: 9.746 min Scan# 1131
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:139 Resp: 215486
Ion Ratio Lower Upper
139 100
65 45.5 39.1 58.7
109 30.4 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 20.429 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:107 Resp: 379144
Ion Ratio Lower Upper
107 100
121 50.7 40.0 60.0
122 85.3 69.1 103.7





#27

Bis(2-Chloroethoxy)methane

Concen: 19.197 ng/ul

RT: 10.046 min Scan# 1182

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion: 93 Resp: 513587

Ion Ratio Lower Upper

93 100

95 32.1 26.1 39.1

123 10.8 8.9 13.3



#29

2,4-Dichlorophenol

Concen: 21.688 ng/ul

RT: 10.275 min Scan# 1222

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:162 Resp: 378584


Ion Ratio Lower Upper

162 100

164 65.6 51.7 77.5

98 33.9 29.2 43.8









#30
Naphthalene
Concen: 21.321 ng/ul
RT: 10.675 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:128 Resp: 1242243
Ion Ratio Lower Upper
128 100
129 10.7 8.9 13.3
127 12.8 10.5 15.7



#32
4-Chloroaniline
Concen: 20.021 ng/ul
RT: 10.793 min Scan# 1310
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:127 Resp: 451935
Ion Ratio Lower Upper
127 100
129 33.0 25.6 38.4





#33

Hexachlorobutadiene

Concen: 26.046 ng/ul

RT: 10.957 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:225 Resp: 304496

Ion Ratio Lower Upper

225 100

223 61.3 49.5 74.3

227 64.0 50.9 76.3



#34

Caprolactam

Concen: 23.808 ng/ul

RT: 11.575 min Scan# 1443

Delta R.T. 0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:113 Resp: 127432


Ion Ratio Lower Upper

113 100

55 148.8 133.7 200.5

56 114.0 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 17.912 ng/ul
RT: 11.922 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:107 Resp: 350256
Ion Ratio Lower Upper
107 100
144 26.6 21.0 31.6
142 80.4 63.6 95.4



#36
2-Methylnaphthalene
Concen: 19.736 ng/ul
RT: 12.287 min Scan# 1564
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:142 Resp: 814668
Ion Ratio Lower Upper
142 100
141 90.0 71.4 107.0





#37

1-Methylnaphthalene

Concen: 19.500 ng/ul

RT: 12.504 min Scan# 1601

Delta R.T. -0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:142 Resp: 812874

Ion Ratio Lower Upper

142 100

141 92.1 73.5 110.3

116 3.8 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 26.061 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. -0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Tgt Ion:216 Resp: 510083




Ion Ratio Lower Upper


216 100

214 77.9 63.9 95.9

179 20.0 16.3 24.5

108 14.4 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 14.023 ng/ul

RT: 12.628 min Scan# 1622

Delta R.T. -0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:237 Resp: 162678

Ion Ratio Lower Upper

237 100

235 62.7 49.4 74.0

272 12.7 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 23.746 ng/ul

RT: 12.898 min Scan# 1668

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:196 Resp: 298981


Ion Ratio Lower Upper

196 100

198 98.0 76.9 115.3

200 31.9 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 22.860 ng/ul

RT: 12.969 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:196 Resp: 312040

Ion Ratio Lower Upper

196 100

198 96.9 77.9 116.9

200 30.5 25.2 37.8



#43

1,1'-Biphenyl

Concen: 22.721 ng/ul

RT: 13.298 min Scan# 1736

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:154 Resp: 1060084


Ion Ratio Lower Upper

154 100

153 39.8 32.8 49.2

76 12.0 10.7 16.1





#44

2-Chloronaphthalene

Concen: 23.113 ng/ul

RT: 13.340 min Scan# 1742

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:162 Resp: 851689

Ion Ratio Lower Upper

162 100

164 32.9 25.7 38.5

127 37.2 30.3 45.5



#45

2-Nitroaniline

Concen: 19.156 ng/ul

RT: 13.551 min Scan# 1779

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion: 65 Resp: 191362


Ion Ratio Lower Upper

65 100

92 73.5 56.9 85.3

138 118.8 88.2 132.2





#47

Dimethylphthalate

Concen: 21.017 ng/ul

RT: 13.928 min Scan# 1843

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:163 Resp: 996096

Ion Ratio Lower Upper

163 100

194 4.2 3.2 4.8

164 10.1 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 21.334 ng/ul

RT: 14.045 min Scan# 1863

Delta R.T. -0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:165 Resp: 191028


Ion Ratio Lower Upper

165 100

89 54.8 46.6 70.0

121 20.6 17.3 25.9









#50
Acenaphthylene
Concen: 21.992 ng/ul
RT: 14.181 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:152 Resp: 1310874
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 13.1 10.6 16.0



#51
3-Nitroaniline
Concen: 21.666 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:138 Resp: 193661
Ion Ratio Lower Upper
138 100
108 10.9 9.5 14.3
92 113.7 93.1 139.7





#52

Acenaphthene

Concen: 21.983 ng/ul

RT: 14.528 min Scan# 1944

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:153 Resp: 866391

Ion Ratio Lower Upper

153 100

152 48.3 38.1 57.1

154 88.4 71.7 107.5



#53

2,4-Dinitrophenol

Concen: 12.643 ng/ul

RT: 14.587 min Scan# 1955

Delta R.T. 0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:184 Resp: 59998


Ion Ratio Lower Upper

184 100

63 48.9 48.2 72.4

154 59.7 49.1 73.7









#55
4-Nitrophenol
Concen: 20.201 ng/ul
RT: 14.687 min Scan# 1971
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:109 Resp: 120619
Ion Ratio Lower Upper
109 100
139 132.1 109.0 163.4
65 125.0 106.2 159.4



#56
Dibenzofuran
Concen: 21.781 ng/ul
RT: 14.863 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:168 Resp: 1193066
Ion Ratio Lower Upper
168 100
139 34.6 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 20.874 ng/ul
RT: 14.834 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:165 Resp: 262922
Ion Ratio Lower Upper
165 100
63 36.9 32.7 49.1
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.660 ng/ul
RT: 15.092 min Scan# 2041
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:232 Resp: 259456
Ion Ratio Lower Upper
232 100
131 33.5 33.8 50.8#
130 2.0 1.8 2.8
166 26.4 24.7 37.1









#59
Diethylphthalate
Concen: 20.579 ng/ul
RT: 15.286 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:149 Resp: 954381
Ion Ratio Lower Upper
149 100
177 23.3 18.2 27.4
150 12.4 9.9 14.9



#61
Fluorene
Concen: 22.048 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:166 Resp: 955633
Ion Ratio Lower Upper
166 100
165 98.3 78.0 117.0
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 22.688 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:204 Resp: 515748
Ion Ratio Lower Upper
204 100
206 33.5 27.1 40.7
141 50.3 45.8 68.8



#63
4-Nitroaniline
Concen: 23.381 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:138 Resp: 199094
Ion Ratio Lower Upper
138 100
92 52.5 42.3 63.5
108 77.3 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 16.422 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:198 Resp: 118281
Ion Ratio Lower Upper
198 100
182 4.5 3.4 5.2
77 22.9 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 20.973 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:169 Resp: 780855
Ion Ratio Lower Upper
169 100
168 67.0 53.1 79.7
167 36.0 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 23.077 ng/ul
RT: 16.404 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:248 Resp: 327338
Ion Ratio Lower Upper
248 100
250 99.2 79.3 118.9
141 61.3 57.2 85.8



#69
Hexachlorobenzene
Concen: 23.578 ng/ul
RT: 16.510 min Scan# 2282
Delta R.T. -0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:284 Resp: 373404
Ion Ratio Lower Upper
284 100
142 29.0 25.8 38.6
249 32.0 24.9 37.3









#70
Atrazine
Concen: 21.697 ng/ul
RT: 16.681 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:200 Resp: 240160
Ion Ratio Lower Upper
200 100
173 25.0 21.0 31.4
215 48.7 38.2 57.2



#71
Pentachlorophenol
Concen: 23.241 ng/ul
RT: 16.863 min Scan# 2342
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:266 Resp: 218900
Ion Ratio Lower Upper
266 100
264 63.2 50.2 75.2
268 63.6 50.5 75.7









#72
Phenanthrene
Concen: 23.672 ng/ul
RT: 17.257 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:178 Resp: 1606315
Ion Ratio Lower Upper
178 100
179 14.9 12.3 18.5
176 18.9 15.4 23.2



#74
Anthracene
Concen: 22.677 ng/ul
RT: 17.345 min Scan# 2424
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:178 Resp: 1542411
Ion Ratio Lower Upper
178 100
179 15.2 12.5 18.7
176 18.2 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 25.323 ng/uL
RT: 13.257 min Scan# 1729
Delta R.T. -0.006 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:216 Resp: 491952
Ion Ratio Lower Upper
216 100
214 78.0 61.4 92.2
218 47.0 39.2 58.8



#76
Pentachlorobenzene
Concen: 23.979 ng/uL
RT: 14.781 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP018910.D
Acq: 18 Dec 2023 17:01

Tgt Ion:250 Resp: 457590
Ion Ratio Lower Upper
250 100
248 62.2 50.6 75.8
252 63.2 51.3 76.9









#77
 Carbazole
 Concen: 24.352 ng/ul
 RT: 17.622 min Scan# 2471
 Delta R.T. -0.000 min
 Lab File: BP018910.D
 Acq: 18 Dec 2023 17:01




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020733


Tgt Ion:167 Resp: 1320094
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.1 25.7
 139 12.0 10.2 15.4



#78
 Di-n-butylphthalate
 Concen: 22.188 ng/ul
 RT: 18.175 min Scan# 2565
 Delta R.T. -0.006 min
 Lab File: BP018910.D
 Acq: 18 Dec 2023 17:01

Tgt Ion:149 Resp: 1601582
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 4.7 3.9 5.9





#80

Fluoranthene

Concen: 19.047 ng/ul

RT: 19.227 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:202 Resp: 1979107

Ion Ratio Lower Upper

202 100

101 9.6 9.1 13.7

100 7.1 7.1 10.7



#82

Pyrene

Concen: 19.117 ng/ul

RT: 19.580 min Scan# 2804

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:202 Resp: 2003816


Ion Ratio Lower Upper

202 100

101 11.0 11.0 16.4

100 8.8 9.1 13.7#





#83

Butylbenzylphthalate

Concen: 19.095 ng/ul

RT: 20.457 min Scan# 2953

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:149 Resp: 732747

Ion Ratio Lower Upper

149 100

91 68.0 56.6 85.0

206 23.5 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 22.001 ng/ul

RT: 21.227 min Scan# 3084

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:252 Resp: 603757


Ion Ratio Lower Upper

252 100

254 64.4 51.0 76.4

126 8.9 7.8 11.6





#85

Benzo(a)anthracene

Concen: 21.638 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. 0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:228 Resp: 1983336

Ion Ratio Lower Upper

228 100

229 19.7 15.7 23.5

226 27.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.980 ng/ul

RT: 21.216 min Scan# 3082

Delta R.T. -0.000 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:149 Resp: 1053797


Ion Ratio Lower Upper

149 100

167 27.7 22.2 33.2

279 4.7 3.3 4.9





#87

Chrysene

Concen: 21.994 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. 0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:228 Resp: 1856445

Ion Ratio Lower Upper

228 100

226 30.0 23.9 35.9

229 20.1 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 18.876 ng/ul




RT: 22.133 min Scan# 3238


Delta R.T. 0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Tgt Ion:149 Resp: 1887734





#90

Benzo(b)fluoranthene

Concen: 20.267 ng/ul

RT: 22.951 min Scan# 3377

Delta R.T. 0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:252 Resp: 2046765

Ion Ratio Lower Upper

252 100

253 22.5 17.3 25.9

125 8.7 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 20.175 ng/ul

RT: 22.998 min Scan# 3385

Delta R.T. 0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:252 Resp: 1971375


Ion Ratio Lower Upper

252 100

253 21.7 17.7 26.5

125 8.3 8.0 12.0





#93

Benzo(a)pyrene

Concen: 20.650 ng/ul

RT: 23.568 min Scan# 3482

Delta R.T. 0.006 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:252 Resp: 1905584

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.8

125 9.6 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.894 ng/ul

RT: 26.133 min Scan# 3918

Delta R.T. 0.012 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01




Tgt Ion:276 Resp: 2413917


Ion Ratio Lower Upper

276 100

138 19.5 19.4 29.2

277 24.6 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 21.971 ng/ul

RT: 26.151 min Scan# 3921

Delta R.T. 0.012 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:278 Resp: 2015135

Ion Ratio Lower Upper

278 100

139 13.2 13.0 19.6

279 24.0 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 21.738 ng/ul

RT: 26.892 min Scan# 4047

Delta R.T. 0.018 min

Lab File: BP018910.D

Acq: 18 Dec 2023 17:01

Tgt Ion:276 Resp: 1928834

Ion Ratio Lower Upper

276 100

138 18.7 18.1 27.1

277 23.3 19.4 29.0

