

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121823\
 Data File : BP018949.D
 Acq On : 19 Dec 2023 15:41
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020735

Quant Time: Dec 19 21:42:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 19 05:28:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	384532	20.000	ng/u1	0.00
20) Naphthalene-d8	10.622	136	1342554	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.463	164	696994	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	1419701	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	1626366	20.000	ng/u1	-0.01
88) Perylene-d12	23.668	264	1960077	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	79074	7.008	ng/uL	0.00
4) Pyridine-d5	3.669	84	551872	18.259	ng/u1	-0.01
7) Phenol-d5	7.005	99	690954	17.905	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.163	67	388953	17.006	ng/u1	0.00
11) 2-Chlorophenol-d4	7.363	132	549116	18.236	ng/u1	0.00
15) 4-Methylphenol-d8	8.552	113	513252	16.435	ng/u1	0.00
21) Nitrobenzene-d5	8.993	128	249523	20.882	ng/u1	0.00
24) 2-Nitrophenol-d4	9.710	143	274604	22.166	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.251	165	503873	19.891	ng/u1	0.00
31) 4-Chloroaniline-d4	10.769	131	602643	18.204	ng/u1	0.00
46) Dimethylphthalate-d6	13.875	166	1152886	18.533	ng/u1	0.00
49) Acenaphthylene-d8	14.157	160	1370889	20.013	ng/u1	0.00
54) 4-Nitrophenol-d4	14.681	143	184918	18.109	ng/u1	0.00
60) Fluorene-d10	15.457	176	964034	18.462	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.586	200	118776	13.953	ng/u1	0.00
73) Anthracene-d10	17.310	188	1482958	19.869	ng/u1	0.00
81) Pyrene-d10	19.551	212	1957707	15.464	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.515	264	2183617	19.107	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.287	88	84223	6.806	ng/uL	98
5) Pyridine	3.693	79	563922	18.475	ng/u1	93
6) Benzaldehyde	6.969	77	358118	17.975	ng/u1	97
8) Phenol	7.034	94	680905	17.993	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.258	93	546155	17.573	ng/u1	98
12) 2-Chlorophenol	7.393	128	552419	18.127	ng/u1	97
13) 2-Methylphenol	8.281	108	493547	17.035	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.357	45	594918	15.086	ng/u1	97
16) Acetophenone	8.657	105	770258	16.550	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.646	70	357381	13.958	ng/u1	96
18) 4-Methylphenol	8.610	108	518071	16.376	ng/u1	99
19) Hexachloroethane	8.899	117	213903	17.121	ng/u1	91
22) Nitrobenzene	9.034	77	570241	19.628	ng/u1	96
23) Isophorone	9.563	82	1005043	16.967	ng/u1	98
25) 2-Nitrophenol	9.746	139	283567	21.378	ng/u1	97
26) 2,4-Dimethylphenol	9.810	107	458929	17.690	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.046	93	638491	17.073	ng/u1	100
29) 2,4-Dichlorophenol	10.275	162	481234	19.723	ng/u1	98
30) Naphthalene	10.675	128	1562668	19.187	ng/u1	100
32) 4-Chloroaniline	10.793	127	577622	18.306	ng/u1	99
33) Hexachlorobutadiene	10.951	225	371807	22.753	ng/u1	99
34) Caprolactam	11.581	113	113989	15.236	ng/u1	89
35) 4-Chloro-3-methylphenol	11.922	107	431517	15.787	ng/u1	98
36) 2-Methylnaphthalene	12.287	142	985533	17.081	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121823\
 Data File : BP018949.D
 Acq On : 19 Dec 2023 15:41
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020735

Quant Time: Dec 19 21:42:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 19 05:28:03 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.504	142	975509	16.742	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.651	216	597450	23.626	ng/ul	99
40) Hexachlorocyclopentadiene	12.628	237	142658	9.518	ng/ul	97
41) 2,4,6-Trichlorophenol	12.898	196	358193	22.019	ng/ul	99
42) 2,4,5-Trichlorophenol	12.969	196	366265	20.768	ng/ul	99
43) 1,1'-Biphenyl	13.298	154	1230724	20.417	ng/ul	98
44) 2-Chloronaphthalene	13.340	162	995722	20.914	ng/ul	98
45) 2-Nitroaniline	13.551	65	236027	18.287	ng/ul	94
47) Dimethylphthalate	13.928	163	1126693	18.400	ng/ul	100
48) 2,6-Dinitrotoluene	14.051	165	224932	19.443	ng/ul	91
50) Acenaphthylene	14.187	152	1515148	19.674	ng/ul	99
51) 3-Nitroaniline	14.375	138	231673	20.061	ng/ul	96
52) Acenaphthene	14.528	153	987297	19.389	ng/ul	99
53) 2,4-Dinitrophenol	14.587	184	65828	10.736	ng/ul	97
55) 4-Nitrophenol	14.692	109	142655	18.492	ng/ul	96
56) Dibenzofuran	14.863	168	1349964	19.075	ng/ul	97
57) 2,4-Dinitrotoluene	14.834	165	303977	18.679	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.092	232	295762	19.993	ng/ul	92
59) Diethylphthalate	15.287	149	1024171	17.092	ng/ul	99
61) Fluorene	15.510	166	1054979	18.839	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.504	204	563754	19.195	ng/ul	95
63) 4-Nitroaniline	15.539	138	238964	21.720	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.598	198	127827	14.045	ng/ul	93
67) N-Nitrosodiphenylamine	15.722	169	866393	18.416	ng/ul	99
68) 4-Bromophenyl-phenylether	16.404	248	348866	19.465	ng/ul	95
69) Hexachlorobenzene	16.516	284	412524	20.615	ng/ul	95
70) Atrazine	16.681	200	248003	17.732	ng/ul	97
71) Pentachlorophenol	16.863	266	261024	21.932	ng/ul	99
72) Phenanthrene	17.257	178	1686742	19.672	ng/ul	100
74) Anthracene	17.345	178	1707652	19.869	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.263	216	577018	23.506	ng/ul	99
76) Pentachlorobenzene	14.781	250	517395	21.457	ng/ul	99
77) Carbazole	17.622	167	1541983	22.511	ng/ul	99
78) Di-n-butylphthalate	18.175	149	1636263	17.939	ng/ul	100
80) Fluoranthene	19.222	202	2209424	14.764	ng/ul	97
82) Pyrene	19.580	202	2323899	15.394	ng/ul#	94
83) Butylbenzylphthalate	20.457	149	838651	15.174	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.227	252	784143	19.841	ng/ul	98
85) Benzo(a)anthracene	21.292	228	2522483	19.108	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.216	149	1178776	15.518	ng/ul#	99
87) Chrysene	21.345	228	2338484	19.236	ng/ul	100
89) Di-n-octyl phthalate	22.127	149	2175093	15.199	ng/ul	100
90) Benzo(b)fluoranthene	22.951	252	2680938	18.551	ng/ul	99
91) Benzo(k)fluoranthene	22.998	252	2497348	17.860	ng/ul	98
93) Benzo(a)pyrene	23.568	252	2469384	18.700	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.133	276	3108351	19.701	ng/ul	95
95) Dibenzo(a,h)anthracene	26.151	278	2578462	19.646	ng/ul	97
96) Benzo(g,h,i)perylene	26.892	276	2523867	19.877	ng/ul#	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121823\
 Data File : BP018949.D
 Acq On : 19 Dec 2023 15:41
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020735

Quant Time: Dec 19 21:42:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 19 05:28:03 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.806 ng/uL
RT: 3.287 min Scan# 34
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion: 88 Resp: 84223
Ion Ratio Lower Upper
88 100
43 24.9 21.8 32.6
58 69.8 56.5 84.7



#5
Pyridine
Concen: 18.475 ng/uL
RT: 3.693 min Scan# 103
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion: 79 Resp: 563922
Ion Ratio Lower Upper
79 100
52 63.5 57.4 86.0
51 31.6 26.2 39.4









#6
Benzaldehyde
Concen: 17.975 ng/ul
RT: 6.969 min Scan# 661
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion: 77 Resp: 358118
Ion Ratio Lower Upper
77 100
105 100.1 77.9 116.9
106 96.6 74.6 111.8



#8
Phenol
Concen: 17.993 ng/ul
RT: 7.034 min Scan# 671
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion: 94 Resp: 680905
Ion Ratio Lower Upper
94 100
65 25.9 21.0 31.6
66 34.7 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 17.573 ng/ul
RT: 7.258 min Scan# 709
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion: 93 Resp: 546155
Ion Ratio Lower Upper
93 100
63 71.5 58.3 87.5
95 31.7 25.9 38.9



#12
2-Chlorophenol
Concen: 18.127 ng/ul
RT: 7.393 min Scan# 732
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion: 128 Resp: 552419
Ion Ratio Lower Upper
128 100
64 44.0 37.0 55.4
130 31.9 26.8 40.2









#13
2-Methylphenol
Concen: 17.035 ng/ul
RT: 8.281 min Scan# 884
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:108 Resp: 493547
Ion Ratio Lower Upper
108 100
107 89.7 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 15.086 ng/ul
RT: 8.357 min Scan# 896
Delta R.T. 0.000 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion: 45 Resp: 594918
Ion Ratio Lower Upper
45 100
77 17.7 12.7 19.1
79 13.9 10.5 15.7









#16
Acetophenone
Concen: 16.550 ng/ul
RT: 8.657 min Scan# 948
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:105 Resp: 770258
Ion Ratio Lower Upper
105 100
77 73.7 59.4 89.0
51 23.0 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 13.958 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion: 70 Resp: 357381
Ion Ratio Lower Upper
70 100
42 47.4 41.0 61.4
101 9.9 7.8 11.8
130 22.3 16.9 25.3









#18
4-Methylphenol
Concen: 16.376 ng/ul
RT: 8.610 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:108 Resp: 518071
Ion Ratio Lower Upper
108 100
107 112.4 90.6 136.0



#19
Hexachloroethane
Concen: 17.121 ng/ul
RT: 8.899 min Scan# 988
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:117 Resp: 213903
Ion Ratio Lower Upper
117 100
201 115.5 83.7 125.5
199 70.7 52.3 78.5









#22
Nitrobenzene
Concen: 19.628 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion: 77 Resp: 570241
Ion Ratio Lower Upper
77 100
123 48.1 35.8 53.8
65 12.6 10.5 15.7



#23
Isophorone
Concen: 16.967 ng/ul
RT: 9.563 min Scan# 1101
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion: 82 Resp: 1005043
Ion Ratio Lower Upper
82 100
95 7.1 5.5 8.3
138 18.0 13.5 20.3









#25
2-Nitrophenol
Concen: 21.378 ng/ul
RT: 9.746 min Scan# 1132
Delta R.T. 0.000 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:139 Resp: 283567
Ion Ratio Lower Upper
139 100
65 45.4 39.1 58.7
109 30.0 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 17.690 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:107 Resp: 458929
Ion Ratio Lower Upper
107 100
121 50.8 40.0 60.0
122 84.9 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.073 ng/ul
 RT: 10.046 min Scan# 1183
 Delta R.T. -0.006 min
 Lab File: BP018949.D
 Acq: 19 Dec 2023 15:41




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020735


Tgt Ion: 93 Resp: 638491
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.1 39.1
 123 11.1 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 19.723 ng/ul
 RT: 10.275 min Scan# 1222
 Delta R.T. -0.012 min
 Lab File: BP018949.D
 Acq: 19 Dec 2023 15:41

Tgt Ion: 162 Resp: 481234
 Ion Ratio Lower Upper
 162 100
 164 65.2 51.7 77.5
 98 34.5 29.2 43.8









#30
Naphthalene
Concen: 19.187 ng/ul
RT: 10.675 min Scan# 1310
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:128 Resp: 1562668
Ion Ratio Lower Upper
128 100
129 10.7 8.9 13.3
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 18.306 ng/ul
RT: 10.793 min Scan# 1310
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:127 Resp: 577622
Ion Ratio Lower Upper
127 100
129 31.2 25.6 38.4









#33
Hexachlorobutadiene
Concen: 22.753 ng/ul
RT: 10.951 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:225 Resp: 371807
Ion Ratio Lower Upper
225 100
223 62.5 49.5 74.3
227 63.0 50.9 76.3



#34
Caprolactam
Concen: 15.236 ng/ul
RT: 11.581 min Scan# 1444
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:113 Resp: 113989
Ion Ratio Lower Upper
113 100
55 150.4 133.7 200.5
56 110.5 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 15.787 ng/ul
RT: 11.922 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:107 Resp: 431517
Ion Ratio Lower Upper
107 100
144 26.3 21.0 31.6
142 81.7 63.6 95.4



#36
2-Methylnaphthalene
Concen: 17.081 ng/ul
RT: 12.287 min Scan# 1564
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:142 Resp: 985533
Ion Ratio Lower Upper
142 100
141 89.3 71.4 107.0





#37

1-Methylnaphthalene

Concen: 16.742 ng/ul

RT: 12.504 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:142 Resp: 975509

Ion Ratio Lower Upper

142 100

141 92.5 73.5 110.3

116 3.7 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 23.626 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Tgt Ion:216 Resp: 597450




Ion Ratio Lower Upper


216 100

214 79.1 63.9 95.9

179 19.9 16.3 24.5

108 14.2 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 9.518 ng/ul

RT: 12.628 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:237 Resp: 142658

Ion Ratio Lower Upper

237 100

235 64.1 49.4 74.0

272 13.0 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 22.019 ng/ul

RT: 12.898 min Scan# 1668

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:196 Resp: 358193


Ion Ratio Lower Upper

196 100

198 95.2 76.9 115.3

200 31.3 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 20.768 ng/ul

RT: 12.969 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:196 Resp: 366265

Ion Ratio Lower Upper

196 100

198 96.4 77.9 116.9

200 29.9 25.2 37.8



#43

1,1'-Biphenyl

Concen: 20.417 ng/ul

RT: 13.298 min Scan# 1736

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:154 Resp: 1230724


Ion Ratio Lower Upper

154 100

153 40.2 32.8 49.2

76 11.6 10.7 16.1





#44

2-Chloronaphthalene

Concen: 20.914 ng/ul

RT: 13.340 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:162 Resp: 995722

Ion Ratio Lower Upper

162 100

164 33.1 25.7 38.5

127 36.3 30.3 45.5



#45

2-Nitroaniline

Concen: 18.287 ng/ul

RT: 13.551 min Scan# 1779

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion: 65 Resp: 236027


Ion Ratio Lower Upper

65 100

92 75.0 56.9 85.3

138 116.9 88.2 132.2









#47
Dimethylphthalate
Concen: 18.400 ng/ul
RT: 13.928 min Scan# 1843
Delta R.T. 0.000 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:163 Resp: 1126693
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.0 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 19.443 ng/ul
RT: 14.051 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:165 Resp: 224932
Ion Ratio Lower Upper
165 100
89 50.7 46.6 70.0
121 18.1 17.3 25.9









#50
Acenaphthylene
Concen: 19.674 ng/ul
RT: 14.187 min Scan# 1887
Delta R.T. 0.000 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:152 Resp: 1515148
Ion Ratio Lower Upper
152 100
151 19.7 15.9 23.9
153 12.9 10.6 16.0



#51
3-Nitroaniline
Concen: 20.061 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:138 Resp: 231673
Ion Ratio Lower Upper
138 100
108 10.5 9.5 14.3
92 112.4 93.1 139.7





#52

Acenaphthene

Concen: 19.389 ng/ul

RT: 14.528 min Scan# 1945

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:153 Resp: 987297

Ion Ratio Lower Upper

153 100

152 47.0 38.1 57.1

154 88.6 71.7 107.5



#53

2,4-Dinitrophenol

Concen: 10.736 ng/ul

RT: 14.587 min Scan# 1955

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:184 Resp: 65828


Ion Ratio Lower Upper

184 100

63 55.9 48.2 72.4

154 61.3 49.1 73.7









#55
4-Nitrophenol
Concen: 18.492 ng/ul
RT: 14.692 min Scan# 1975
Delta R.T. -0.012 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:109 Resp: 142655
Ion Ratio Lower Upper
109 100
139 140.2 109.0 163.4
65 126.1 106.2 159.4



#56
Dibenzofuran
Concen: 19.075 ng/ul
RT: 14.863 min Scan# 2002
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:168 Resp: 1349964
Ion Ratio Lower Upper
168 100
139 34.6 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 18.679 ng/ul
RT: 14.834 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:165 Resp: 303977
Ion Ratio Lower Upper
165 100
63 39.3 32.7 49.1
182 2.9 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.993 ng/ul
RT: 15.092 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:232 Resp: 295762
Ion Ratio Lower Upper
232 100
131 36.2 33.8 50.8
130 1.9 1.8 2.8
166 27.3 24.7 37.1









#59
Diethylphthalate
Concen: 17.092 ng/ul
RT: 15.287 min Scan# 2075
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:149 Resp: 1024171
Ion Ratio Lower Upper
149 100
177 23.4 18.2 27.4
150 12.4 9.9 14.9



#61
Fluorene
Concen: 18.839 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:166 Resp: 1054979
Ion Ratio Lower Upper
166 100
165 98.2 78.0 117.0
167 13.4 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 19.195 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:204 Resp: 563754
Ion Ratio Lower Upper
204 100
206 32.2 27.1 40.7
141 52.2 45.8 68.8



#63
4-Nitroaniline
Concen: 21.720 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:138 Resp: 238964
Ion Ratio Lower Upper
138 100
92 50.6 42.3 63.5
108 75.8 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 14.045 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:198 Resp: 127827
Ion Ratio Lower Upper
198 100
182 4.1 3.4 5.2
77 21.8 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 18.416 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:169 Resp: 866393
Ion Ratio Lower Upper
169 100
168 67.7 53.1 79.7
167 35.8 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 19.465 ng/ul
RT: 16.404 min Scan# 2264
Delta R.T. 0.000 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:248 Resp: 348866
Ion Ratio Lower Upper
248 100
250 98.9 79.3 118.9
141 60.9 57.2 85.8



#69
Hexachlorobenzene
Concen: 20.615 ng/ul
RT: 16.516 min Scan# 2283
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:284 Resp: 412524
Ion Ratio Lower Upper
284 100
142 28.3 25.8 38.6
249 30.0 24.9 37.3





#70

Atrazine

Concen: 17.732 ng/ul

RT: 16.681 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:200 Resp: 248003

Ion Ratio Lower Upper

200 100

173 24.0 21.0 31.4

215 49.5 38.2 57.2



#71

Pentachlorophenol

Concen: 21.932 ng/ul

RT: 16.863 min Scan# 2342

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:266 Resp: 261024


Ion Ratio Lower Upper

266 100

264 62.3 50.2 75.2

268 63.5 50.5 75.7





#72

Phenanthrene

Concen: 19.672 ng/ul

RT: 17.257 min Scan# 2410

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:178 Resp: 1686742

Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 19.3 15.4 23.2



#74

Anthracene

Concen: 19.869 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:178 Resp: 1707652


Ion Ratio Lower Upper

178 100

179 15.7 12.5 18.7

176 18.7 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 23.506 ng/uL

RT: 13.263 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:216 Resp: 577018

Ion Ratio Lower Upper

216 100

214 78.7 61.4 92.2

218 48.8 39.2 58.8



#76

Pentachlorobenzene

Concen: 21.457 ng/uL

RT: 14.781 min Scan# 1988

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:250 Resp: 517395


Ion Ratio Lower Upper

250 100

248 62.8 50.6 75.8

252 63.6 51.3 76.9





#77

Carbazole

Concen: 22.511 ng/ul

RT: 17.622 min Scan# 2472

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:167 Resp: 1541983

Ion Ratio Lower Upper

167 100

166 21.7 17.1 25.7

139 12.3 10.2 15.4



#78

Di-n-butylphthalate

Concen: 17.939 ng/ul

RT: 18.175 min Scan# 2565

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:149 Resp: 1636263


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2

104 4.9 3.9 5.9









#80
Fluoranthene
Concen: 14.764 ng/ul
RT: 19.222 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:202 Resp: 2209424
Ion Ratio Lower Upper
202 100
101 10.1 9.1 13.7
100 7.6 7.1 10.7



#82
Pyrene
Concen: 15.394 ng/ul
RT: 19.580 min Scan# 2804
Delta R.T. -0.006 min
Lab File: BP018949.D
Acq: 19 Dec 2023 15:41

Tgt Ion:202 Resp: 2323899
Ion Ratio Lower Upper
202 100
101 11.3 11.0 16.4
100 9.1 9.1 13.7#





#83

Butylbenzylphthalate

Concen: 15.174 ng/ul

RT: 20.457 min Scan# 2954

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:149 Resp: 838651

Ion Ratio Lower Upper

149 100

91 64.8 56.6 85.0

206 24.1 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 19.841 ng/ul

RT: 21.227 min Scan# 3084

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:252 Resp: 784143


Ion Ratio Lower Upper

252 100

254 65.3 51.0 76.4

126 8.7 7.8 11.6





#85

Benzo(a)anthracene

Concen: 19.108 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:228 Resp: 2522483

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 27.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 15.518 ng/ul

RT: 21.216 min Scan# 3082

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:149 Resp: 1178776


Ion Ratio Lower Upper

149 100

167 28.2 22.2 33.2

279 5.1 3.3 4.9#





#87

Chrysene

Concen: 19.236 ng/ul

RT: 21.345 min Scan# 3105

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:228 Resp: 2338484

Ion Ratio Lower Upper

228 100

226 30.2 23.9 35.9

229 19.6 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 15.199 ng/ul




RT: 22.127 min Scan# 3237


Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Tgt Ion:149 Resp: 2175093





#90

Benzo(b)fluoranthene

Concen: 18.551 ng/ul

RT: 22.951 min Scan# 3378

Delta R.T. -0.006 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:252 Resp: 2680938

Ion Ratio Lower Upper

252 100

253 21.9 17.3 25.9

125 8.4 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 17.860 ng/ul

RT: 22.998 min Scan# 3385

Delta R.T. -0.012 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:252 Resp: 2497348


Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 8.1 8.0 12.0





#93

Benzo(a)pyrene

Concen: 18.700 ng/ul

RT: 23.568 min Scan# 34

Delta R.T. -0.012 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:252 Resp: 2469384

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8

125 9.2 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.701 ng/ul

RT: 26.133 min Scan# 3918

Delta R.T. -0.023 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41




Tgt Ion:276 Resp: 3108351


Ion Ratio Lower Upper

276 100

138 20.3 19.4 29.2

277 25.7 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 19.646 ng/ul

RT: 26.151 min Scan# 3925

Delta R.T. -0.023 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:278 Resp: 2578462

Ion Ratio Lower Upper

278 100

139 13.2 13.0 19.6

279 23.2 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 19.877 ng/ul

RT: 26.892 min Scan# 4047

Delta R.T. -0.023 min

Lab File: BP018949.D

Acq: 19 Dec 2023 15:41

Tgt Ion:276 Resp: 2523867

Ion Ratio Lower Upper

276 100

138 17.9 18.1 27.1#

277 23.7 19.4 29.0

