

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121824\
 Data File : BP023521.D
 Acq On : 19 Dec 2024 07:09
 Operator : RC/JU
 Sample : PB165677BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS677

Quant Time: Dec 19 08:13:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 00:30:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.528	152	231074	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.263	136	957520	20.000	ng/u1	#-0.02
38) Acenaphthene-d10	14.134	164	693199	20.000	ng/u1	-0.01
64) Phenanthrene-d10	16.939	188	1465229	20.000	ng/u1	#-0.02
79) Chrysene-d12	21.374	240	1556254	20.000	ng/u1	#-0.01
88) Perylene-d12	24.551	264	1639969	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.099	96	31933	5.088	ng/uL	0.00
4) Pyridine-d5	3.511	84	500183	30.991	ng/u1	0.00
7) Phenol-d5	6.740	99	596040	32.875	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.881	67	352388	30.806	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.075	132	414531	30.978	ng/u1	-0.01
15) 4-Methylphenol-d8	8.240	113	462808	31.628	ng/u1	-0.02
21) Nitrobenzene-d5	8.669	128	189180	26.458	ng/u1	-0.02
24) 2-Nitrophenol-d4	9.381	143	225192	27.327	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	9.916	165	515231	29.624	ng/u1	-0.02
31) 4-Chloroaniline-d4	10.422	131	529616	27.182	ng/u1	-0.02
46) Dimethylphthalate-d6	13.540	166	1496790	27.814	ng/u1	-0.02
49) Acenaphthylene-d8	13.822	160	1592975	28.065	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.416	143	199862	30.308	ng/u1	0.00
60) Fluorene-d10	15.151	176	1163573	24.583	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.310	200	252227	28.483	ng/u1	0.00
73) Anthracene-d10	17.039	188	1942336	29.046	ng/u1	-0.01
81) Pyrene-d10	19.427	212	2321416	28.551	ng/u1	-0.02
92) Benzo(a)pyrene-d12	24.310	264	2357330	26.852	ng/u1	-0.04
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.134	88	66119	9.521	ng/uL	95
5) Pyridine	3.528	79	501989	30.449	ng/u1	90
6) Benzaldehyde	6.710	77	48979	4.725	ng/u1	94
8) Phenol	6.763	94	623644	33.040	ng/u1#	86
10) Bis(2-Chloroethyl)ether	6.969	93	456766	30.657	ng/u1	99
12) 2-Chlorophenol	7.105	128	428547	30.549	ng/u1	99
13) 2-Methylphenol	7.981	108	436328	31.170	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.028	45	576213	31.802	ng/u1	98
16) Acetophenone	8.340	105	670875	27.239	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.316	70	373341	27.820	ng/u1	93
18) 4-Methylphenol	8.304	108	476133	31.513	ng/u1	98
19) Hexachloroethane	8.569	117	180942	26.491	ng/u1	92
22) Nitrobenzene	8.710	77	555756	25.544	ng/u1	98
23) Isophorone	9.222	82	1047226	27.176	ng/u1	98
25) 2-Nitrophenol	9.416	139	236253	26.839	ng/u1#	84
26) 2,4-Dimethylphenol	9.481	107	509062	26.117	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.699	93	612715	28.143	ng/u1	98
29) 2,4-Dichlorophenol	9.940	162	509533	29.560	ng/u1	99
30) Naphthalene	10.316	128	1361886	27.169	ng/u1	100
32) 4-Chloroaniline	10.451	127	335712	17.769	ng/u1	99
33) Hexachlorobutadiene	10.587	225	343812	22.486	ng/u1	97
34) Caprolactam	11.240	113	157485	31.312	ng/u1	95
35) 4-Chloro-3-methylphenol	11.593	107	552773	29.533	ng/u1	92
36) 2-Methylnaphthalene	11.928	142	975004	26.229	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121824\
 Data File : BP023521.D
 Acq On : 19 Dec 2024 07:09
 Operator : RC/JU
 Sample : PB165677BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS677

Quant Time: Dec 19 08:13:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 00:30:30 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.145	142	981878	25.921	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.304	216	644211	25.467	ng/ul	98
40) Hexachlorocyclopentadiene	12.269	237	227750	18.738	ng/ul	97
41) 2,4,6-Trichlorophenol	12.563	196	452437	30.223	ng/ul	98
42) 2,4,5-Trichlorophenol	12.657	196	493485	30.853	ng/ul	99
43) 1,1'-Biphenyl	12.951	154	1363351	27.758	ng/ul#	96
44) 2-Chloronaphthalene	12.998	162	1102012	27.739	ng/ul	96
45) 2-Nitroaniline	13.228	65	367164	31.898	ng/ul	92
47) Dimethylphthalate	13.592	163	1511789	28.519	ng/ul	99
48) 2,6-Dinitrotoluene	13.740	165	308067	30.330	ng/ul	86
50) Acenaphthylene	13.851	152	1635705	28.298	ng/ul	100
51) 3-Nitroaniline	14.081	138	272459	32.006	ng/ul	99
52) Acenaphthene	14.192	153	1155880	27.359	ng/ul	97
53) 2,4-Dinitrophenol	14.316	184	183293	27.984	ng/ul	88
55) 4-Nitrophenol	14.428	109	224004	25.214	ng/ul#	69
56) Dibenzofuran	14.545	168	1643235	25.610	ng/ul	95
57) 2,4-Dinitrotoluene	14.539	165	439304	26.840	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	14.792	232	385219	24.260	ng/ul	96
59) Diethylphthalate	14.981	149	1447046	27.371	ng/ul	100
61) Fluorene	15.204	166	1279222	24.485	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.192	204	691242	22.812	ng/ul	97
63) 4-Nitroaniline	15.263	138	260960	34.853	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.328	198	277450	29.535	ng/ul	99
67) N-Nitrosodiphenylamine	15.422	169	1143880	30.571	ng/ul	99
68) 4-Bromophenyl-phenylether	16.116	248	447853	25.892	ng/ul	98
69) Hexachlorobenzene	16.228	284	505494	23.855	ng/ul	97
70) Atrazine	16.416	200	467497	28.713	ng/ul	99
71) Pentachlorophenol	16.592	266	326530	26.400	ng/ul	97
72) Phenanthrene	16.980	178	2210198	29.177	ng/ul	99
74) Anthracene	17.081	178	2157859	28.489	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.916	216	638268	30.220	ng/uL	99
76) Pentachlorobenzene	14.463	250	619388	26.606	ng/uL	100
77) Carbazole	17.369	167	2007788	32.536	ng/ul	98
78) Di-n-butylphthalate	17.945	149	2358946	31.112	ng/ul	99
80) Fluoranthene	19.080	202	2781303	29.853	ng/ul#	93
82) Pyrene	19.457	202	2914752	29.482	ng/ul#	91
83) Butylbenzylphthalate	20.416	149	1115695	32.583	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.280	252	952308	26.668	ng/ul#	98
85) Benzo(a)anthracene	21.357	228	2867495	27.477	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.280	149	1578882	32.109	ng/ul	98
87) Chrysene	21.421	228	2690081	26.899	ng/ul	100
89) Di-n-octyl phthalate	22.480	149	2785119	34.772	ng/ul	100
90) Benzo(b)fluoranthene	23.539	252	2921957	28.358	ng/ul#	98
91) Benzo(k)fluoranthene	23.610	252	2806210	27.373	ng/ul#	98
93) Benzo(a)pyrene	24.386	252	2627051	28.120	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	28.127	276	3282982	27.345	ng/ul#	94
95) Dibenzo(a,h)anthracene	28.180	278	2596135	27.021	ng/ul#	96
96) Benzo(g,h,i)perylene	29.227	276	2553589	27.780	ng/ul#	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121824\
 Data File : BP023521.D
 Acq On : 19 Dec 2024 07:09
 Operator : RC/JU
 Sample : PB165677BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS677

Quant Time: Dec 19 08:13:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 00:30:30 2024
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.521 ng/uL
RT: 3.134 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion: 88 Resp: 66119
Ion Ratio Lower Upper
88 100
43 27.1 22.9 34.3
58 71.7 53.1 79.7



#5
Pyridine
Concen: 30.449 ng/uL
RT: 3.528 min Scan# 92
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion: 79 Resp: 501989
Ion Ratio Lower Upper
79 100
52 68.8 60.8 91.2
51 31.2 30.7 46.1









#6
Benzaldehyde
Concen: 4.725 ng/ul
RT: 6.710 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion: 77 Resp: 48979
Ion Ratio Lower Upper
77 100
105 93.2 70.6 106.0
106 90.0 67.4 101.0



#8
Phenol
Concen: 33.040 ng/ul
RT: 6.763 min Scan# 642
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion: 94 Resp: 623644
Ion Ratio Lower Upper
94 100
65 28.8 28.0 42.0
66 40.8 42.0 63.0#









#10
Bis(2-Chloroethyl)ether
Concen: 30.657 ng/ul
RT: 6.969 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion: 93 Resp: 456766
Ion Ratio Lower Upper
93 100
63 76.0 61.3 91.9
95 32.0 25.4 38.0



#12
2-Chlorophenol
Concen: 30.549 ng/ul
RT: 7.105 min Scan# 700
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion: 128 Resp: 428547
Ion Ratio Lower Upper
128 100
64 56.2 44.8 67.2
130 32.2 24.7 37.1





#13

2-Methylphenol

Concen: 31.170 ng/ul

RT: 7.981 min Scan# 849

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :





SLCS677

Tgt Ion:108 Resp: 436328

Ion Ratio Lower Upper

108 100

107 91.2 73.8 110.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 31.802 ng/ul

RT: 8.028 min Scan# 857

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion: 45 Resp: 576213


Ion Ratio Lower Upper

45 100

77 18.3 13.9 20.9

79 13.0 11.3 16.9









#16
Acetophenone
Concen: 27.239 ng/ul
RT: 8.340 min Scan# 910
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:105 Resp: 670875
Ion Ratio Lower Upper
105 100
77 85.3 71.2 106.8
51 26.1 22.9 34.3



#17
N-Nitroso-di-n-propylamine
Concen: 27.820 ng/ul
RT: 8.316 min Scan# 906
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion: 70 Resp: 373341
Ion Ratio Lower Upper
70 100
42 47.3 43.5 65.3
101 9.3 6.8 10.2
130 16.9 14.0 21.0









#18
4-Methylphenol
Concen: 31.513 ng/ul
RT: 8.304 min Scan# 905
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:108 Resp: 476133
Ion Ratio Lower Upper
108 100
107 116.4 95.1 142.7



#19
Hexachloroethane
Concen: 26.491 ng/ul
RT: 8.569 min Scan# 949
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:117 Resp: 180942
Ion Ratio Lower Upper
117 100
201 121.4 107.0 160.6
199 74.1 62.1 93.1









#22
Nitrobenzene
Concen: 25.544 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion: 77 Resp: 555756
Ion Ratio Lower Upper
77 100
123 38.4 29.8 44.6
65 12.9 11.0 16.6



#23
Isophorone
Concen: 27.176 ng/ul
RT: 9.222 min Scan# 1060
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion: 82 Resp: 1047226
Ion Ratio Lower Upper
82 100
95 7.2 7.0 10.4
138 14.5 12.0 18.0









#25
2-Nitrophenol
Concen: 26.839 ng/ul
RT: 9.416 min Scan# 1092
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:139 Resp: 236253
Ion Ratio Lower Upper
139 100
65 59.4 55.1 82.7
109 33.9 39.1 58.7#



#26
2,4-Dimethylphenol
Concen: 26.117 ng/ul
RT: 9.481 min Scan# 1104
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:107 Resp: 509062
Ion Ratio Lower Upper
107 100
121 46.1 36.6 55.0
122 76.6 58.6 87.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 28.143 ng/ul
 RT: 9.699 min Scan# 1141
 Delta R.T. -0.012 min
 Lab File: BP023521.D
 Acq: 19 Dec 2024 07:09




Instrument :
 BNA_P
 ClientSampleId :
 SLCS677


Tgt Ion: 93 Resp: 612715
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.0 37.4
 123 9.5 7.8 11.6



#29
 2,4-Dichlorophenol
 Concen: 29.560 ng/ul
 RT: 9.940 min Scan# 1182
 Delta R.T. -0.024 min
 Lab File: BP023521.D
 Acq: 19 Dec 2024 07:09

Tgt Ion:162 Resp: 509533
 Ion Ratio Lower Upper
 162 100
 164 66.7 53.0 79.6
 98 38.9 30.3 45.5









#30
Naphthalene
Concen: 27.169 ng/ul
RT: 10.316 min Scan# 1246
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:128 Resp: 1361886
Ion Ratio Lower Upper
128 100
129 10.9 8.5 12.7
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 17.769 ng/ul
RT: 10.451 min Scan# 1269
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:127 Resp: 335712
Ion Ratio Lower Upper
127 100
129 33.0 25.9 38.9









#33
Hexachlorobutadiene
Concen: 22.486 ng/ul
RT: 10.587 min Scan# 1292
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:225 Resp: 343812
Ion Ratio Lower Upper
225 100
223 63.1 51.4 77.2
227 63.0 53.0 79.4



#34
Caprolactam
Concen: 31.312 ng/ul
RT: 11.240 min Scan# 1403
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:113 Resp: 157485
Ion Ratio Lower Upper
113 100
55 169.8 127.5 191.3
56 126.8 100.7 151.1









#35
4-Chloro-3-methylphenol
Concen: 29.533 ng/ul
RT: 11.593 min Scan# 1463
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:107 Resp: 552773
Ion Ratio Lower Upper
107 100
144 23.6 17.9 26.9
142 74.3 53.0 79.4



#36
2-Methylnaphthalene
Concen: 26.229 ng/ul
RT: 11.928 min Scan# 1520
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:142 Resp: 975004
Ion Ratio Lower Upper
142 100
141 89.4 69.0 103.6





#37

1-Methylnaphthalene

Concen: 25.921 ng/ul

RT: 12.145 min Scan# 1557

Delta R.T. -0.024 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:142 Resp: 981878

Ion Ratio Lower Upper

142 100

141 92.9 71.1 106.7

116 4.2 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 25.467 ng/ul

RT: 12.304 min Scan# 1584

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Tgt Ion:216 Resp: 644211




Ion Ratio Lower Upper


216 100

214 78.5 64.4 96.6

179 20.9 16.4 24.6

108 15.1 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 18.738 ng/ul
RT: 12.269 min Scan# 1578
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:237 Resp: 227750
Ion Ratio Lower Upper
237 100
235 60.3 50.5 75.7
272 12.5 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 30.223 ng/ul
RT: 12.563 min Scan# 1628
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:196 Resp: 452437
Ion Ratio Lower Upper
196 100
198 93.9 77.5 116.3
200 31.0 25.0 37.4





#42

2,4,5-Trichlorophenol

Concen: 30.853 ng/ul

RT: 12.657 min Scan# 1644

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:196 Resp: 493485

Ion Ratio Lower Upper

196 100

198 95.7 76.3 114.5

200 32.1 24.6 36.8



#43

1,1'-Biphenyl

Concen: 27.758 ng/ul

RT: 12.951 min Scan# 1694

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:154 Resp: 1363351


Ion Ratio Lower Upper

154 100

153 39.4 32.8 49.2

76 15.6 10.0 15.0#





#44

2-Chloronaphthalene

Concen: 27.739 ng/ul

RT: 12.998 min Scan# 1702

Delta R.T. -0.006 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:162 Resp: 1102012

Ion Ratio Lower Upper

162 100

164 33.4 25.8 38.6

127 35.6 31.2 46.8



#45

2-Nitroaniline

Concen: 31.898 ng/ul

RT: 13.228 min Scan# 1741

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion: 65 Resp: 367164


Ion Ratio Lower Upper

65 100

92 68.8 50.5 75.7

138 93.0 68.0 102.0









#47
Dimethylphthalate
Concen: 28.519 ng/ul
RT: 13.592 min Scan# 1803
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:163 Resp: 1511789
Ion Ratio Lower Upper
163 100
194 4.3 3.3 4.9
164 10.3 8.6 13.0



#48
2,6-Dinitrotoluene
Concen: 30.330 ng/ul
RT: 13.740 min Scan# 1828
Delta R.T. -0.000 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:165 Resp: 308067
Ion Ratio Lower Upper
165 100
89 55.9 55.4 83.0
121 19.3 18.2 27.2





#50

Acenaphthylene

Concen: 28.298 ng/ul

RT: 13.851 min Scan# 1847

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:152 Resp: 1635705

Ion Ratio Lower Upper

152 100

151 19.4 15.7 23.5

153 13.1 10.5 15.7



#51

3-Nitroaniline

Concen: 32.006 ng/ul

RT: 14.081 min Scan# 1886

Delta R.T. -0.006 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:138 Resp: 272459


Ion Ratio Lower Upper

138 100

108 11.9 9.8 14.8

92 132.9 105.8 158.8





#52

Acenaphthene

Concen: 27.359 ng/ul

RT: 14.192 min Scan# 1905

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:153 Resp: 1155880

Ion Ratio Lower Upper

153 100

152 46.0 39.0 58.6

154 86.4 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 27.984 ng/ul

RT: 14.316 min Scan# 1926

Delta R.T. -0.000 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:184 Resp: 183293


Ion Ratio Lower Upper

184 100

63 56.6 50.4 75.6

154 60.0 57.9 86.9









#55
4-Nitrophenol
Concen: 25.214 ng/ul
RT: 14.428 min Scan# 1945
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:109 Resp: 224004
Ion Ratio Lower Upper
109 100
139 114.9 66.7 100.1#
65 136.9 86.1 129.1#



#56
Dibenzofuran
Concen: 25.610 ng/ul
RT: 14.545 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:168 Resp: 1643235
Ion Ratio Lower Upper
168 100
139 35.6 31.0 46.4









#57
2,4-Dinitrotoluene
Concen: 26.840 ng/ul
RT: 14.539 min Scan# 1964
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:165 Resp: 439304
Ion Ratio Lower Upper
165 100
63 61.4 52.2 78.2
182 2.7 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 24.260 ng/ul
RT: 14.792 min Scan# 2007
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:232 Resp: 385219
Ion Ratio Lower Upper
232 100
131 36.4 32.2 48.2
130 2.1 2.1 3.1
166 29.7 24.0 36.0





#59

Diethylphthalate

Concen: 27.371 ng/ul

RT: 14.981 min Scan# 2039

Delta R.T. -0.000 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:149 Resp: 1447046

Ion Ratio Lower Upper

149 100

177 22.8 18.4 27.6

150 12.4 9.8 14.8



#61

Fluorene

Concen: 24.485 ng/ul

RT: 15.204 min Scan# 2077

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:166 Resp: 1279222


Ion Ratio Lower Upper

166 100

165 97.3 78.5 117.7

167 13.1 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 22.812 ng/ul
RT: 15.192 min Scan# 2075
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:204 Resp: 691242
Ion Ratio Lower Upper
204 100
206 32.3 26.9 40.3
141 50.8 42.4 63.6



#63
4-Nitroaniline
Concen: 34.853 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:138 Resp: 260960
Ion Ratio Lower Upper
138 100
92 55.0 53.0 79.4
108 88.9 96.2 144.2#









#66
4,6-Dinitro-2-methylphenol
Concen: 29.535 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. -0.000 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:198 Resp: 277450
Ion Ratio Lower Upper
198 100
182 4.4 3.7 5.5
77 24.2 19.1 28.7



#67
N-Nitrosodiphenylamine
Concen: 30.571 ng/ul
RT: 15.422 min Scan# 2114
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:169 Resp: 1143880
Ion Ratio Lower Upper
169 100
168 67.8 53.3 79.9
167 36.5 28.6 42.8









#68
4-Bromophenyl-phenylether
Concen: 25.892 ng/ul
RT: 16.116 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:248 Resp: 447853
Ion Ratio Lower Upper
248 100
250 97.3 78.5 117.7
141 62.5 52.5 78.7



#69
Hexachlorobenzene
Concen: 23.855 ng/ul
RT: 16.228 min Scan# 2251
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:284 Resp: 505494
Ion Ratio Lower Upper
284 100
142 31.1 23.3 34.9
249 32.9 27.2 40.8





#70

Atrazine

Concen: 28.713 ng/ul

RT: 16.416 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:200 Resp: 467497

Ion Ratio Lower Upper

200 100

173 24.5 18.9 28.3

215 47.4 38.0 57.0



#71

Pentachlorophenol

Concen: 26.400 ng/ul

RT: 16.592 min Scan# 2313

Delta R.T. -0.024 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:266 Resp: 326530


Ion Ratio Lower Upper

266 100

264 61.6 47.9 71.9

268 63.2 48.2 72.4





#72

Phenanthrene

Concen: 29.177 ng/ul

RT: 16.980 min Scan# 21

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:178 Resp: 2210198

Ion Ratio Lower Upper

178 100

179 15.4 12.1 18.1

176 18.9 15.4 23.0



#74

Anthracene

Concen: 28.489 ng/ul

RT: 17.081 min Scan# 2396

Delta R.T. -0.006 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:178 Resp: 2157859


Ion Ratio Lower Upper

178 100

179 15.7 12.3 18.5

176 18.9 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 30.220 ng/uL
RT: 12.916 min Scan# 1688
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:216 Resp: 638268
Ion Ratio Lower Upper
216 100
214 77.3 62.9 94.3
218 48.0 39.2 58.8



#76
Pentachlorobenzene
Concen: 26.606 ng/uL
RT: 14.463 min Scan# 1951
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:250 Resp: 619388
Ion Ratio Lower Upper
250 100
248 62.0 50.1 75.1
252 63.6 51.0 76.4





#77

Carbazole

Concen: 32.536 ng/ul

RT: 17.369 min Scan# 2447

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:167 Resp: 2007788

Ion Ratio Lower Upper

167 100

166 21.1 17.3 25.9

139 11.9 10.3 15.5



#78

Di-n-butylphthalate

Concen: 31.112 ng/ul

RT: 17.945 min Scan# 2543

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:149 Resp: 2358946


Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.0

104 5.5 5.2 7.8





#80

Fluoranthene

Concen: 29.853 ng/u1

RT: 19.080 min Scan# 21

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:202 Resp: 2781303

Ion Ratio Lower Upper

202 100

101 9.6 5.6 8.4#

100 7.8 4.6 7.0#



#82

Pyrene

Concen: 29.482 ng/u1

RT: 19.457 min Scan# 2800

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:202 Resp: 2914752


Ion Ratio Lower Upper

202 100

101 12.0 6.7 10.1#

100 10.0 5.8 8.8#





#83

Butylbenzylphthalate

Concen: 32.583 ng/ul

RT: 20.416 min Scan# 2963

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:149 Resp: 1115695

Ion Ratio Lower Upper

149 100

91 75.4 59.0 88.4

206 24.1 20.7 31.1



#84

3,3'-Dichlorobenzidine

Concen: 26.668 ng/ul

RT: 21.280 min Scan# 3110

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:252 Resp: 952308


Ion Ratio Lower Upper

252 100

254 63.6 51.8 77.8

126 8.2 5.0 7.6#





#85

Benzo(a)anthracene

Concen: 27.477 ng/ul

RT: 21.357 min Scan# 3110

Delta R.T. -0.012 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:228 Resp: 2867495

Ion Ratio Lower Upper

228 100

229 19.8 15.7 23.5

226 27.1 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 32.109 ng/ul

RT: 21.280 min Scan# 3110

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:149 Resp: 1578882


Ion Ratio Lower Upper

149 100

167 29.5 22.5 33.7

279 4.9 3.7 5.5





#87

Chrysene

Concen: 26.899 ng/ul

RT: 21.421 min Scan# 3134

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:228 Resp: 2690081

Ion Ratio Lower Upper

228 100

226 30.4 24.2 36.2

229 19.7 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 34.772 ng/ul




RT: 22.480 min Scan# 3314


Delta R.T. -0.024 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Tgt Ion:149 Resp: 2785119









#90
Benzo(b)fluoranthene
Concen: 28.358 ng/ul
RT: 23.539 min Scan# 3494
Delta R.T. -0.018 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09




Instrument :
BNA_P
ClientSampleId :
SLCS677


Tgt Ion:252 Resp: 2921957
Ion Ratio Lower Upper
252 100
253 21.8 17.3 25.9
125 7.8 4.4 6.6#



#91
Benzo(k)fluoranthene
Concen: 27.373 ng/ul
RT: 23.610 min Scan# 3506
Delta R.T. -0.012 min
Lab File: BP023521.D
Acq: 19 Dec 2024 07:09

Tgt Ion:252 Resp: 2806210
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 7.9 4.3 6.5#





#93

Benzo(a)pyrene

Concen: 28.120 ng/ul

RT: 24.386 min Scan# 30

Delta R.T. -0.024 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09

Instrument :

BNA_P

ClientSampleId :

SLCS677





Tgt Ion:252 Resp: 2627051

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 9.1 5.0 7.4#



#94

Indeno(1,2,3-cd)pyrene

Concen: 27.345 ng/ul

RT: 28.127 min Scan# 4274

Delta R.T. -0.018 min

Lab File: BP023521.D

Acq: 19 Dec 2024 07:09




Tgt Ion:276 Resp: 3282982


Ion Ratio Lower Upper

276 100

138 15.2 7.8 11.6#

277 23.8 19.8 29.8









#95
 Dibenzo(a,h)anthracene
 Concen: 27.021 ng/ul
 RT: 28.180 min Scan# 41
 Delta R.T. -0.035 min
 Lab File: BP023521.D
 Acq: 19 Dec 2024 07:09

Instrument :
 BNA_P
 ClientSampleId :
 SLCS677

Tgt Ion:278 Resp: 2596135
 Ion Ratio Lower Upper
 278 100
 139 12.0 6.0 9.0#
 279 23.5 19.2 28.8



#96
 Benzo(g,h,i)perylene
 Concen: 27.780 ng/ul
 RT: 29.227 min Scan# 4461
 Delta R.T. -0.024 min
 Lab File: BP023521.D
 Acq: 19 Dec 2024 07:09

Tgt Ion:276 Resp: 2553589
 Ion Ratio Lower Upper
 276 100
 138 17.1 8.5 12.7#
 277 23.8 18.9 28.3

