

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122023\
 Data File : BP018951.D
 Acq On : 20 Dec 2023 09:13
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736

Quant Time: Dec 20 21:45:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 19 05:28:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	179362	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.622	136	661015	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.463	164	410826	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.216	188	920087	20.000	ng/ul	0.00
79) Chrysene-d12	21.304	240	948837	20.000	ng/ul	-0.01
88) Perylene-d12	23.669	264	1151649	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	40338	7.664	ng/uL	0.00
4) Pyridine-d5	3.675	84	283772	20.128	ng/ul	0.00
7) Phenol-d5	7.005	99	343916	19.107	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.164	67	195851	18.358	ng/ul	0.00
11) 2-Chlorophenol-d4	7.358	132	269397	19.181	ng/ul	-0.01
15) 4-Methylphenol-d8	8.546	113	260002	17.849	ng/ul	-0.01
21) Nitrobenzene-d5	8.987	128	124755	21.205	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.710	143	139313	22.840	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.246	165	267914	21.481	ng/ul	-0.01
31) 4-Chloroaniline-d4	10.763	131	328348	20.144	ng/ul	-0.01
46) Dimethylphthalate-d6	13.881	166	749350	20.437	ng/ul	0.00
49) Acenaphthylene-d8	14.151	160	847758	20.997	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.675	143	130600	21.699	ng/ul	-0.01
60) Fluorene-d10	15.457	176	643721	20.915	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	123550	22.395	ng/ul	-0.01
73) Anthracene-d10	17.310	188	1015084	20.986	ng/ul	0.00
81) Pyrene-d10	19.551	212	1231307	16.671	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.516	264	1377183	20.509	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	44710	7.746	ng/uL	93
5) Pyridine	3.693	79	289512	20.335	ng/ul	97
6) Benzaldehyde	6.969	77	189242	20.363	ng/ul	97
8) Phenol	7.028	94	337216	19.104	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.258	93	271859	18.753	ng/ul	99
12) 2-Chlorophenol	7.393	128	274552	19.315	ng/ul	96
13) 2-Methylphenol	8.275	108	251093	18.581	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.352	45	309751	16.840	ng/ul	98
16) Acetophenone	8.652	105	396550	18.266	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.640	70	189547	15.871	ng/ul	97
18) 4-Methylphenol	8.605	108	262100	17.762	ng/ul	100
19) Hexachloroethane	8.899	117	113105	19.409	ng/ul	89
22) Nitrobenzene	9.028	77	290344	20.298	ng/ul	99
23) Isophorone	9.557	82	539123	18.485	ng/ul	98
25) 2-Nitrophenol	9.740	139	143985	22.047	ng/ul	98
26) 2,4-Dimethylphenol	9.805	107	253940	19.881	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.040	93	338057	18.360	ng/ul	99
29) 2,4-Dichlorophenol	10.275	162	256450	21.347	ng/ul	98
30) Naphthalene	10.669	128	825864	20.595	ng/ul	100
32) 4-Chloroaniline	10.787	127	317555	20.441	ng/ul	99
33) Hexachlorobutadiene	10.952	225	194740	24.204	ng/ul	100
34) Caprolactam	11.575	113	83023	22.538	ng/ul	92
35) 4-Chloro-3-methylphenol	11.922	107	253817	18.860	ng/ul	99
36) 2-Methylnaphthalene	12.281	142	558918	19.674	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122023\
 Data File : BP018951.D
 Acq On : 20 Dec 2023 09:13
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736

Quant Time: Dec 20 21:45:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 19 05:28:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.504	142	561504	19.572	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.651	216	343450	23.042	ng/ul	97
40) Hexachlorocyclopentadiene	12.628	237	150278	17.010	ng/ul	96
41) 2,4,6-Trichlorophenol	12.898	196	213185	22.233	ng/ul	99
42) 2,4,5-Trichlorophenol	12.969	196	221928	21.349	ng/ul	99
43) 1,1'-Biphenyl	13.298	154	739863	20.823	ng/ul	98
44) 2-Chloronaphthalene	13.340	162	601301	21.427	ng/ul	98
45) 2-Nitroaniline	13.551	65	150137	19.735	ng/ul	94
47) Dimethylphthalate	13.922	163	737765	20.441	ng/ul	99
48) 2,6-Dinitrotoluene	14.045	165	147012	21.559	ng/ul	96
50) Acenaphthylene	14.181	152	947652	20.876	ng/ul	100
51) 3-Nitroaniline	14.375	138	145919	21.436	ng/ul	99
52) Acenaphthene	14.522	153	625887	20.853	ng/ul	98
53) 2,4-Dinitrophenol	14.587	184	78134	21.620	ng/ul	94
55) 4-Nitrophenol	14.693	109	100945	22.200	ng/ul	94
56) Dibenzofuran	14.863	168	883507	21.180	ng/ul	97
57) 2,4-Dinitrotoluene	14.834	165	213005	22.206	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.092	232	197023	22.595	ng/ul	91
59) Diethylphthalate	15.287	149	717709	20.321	ng/ul	100
61) Fluorene	15.510	166	714793	21.655	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.510	204	382920	22.119	ng/ul	94
63) 4-Nitroaniline	15.540	138	144401	22.268	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.598	198	130481	22.122	ng/ul	92
67) N-Nitrosodiphenylamine	15.722	169	606531	19.894	ng/ul	99
68) 4-Bromophenyl-phenylether	16.404	248	243082	20.927	ng/ul	93
69) Hexachlorobenzene	16.516	284	287681	22.182	ng/ul	97
70) Atrazine	16.681	200	167772	18.509	ng/ul	100
71) Pentachlorophenol	16.863	266	172433	22.356	ng/ul	100
72) Phenanthrene	17.257	178	1149487	20.686	ng/ul	100
74) Anthracene	17.345	178	1181738	21.216	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.257	216	341747	21.481	ng/ul	98
76) Pentachlorobenzene	14.781	250	337293	21.583	ng/ul	99
77) Carbazole	17.622	167	1046682	23.578	ng/ul	99
78) Di-n-butylphthalate	18.175	149	1236783	20.923	ng/ul	99
80) Fluoranthene	19.228	202	1399788	16.033	ng/ul	97
82) Pyrene	19.580	202	1470586	16.697	ng/ul	94
83) Butylbenzylphthalate	20.457	149	570628	17.698	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.227	252	510504	22.140	ng/ul	99
85) Benzo(a)anthracene	21.292	228	1569409	20.378	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.216	149	834077	18.821	ng/ul	99
87) Chrysene	21.339	228	1460700	20.596	ng/ul	100
89) Di-n-octyl phthalate	22.127	149	1453315	17.284	ng/ul	100
90) Benzo(b)fluoranthene	22.951	252	1691563	19.921	ng/ul	98
91) Benzo(k)fluoranthene	22.998	252	1576487	19.189	ng/ul	98
93) Benzo(a)pyrene	23.563	252	1562570	20.140	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.127	276	2007335	21.654	ng/ul	96
95) Dibenzo(a,h)anthracene	26.145	278	1671888	21.681	ng/ul	97
96) Benzo(g,h,i)perylene	26.886	276	1625306	21.786	ng/ul#	94





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.746 ng/uL
RT: 3.293 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion: 88 Resp: 44710
Ion Ratio Lower Upper
88 100
43 25.3 21.8 32.6
58 64.1 56.5 84.7



#5
Pyridine
Concen: 20.335 ng/uL
RT: 3.693 min Scan# 103
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion: 79 Resp: 289512
Ion Ratio Lower Upper
79 100
52 68.6 57.4 86.0
51 32.1 26.2 39.4









#6
Benzaldehyde
Concen: 20.363 ng/ul
RT: 6.969 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion: 77 Resp: 189242
Ion Ratio Lower Upper
77 100
105 101.4 77.9 116.9
106 95.7 74.6 111.8



#8
Phenol
Concen: 19.104 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion: 94 Resp: 337216
Ion Ratio Lower Upper
94 100
65 26.2 21.0 31.6
66 35.5 28.6 43.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 18.753 ng/ul
 RT: 7.258 min Scan# 709
 Delta R.T. -0.006 min
 Lab File: BP018951.D
 Acq: 20 Dec 2023 09:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736


Tgt Ion: 93 Resp: 271859
 Ion Ratio Lower Upper
 93 100
 63 72.0 58.3 87.5
 95 32.5 25.9 38.9



#12
 2-Chlorophenol
 Concen: 19.315 ng/ul
 RT: 7.393 min Scan# 732
 Delta R.T. -0.006 min
 Lab File: BP018951.D
 Acq: 20 Dec 2023 09:13

Tgt Ion: 128 Resp: 274552
 Ion Ratio Lower Upper
 128 100
 64 42.8 37.0 55.4
 130 31.9 26.8 40.2





#13

2-Methylphenol

Concen: 18.581 ng/ul

RT: 8.275 min Scan# 882

Delta R.T. -0.012 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :





SSTD020736

Tgt Ion:108 Resp: 251093

Ion Ratio Lower Upper

108 100

107 87.7 70.6 105.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 16.840 ng/ul

RT: 8.352 min Scan# 895

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13




Tgt Ion: 45 Resp: 309751


Ion Ratio Lower Upper

45 100

77 17.4 12.7 19.1

79 12.8 10.5 15.7









#16
Acetophenone
Concen: 18.266 ng/ul
RT: 8.652 min Scan# 946
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:105 Resp: 396550
Ion Ratio Lower Upper
105 100
77 74.9 59.4 89.0
51 23.6 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 15.871 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion: 70 Resp: 189547
Ion Ratio Lower Upper
70 100
42 49.2 41.0 61.4
101 10.8 7.8 11.8
130 22.5 16.9 25.3









#18
4-Methylphenol
Concen: 17.762 ng/ul
RT: 8.605 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:108 Resp: 262100
Ion Ratio Lower Upper
108 100
107 113.6 90.6 136.0



#19
Hexachloroethane
Concen: 19.409 ng/ul
RT: 8.899 min Scan# 988
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:117 Resp: 113105
Ion Ratio Lower Upper
117 100
201 116.9 83.7 125.5
199 72.6 52.3 78.5









#22
Nitrobenzene
Concen: 20.298 ng/ul
RT: 9.028 min Scan# 1010
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion: 77 Resp: 290344
Ion Ratio Lower Upper
77 100
123 45.7 35.8 53.8
65 13.1 10.5 15.7



#23
Isophorone
Concen: 18.485 ng/ul
RT: 9.557 min Scan# 1100
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion: 82 Resp: 539123
Ion Ratio Lower Upper
82 100
95 7.3 5.5 8.3
138 17.6 13.5 20.3









#25
2-Nitrophenol
Concen: 22.047 ng/ul
RT: 9.740 min Scan# 1131
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:139 Resp: 143985
Ion Ratio Lower Upper
139 100
65 48.0 39.1 58.7
109 31.7 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 19.881 ng/ul
RT: 9.805 min Scan# 1142
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:107 Resp: 253940
Ion Ratio Lower Upper
107 100
121 51.3 40.0 60.0
122 83.7 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.360 ng/ul
 RT: 10.040 min Scan# 1182
 Delta R.T. -0.012 min
 Lab File: BP018951.D
 Acq: 20 Dec 2023 09:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736


Tgt Ion: 93 Resp: 338057
 Ion Ratio Lower Upper
 93 100
 95 31.7 26.1 39.1
 123 10.8 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 21.347 ng/ul
 RT: 10.275 min Scan# 1222
 Delta R.T. -0.012 min
 Lab File: BP018951.D
 Acq: 20 Dec 2023 09:13

Tgt Ion: 162 Resp: 256450
 Ion Ratio Lower Upper
 162 100
 164 64.7 51.7 77.5
 98 33.8 29.2 43.8









#30
Naphthalene
Concen: 20.595 ng/ul
RT: 10.669 min Scan# 1289
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:128 Resp: 825864
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 12.9 10.5 15.7



#32
4-Chloroaniline
Concen: 20.441 ng/ul
RT: 10.787 min Scan# 1309
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:127 Resp: 317555
Ion Ratio Lower Upper
127 100
129 31.2 25.6 38.4









#33
Hexachlorobutadiene
Concen: 24.204 ng/ul
RT: 10.952 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:225 Resp: 194740
Ion Ratio Lower Upper
225 100
223 62.1 49.5 74.3
227 63.1 50.9 76.3



#34
Caprolactam
Concen: 22.538 ng/ul
RT: 11.575 min Scan# 1443
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:113 Resp: 83023
Ion Ratio Lower Upper
113 100
55 152.8 133.7 200.5
56 116.1 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 18.860 ng/ul
RT: 11.922 min Scan# 1502
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:107 Resp: 253817
Ion Ratio Lower Upper
107 100
144 26.3 21.0 31.6
142 81.0 63.6 95.4



#36
2-Methylnaphthalene
Concen: 19.674 ng/ul
RT: 12.281 min Scan# 1563
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:142 Resp: 558918
Ion Ratio Lower Upper
142 100
141 90.3 71.4 107.0









#37
 1-Methylnaphthalene
 Concen: 19.572 ng/ul
 RT: 12.504 min Scan# 1601
 Delta R.T. -0.006 min
 Lab File: BP018951.D
 Acq: 20 Dec 2023 09:13

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736

Tgt Ion:142 Resp: 561504




Ion	Ratio	Lower	Upper
142	100		
141	91.8	73.5	110.3
116	3.8	2.8	4.2




#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 23.042 ng/ul
 RT: 12.651 min Scan# 1626
 Delta R.T. -0.006 min
 Lab File: BP018951.D
 Acq: 20 Dec 2023 09:13

Tgt Ion:216 Resp: 343450

Ion	Ratio	Lower	Upper
216	100		
214	77.5	63.9	95.9
179	19.3	16.3	24.5
108	13.8	12.6	19.0









#40
Hexachlorocyclopentadiene
Concen: 17.010 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:237 Resp: 150278
Ion Ratio Lower Upper
237 100
235 64.9 49.4 74.0
272 13.9 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 22.233 ng/ul
RT: 12.898 min Scan# 1668
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:196 Resp: 213185
Ion Ratio Lower Upper
196 100
198 97.1 76.9 115.3
200 32.8 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 21.349 ng/ul

RT: 12.969 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:196 Resp: 221928

Ion Ratio Lower Upper

196 100

198 98.2 77.9 116.9

200 30.7 25.2 37.8



#43

1,1'-Biphenyl

Concen: 20.823 ng/ul

RT: 13.298 min Scan# 1736

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13




Tgt Ion:154 Resp: 739863


Ion Ratio Lower Upper

154 100

153 40.1 32.8 49.2

76 11.9 10.7 16.1









#44
2-Chloronaphthalene
Concen: 21.427 ng/ul
RT: 13.340 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:162 Resp: 601301
Ion Ratio Lower Upper
162 100
164 32.1 25.7 38.5
127 36.0 30.3 45.5



#45
2-Nitroaniline
Concen: 19.735 ng/ul
RT: 13.551 min Scan# 1779
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion: 65 Resp: 150137
Ion Ratio Lower Upper
65 100
92 75.0 56.9 85.3
138 117.7 88.2 132.2









#47
Dimethylphthalate
Concen: 20.441 ng/ul
RT: 13.922 min Scan# 1842
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:163 Resp: 737765
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.2 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.559 ng/ul
RT: 14.045 min Scan# 1863
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:165 Resp: 147012
Ion Ratio Lower Upper
165 100
89 55.1 46.6 70.0
121 20.6 17.3 25.9









#50
Acenaphthylene
Concen: 20.876 ng/ul
RT: 14.181 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:152 Resp: 947652
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9
153 13.6 10.6 16.0



#51
3-Nitroaniline
Concen: 21.436 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:138 Resp: 145919
Ion Ratio Lower Upper
138 100
108 10.6 9.5 14.3
92 116.2 93.1 139.7





#52

Acenaphthene

Concen: 20.853 ng/ul

RT: 14.522 min Scan# 1944

Delta R.T. -0.012 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:153 Resp: 625887

Ion Ratio Lower Upper

153 100

152 48.1 38.1 57.1

154 87.3 71.7 107.5



#53

2,4-Dinitrophenol

Concen: 21.620 ng/ul

RT: 14.587 min Scan# 1955

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13




Tgt Ion:184 Resp: 78134


Ion Ratio Lower Upper

184 100

63 54.0 48.2 72.4

154 58.5 49.1 73.7









#55
4-Nitrophenol
Concen: 22.200 ng/ul
RT: 14.693 min Scan# 1973
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:109 Resp: 100945
Ion Ratio Lower Upper
109 100
139 131.9 109.0 163.4
65 122.5 106.2 159.4



#56
Dibenzofuran
Concen: 21.180 ng/ul
RT: 14.863 min Scan# 2002
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:168 Resp: 883507
Ion Ratio Lower Upper
168 100
139 34.7 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 22.206 ng/ul
RT: 14.834 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:165 Resp: 213005
Ion Ratio Lower Upper
165 100
63 38.6 32.7 49.1
182 2.9 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.595 ng/ul
RT: 15.092 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:232 Resp: 197023
Ion Ratio Lower Upper
232 100
131 35.6 33.8 50.8
130 2.1 1.8 2.8
166 27.0 24.7 37.1









#59
Diethylphthalate
Concen: 20.321 ng/ul
RT: 15.287 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:149 Resp: 717709
Ion Ratio Lower Upper
149 100
177 23.0 18.2 27.4
150 12.3 9.9 14.9



#61
Fluorene
Concen: 21.655 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:166 Resp: 714793
Ion Ratio Lower Upper
166 100
165 96.1 78.0 117.0
167 13.2 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 22.119 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:204 Resp: 382920
Ion Ratio Lower Upper
204 100
206 32.9 27.1 40.7
141 51.5 45.8 68.8



#63
4-Nitroaniline
Concen: 22.268 ng/ul
RT: 15.540 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:138 Resp: 144401
Ion Ratio Lower Upper
138 100
92 52.3 42.3 63.5
108 81.7 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 22.122 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:198 Resp: 130481
Ion Ratio Lower Upper
198 100
182 4.7 3.4 5.2
77 21.1 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 19.894 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:169 Resp: 606531
Ion Ratio Lower Upper
169 100
168 67.6 53.1 79.7
167 36.6 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 20.927 ng/ul
RT: 16.404 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:248 Resp: 243082
Ion Ratio Lower Upper
248 100
250 96.5 79.3 118.9
141 59.9 57.2 85.8



#69
Hexachlorobenzene
Concen: 22.182 ng/ul
RT: 16.516 min Scan# 2283
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:284 Resp: 287681
Ion Ratio Lower Upper
284 100
142 29.7 25.8 38.6
249 30.2 24.9 37.3









#70
Atrazine
Concen: 18.509 ng/ul
RT: 16.681 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:200 Resp: 167772
Ion Ratio Lower Upper
200 100
173 25.9 21.0 31.4
215 47.7 38.2 57.2



#71
Pentachlorophenol
Concen: 22.356 ng/ul
RT: 16.863 min Scan# 2342
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:266 Resp: 172433
Ion Ratio Lower Upper
266 100
264 63.1 50.2 75.2
268 63.4 50.5 75.7





#72

Phenanthrene

Concen: 20.686 ng/ul

RT: 17.257 min Scan# 2409

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:178 Resp: 1149487

Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 19.3 15.4 23.2



#74

Anthracene

Concen: 21.216 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13




Tgt Ion:178 Resp: 1181738


Ion Ratio Lower Upper

178 100

179 15.5 12.5 18.7

176 18.7 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.481 ng/uL
RT: 13.257 min Scan# 1729
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:216 Resp: 341747
Ion Ratio Lower Upper
216 100
214 79.4 61.4 92.2
218 48.8 39.2 58.8



#76
Pentachlorobenzene
Concen: 21.583 ng/uL
RT: 14.781 min Scan# 1988
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:250 Resp: 337293
Ion Ratio Lower Upper
250 100
248 61.9 50.6 75.8
252 64.9 51.3 76.9





#77

Carbazole

Concen: 23.578 ng/ul

RT: 17.622 min Scan# 2471

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :





BNA_P

ClientSampleId :

SSTD020736

Tgt Ion:167 Resp: 1046682

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.1	25.7
139	12.2	10.2	15.4



#78

Di-n-butylphthalate

Concen: 20.923 ng/ul

RT: 18.175 min Scan# 2565




Delta R.T. -0.006 min


Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Tgt Ion:149 Resp: 1236783

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.4	11.2
104	4.8	3.9	5.9





#80

Fluoranthene

Concen: 16.033 ng/ul

RT: 19.228 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:202 Resp: 1399788

Ion Ratio Lower Upper

202 100

101 10.0 9.1 13.7

100 7.7 7.1 10.7



#82

Pyrene

Concen: 16.697 ng/ul

RT: 19.580 min Scan# 2804

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13




Tgt Ion:202 Resp: 1470586


Ion Ratio Lower Upper

202 100

101 11.5 11.0 16.4

100 9.3 9.1 13.7





#83

Butylbenzylphthalate

Concen: 17.698 ng/ul

RT: 20.457 min Scan# 2953

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:149 Resp: 570628

Ion Ratio Lower Upper

149 100

91 65.9 56.6 85.0

206 23.5 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 22.140 ng/ul

RT: 21.227 min Scan# 3084

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13




Tgt Ion:252 Resp: 510504


Ion Ratio Lower Upper

252 100

254 64.9 51.0 76.4

126 9.3 7.8 11.6





#85

Benzo(a)anthracene

Concen: 20.378 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:228 Resp: 1569409

Ion Ratio Lower Upper

228 100

229 19.5 15.7 23.5

226 27.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.821 ng/ul

RT: 21.216 min Scan# 3082

Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13




Tgt Ion:149 Resp: 834077


Ion Ratio Lower Upper

149 100

167 27.4 22.2 33.2

279 4.7 3.3 4.9





#87

Chrysene

Concen: 20.596 ng/ul

RT: 21.339 min Scan# 3103

Delta R.T. -0.012 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:228 Resp: 1460700

Ion Ratio Lower Upper

228 100

226 30.0 23.9 35.9

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 17.284 ng/ul




RT: 22.127 min Scan# 3237


Delta R.T. -0.006 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Tgt Ion:149 Resp: 1453315









#90
Benzo(b)fluoranthene
Concen: 19.921 ng/ul
RT: 22.951 min Scan# 3377
Delta R.T. -0.006 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:252 Resp: 1691563
Ion Ratio Lower Upper
252 100
253 22.1 17.3 25.9
125 8.1 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 19.189 ng/ul
RT: 22.998 min Scan# 3385
Delta R.T. -0.012 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:252 Resp: 1576487
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 8.2 8.0 12.0





#93

Benzo(a)pyrene

Concen: 20.140 ng/ul

RT: 23.563 min Scan# 3481

Delta R.T. -0.017 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:252 Resp: 1562570

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.8

125 9.7 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.654 ng/ul

RT: 26.127 min Scan# 3917

Delta R.T. -0.029 min

Lab File: BP018951.D

Acq: 20 Dec 2023 09:13




Tgt Ion:276 Resp: 2007335


Ion Ratio Lower Upper

276 100

138 20.6 19.4 29.2

277 25.5 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 21.681 ng/ul
RT: 26.145 min Scan# 3920
Delta R.T. -0.029 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Instrument :
BNA_P
ClientSampleId :
SSTD020736

Tgt Ion:278 Resp: 1671888
Ion Ratio Lower Upper
278 100
139 13.2 13.0 19.6
279 23.8 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 21.786 ng/ul
RT: 26.886 min Scan# 4046
Delta R.T. -0.029 min
Lab File: BP018951.D
Acq: 20 Dec 2023 09:13

Tgt Ion:276 Resp: 1625306
Ion Ratio Lower Upper
276 100
138 17.6 18.1 27.1#
277 23.7 19.4 29.0

