

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122023\
 Data File : BP018981.D
 Acq On : 21 Dec 2023 02:48
 Operator : MA/JU
 Sample : PB157888BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS888

Quant Time: Dec 21 03:20:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 19 05:28:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	113716	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.616	136	454906	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.457	164	312761	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.210	188	758983	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.298	240	808453	20.000	ng/u1	#-0.02
88) Perylene-d12	23.663	264	926646	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	18111	5.427	ng/uL	0.00
4) Pyridine-d5	3.670	84	257415	28.799	ng/u1	-0.01
7) Phenol-d5	6.999	99	323556	28.353	ng/u1	-0.02
9) Bis-(2-Chloroethyl)eth...	7.164	67	180972	26.757	ng/u1	0.00
11) 2-Chlorophenol-d4	7.358	132	257979	28.971	ng/u1	-0.01
15) 4-Methylphenol-d8	8.540	113	260039	28.157	ng/u1	-0.02
21) Nitrobenzene-d5	8.987	128	126294	31.193	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.705	143	143915	34.284	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.246	165	287005	33.437	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.763	131	331282	29.533	ng/u1	-0.01
46) Dimethylphthalate-d6	13.875	166	882540	31.616	ng/u1	0.00
49) Acenaphthylene-d8	14.151	160	967400	31.473	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.675	143	153757	33.557	ng/u1	-0.01
60) Fluorene-d10	15.451	176	755093	32.226	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	167010	36.698	ng/u1	-0.01
73) Anthracene-d10	17.310	188	1223562	30.665	ng/u1	0.00
81) Pyrene-d10	19.545	212	1562454	24.828	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.510	264	1713683	31.717	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.287	88	39245	10.724	ng/uL	97
5) Pyridine	3.687	79	271083	30.032	ng/u1	99
6) Benzaldehyde	6.969	77	156036	26.483	ng/u1	95
8) Phenol	7.028	94	344562	30.789	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.252	93	266732	29.021	ng/u1	98
12) 2-Chlorophenol	7.387	128	276019	30.627	ng/u1	97
13) 2-Methylphenol	8.275	108	260925	30.454	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.352	45	312422	26.790	ng/u1	98
16) Acetophenone	8.652	105	421316	30.611	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.640	70	203290	26.848	ng/u1	98
18) 4-Methylphenol	8.605	108	280410	29.973	ng/u1	99
19) Hexachloroethane	8.893	117	114119	30.888	ng/u1	88
22) Nitrobenzene	9.028	77	311898	31.684	ng/u1	98
23) Isophorone	9.558	82	603998	30.093	ng/u1	97
25) 2-Nitrophenol	9.740	139	158648	35.299	ng/u1	95
26) 2,4-Dimethylphenol	9.805	107	266304	30.296	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.040	93	365154	28.817	ng/u1	99
29) 2,4-Dichlorophenol	10.269	162	291847	35.300	ng/u1	98
30) Naphthalene	10.669	128	876909	31.776	ng/u1	99
32) 4-Chloroaniline	10.787	127	275146	25.735	ng/u1	99
33) Hexachlorobutadiene	10.952	225	219774	39.692	ng/u1	98
34) Caprolactam	11.575	113	93458	36.866	ng/u1	91
35) 4-Chloro-3-methylphenol	11.916	107	300894	32.488	ng/u1	100
36) 2-Methylnaphthalene	12.281	142	620918	31.760	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122023\
 Data File : BP018981.D
 Acq On : 21 Dec 2023 02:48
 Operator : MA/JU
 Sample : PB157888BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS888

Quant Time: Dec 21 03:20:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 19 05:28:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.499	142	610083	30.900	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.652	216	408543	36.003	ng/ul	97
40) Hexachlorocyclopentadiene	12.622	237	206028	30.632	ng/ul	99
41) 2,4,6-Trichlorophenol	12.893	196	248619	34.059	ng/ul	99
42) 2,4,5-Trichlorophenol	12.969	196	276291	34.912	ng/ul	98
43) 1,1'-Biphenyl	13.293	154	857924	31.717	ng/ul	98
44) 2-Chloronaphthalene	13.334	162	698431	32.692	ng/ul	98
45) 2-Nitroaniline	13.546	65	189402	32.703	ng/ul	98
47) Dimethylphthalate	13.922	163	926247	33.709	ng/ul	99
48) 2,6-Dinitrotoluene	14.046	165	190488	36.693	ng/ul	93
50) Acenaphthylene	14.181	152	1116480	32.307	ng/ul	99
51) 3-Nitroaniline	14.375	138	175579	33.881	ng/ul	97
52) Acenaphthene	14.522	153	742997	32.517	ng/ul	99
53) 2,4-Dinitrophenol	14.581	184	113677	41.317	ng/ul	94
55) 4-Nitrophenol	14.687	109	138571	40.029	ng/ul	82
56) Dibenzofuran	14.857	168	1070391	33.705	ng/ul	100
57) 2,4-Dinitrotoluene	14.834	165	278708	38.167	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.087	232	243141	36.627	ng/ul	92
59) Diethylphthalate	15.287	149	918605	34.164	ng/ul	99
61) Fluorene	15.510	166	874526	34.802	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.504	204	483670	36.699	ng/ul	97
63) 4-Nitroaniline	15.540	138	195400	39.580	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.593	198	186321	38.294	ng/ul	94
67) N-Nitrosodiphenylamine	15.722	169	758755	30.169	ng/ul	99
68) 4-Bromophenyl-phenylether	16.398	248	330497	34.492	ng/ul	95
69) Hexachlorobenzene	16.510	284	396645	37.076	ng/ul	95
70) Atrazine	16.681	200	163245	21.832	ng/ul	98
71) Pentachlorophenol	16.857	266	222997	35.049	ng/ul	99
72) Phenanthrene	17.251	178	1488320	32.468	ng/ul	100
74) Anthracene	17.345	178	1486053	32.342	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.257	216	414002	31.547	ng/ul	99
76) Pentachlorobenzene	14.775	250	438725	34.033	ng/ul	99
77) Carbazole	17.622	167	1337224	36.516	ng/ul	99
78) Di-n-butylphthalate	18.175	149	1626001	33.346	ng/ul	100
80) Fluoranthene	19.222	202	1905376	25.614	ng/ul#	94
82) Pyrene	19.575	202	1968099	26.226	ng/ul#	92
83) Butylbenzylphthalate	20.451	149	779939	28.389	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.222	252	705349	35.903	ng/ul	100
85) Benzo(a)anthracene	21.286	228	2104287	32.067	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.210	149	1153067	30.537	ng/ul	99
87) Chrysene	21.339	228	1962210	32.471	ng/ul	100
89) Di-n-octyl phthalate	22.122	149	2035731	30.089	ng/ul	100
90) Benzo(b)fluoranthene	22.945	252	2145763	31.406	ng/ul#	97
91) Benzo(k)fluoranthene	22.992	252	2189465	33.120	ng/ul#	97
93) Benzo(a)pyrene	23.557	252	2040204	32.681	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.121	276	2584734	34.653	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.133	278	2143823	34.551	ng/ul#	95
96) Benzo(g,h,i)perylene	26.874	276	2077128	34.603	ng/ul#	93





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.724 ng/uL
RT: 3.287 min Scan# 34
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion: 88 Resp: 39245
Ion Ratio Lower Upper
88 100
43 25.8 21.8 32.6
58 67.7 56.5 84.7



#5
Pyridine
Concen: 30.032 ng/uL
RT: 3.687 min Scan# 102
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion: 79 Resp: 271083
Ion Ratio Lower Upper
79 100
52 71.4 57.4 86.0
51 34.7 26.2 39.4









#6
Benzaldehyde
Concen: 26.483 ng/ul
RT: 6.969 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion: 77 Resp: 156036
Ion Ratio Lower Upper
77 100
105 102.8 77.9 116.9
106 98.2 74.6 111.8



#8
Phenol
Concen: 30.789 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion: 94 Resp: 344562
Ion Ratio Lower Upper
94 100
65 27.5 21.0 31.6
66 38.2 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 29.021 ng/ul
RT: 7.252 min Scan# 708
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion: 93 Resp: 266732
Ion Ratio Lower Upper
93 100
63 71.0 58.3 87.5
95 32.8 25.9 38.9



#12
2-Chlorophenol
Concen: 30.627 ng/ul
RT: 7.387 min Scan# 731
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:128 Resp: 276019
Ion Ratio Lower Upper
128 100
64 44.0 37.0 55.4
130 32.5 26.8 40.2









#13
2-Methylphenol
Concen: 30.454 ng/ul
RT: 8.275 min Scan# 881
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:108 Resp: 260925
Ion Ratio Lower Upper
108 100
107 89.6 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 26.790 ng/ul
RT: 8.352 min Scan# 895
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion: 45 Resp: 312422
Ion Ratio Lower Upper
45 100
77 17.4 12.7 19.1
79 13.0 10.5 15.7









#16
Acetophenone
Concen: 30.611 ng/ul
RT: 8.652 min Scan# 945
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:105 Resp: 421316
Ion Ratio Lower Upper
105 100
77 76.0 59.4 89.0
51 23.8 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 26.848 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion: 70 Resp: 203290
Ion Ratio Lower Upper
70 100
42 49.7 41.0 61.4
101 10.5 7.8 11.8
130 22.8 16.9 25.3









#18
4-Methylphenol
Concen: 29.973 ng/ul
RT: 8.605 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:108 Resp: 280410
Ion Ratio Lower Upper
108 100
107 112.7 90.6 136.0



#19
Hexachloroethane
Concen: 30.888 ng/ul
RT: 8.893 min Scan# 987
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:117 Resp: 114119
Ion Ratio Lower Upper
117 100
201 118.8 83.7 125.5
199 72.5 52.3 78.5









#22
Nitrobenzene
Concen: 31.684 ng/ul
RT: 9.028 min Scan# 1100
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion: 77 Resp: 311898
Ion Ratio Lower Upper
77 100
123 46.2 35.8 53.8
65 13.1 10.5 15.7



#23
Isophorone
Concen: 30.093 ng/ul
RT: 9.558 min Scan# 1100
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion: 82 Resp: 603998
Ion Ratio Lower Upper
82 100
95 8.0 5.5 8.3
138 18.2 13.5 20.3









#25
2-Nitrophenol
Concen: 35.299 ng/ul
RT: 9.740 min Scan# 1131
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:139 Resp: 158648
Ion Ratio Lower Upper
139 100
65 47.0 39.1 58.7
109 35.0 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 30.296 ng/ul
RT: 9.805 min Scan# 1142
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:107 Resp: 266304
Ion Ratio Lower Upper
107 100
121 49.4 40.0 60.0
122 84.3 69.1 103.7









#27
Bis(2-Chloroethoxy)methane
Concen: 28.817 ng/ul
RT: 10.040 min Scan# 1182
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion: 93 Resp: 365154
Ion Ratio Lower Upper
93 100
95 31.7 26.1 39.1
123 11.4 8.9 13.3



#29
2,4-Dichlorophenol
Concen: 35.300 ng/ul
RT: 10.269 min Scan# 1221
Delta R.T. -0.017 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion: 162 Resp: 291847
Ion Ratio Lower Upper
162 100
164 64.8 51.7 77.5
98 34.0 29.2 43.8









#30
Naphthalene
Concen: 31.776 ng/ul
RT: 10.669 min Scan# 1289
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:128 Resp: 876909
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.5 10.5 15.7



#32
4-Chloroaniline
Concen: 25.735 ng/ul
RT: 10.787 min Scan# 1309
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:127 Resp: 275146
Ion Ratio Lower Upper
127 100
129 32.7 25.6 38.4









#33
Hexachlorobutadiene
Concen: 39.692 ng/ul
RT: 10.952 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:225 Resp: 219774
Ion Ratio Lower Upper
225 100
223 63.4 49.5 74.3
227 64.7 50.9 76.3



#34
Caprolactam
Concen: 36.866 ng/ul
RT: 11.575 min Scan# 1443
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:113 Resp: 93458
Ion Ratio Lower Upper
113 100
55 149.9 133.7 200.5
56 117.1 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 32.488 ng/ul
RT: 11.916 min Scan# 1501
Delta R.T. -0.017 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:107 Resp: 300894
Ion Ratio Lower Upper
107 100
144 26.0 21.0 31.6
142 79.8 63.6 95.4



#36
2-Methylnaphthalene
Concen: 31.760 ng/ul
RT: 12.281 min Scan# 1563
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:142 Resp: 620918
Ion Ratio Lower Upper
142 100
141 88.4 71.4 107.0









#37
1-Methylnaphthalene
Concen: 30.900 ng/ul
RT: 12.499 min Scan# 1600
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:142 Resp: 610083
Ion Ratio Lower Upper
142 100
141 93.9 73.5 110.3
116 4.0 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 36.003 ng/ul
RT: 12.652 min Scan# 1626
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:216 Resp: 408543
Ion Ratio Lower Upper
216 100
214 77.9 63.9 95.9
179 19.2 16.3 24.5
108 13.1 12.6 19.0









#40
Hexachlorocyclopentadiene
Concen: 30.632 ng/ul
RT: 12.622 min Scan# 1621
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:237 Resp: 206028
Ion Ratio Lower Upper
237 100
235 62.5 49.4 74.0
272 13.4 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 34.059 ng/ul
RT: 12.893 min Scan# 1667
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:196 Resp: 248619
Ion Ratio Lower Upper
196 100
198 97.6 76.9 115.3
200 30.8 25.0 37.6









#42
 2,4,5-Trichlorophenol
 Concen: 34.912 ng/ul
 RT: 12.969 min Scan# 1680
 Delta R.T. -0.012 min
 Lab File: BP018981.D
 Acq: 21 Dec 2023 02:48




Instrument :
 BNA_P
 ClientSampleId :
 SLCS888


Tgt Ion:	196	Resp:	276291
Ion Ratio	Lower	Upper	
196	100		
198	99.5	77.9	116.9
200	32.2	25.2	37.8



#43
 1,1'-Biphenyl
 Concen: 31.717 ng/ul
 RT: 13.293 min Scan# 1735
 Delta R.T. -0.012 min
 Lab File: BP018981.D
 Acq: 21 Dec 2023 02:48

Tgt Ion:	154	Resp:	857924
Ion Ratio	Lower	Upper	
154	100		
153	40.4	32.8	49.2
76	11.3	10.7	16.1









#44
2-Chloronaphthalene
Concen: 32.692 ng/ul
RT: 13.334 min Scan# 1742
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:162 Resp: 698431
Ion Ratio Lower Upper
162 100
164 33.2 25.7 38.5
127 37.0 30.3 45.5



#45
2-Nitroaniline
Concen: 32.703 ng/ul
RT: 13.546 min Scan# 1778
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion: 65 Resp: 189402
Ion Ratio Lower Upper
65 100
92 72.9 56.9 85.3
138 111.9 88.2 132.2





#47

Dimethylphthalate

Concen: 33.709 ng/ul

RT: 13.922 min Scan# 1842

Delta R.T. -0.006 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:163 Resp: 926247

Ion Ratio Lower Upper

163 100

194 4.2 3.2 4.8

164 10.4 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 36.693 ng/ul

RT: 14.046 min Scan# 1863

Delta R.T. -0.006 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:165 Resp: 190488


Ion Ratio Lower Upper

165 100

89 52.6 46.6 70.0

121 18.4 17.3 25.9









#50
Acenaphthylene
Concen: 32.307 ng/ul
RT: 14.181 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument : BNA_P
ClientSampleId : SLCS888


Tgt Ion:152 Resp: 1116480
Ion Ratio Lower Upper
152 100
151 20.6 15.9 23.9
153 13.2 10.6 16.0



#51
3-Nitroaniline
Concen: 33.881 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:138 Resp: 175579
Ion Ratio Lower Upper
138 100
108 11.0 9.5 14.3
92 113.0 93.1 139.7









#52
Acenaphthene
Concen: 32.517 ng/ul
RT: 14.522 min Scan# 1944
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:153 Resp: 742997
Ion Ratio Lower Upper
153 100
152 47.5 38.1 57.1
154 87.7 71.7 107.5



#53
2,4-Dinitrophenol
Concen: 41.317 ng/ul
RT: 14.581 min Scan# 1954
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:184 Resp: 113677
Ion Ratio Lower Upper
184 100
63 53.8 48.2 72.4
154 63.7 49.1 73.7









#55
4-Nitrophenol
Concen: 40.029 ng/ul
RT: 14.687 min Scan# 1973
Delta R.T. -0.017 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:109 Resp: 138571
Ion Ratio Lower Upper
109 100
139 116.9 109.0 163.4
65 110.3 106.2 159.4



#56
Dibenzofuran
Concen: 33.705 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:168 Resp: 1070391
Ion Ratio Lower Upper
168 100
139 36.5 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 38.167 ng/ul
RT: 14.834 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:165 Resp: 278708
Ion Ratio Lower Upper
165 100
63 38.5 32.7 49.1
182 2.8 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 36.627 ng/ul
RT: 15.087 min Scan# 2040
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:232 Resp: 243141
Ion Ratio Lower Upper
232 100
131 36.4 33.8 50.8
130 2.5 1.8 2.8
166 27.7 24.7 37.1





#59

Diethylphthalate

Concen: 34.164 ng/ul

RT: 15.287 min Scan# 2074

Delta R.T. -0.006 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:149 Resp: 918605

Ion Ratio Lower Upper

149 100

177 23.2 18.2 27.4

150 12.5 9.9 14.9



#61

Fluorene

Concen: 34.802 ng/ul

RT: 15.510 min Scan# 2112

Delta R.T. -0.006 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:166 Resp: 874526


Ion Ratio Lower Upper

166 100

165 98.6 78.0 117.0

167 13.1 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 36.699 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:204 Resp: 483670
Ion Ratio Lower Upper
204 100
206 33.8 27.1 40.7
141 53.7 45.8 68.8



#63
4-Nitroaniline
Concen: 39.580 ng/ul
RT: 15.540 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:138 Resp: 195400
Ion Ratio Lower Upper
138 100
92 52.1 42.3 63.5
108 89.5 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 38.294 ng/ul
RT: 15.593 min Scan# 2127
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:198 Resp: 186321
Ion Ratio Lower Upper
198 100
182 3.9 3.4 5.2
77 22.5 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 30.169 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:169 Resp: 758755
Ion Ratio Lower Upper
169 100
168 67.0 53.1 79.7
167 36.1 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 34.492 ng/ul
RT: 16.398 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:248 Resp: 330497
Ion Ratio Lower Upper
248 100
250 97.5 79.3 118.9
141 64.3 57.2 85.8



#69
Hexachlorobenzene
Concen: 37.076 ng/ul
RT: 16.510 min Scan# 2282
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:284 Resp: 396645
Ion Ratio Lower Upper
284 100
142 27.4 25.8 38.6
249 29.9 24.9 37.3





#70

Atrazine

Concen: 21.832 ng/ul

RT: 16.681 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:200 Resp: 163245

Ion Ratio Lower Upper

200 100

173 23.7 21.0 31.4

215 47.4 38.2 57.2



#71

Pentachlorophenol

Concen: 35.049 ng/ul

RT: 16.857 min Scan# 2341

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:266 Resp: 222997


Ion Ratio Lower Upper

266 100

264 63.5 50.2 75.2

268 64.1 50.5 75.7





#72

Phenanthrene

Concen: 32.468 ng/ul

RT: 17.251 min Scan# 2409

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:178 Resp: 1488320

Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 19.0 15.4 23.2



#74

Anthracene

Concen: 32.342 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. -0.006 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:178 Resp: 1486053


Ion Ratio Lower Upper

178 100

179 14.9 12.5 18.7

176 18.9 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 31.547 ng/uL

RT: 13.257 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:216 Resp: 414002

Ion Ratio Lower Upper

216 100

214 78.2 61.4 92.2

218 48.1 39.2 58.8



#76

Pentachlorobenzene

Concen: 34.033 ng/uL

RT: 14.775 min Scan# 1987

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:250 Resp: 438725


Ion Ratio Lower Upper

250 100

248 64.1 50.6 75.8

252 65.0 51.3 76.9









#77
 Carbazole
 Concen: 36.516 ng/ul
 RT: 17.622 min Scan# 2471
 Delta R.T. -0.006 min
 Lab File: BP018981.D
 Acq: 21 Dec 2023 02:48




Instrument :
 BNA_P
 ClientSampleId :
 SLCS888


Tgt Ion:167 Resp: 1337224
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.1 25.7
 139 12.2 10.2 15.4



#78
 Di-n-butylphthalate
 Concen: 33.346 ng/ul
 RT: 18.175 min Scan# 2565
 Delta R.T. -0.006 min
 Lab File: BP018981.D
 Acq: 21 Dec 2023 02:48

Tgt Ion:149 Resp: 1626001
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 4.9 3.9 5.9









#80
Fluoranthene
Concen: 25.614 ng/u1
RT: 19.222 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48




Instrument :
BNA_P
ClientSampleId :
SLCS888


Tgt Ion:202 Resp: 1905376
Ion Ratio Lower Upper
202 100
101 8.9 9.1 13.7#
100 6.8 7.1 10.7#



#82
Pyrene
Concen: 26.226 ng/u1
RT: 19.575 min Scan# 2803
Delta R.T. -0.012 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:202 Resp: 1968099
Ion Ratio Lower Upper
202 100
101 10.2 11.0 16.4#
100 8.8 9.1 13.7#





#83

Butylbenzylphthalate

Concen: 28.389 ng/ul

RT: 20.451 min Scan# 2953

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:149 Resp: 779939

Ion Ratio Lower Upper

149 100

91 67.8 56.6 85.0

206 22.9 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 35.903 ng/ul

RT: 21.222 min Scan# 3083

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:252 Resp: 705349


Ion Ratio Lower Upper

252 100

254 63.8 51.0 76.4

126 8.9 7.8 11.6





#85

Benzo(a)anthracene

Concen: 32.067 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:228 Resp: 2104287

Ion Ratio Lower Upper

228 100

229 19.8 15.7 23.5

226 27.5 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 30.537 ng/ul

RT: 21.210 min Scan# 3081

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:149 Resp: 1153067


Ion Ratio Lower Upper

149 100

167 27.4 22.2 33.2

279 4.8 3.3 4.9





#87

Chrysene

Concen: 32.471 ng/ul

RT: 21.339 min Scan# 3104

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:228 Resp: 1962210

Ion Ratio Lower Upper

228 100

226 29.9 23.9 35.9

229 19.3 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 30.089 ng/ul




RT: 22.122 min Scan# 3236


Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Tgt Ion:149 Resp: 2035731





#90

Benzo(b)fluoranthene

Concen: 31.406 ng/ul

RT: 22.945 min Scan# 3376

Delta R.T. -0.012 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:252 Resp: 2145763

Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 7.2 7.8 11.6#



#91

Benzo(k)fluoranthene

Concen: 33.120 ng/ul

RT: 22.992 min Scan# 3384

Delta R.T. -0.017 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:252 Resp: 2189465


Ion Ratio Lower Upper

252 100

253 21.5 17.7 26.5

125 7.4 8.0 12.0#





#93

Benzo(a)pyrene

Concen: 32.681 ng/ul

RT: 23.557 min Scan# 34

Delta R.T. -0.023 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48

Instrument :

BNA_P

ClientSampleId :

SLCS888





Tgt Ion:252 Resp: 2040204

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.8

125 8.5 9.2 13.8#



#94

Indeno(1,2,3-cd)pyrene

Concen: 34.653 ng/ul

RT: 26.121 min Scan# 3916

Delta R.T. -0.035 min

Lab File: BP018981.D

Acq: 21 Dec 2023 02:48




Tgt Ion:276 Resp: 2584734


Ion Ratio Lower Upper

276 100

138 18.3 19.4 29.2#

277 25.7 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 34.551 ng/ul
RT: 26.133 min Scan# 3918
Delta R.T. -0.041 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Instrument :
BNA_P
ClientSampleId :
SLCS888

Tgt Ion:278 Resp: 2143823
Ion Ratio Lower Upper
278 100
139 12.0 13.0 19.6#
279 24.4 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 34.603 ng/ul
RT: 26.874 min Scan# 4044
Delta R.T. -0.041 min
Lab File: BP018981.D
Acq: 21 Dec 2023 02:48

Tgt Ion:276 Resp: 2077128
Ion Ratio Lower Upper
276 100
138 16.2 18.1 27.1#
277 23.8 19.4 29.0

