

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122023\
 Data File : BP018986.D
 Acq On : 21 Dec 2023 06:11
 Operator : MA/JU
 Sample : PB157899BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK899

Quant Time: Dec 21 06:41:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 21 04:46:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	130605	20.000	ng/u1	0.00
20) Naphthalene-d8	10.616	136	463280	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	281457	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.204	188	671799	20.000	ng/u1	0.00
79) Chrysene-d12	21.298	240	859687	20.000	ng/u1	# 0.00
88) Perylene-d12	23.657	264	1013164	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	21888	5.711	ng/uL	0.00
4) Pyridine-d5	3.669	84	288176	28.071	ng/u1	0.00
7) Phenol-d5	6.999	99	334400	25.514	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.157	67	195703	25.193	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	272937	26.688	ng/u1	0.00
15) 4-Methylphenol-d8	8.540	113	257115	24.240	ng/u1	0.00
21) Nitrobenzene-d5	8.987	128	125869	30.526	ng/u1	0.00
24) 2-Nitrophenol-d4	9.704	143	140466	32.858	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.240	165	255879	29.272	ng/u1	0.00
31) 4-Chloroaniline-d4	10.763	131	334716	29.300	ng/u1	0.00
46) Dimethylphthalate-d6	13.875	166	756310	30.108	ng/u1	0.00
49) Acenaphthylene-d8	14.151	160	834860	30.182	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	122242	29.646	ng/u1	0.00
60) Fluorene-d10	15.451	176	649138	30.785	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	119958	29.780	ng/u1	0.00
73) Anthracene-d10	17.304	188	1105793	31.310	ng/u1	0.00
81) Pyrene-d10	19.545	212	1498183	22.388	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.510	264	1866676	31.599	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122023\
Data File : BP018986.D
Acq On : 21 Dec 2023 06:11
Operator : MA/JU
Sample : PB157899BL
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK899

Quant Time: Dec 21 06:41:21 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 21 04:46:12 2023
Response via : Initial Calibration

