

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122123\
 Data File : BP019025.D
 Acq On : 22 Dec 2023 06:22
 Operator : MA/JU
 Sample : 05808-15
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0B09

Quant Time: Dec 22 08:53:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 21 04:46:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	143185	20.000	ng/u1	0.00
20) Naphthalene-d8	10.616	136	550773	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	368367	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.204	188	877681	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.298	240	915959	20.000	ng/u1	# 0.00
88) Perylene-d12	23.657	264	1081305	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	15121	3.599	ng/uL	0.00
4) Pyridine-d5	3.675	84	67660	6.012	ng/u1	0.00
7) Phenol-d5	6.999	99	316351	22.016	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.158	67	171327	20.117	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	244036	21.765	ng/u1	0.00
15) 4-Methylphenol-d8	8.534	113	234832	20.194	ng/u1	0.00
21) Nitrobenzene-d5	8.981	128	121424	24.770	ng/u1	0.00
24) 2-Nitrophenol-d4	9.704	143	143421	28.220	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.240	165	279264	26.872	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	294252	21.666	ng/u1	0.00
46) Dimethylphthalate-d6	13.869	166	925372	28.146	ng/u1	0.00
49) Acenaphthylene-d8	14.145	160	1019907	28.173	ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	163958	30.381	ng/u1	0.00
60) Fluorene-d10	15.445	176	817370	29.618	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	147228	27.976	ng/u1	0.00
73) Anthracene-d10	17.304	188	1306776	28.321	ng/u1	0.00
81) Pyrene-d10	19.539	212	1714451	24.046	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.504	264	1814877	28.786	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122123\
Data File : BP019025.D
Acq On : 22 Dec 2023 06:22
Operator : MA/JU
Sample : 05808-15
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0B09

Quant Time: Dec 22 08:53:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 21 04:46:12 2023
Response via : Initial Calibration

