

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122123\
 Data File : BP019057.D
 Acq On : 23 Dec 2023 01:41
 Operator : MA/JU
 Sample : 05835-05
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0B37

Quant Time: Dec 23 02:39:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 21 04:46:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	126120	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	496087	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	326428	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.204	188	772413	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.292	240	809937	20.000	ng/u1	0.00
88) Perylene-d12	23.651	264	927022	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	20039	5.414	ng/uL	0.00
4) Pyridine-d5	3.670	84	168492	16.996	ng/u1	0.00
7) Phenol-d5	6.993	99	339889	26.855	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	195292	26.034	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.346	132	266397	26.974	ng/u1	0.00
15) 4-Methylphenol-d8	8.534	113	266745	26.042	ng/u1	0.00
21) Nitrobenzene-d5	8.981	128	128714	29.152	ng/u1	0.00
24) 2-Nitrophenol-d4	9.699	143	146109	31.918	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.240	165	271454	29.000	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	302391	24.719	ng/u1	0.00
46) Dimethylphthalate-d6	13.869	166	838728	28.789	ng/u1	0.00
49) Acenaphthylene-d8	14.145	160	948672	29.572	ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	125902	26.327	ng/u1	0.00
60) Fluorene-d10	15.445	176	756634	30.940	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	111858	24.152	ng/u1	0.00
73) Anthracene-d10	17.298	188	1207390	29.733	ng/u1	0.00
81) Pyrene-d10	19.539	212	1529165	24.254	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.498	264	1608303	29.755	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122123\
Data File : BP019057.D
Acq On : 23 Dec 2023 01:41
Operator : MA/JU
Sample : 05835-05
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0B37

Quant Time: Dec 23 02:39:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 21 04:46:12 2023
Response via : Initial Calibration

