

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122123\
 Data File : BP019068.D
 Acq On : 23 Dec 2023 08:28
 Operator : MA/JU
 Sample : 05869-05
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0B36

Quant Time: Dec 24 23:58:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 21 04:46:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	211893	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	825295	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.445	164	542677	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.198	188	1260187	20.000	ng/u1	-0.01
79) Chrysene-d12	21.292	240	1285980	20.000	ng/u1	# 0.00
88) Perylene-d12	23.651	264	1493945	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	21091	3.392	ng/uL	0.00
4) Pyridine-d5	3.670	84	194419	11.673	ng/u1	0.00
7) Phenol-d5	6.993	99	400074	18.814	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	228464	18.128	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.346	132	309005	18.623	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	327313	19.020	ng/u1	-0.01
21) Nitrobenzene-d5	8.975	128	151689	20.651	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.693	143	174417	22.903	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.234	165	344657	22.133	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.752	131	425783	20.922	ng/u1	-0.01
46) Dimethylphthalate-d6	13.869	166	1185004	24.466	ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	1284275	24.080	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.663	143	200863	25.265	ng/u1	0.00
60) Fluorene-d10	15.439	176	1034277	25.440	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.569	200	175521	23.229	ng/u1	0.00
73) Anthracene-d10	17.298	188	1712671	25.852	ng/u1	0.00
81) Pyrene-d10	19.539	212	2120582	21.184	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.498	264	2247953	25.807	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122123\
Data File : BP019068.D
Acq On : 23 Dec 2023 08:28
Operator : MA/JU
Sample : 05869-05
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0B36

Quant Time: Dec 24 23:58:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 21 04:46:12 2023
Response via : Initial Calibration

