

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
 Data File : BP013247.D
 Acq On : 22 Dec 2022 15:43
 Operator : CG/JU
 Sample : N6015-06
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Dec 22 21:47:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 23:38:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	520291	20.000	ng/ul	0.00
20) Naphthalene-d8	10.763	136	1979962	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.587	164	1078759	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.345	188	2160387	20.000	ng/ul	0.00
79) Chrysene-d12	21.433	240	2234392	20.000	ng/ul	0.00
88) Perylene-d12	23.910	264	2591114	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	68919	5.664	ng/uL	0.00
4) Pyridine-d5	3.764	84	760101	21.989	ng/ul	0.00
7) Phenol-d5	7.105	99	1274520	30.074	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	820455	32.093	ng/ul	0.00
11) 2-Chlorophenol-d4	7.475	132	1066174	31.959	ng/ul	0.00
15) 4-Methylphenol-d8	8.658	113	943852	27.181	ng/ul	0.00
21) Nitrobenzene-d5	9.116	128	501350	34.463	ng/ul	0.00
24) 2-Nitrophenol-d4	9.840	143	568756	34.728	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	1003557	31.548	ng/ul	0.00
31) 4-Chloroaniline-d4	10.899	131	997902	22.221	ng/ul	0.00
46) Dimethylphthalate-d6	13.993	166	2601337	31.376	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	3061374	33.604	ng/ul	0.00
54) 4-Nitrophenol-d4	14.787	143	335779	22.515	ng/ul	0.00
60) Fluorene-d10	15.581	176	2234217	31.854	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	313943	22.582	ng/ul	0.00
73) Anthracene-d10	17.445	188	3245517	33.299	ng/ul	0.00
81) Pyrene-d10	19.675	212	4033597	31.450	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.751	264	4375550	33.879	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
Data File : BP013247.D
Acq On : 22 Dec 2022 15:43
Operator : CG/JU
Sample : N6015-06
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Dec 22 21:47:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 21 23:38:37 2022
Response via : Initial Calibration

