

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
 Data File : BPCCC252.D
 Acq On : 22 Dec 2022 19:00
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Dec 22 21:49:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 23:38:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	641606	20.000	ng/ul	0.00
20) Naphthalene-d8	10.769	136	2742003	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.598	164	1700133	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.357	188	3446084	20.000	ng/ul	0.01
79) Chrysene-d12	21.439	240	2357064	20.000	ng/ul	0.00
88) Perylene-d12	23.921	264	2680877	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	117427	7.825	ng/uL	0.00
4) Pyridine-d5	3.764	84	893187	20.953	ng/ul	0.00
7) Phenol-d5	7.110	99	1116855	21.371	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	681572	21.620	ng/ul	0.00
11) 2-Chlorophenol-d4	7.481	132	894767	21.749	ng/ul	0.00
15) 4-Methylphenol-d8	8.669	113	899820	21.014	ng/ul	0.00
21) Nitrobenzene-d5	9.122	128	442975	21.988	ng/ul	0.00
24) 2-Nitrophenol-d4	9.846	143	509596	22.468	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	966289	21.934	ng/ul	0.00
31) 4-Chloroaniline-d4	10.904	131	1238962	19.921	ng/ul	0.00
46) Dimethylphthalate-d6	14.004	166	2621302	20.062	ng/ul	0.00
49) Acenaphthylene-d8	14.292	160	3076547	21.428	ng/ul	0.00
54) 4-Nitrophenol-d4	14.798	143	422092	17.959	ng/ul	0.01
60) Fluorene-d10	15.592	176	2255862	20.407	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	403713	18.205	ng/ul	0.00
73) Anthracene-d10	17.457	188	3274178	21.060	ng/ul	0.01
81) Pyrene-d10	19.686	212	3185637	23.546	ng/ul	0.01
92) Benzo(a)pyrene-d12	23.763	264	2784516	20.838	ng/ul	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.370	88	123578	7.889	ng/uL	94
5) Pyridine	3.781	79	892913	21.076	ng/ul	94
6) Benzaldehyde	7.087	77	437159	21.363	ng/ul	95
8) Phenol	7.140	94	1135232	21.484	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.369	93	908045	20.938	ng/ul	97
12) 2-Chlorophenol	7.510	128	910862	21.591	ng/ul	96
13) 2-Methylphenol	8.399	108	873531	21.205	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.487	45	1354980	22.768	ng/ul	98
16) Acetophenone	8.781	105	1403695	21.379	ng/ul	96
17) N-Nitroso-di-n-propyla...	8.769	70	711538	21.551	ng/ul	96
18) 4-Methylphenol	8.728	108	954804	20.925	ng/ul	99
19) Hexachloroethane	9.034	117	349282	20.718	ng/ul	94
22) Nitrobenzene	9.163	77	1054605	22.308	ng/ul	97
23) Isophorone	9.693	82	1975122	20.662	ng/ul	97
25) 2-Nitrophenol	9.875	139	531338	21.621	ng/ul	94
26) 2,4-Dimethylphenol	9.940	107	1036732	21.621	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.175	93	1241205	20.307	ng/ul	100
29) 2,4-Dichlorophenol	10.416	162	950508	22.026	ng/ul	98
30) Naphthalene	10.816	128	2983232	20.803	ng/ul	100
32) 4-Chloroaniline	10.928	127	1232537	20.027	ng/ul	100
33) Hexachlorobutadiene	11.099	225	648604	21.890	ng/ul	98
34) Caprolactam	11.716	113	220908	15.013	ng/ul	94
35) 4-Chloro-3-methylphenol	12.051	107	930494	21.128	ng/ul	99
36) 2-Methylnaphthalene	12.422	142	2015178	20.484	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
 Data File : BPCCC252.D
 Acq On : 22 Dec 2022 19:00
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Dec 22 21:49:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 23:38:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.640	142	2065404	20.415	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.787	216	1234369	22.548	ng/ul	98
40) Hexachlorocyclopentadiene	12.763	237	348752	9.789	ng/ul	99
41) 2,4,6-Trichlorophenol	13.028	196	791343	22.492	ng/ul	99
42) 2,4,5-Trichlorophenol	13.104	196	852586	22.560	ng/ul	100
43) 1,1'-Biphenyl	13.428	154	2741093	21.467	ng/ul	99
44) 2-Chloronaphthalene	13.475	162	2181084	21.704	ng/ul	100
45) 2-Nitroaniline	13.681	65	584035	23.970	ng/ul	95
47) Dimethylphthalate	14.051	163	2602085	20.093	ng/ul	100
48) 2,6-Dinitrotoluene	14.175	165	525148	21.820	ng/ul	100
50) Acenaphthylene	14.322	152	3280600	21.253	ng/ul	100
51) 3-Nitroaniline	14.504	138	473234	20.824	ng/ul	96
52) Acenaphthene	14.663	153	2243261	20.972	ng/ul	99
53) 2,4-Dinitrophenol	14.710	184	241916	14.705	ng/ul	93
55) 4-Nitrophenol	14.810	109	313848	19.570	ng/ul	94
56) Dibenzofuran	14.992	168	3166634	20.822	ng/ul	100
57) 2,4-Dinitrotoluene	14.963	165	723832	20.705	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.222	232	711730	20.637	ng/ul	98
59) Diethylphthalate	15.416	149	2480711	19.705	ng/ul	99
61) Fluorene	15.645	166	2532970	20.767	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.639	204	1352317	20.966	ng/ul	99
63) 4-Nitroaniline	15.669	138	399004	19.953	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.728	198	398782	18.401	ng/ul	98
67) N-Nitrosodiphenylamine	15.851	169	2134439	22.341	ng/ul	100
68) 4-Bromophenyl-phenylether	16.539	248	824722	22.037	ng/ul	99
69) Hexachlorobenzene	16.657	284	969797	21.824	ng/ul	98
70) Atrazine	16.810	200	789990	21.219	ng/ul	99
71) Pentachlorophenol	17.004	266	529658	19.682	ng/ul	97
72) Phenanthrene	17.398	178	3743199	20.854	ng/ul	100
74) Anthracene	17.492	178	3781075	21.097	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.398	216	1200810	24.424	ng/uL	100
76) Pentachlorobenzene	14.916	250	1176796	23.476	ng/uL	99
77) Carbazole	17.757	167	3132550	20.754	ng/ul	99
78) Di-n-butylphthalate	18.298	149	3907088	21.260	ng/ul	100
80) Fluoranthene	19.363	202	3833042	24.961	ng/ul	99
82) Pyrene	19.716	202	3840131	24.243	ng/ul	98
83) Butylbenzylphthalate	20.580	149	1374754	22.465	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.351	252	830999	15.606	ng/ul	100
85) Benzo(a)anthracene	21.427	228	3344233	21.372	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.333	149	1802450	20.881	ng/ul	99
87) Chrysene	21.480	228	3160285	21.357	ng/ul	99
89) Di-n-octyl phthalate	22.286	149	2740444	18.246	ng/ul	100
90) Benzo(b)fluoranthene	23.163	252	3449484	20.178	ng/ul	100
91) Benzo(k)fluoranthene	23.210	252	3409702	20.114	ng/ul	100
93) Benzo(a)pyrene	23.815	252	3086490	20.735	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.521	276	4286716	23.120	ng/ul	100
95) Dibenzo(a,h)anthracene	26.539	278	3528615	22.986	ng/ul	99
96) Benzo(g,h,i)perylene	27.321	276	3503631	23.538	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 22 21:49:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 21 23:38:37 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.889 ng/uL
 RT: 3.370 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: BPCCC252.D
 Acq: 22 Dec 2022 19:00




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 88 Resp: 123578
 Ion Ratio Lower Upper
 88 100
 43 32.1 22.4 33.6
 58 77.5 58.2 87.4



#5
 Pyridine
 Concen: 21.076 ng/uL
 RT: 3.781 min Scan# 101
 Delta R.T. -0.000 min
 Lab File: BPCCC252.D
 Acq: 22 Dec 2022 19:00

Tgt Ion: 79 Resp: 892913
 Ion Ratio Lower Upper
 79 100
 52 81.8 60.7 91.1
 51 38.5 28.9 43.3









#6
Benzaldehyde
Concen: 21.363 ng/ul
RT: 7.087 min Scan# 663
Delta R.T. -0.000 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 437159
Ion Ratio Lower Upper
77 100
105 95.5 81.5 122.3
106 93.0 76.9 115.3



#8
Phenol
Concen: 21.484 ng/ul
RT: 7.140 min Scan# 672
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00


Tgt Ion: 94 Resp: 1135232
Ion Ratio Lower Upper
94 100
65 28.2 21.1 31.7
66 35.3 27.0 40.6








#10
 Bis(2-Chloroethyl)ether
 Concen: 20.938 ng/ul
 RT: 7.369 min Scan# 711
 Delta R.T. -0.000 min
 Lab File: BPCCC252.D
 Acq: 22 Dec 2022 19:00

Instrument :
 BNA_P
 ClientSampleId :



Tgt Ion: 93 Resp: 908045




Ion	Ratio	Lower	Upper
93	100		
63	77.9	59.8	89.8
95	32.7	26.4	39.6




#12
 2-Chlorophenol
 Concen: 21.591 ng/ul
 RT: 7.510 min Scan# 735
 Delta R.T. -0.000 min
 Lab File: BPCCC252.D
 Acq: 22 Dec 2022 19:00

Tgt Ion:128 Resp: 910862

Ion	Ratio	Lower	Upper
128	100		
64	47.3	34.9	52.3
130	31.4	25.8	38.6









#13
2-Methylphenol
Concen: 21.205 ng/ul
RT: 8.399 min Scan# 885
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 873531
Ion Ratio Lower Upper
108 100
107 87.1 69.1 103.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 22.768 ng/ul
RT: 8.487 min Scan# 901
Delta R.T. 0.012 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion: 45 Resp: 1354980
Ion Ratio Lower Upper
45 100
77 13.6 11.0 16.6
79 10.0 9.3 13.9









#16
Acetophenone
Concen: 21.379 ng/ul
RT: 8.781 min Scan# 910
Delta R.T. -0.000 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 1403695
Ion Ratio Lower Upper
105 100
77 76.2 59.3 88.9
51 28.6 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 21.551 ng/ul
RT: 8.769 min Scan# 949
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion: 70 Resp: 711538
Ion Ratio Lower Upper
70 100
42 61.6 45.5 68.3
101 9.9 7.9 11.9
130 21.7 17.4 26.2





#18

4-Methylphenol

Concen: 20.925 ng/ul

RT: 8.728 min Scan# 941

Delta R.T. 0.006 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 954804

Ion Ratio Lower Upper

108 100

107 110.1 87.4 131.2



#19

Hexachloroethane

Concen: 20.718 ng/ul

RT: 9.034 min Scan# 994

Delta R.T. -0.000 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00




Tgt Ion:117 Resp: 349282


Ion Ratio Lower Upper

117 100

201 111.2 93.7 140.5

199 68.3 58.6 88.0









#22
Nitrobenzene
Concen: 22.308 ng/ul
RT: 9.163 min Scan# 1015
Delta R.T. -0.000 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 1054605
Ion Ratio Lower Upper
77 100
123 45.1 38.3 57.5
65 13.3 10.3 15.5



#23
Isophorone
Concen: 20.662 ng/ul
RT: 9.693 min Scan# 1106
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion: 82 Resp: 1975122
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 17.6 15.3 22.9









#25
2-Nitrophenol
Concen: 21.621 ng/ul
RT: 9.875 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 531338
Ion Ratio Lower Upper
139 100
65 51.1 36.6 54.8
109 28.1 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 21.621 ng/ul
RT: 9.940 min Scan# 1148
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:107 Resp: 1036732
Ion Ratio Lower Upper
107 100
121 50.9 41.8 62.6
122 86.3 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.307 ng/ul
 RT: 10.175 min Scan# 1187
 Delta R.T. 0.006 min
 Lab File: BPCCC252.D
 Acq: 22 Dec 2022 19:00




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 93 Resp: 1241205
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.0 39.0
 123 11.5 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 22.026 ng/ul
 RT: 10.416 min Scan# 1229
 Delta R.T. 0.006 min
 Lab File: BPCCC252.D
 Acq: 22 Dec 2022 19:00

Tgt Ion:162 Resp: 950508
 Ion Ratio Lower Upper
 162 100
 164 64.6 52.5 78.7
 98 31.9 26.6 39.8









#30
Naphthalene
Concen: 20.803 ng/ul
RT: 10.816 min Scan# 1296
Delta R.T. -0.000 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 2983232
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 20.027 ng/ul
RT: 10.928 min Scan# 1316
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00


Tgt Ion:127 Resp: 1232537
Ion Ratio Lower Upper
127 100
129 31.9 25.6 38.4








#33
Hexachlorobutadiene
Concen: 21.890 ng/ul
RT: 11.099 min Scan# 11
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00


Instrument :
BNA_P
ClientSampleId :





Tgt Ion:225 Resp: 648604
Ion Ratio Lower Upper
225 100
223 62.6 49.2 73.8
227 64.5 50.2 75.2




#34
Caprolactam
Concen: 15.013 ng/ul
RT: 11.716 min Scan# 1450
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00



Tgt Ion:113 Resp: 220908
Ion Ratio Lower Upper
113 100
55 181.7 141.9 212.9
56 136.4 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 21.128 ng/ul
RT: 12.051 min Scan# 1507
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 930494
Ion Ratio Lower Upper
107 100
144 26.9 21.4 32.2
142 81.4 65.8 98.6



#36
2-Methylnaphthalene
Concen: 20.484 ng/ul
RT: 12.422 min Scan# 1570
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:142 Resp: 2015178
Ion Ratio Lower Upper
142 100
141 88.2 70.8 106.2









#37
1-Methylnaphthalene
Concen: 20.415 ng/ul
RT: 12.640 min Scan# 1606
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 2065404
Ion Ratio Lower Upper
142 100
141 91.4 72.7 109.1
116 3.9 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.548 ng/ul
RT: 12.787 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:216 Resp: 1234369
Ion Ratio Lower Upper
216 100
214 78.8 61.9 92.9
179 19.6 15.0 22.6
108 14.9 11.0 16.4









#40
Hexachlorocyclopentadiene
Concen: 9.789 ng/ul
RT: 12.763 min Scan# 1628
Delta R.T. -0.000 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 348752
Ion Ratio Lower Upper
237 100
235 62.4 49.7 74.5
272 13.0 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 22.492 ng/ul
RT: 13.028 min Scan# 1673
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:196 Resp: 791343
Ion Ratio Lower Upper
196 100
198 95.8 77.0 115.6
200 30.8 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 22.560 ng/ul
RT: 13.104 min Scan# 1
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:196 Resp: 852586
Ion Ratio Lower Upper
196 100
198 96.6 77.1 115.7
200 31.6 25.4 38.2



#43
1,1'-Biphenyl
Concen: 21.467 ng/ul
RT: 13.428 min Scan# 1741
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:154 Resp: 2741093
Ion Ratio Lower Upper
154 100
153 40.1 31.5 47.3
76 12.7 9.6 14.4





#44

2-Chloronaphthalene

Concen: 21.704 ng/ul

RT: 13.475 min Scan# 1

Delta R.T. 0.006 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:162 Resp: 2181084

Ion Ratio Lower Upper

162 100

164 32.9 26.3 39.5

127 36.3 29.3 43.9



#45

2-Nitroaniline

Concen: 23.970 ng/ul

RT: 13.681 min Scan# 1784

Delta R.T. 0.006 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00




Tgt Ion: 65 Resp: 584035


Ion Ratio Lower Upper

65 100

92 68.3 56.4 84.6

138 106.1 90.3 135.5









#47
Dimethylphthalate
Concen: 20.093 ng/ul
RT: 14.051 min Scan# 1847
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 2602085
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.2 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 21.820 ng/ul
RT: 14.175 min Scan# 1868
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:165 Resp: 525148
Ion Ratio Lower Upper
165 100
89 53.5 43.0 64.4
121 20.4 16.1 24.1









#50
Acenaphthylene
Concen: 21.253 ng/ul
RT: 14.322 min Scan# 1893
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 3280600
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.0 10.6 16.0



#51
3-Nitroaniline
Concen: 20.824 ng/ul
RT: 14.504 min Scan# 1924
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:138 Resp: 473234
Ion Ratio Lower Upper
138 100
108 10.5 8.6 13.0
92 115.5 88.4 132.6









#52
Acenaphthene
Concen: 20.972 ng/ul
RT: 14.663 min Scan# 1951
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 2243261
Ion Ratio Lower Upper
153 100
152 47.2 38.2 57.4
154 87.1 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 14.705 ng/ul
RT: 14.710 min Scan# 1959
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:184 Resp: 241916
Ion Ratio Lower Upper
184 100
63 66.9 48.4 72.6
154 60.5 45.5 68.3









#55
4-Nitrophenol
Concen: 19.570 ng/ul
RT: 14.810 min Scan# 1976
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 313848
Ion Ratio Lower Upper
109 100
139 138.4 120.8 181.2
65 134.1 107.9 161.9



#56
Dibenzofuran
Concen: 20.822 ng/ul
RT: 14.992 min Scan# 2007
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:168 Resp: 3166634
Ion Ratio Lower Upper
168 100
139 35.0 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 20.705 ng/ul
RT: 14.963 min Scan# 2000
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 723832
Ion Ratio Lower Upper
165 100
63 41.5 31.8 47.6
182 2.9 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.637 ng/ul
RT: 15.222 min Scan# 2046
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:232 Resp: 711730
Ion Ratio Lower Upper
232 100
131 38.2 31.8 47.6
130 2.1 1.8 2.6
166 27.6 22.9 34.3









#59
Diethylphthalate
Concen: 19.705 ng/ul
RT: 15.416 min Scan# 2079
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 2480711
Ion Ratio Lower Upper
149 100
177 23.4 18.4 27.6
150 12.2 9.9 14.9



#61
Fluorene
Concen: 20.767 ng/ul
RT: 15.645 min Scan# 2118
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:166 Resp: 2532970
Ion Ratio Lower Upper
166 100
165 98.4 77.4 116.2
167 13.7 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.966 ng/ul
RT: 15.639 min Scan# 2117
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 1352317
Ion Ratio Lower Upper
204 100
206 33.4 26.3 39.5
141 49.8 40.4 60.6



#63
4-Nitroaniline
Concen: 19.953 ng/ul
RT: 15.669 min Scan# 2122
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:138 Resp: 399004
Ion Ratio Lower Upper
138 100
92 51.1 39.0 58.4
108 74.8 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.401 ng/ul
RT: 15.728 min Scan# 2132
Delta R.T. 0.012 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 398782
Ion Ratio Lower Upper
198 100
182 4.0 3.3 4.9
77 22.0 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 22.341 ng/ul
RT: 15.851 min Scan# 2153
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:169 Resp: 2134439
Ion Ratio Lower Upper
169 100
168 66.3 53.0 79.6
167 36.4 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 22.037 ng/ul
RT: 16.539 min Scan# 21
Delta R.T. 0.012 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 824722
Ion Ratio Lower Upper
248 100
250 97.5 77.4 116.0
141 60.1 48.3 72.5



#69
Hexachlorobenzene
Concen: 21.824 ng/ul
RT: 16.657 min Scan# 2290
Delta R.T. 0.012 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:284 Resp: 969797
Ion Ratio Lower Upper
284 100
142 27.0 22.6 34.0
249 29.5 24.6 37.0





#70

Atrazine

Concen: 21.219 ng/ul

RT: 16.810 min Scan# 21

Delta R.T. 0.012 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 789990

Ion Ratio Lower Upper

200 100

173 25.2 20.5 30.7

215 50.9 39.9 59.9



#71

Pentachlorophenol

Concen: 19.682 ng/ul

RT: 17.004 min Scan# 2349

Delta R.T. 0.012 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00




Tgt Ion:266 Resp: 529658


Ion Ratio Lower Upper

266 100

264 61.5 50.9 76.3

268 62.4 51.4 77.2





#72

Phenanthrene

Concen: 20.854 ng/ul

RT: 17.398 min Scan# 2414

Delta R.T. 0.006 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 3743199

Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 19.1 15.4 23.0



#74

Anthracene

Concen: 21.097 ng/ul

RT: 17.492 min Scan# 2432

Delta R.T. 0.012 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00




Tgt Ion:178 Resp: 3781075


Ion Ratio Lower Upper

178 100

179 15.7 12.5 18.7

176 19.0 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 24.424 ng/uL
RT: 13.398 min Scan# 1736
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 1200810
Ion Ratio Lower Upper
216 100
214 77.6 62.2 93.2
218 48.5 38.9 58.3



#76
Pentachlorobenzene
Concen: 23.476 ng/uL
RT: 14.916 min Scan# 1994
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:250 Resp: 1176796
Ion Ratio Lower Upper
250 100
248 62.3 50.6 76.0
252 62.9 51.1 76.7









#77
 Carbazole
 Concen: 20.754 ng/ul
 RT: 17.757 min Scan# 2476
 Delta R.T. 0.006 min
 Lab File: BPCCC252.D
 Acq: 22 Dec 2022 19:00




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:167 Resp: 3132550
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.2 25.8
 139 12.4 9.4 14.2



#78
 Di-n-butylphthalate
 Concen: 21.260 ng/ul
 RT: 18.298 min Scan# 2569
 Delta R.T. 0.006 min
 Lab File: BPCCC252.D
 Acq: 22 Dec 2022 19:00

Tgt Ion:149 Resp: 3907088
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 5.0 3.8 5.6









#80
Fluoranthene
Concen: 24.961 ng/ul
RT: 19.363 min Scan# 21
Delta R.T. 0.012 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 3833042
Ion Ratio Lower Upper
202 100
101 9.5 8.2 12.2
100 7.3 6.1 9.1



#82
Pyrene
Concen: 24.243 ng/ul
RT: 19.716 min Scan# 2810
Delta R.T. 0.012 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:202 Resp: 3840131
Ion Ratio Lower Upper
202 100
101 10.9 9.4 14.2
100 9.2 7.8 11.8





#83

Butylbenzylphthalate

Concen: 22.465 ng/ul

RT: 20.580 min Scan# 2955

Delta R.T. 0.012 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 1374754

Ion Ratio Lower Upper

149 100

91 65.5 53.4 80.0

206 23.8 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 15.606 ng/ul

RT: 21.351 min Scan# 3088

Delta R.T. 0.006 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00




Tgt Ion:252 Resp: 830999


Ion Ratio Lower Upper

252 100

254 64.8 51.8 77.8

126 10.0 7.4 11.2









#85
Benzo(a)anthracene
Concen: 21.372 ng/ul
RT: 21.427 min Scan# 3101
Delta R.T. 0.012 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 3344233
Ion Ratio Lower Upper
228 100
229 19.5 16.1 24.1
226 27.3 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.881 ng/ul
RT: 21.333 min Scan# 3085
Delta R.T. 0.006 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:149 Resp: 1802450
Ion Ratio Lower Upper
149 100
167 28.4 22.5 33.7
279 4.7 3.8 5.8





#87

Chrysene

Concen: 21.357 ng/ul

RT: 21.480 min Scan# 3110

Delta R.T. 0.012 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 3160285

Ion Ratio Lower Upper

228 100

226 30.1 24.5 36.7

229 19.8 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 18.246 ng/ul




RT: 22.286 min Scan# 3247


Delta R.T. 0.012 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Tgt Ion:149 Resp: 2740444





#90

Benzo(b)fluoranthene

Concen: 20.178 ng/ul

RT: 23.163 min Scan# 31

Delta R.T. 0.012 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 3449484

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 8.3 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.114 ng/ul

RT: 23.210 min Scan# 3404

Delta R.T. 0.012 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00




Tgt Ion:252 Resp: 3409702


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 8.5 6.8 10.2









#93
Benzo(a)pyrene
Concen: 20.735 ng/ul
RT: 23.815 min Scan# 311
Delta R.T. 0.018 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 3086490
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 9.4 7.5 11.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 23.120 ng/ul
RT: 26.521 min Scan# 3967
Delta R.T. 0.018 min
Lab File: BPCCC252.D
Acq: 22 Dec 2022 19:00

Tgt Ion:276 Resp: 4286716
Ion Ratio Lower Upper
276 100
138 21.2 17.1 25.7
277 24.9 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 22.986 ng/ul

RT: 26.539 min Scan# 3966

Delta R.T. 0.018 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 3528615

Ion Ratio Lower Upper

278 100

139 13.2 11.1 16.7

279 23.8 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 23.538 ng/ul

RT: 27.321 min Scan# 4103

Delta R.T. 0.023 min

Lab File: BPCCC252.D

Acq: 22 Dec 2022 19:00

Tgt Ion:276 Resp: 3503631

Ion Ratio Lower Upper

276 100

138 17.9 15.4 23.2

277 23.8 19.1 28.7

