

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
 Data File : BP013261.D
 Acq On : 23 Dec 2022 00:43
 Operator : CG/JU
 Sample : N6015-20
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DBXZ9

Quant Time: Dec 23 03:12:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 22 22:43:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	650507	20.000	ng/u1	0.00
20) Naphthalene-d8	10.757	136	2622343	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.587	164	1507289	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.345	188	2742138	20.000	ng/u1	-0.01
79) Chrysene-d12	21.433	240	2422170	20.000	ng/u1	0.00
88) Perylene-d12	23.916	264	2918876	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	59706	3.924	ng/uL	0.00
4) Pyridine-d5	3.764	84	763492	17.666	ng/u1	0.00
7) Phenol-d5	7.105	99	1146889	21.645	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	747397	23.383	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.475	132	974958	23.374	ng/u1	0.00
15) 4-Methylphenol-d8	8.658	113	767236	17.672	ng/u1	-0.01
21) Nitrobenzene-d5	9.116	128	479827	24.904	ng/u1	0.00
24) 2-Nitrophenol-d4	9.840	143	548087	25.268	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	1013398	24.053	ng/u1	0.00
31) 4-Chloroaniline-d4	10.899	131	1013329	17.037	ng/u1	0.00
46) Dimethylphthalate-d6	13.993	166	2894805	24.989	ng/u1	-0.01
49) Acenaphthylene-d8	14.281	160	3291357	25.857	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.787	143	313398	15.040	ng/u1	-0.01
60) Fluorene-d10	15.581	176	2397052	24.459	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.704	200	255773	14.494	ng/u1	0.00
73) Anthracene-d10	17.445	188	3108373	25.126	ng/u1	-0.01
81) Pyrene-d10	19.680	212	3494956	25.138	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.757	264	3748644	25.766	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
Data File : BP013261.D
Acq On : 23 Dec 2022 00:43
Operator : CG/JU
Sample : N6015-20
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DBXZ9

Quant Time: Dec 23 03:12:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 22 22:43:06 2022
Response via : Initial Calibration

