

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
 Data File : BP013276.D
 Acq On : 23 Dec 2022 13:09
 Operator : CG/JU
 Sample : N6015-21
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DBYA4

Quant Time: Dec 24 00:24:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 00:19:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	683107	20.000	ng/u1	0.00
20) Naphthalene-d8	10.757	136	2882365	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.592	164	1813679	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.351	188	3475155	20.000	ng/u1	0.00
79) Chrysene-d12	21.439	240	2554324	20.000	ng/u1	0.00
88) Perylene-d12	23.927	264	2953118	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	112566	7.046	ng/uL	0.00
4) Pyridine-d5	3.758	84	1775746	39.127	ng/u1	0.00
7) Phenol-d5	7.110	99	2516791	45.233	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	1484416	44.226	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	2002423	45.716	ng/u1	0.00
15) 4-Methylphenol-d8	8.657	113	1992094	43.695	ng/u1	0.00
21) Nitrobenzene-d5	9.116	128	1030383	48.655	ng/u1	0.00
24) 2-Nitrophenol-d4	9.840	143	1215711	50.991	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	2206512	47.647	ng/u1	0.00
31) 4-Chloroaniline-d4	10.898	131	2608789	39.904	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	6535698	46.888	ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	7456318	48.681	ng/u1	0.00
54) 4-Nitrophenol-d4	14.792	143	938619	37.435	ng/u1	0.00
60) Fluorene-d10	15.586	176	5485197	46.515	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	830921	37.155	ng/u1	0.00
73) Anthracene-d10	17.451	188	7430606	47.395	ng/u1	0.00
81) Pyrene-d10	19.680	212	7433623	50.701	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.768	264	7244608	49.218	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
Data File : BP013276.D
Acq On : 23 Dec 2022 13:09
Operator : CG/JU
Sample : N6015-21
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DBYA4

Quant Time: Dec 24 00:24:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Dec 24 00:19:17 2022
Response via : Initial Calibration

