

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
 Data File : BP013274.D
 Acq On : 23 Dec 2022 12:00
 Operator : CG/JU
 Sample : N6018-06MS
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DBYH7MS

Quant Time: Dec 24 00:23:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 00:19:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	651905	20.000	ng/u1	0.00
20) Naphthalene-d8	10.757	136	2735600	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.593	164	1702333	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.351	188	3353279	20.000	ng/u1	0.00
79) Chrysene-d12	21.439	240	2378330	20.000	ng/u1	0.00
88) Perylene-d12	23.927	264	2762464	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	65123	4.271	ng/uL	0.00
4) Pyridine-d5	3.758	84	988430	22.821	ng/u1	0.00
7) Phenol-d5	7.111	99	1489083	28.043	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	893178	27.884	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	1209405	28.933	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	1134527	26.076	ng/u1	0.00
21) Nitrobenzene-d5	9.116	128	610520	30.375	ng/u1	0.00
24) 2-Nitrophenol-d4	9.840	143	716078	31.646	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	1353046	30.785	ng/u1	0.00
31) 4-Chloroaniline-d4	10.899	131	1467088	23.645	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	3733851	28.539	ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	4327102	30.099	ng/u1	0.00
54) 4-Nitrophenol-d4	14.798	143	581330	24.702	ng/u1	0.00
60) Fluorene-d10	15.587	176	3194399	28.860	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	571229	26.471	ng/u1	0.00
73) Anthracene-d10	17.451	188	4488498	29.670	ng/u1	0.00
81) Pyrene-d10	19.681	212	4379867	32.083	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.769	264	4166810	30.262	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.370	88	186616	11.725	ng/uL	94
5) Pyridine	3.781	79	1409249	32.739	ng/u1	99
6) Benzaldehyde	7.081	77	1160046	55.793	ng/u1	96
8) Phenol	7.134	94	1994921	37.158	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.364	93	1520560	34.507	ng/u1	97
12) 2-Chlorophenol	7.511	128	1550217	36.165	ng/u1	97
13) 2-Methylphenol	8.393	108	1519521	36.303	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.481	45	2266065	37.476	ng/u1	99
16) Acetophenone	8.775	105	2347303	35.186	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.763	70	1193774	35.585	ng/u1	95
18) 4-Methylphenol	8.728	108	1657929	35.761	ng/u1	99
19) Hexachloroethane	9.028	117	583728	34.077	ng/u1	97
22) Nitrobenzene	9.158	77	1783179	37.808	ng/u1	98
23) Isophorone	9.687	82	3404591	35.700	ng/u1	98
25) 2-Nitrophenol	9.869	139	927171	37.816	ng/u1	94
26) 2,4-Dimethylphenol	9.934	107	1780178	37.213	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.169	93	2089337	34.263	ng/u1	99
29) 2,4-Dichlorophenol	10.410	162	1628082	37.815	ng/u1	100
30) Naphthalene	10.810	128	5034412	35.189	ng/u1	100
32) 4-Chloroaniline	10.922	127	1867420	30.414	ng/u1	99
33) Hexachlorobutadiene	11.093	225	1096439	37.091	ng/u1	99
34) Caprolactam	11.722	113	483652	32.946	ng/u1	97
35) 4-Chloro-3-methylphenol	12.046	107	1650685	37.569	ng/u1	98
36) 2-Methylnaphthalene	12.416	142	3476124	35.417	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122222\
 Data File : BP013274.D
 Acq On : 23 Dec 2022 12:00
 Operator : CG/JU
 Sample : N6018-06MS
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DBYH7MS

Quant Time: Dec 24 00:23:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 00:19:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.634	142	3442908	34.110	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.781	216	2105433	38.410	ng/ul	99
40) Hexachlorocyclopentadiene	12.757	237	569596	15.968	ng/ul	99
41) 2,4,6-Trichlorophenol	13.022	196	1396187	39.633	ng/ul	98
42) 2,4,5-Trichlorophenol	13.099	196	1488340	39.332	ng/ul	99
43) 1,1'-Biphenyl	13.422	154	4647626	36.352	ng/ul	99
44) 2-Chloronaphthalene	13.469	162	3695753	36.728	ng/ul	99
45) 2-Nitroaniline	13.675	65	1038320	42.560	ng/ul	97
47) Dimethylphthalate	14.046	163	4497061	34.680	ng/ul	100
48) 2,6-Dinitrotoluene	14.169	165	956033	39.672	ng/ul	98
50) Acenaphthylene	14.316	152	5584299	36.131	ng/ul	100
51) 3-Nitroaniline	14.498	138	912683	40.109	ng/ul	98
52) Acenaphthene	14.657	153	3836416	35.819	ng/ul	100
53) 2,4-Dinitrophenol	14.704	184	471503	28.623	ng/ul	98
55) 4-Nitrophenol	14.810	109	548853	34.179	ng/ul	96
56) Dibenzofuran	14.987	168	5376506	35.307	ng/ul	100
57) 2,4-Dinitrotoluene	14.957	165	1295563	37.011	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.216	232	1259038	36.459	ng/ul	99
59) Diethylphthalate	15.410	149	4338877	34.420	ng/ul	99
61) Fluorene	15.640	166	4265274	34.925	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.634	204	2301100	35.630	ng/ul	99
63) 4-Nitroaniline	15.669	138	892732	44.584	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.722	198	692421	32.835	ng/ul	98
67) N-Nitrosodiphenylamine	15.851	169	3651289	39.275	ng/ul	100
68) 4-Bromophenyl-phenylether	16.534	248	1460070	40.094	ng/ul	98
69) Hexachlorobenzene	16.651	284	1651946	38.204	ng/ul	99
70) Atrazine	16.810	200	1369781	37.810	ng/ul	99
71) Pentachlorophenol	16.998	266	935264	35.716	ng/ul	99
72) Phenanthrene	17.392	178	6365508	36.445	ng/ul	100
74) Anthracene	17.487	178	6263787	35.917	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.387	216	2144055	44.815	ng/ul	99
76) Pentachlorobenzene	14.910	250	1970583	40.400	ng/ul	100
77) Carbazole	17.757	167	5269405	35.878	ng/ul	100
78) Di-n-butylphthalate	18.298	149	6940489	38.812	ng/ul	100
80) Fluoranthene	19.357	202	6438186	41.552	ng/ul	99
82) Pyrene	19.710	202	6335141	39.637	ng/ul	99
83) Butylbenzylphthalate	20.575	149	2374218	38.451	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.351	252	1386248	25.801	ng/ul	100
85) Benzo(a)anthracene	21.422	228	5918759	37.488	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.333	149	3245170	37.258	ng/ul	99
87) Chrysene	21.480	228	5568205	37.293	ng/ul	100
89) Di-n-octyl phthalate	22.286	149	5330570	34.444	ng/ul	100
90) Benzo(b)fluoranthene	23.169	252	6472380	36.742	ng/ul	99
91) Benzo(k)fluoranthene	23.216	252	5941141	34.012	ng/ul	100
93) Benzo(a)pyrene	23.816	252	5602328	36.525	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.527	276	7546749	39.500	ng/ul	98
95) Dibenzo(a,h)anthracene	26.551	278	6304458	39.855	ng/ul	99
96) Benzo(g,h,i)perylene	27.333	276	6236977	40.663	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.725 ng/uL
RT: 3.370 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion: 88 Resp: 186616
Ion Ratio Lower Upper
88 100
43 32.8 22.4 33.6
58 76.9 58.2 87.4



#5
Pyridine
Concen: 32.739 ng/uL
RT: 3.781 min Scan# 101
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion: 79 Resp: 1409249
Ion Ratio Lower Upper
79 100
52 76.4 60.7 91.1
51 37.5 28.9 43.3









#6
Benzaldehyde
Concen: 55.793 ng/ul
RT: 7.081 min Scan# 60
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion: 77 Resp: 1160046
Ion Ratio Lower Upper
77 100
105 96.9 81.5 122.3
106 93.8 76.9 115.3



#8
Phenol
Concen: 37.158 ng/ul
RT: 7.134 min Scan# 671
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion: 94 Resp: 1994921
Ion Ratio Lower Upper
94 100
65 28.3 21.1 31.7
66 36.3 27.0 40.6





#10

Bis(2-Chloroethyl)ether

Concen: 34.507 ng/ul

RT: 7.364 min Scan# 710

Delta R.T. 0.000 min

Lab File: BP013274.D

Acq: 23 Dec 2022 12:00

Instrument :





BNA_P

ClientSampleId :

DBYH7MS

Tgt Ion: 93 Resp: 1520560

Ion	Ratio	Lower	Upper
93	100		
63	77.5	59.8	89.8
95	31.8	26.4	39.6



#12

2-Chlorophenol

Concen: 36.165 ng/ul

RT: 7.511 min Scan# 735




Delta R.T. 0.006 min


Lab File: BP013274.D

Acq: 23 Dec 2022 12:00

Tgt Ion:128 Resp: 1550217

Ion	Ratio	Lower	Upper
128	100		
64	46.5	34.9	52.3
130	32.0	25.8	38.6





#13

2-Methylphenol

Concen: 36.303 ng/ul

RT: 8.393 min Scan# 885

Delta R.T. 0.000 min

Lab File: BP013274.D

Acq: 23 Dec 2022 12:00

Instrument :

BNA_P

ClientSampleId :





DBYH7MS

Tgt Ion:108 Resp: 1519521

Ion Ratio Lower Upper

108 100

107 89.5 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 37.476 ng/ul

RT: 8.481 min Scan# 900

Delta R.T. 0.000 min

Lab File: BP013274.D

Acq: 23 Dec 2022 12:00




Tgt Ion: 45 Resp: 2266065


Ion Ratio Lower Upper

45 100

77 13.7 11.0 16.6

79 10.6 9.3 13.9









#16
Acetophenone
Concen: 35.186 ng/ul
RT: 8.775 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:105 Resp: 2347303
Ion Ratio Lower Upper
105 100
77 75.2 59.3 88.9
51 28.6 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 35.585 ng/ul
RT: 8.763 min Scan# 948
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion: 70 Resp: 1193774
Ion Ratio Lower Upper
70 100
42 62.2 45.5 68.3
101 9.9 7.9 11.9
130 22.1 17.4 26.2





#18

4-Methylphenol

Concen: 35.761 ng/ul

RT: 8.728 min Scan# 941

Delta R.T. 0.006 min

Lab File: BP013274.D

Acq: 23 Dec 2022 12:00

Instrument :

BNA_P

ClientSampleId :





DBYH7MS

Tgt Ion:108 Resp: 1657929

Ion Ratio Lower Upper

108 100

107 110.1 87.4 131.2



#19

Hexachloroethane

Concen: 34.077 ng/ul

RT: 9.028 min Scan# 993

Delta R.T. 0.000 min

Lab File: BP013274.D

Acq: 23 Dec 2022 12:00




Tgt Ion:117 Resp: 583728


Ion Ratio Lower Upper

117 100

201 113.8 93.7 140.5

199 71.5 58.6 88.0









#22
Nitrobenzene
Concen: 37.808 ng/u1
RT: 9.158 min Scan# 1015
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion: 77 Resp: 1783179
Ion Ratio Lower Upper
77 100
123 46.5 38.3 57.5
65 13.4 10.3 15.5



#23
Isophorone
Concen: 35.700 ng/u1
RT: 9.687 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion: 82 Resp: 3404591
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.2
138 18.0 15.3 22.9









#25
2-Nitrophenol
Concen: 37.816 ng/ul
RT: 9.869 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:139 Resp: 927171
Ion Ratio Lower Upper
139 100
65 50.5 36.6 54.8
109 28.3 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 37.213 ng/ul
RT: 9.934 min Scan# 1147
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:107 Resp: 1780178
Ion Ratio Lower Upper
107 100
121 51.7 41.8 62.6
122 88.0 72.6 108.8









#27
Bis(2-Chloroethoxy)methane
Concen: 34.263 ng/ul
RT: 10.169 min Scan# 1187
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion: 93 Resp: 2089337
Ion Ratio Lower Upper
93 100
95 32.7 26.0 39.0
123 11.1 9.2 13.8



#29
2,4-Dichlorophenol
Concen: 37.815 ng/ul
RT: 10.410 min Scan# 1228
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion: 162 Resp: 1628082
Ion Ratio Lower Upper
162 100
164 65.8 52.5 78.7
98 32.6 26.6 39.8









#30
Naphthalene
Concen: 35.189 ng/ul
RT: 10.810 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:128 Resp: 5034412
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.1 10.4 15.6



#32
4-Chloroaniline
Concen: 30.414 ng/ul
RT: 10.922 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:127 Resp: 1867420
Ion Ratio Lower Upper
127 100
129 32.7 25.6 38.4









#33
Hexachlorobutadiene
Concen: 37.091 ng/ul
RT: 11.093 min Scan# 11093
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:225 Resp: 1096439
Ion Ratio Lower Upper
225 100
223 61.4 49.2 73.8
227 63.8 50.2 75.2



#34
Caprolactam
Concen: 32.946 ng/ul
RT: 11.722 min Scan# 1451
Delta R.T. 0.012 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:113 Resp: 483652
Ion Ratio Lower Upper
113 100
55 178.3 141.9 212.9
56 131.8 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 37.569 ng/ul
RT: 12.046 min Scan# 1506
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:107 Resp: 1650685
Ion Ratio Lower Upper
107 100
144 25.9 21.4 32.2
142 80.2 65.8 98.6



#36
2-Methylnaphthalene
Concen: 35.417 ng/ul
RT: 12.416 min Scan# 1569
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:142 Resp: 3476124
Ion Ratio Lower Upper
142 100
141 88.1 70.8 106.2









#37
1-Methylnaphthalene
Concen: 34.110 ng/ul
RT: 12.634 min Scan# 1606
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:142 Resp: 3442908
Ion Ratio Lower Upper
142 100
141 92.1 72.7 109.1
116 3.7 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 38.410 ng/ul
RT: 12.781 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:216 Resp: 2105433
Ion Ratio Lower Upper
216 100
214 78.2 61.9 92.9
179 19.9 15.0 22.6
108 14.4 11.0 16.4









#40
Hexachlorocyclopentadiene
Concen: 15.968 ng/ul
RT: 12.757 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:237 Resp: 569596
Ion Ratio Lower Upper
237 100
235 61.3 49.7 74.5
272 12.7 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 39.633 ng/ul
RT: 13.022 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:196 Resp: 1396187
Ion Ratio Lower Upper
196 100
198 93.9 77.0 115.6
200 30.6 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 39.332 ng/ul
RT: 13.099 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:196 Resp: 1488340
Ion Ratio Lower Upper
196 100
198 97.2 77.1 115.7
200 30.8 25.4 38.2



#43
1,1'-Biphenyl
Concen: 36.352 ng/ul
RT: 13.422 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:154 Resp: 4647626
Ion Ratio Lower Upper
154 100
153 40.0 31.5 47.3
76 12.3 9.6 14.4









#44
2-Chloronaphthalene
Concen: 36.728 ng/ul
RT: 13.469 min Scan# 1747
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:162 Resp: 3695753
Ion Ratio Lower Upper
162 100
164 33.6 26.3 39.5
127 36.5 29.3 43.9



#45
2-Nitroaniline
Concen: 42.560 ng/ul
RT: 13.675 min Scan# 1783
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion: 65 Resp: 1038320
Ion Ratio Lower Upper
65 100
92 68.9 56.4 84.6
138 108.6 90.3 135.5









#47
Dimethylphthalate
Concen: 34.680 ng/ul
RT: 14.046 min Scan# 1846
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:163 Resp: 4497061
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.0
164 10.4 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 39.672 ng/ul
RT: 14.169 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:165 Resp: 956033
Ion Ratio Lower Upper
165 100
89 52.5 43.0 64.4
121 19.3 16.1 24.1









#50
Acenaphthylene
Concen: 36.131 ng/ul
RT: 14.316 min Scan# 1891
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:152 Resp: 5584299
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.6
153 13.2 10.6 16.0



#51
3-Nitroaniline
Concen: 40.109 ng/ul
RT: 14.498 min Scan# 1923
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:138 Resp: 912683
Ion Ratio Lower Upper
138 100
108 10.1 8.6 13.0
92 112.4 88.4 132.6









#52
Acenaphthene
Concen: 35.819 ng/ul
RT: 14.657 min Scan# 1949
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:153 Resp: 3836416
Ion Ratio Lower Upper
153 100
152 47.5 38.2 57.4
154 88.3 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 28.623 ng/ul
RT: 14.704 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:184 Resp: 471503
Ion Ratio Lower Upper
184 100
63 62.3 48.4 72.6
154 58.5 45.5 68.3









#55
4-Nitrophenol
Concen: 34.179 ng/ul
RT: 14.810 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:109 Resp: 548853
Ion Ratio Lower Upper
109 100
139 143.3 120.8 181.2
65 135.9 107.9 161.9



#56
Dibenzofuran
Concen: 35.307 ng/ul
RT: 14.987 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:168 Resp: 5376506
Ion Ratio Lower Upper
168 100
139 35.0 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 37.011 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:165 Resp: 1295563
Ion Ratio Lower Upper
165 100
63 41.2 31.8 47.6
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 36.459 ng/ul
RT: 15.216 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:232 Resp: 1259038
Ion Ratio Lower Upper
232 100
131 38.5 31.8 47.6
130 2.1 1.8 2.6
166 28.7 22.9 34.3









#59
Diethylphthalate
Concen: 34.420 ng/ul
RT: 15.410 min Scan# 2077
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:149 Resp: 4338877
Ion Ratio Lower Upper
149 100
177 23.9 18.4 27.6
150 12.6 9.9 14.9



#61
Fluorene
Concen: 34.925 ng/ul
RT: 15.640 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:166 Resp: 4265274
Ion Ratio Lower Upper
166 100
165 97.4 77.4 116.2
167 13.7 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 35.630 ng/ul
RT: 15.634 min Scan# 2116
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:204 Resp: 2301100
Ion Ratio Lower Upper
204 100
206 33.5 26.3 39.5
141 50.8 40.4 60.6



#63
4-Nitroaniline
Concen: 44.584 ng/ul
RT: 15.669 min Scan# 2122
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:138 Resp: 892732
Ion Ratio Lower Upper
138 100
92 49.1 39.0 58.4
108 73.2 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 32.835 ng/ul
RT: 15.722 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:198 Resp: 692421
Ion Ratio Lower Upper
198 100
182 4.0 3.3 4.9
77 21.8 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 39.275 ng/ul
RT: 15.851 min Scan# 2153
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:169 Resp: 3651289
Ion Ratio Lower Upper
169 100
168 66.4 53.0 79.6
167 36.5 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 40.094 ng/ul
RT: 16.534 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:248 Resp: 1460070
Ion Ratio Lower Upper
248 100
250 97.2 77.4 116.0
141 57.7 48.3 72.5



#69
Hexachlorobenzene
Concen: 38.204 ng/ul
RT: 16.651 min Scan# 2289
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:284 Resp: 1651946
Ion Ratio Lower Upper
284 100
142 27.6 22.6 34.0
249 30.7 24.6 37.0









#70
Atrazine
Concen: 37.810 ng/ul
RT: 16.810 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:200 Resp: 1369781
Ion Ratio Lower Upper
200 100
173 25.1 20.5 30.7
215 50.3 39.9 59.9



#71
Pentachlorophenol
Concen: 35.716 ng/ul
RT: 16.998 min Scan# 2348
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:266 Resp: 935264
Ion Ratio Lower Upper
266 100
264 62.6 50.9 76.3
268 64.6 51.4 77.2









#72
Phenanthrene
Concen: 36.445 ng/u1
RT: 17.392 min Scan# 2414
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:178 Resp: 6365508
Ion Ratio Lower Upper
178 100
179 15.6 12.5 18.7
176 19.4 15.4 23.0



#74
Anthracene
Concen: 35.917 ng/u1
RT: 17.487 min Scan# 2431
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:178 Resp: 6263787
Ion Ratio Lower Upper
178 100
179 15.9 12.5 18.7
176 18.8 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 44.815 ng/uL
RT: 13.387 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:216 Resp: 2144055
Ion Ratio Lower Upper
216 100
214 78.3 62.2 93.2
218 48.1 38.9 58.3



#76
Pentachlorobenzene
Concen: 40.400 ng/uL
RT: 14.910 min Scan# 1993
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:250 Resp: 1970583
Ion Ratio Lower Upper
250 100
248 63.8 50.6 76.0
252 64.1 51.1 76.7









#77
Carbazole
Concen: 35.878 ng/ul
RT: 17.757 min Scan# 2476
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:167 Resp: 5269405
Ion Ratio Lower Upper
167 100
166 21.4 17.2 25.8
139 12.0 9.4 14.2



#78
Di-n-butylphthalate
Concen: 38.812 ng/ul
RT: 18.298 min Scan# 2569
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:149 Resp: 6940489
Ion Ratio Lower Upper
149 100
150 9.2 7.5 11.3
104 4.6 3.8 5.6









#80
Fluoranthene
Concen: 41.552 ng/ul
RT: 19.357 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:202 Resp: 6438186
Ion Ratio Lower Upper
202 100
101 9.9 8.2 12.2
100 7.4 6.1 9.1



#82
Pyrene
Concen: 39.637 ng/ul
RT: 19.710 min Scan# 2809
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:202 Resp: 6335141
Ion Ratio Lower Upper
202 100
101 12.2 9.4 14.2
100 9.8 7.8 11.8









#83
Butylbenzylphthalate
Concen: 38.451 ng/ul
RT: 20.575 min Scan# 2955
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:149 Resp: 2374218
Ion Ratio Lower Upper
149 100
91 70.2 53.4 80.0
206 22.9 17.9 26.9



#84
3,3'-Dichlorobenzidine
Concen: 25.801 ng/ul
RT: 21.351 min Scan# 3088
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:252 Resp: 1386248
Ion Ratio Lower Upper
252 100
254 64.9 51.8 77.8
126 9.9 7.4 11.2









#85
Benzo(a)anthracene
Concen: 37.488 ng/ul
RT: 21.422 min Scan# 3100
Delta R.T. 0.000 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:228 Resp: 5918759
Ion Ratio Lower Upper
228 100
229 20.2 16.1 24.1
226 27.9 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 37.258 ng/ul
RT: 21.333 min Scan# 3085
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:149 Resp: 3245170
Ion Ratio Lower Upper
149 100
167 28.7 22.5 33.7
279 5.0 3.8 5.8









#87
Chrysene
Concen: 37.293 ng/ul
RT: 21.480 min Scan# 3109
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:228 Resp: 5568205
Ion Ratio Lower Upper
228 100
226 30.4 24.5 36.7
229 20.2 16.0 24.0



#89
Di-n-octyl phthalate
Concen: 34.444 ng/ul
RT: 22.286 min Scan# 3247
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:149 Resp: 5330570









#90
Benzo(b)fluoranthene
Concen: 36.742 ng/ul
RT: 23.169 min Scan# 31
Delta R.T. 0.012 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:252 Resp: 6472380
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 8.3 6.8 10.2



#91
Benzo(k)fluoranthene
Concen: 34.012 ng/ul
RT: 23.216 min Scan# 3405
Delta R.T. 0.012 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:252 Resp: 5941141
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 8.6 6.8 10.2









#93
Benzo(a)pyrene
Concen: 36.525 ng/ul
RT: 23.816 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00




Instrument :
BNA_P
ClientSampleId :
DBYH7MS


Tgt Ion:252 Resp: 5602328
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 9.5 7.5 11.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 39.500 ng/ul
RT: 26.527 min Scan# 3968
Delta R.T. 0.012 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:276 Resp: 7546749
Ion Ratio Lower Upper
276 100
138 19.4 17.1 25.7
277 24.9 20.0 30.0









#95
Dibenzo(a,h)anthracene
Concen: 39.855 ng/ul
RT: 26.551 min Scan# 3968
Delta R.T. 0.024 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Instrument :
BNA_P
ClientSampleId :
DBYH7MS

Tgt Ion:278 Resp: 6304458
Ion Ratio Lower Upper
278 100
139 13.4 11.1 16.7
279 24.3 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 40.663 ng/ul
RT: 27.333 min Scan# 4105
Delta R.T. 0.024 min
Lab File: BP013274.D
Acq: 23 Dec 2022 12:00

Tgt Ion:276 Resp: 6236977
Ion Ratio Lower Upper
276 100
138 18.1 15.4 23.2
277 23.5 19.1 28.7

