

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122320\
 Data File : BP004422.D
 Acq On : 23 Dec 2020 13:59
 Operator : CG/JU
 Sample : L5059-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CALTANK2-20201209

Quant Time: Dec 23 14:37:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP121120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 10:27:21 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	268542	20.00	ng	0.00
21) Naphthalene-d8	10.622	136	1065417	20.00	ng	# 0.00
39) Acenaphthene-d10	14.475	164	640670	20.00	ng	0.00
64) Phenanthrene-d10	17.239	188	1423920	20.00	ng	# 0.00
76) Chrysene-d12	21.327	240	1560790	20.00	ng	0.00
86) Perylene-d12	23.692	264	1684618	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.416	112	827273	50.48	ng	0.00
7) Phenol-d6	6.999	99	642698	27.85	ng	0.00
23) Nitrobenzene-d5	8.987	82	1798371	80.24	ng	0.00
42) 2,4,6-Tribromophenol	15.975	330	1217959	142.48	ng	0.00
45) 2-Fluorobiphenyl	13.098	172	3987333	78.88	ng	0.00
79) Terphenyl-d14	19.804	244	5625254	69.07	ng	0.00
Target Compounds						
10) Phenol	7.022	94	65430	2.87	ng	88
50) Dimethylphthalate	13.928	163	190649	3.62	ng	99
84) Bis(2-ethylhexyl)phtha...	21.239	149	38654	3.99	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122320\
Data File : BP004422.D
Acq On : 23 Dec 2020 13:59
Operator : CG/JU
Sample : L5059-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CALTANK2-20201209

Quant Time: Dec 23 14:37:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP121120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 10:27:21 2020
Response via : Initial Calibration

