

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP122420\
 Data File : BP004429.D
 Acq On : 24 Dec 2020 11:47
 Operator : CG/JU
 Sample : PB133749BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK49

Quant Time: Dec 24 13:56:10 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP121820MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 24 13:54:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	180898	20.00	ng/ul	0.00
18) Naphthalene-d8	10.62	136	686794	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.47	164	382652	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.23	188	814502	20.00	ng/ul	0.00
78) Chrysene-d12	21.32	240	940755	20.00	ng/ul	0.00
86) Perylene-d12	23.68	264	1034111	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.32	96	29916	6.88	ng/uL	0.00
5) Phenol-d5	6.99	99	514047	32.73	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.16	67	293790	33.35	ng/ul	0.00
9) 2-Chlorophenol-d4	7.36	132	424769	33.07	ng/ul	0.00
13) 4-Methylphenol-d8	8.53	113	383673	30.68	ng/ul	0.00
19) Nitrobenzene-d5	8.98	128	196960	33.13	ng/ul	0.00
22) 2-Nitrophenol-d4	9.70	143	209095	32.25	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.24	165	377614	31.41	ng/ul	0.00
29) 4-Chloroaniline-d4	10.75	131	507358	37.22	ng/ul	0.00
44) Dimethylphthalate-d6	13.87	166	1058345	34.56	ng/ul	0.00
47) Acenaphthylene-d8	14.16	160	1406611	36.98	ng/ul	0.00
52) 4-Nitrophenol-d4	14.67	143	153075	27.66	ng/ul	0.00
58) Fluorene-d10	15.47	176	930065	34.71	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.59	200	120301	22.89	ng/ul	0.00
71) Anthracene-d10	17.33	188	1534733	38.14	ng/ul	0.00
79) Pyrene-d10	19.57	212	1840017	36.61	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.53	264	2252129	39.54	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP122420\
Data File : BP004429.D
Acq On : 24 Dec 2020 11:47
Operator : CG/JU
Sample : PB133749BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK49

Quant Time: Dec 24 13:56:10 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP121820MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 24 13:54:26 2020
Response via : Initial Calibration

