

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122623\
 Data File : BP019101.D
 Acq On : 27 Dec 2023 01:42
 Operator : MA/JU
 Sample : PB158031BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS031

Quant Time: Dec 27 02:30:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 26 22:20:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.810	152	184641	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	679526	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	395095	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.198	188	875717	20.000	ng/u1	0.00
79) Chrysene-d12	21.292	240	955426	20.000	ng/u1	# 0.00
88) Perylene-d12	23.645	264	1136001	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.246	96	32202	5.943	ng/uL	0.00
4) Pyridine-d5	3.664	84	422175	29.089	ng/u1	0.00
7) Phenol-d5	6.987	99	490568	26.475	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	294177	26.787	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	395834	27.377	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	376220	25.089	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	182203	30.126	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	198714	31.691	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	395516	30.848	ng/u1	0.00
31) 4-Chloroaniline-d4	10.751	131	448792	26.783	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	1010799	28.665	ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	1190421	30.658	ng/u1	0.00
54) 4-Nitrophenol-d4	14.663	143	176069	30.418	ng/u1	0.00
60) Fluorene-d10	15.439	176	888858	30.030	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	175685	33.459	ng/u1	0.00
73) Anthracene-d10	17.292	188	1382880	30.038	ng/u1	0.00
81) Pyrene-d10	19.533	212	1747090	23.491	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.492	264	2027050	30.603	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.281	88	65163	10.967	ng/uL	95
5) Pyridine	3.681	79	423538	28.898	ng/u1	97
6) Benzaldehyde	6.958	77	263239	27.516	ng/u1	96
8) Phenol	7.016	94	505052	27.794	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.246	93	411709	27.588	ng/u1	97
12) 2-Chlorophenol	7.375	128	404516	27.644	ng/u1	98
13) 2-Methylphenol	8.263	108	366394	26.338	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.334	45	478588	25.275	ng/u1	98
16) Acetophenone	8.640	105	583319	26.101	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.628	70	273297	22.230	ng/u1	98
18) 4-Methylphenol	8.593	108	389511	25.641	ng/u1	99
19) Hexachloroethane	8.881	117	173467	28.916	ng/u1	88
22) Nitrobenzene	9.016	77	439226	29.870	ng/u1	99
23) Isophorone	9.546	82	766210	25.556	ng/u1	97
25) 2-Nitrophenol	9.728	139	213863	31.855	ng/u1	96
26) 2,4-Dimethylphenol	9.793	107	374903	28.552	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.028	93	508455	26.862	ng/u1	100
29) 2,4-Dichlorophenol	10.257	162	378694	30.664	ng/u1	98
30) Naphthalene	10.651	128	1213955	29.449	ng/u1	100
32) 4-Chloroaniline	10.775	127	405710	25.404	ng/u1	100
33) Hexachlorobutadiene	10.934	225	292903	35.413	ng/u1	98
34) Caprolactam	11.563	113	101377	26.771	ng/u1	92
35) 4-Chloro-3-methylphenol	11.904	107	359657	25.996	ng/u1	99
36) 2-Methylnaphthalene	12.269	142	794614	27.209	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122623\
 Data File : BP019101.D
 Acq On : 27 Dec 2023 01:42
 Operator : MA/JU
 Sample : PB158031BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS031

Quant Time: Dec 27 02:30:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 26 22:20:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	780534	26.466	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.634	216	498435	34.771	ng/ul	99
40) Hexachlorocyclopentadiene	12.610	237	272310	32.050	ng/ul	99
41) 2,4,6-Trichlorophenol	12.881	196	298147	32.332	ng/ul	98
42) 2,4,5-Trichlorophenol	12.951	196	317853	31.794	ng/ul	98
43) 1,1'-Biphenyl	13.281	154	1039114	30.410	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	843125	31.241	ng/ul	98
45) 2-Nitroaniline	13.534	65	211870	28.959	ng/ul	99
47) Dimethylphthalate	13.910	163	1013318	29.193	ng/ul	99
48) 2,6-Dinitrotoluene	14.034	165	198320	30.241	ng/ul	93
50) Acenaphthylene	14.169	152	1320771	30.254	ng/ul	99
51) 3-Nitroaniline	14.363	138	185922	28.401	ng/ul	99
52) Acenaphthene	14.510	153	869744	30.132	ng/ul	99
53) 2,4-Dinitrophenol	14.569	184	112993	32.510	ng/ul	97
55) 4-Nitrophenol	14.675	109	141510	32.360	ng/ul	92
56) Dibenzofuran	14.845	168	1219077	30.388	ng/ul	99
57) 2,4-Dinitrotoluene	14.822	165	298357	32.343	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.075	232	277054	33.039	ng/ul#	92
59) Diethylphthalate	15.275	149	977349	28.774	ng/ul	99
61) Fluorene	15.492	166	978402	30.822	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.492	204	525336	31.554	ng/ul	95
63) 4-Nitroaniline	15.528	138	205878	33.012	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.581	198	189729	33.797	ng/ul	95
67) N-Nitrosodiphenylamine	15.710	169	830736	28.628	ng/ul	99
68) 4-Bromophenyl-phenylether	16.386	248	334539	30.260	ng/ul	94
69) Hexachlorobenzene	16.498	284	402423	32.602	ng/ul	95
70) Atrazine	16.669	200	301237	34.917	ng/ul	98
71) Pentachlorophenol	16.845	266	243688	33.195	ng/ul	99
72) Phenanthrene	17.239	178	1602455	30.298	ng/ul	99
74) Anthracene	17.333	178	1618683	30.533	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.245	216	488159	32.239	ng/ul	99
76) Pentachlorobenzene	14.763	250	466660	31.375	ng/ul	99
77) Carbazole	17.610	167	1460380	34.563	ng/ul	99
78) Di-n-butylphthalate	18.163	149	1609557	28.609	ng/ul	99
80) Fluoranthene	19.210	202	2037708	23.179	ng/ul	96
82) Pyrene	19.563	202	2127803	23.992	ng/ul	94
83) Butylbenzylphthalate	20.445	149	761962	23.469	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.210	252	740232	31.882	ng/ul	100
85) Benzo(a)anthracene	21.274	228	2308846	29.772	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.204	149	1088842	24.400	ng/ul#	99
87) Chrysene	21.327	228	2184705	30.592	ng/ul	100
89) Di-n-octyl phthalate	22.110	149	1924960	23.208	ng/ul	100
90) Benzo(b)fluoranthene	22.927	252	2538203	30.304	ng/ul	98
91) Benzo(k)fluoranthene	22.974	252	2383986	29.417	ng/ul	98
93) Benzo(a)pyrene	23.539	252	2358828	30.821	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.092	276	3049896	33.354	ng/ul	95
95) Dibenzo(a,h)anthracene	26.109	278	2524378	33.186	ng/ul	97
96) Benzo(g,h,i)perylene	26.845	276	2481076	33.715	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 10.967 ng/uL
 RT: 3.281 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: BP019101.D
 Acq: 27 Dec 2023 01:42




Instrument :
 BNA_P
 ClientSampleId :
 SLCS031


Tgt Ion: 88 Resp: 65163
 Ion Ratio Lower Upper
 88 100
 43 22.7 21.8 32.6
 58 67.1 56.5 84.7



#5
 Pyridine
 Concen: 28.898 ng/uL
 RT: 3.681 min Scan# 101
 Delta R.T. -0.000 min
 Lab File: BP019101.D
 Acq: 27 Dec 2023 01:42

Tgt Ion: 79 Resp: 423538
 Ion Ratio Lower Upper
 79 100
 52 68.1 57.4 86.0
 51 32.8 26.2 39.4









#6
Benzaldehyde
Concen: 27.516 ng/ul
RT: 6.958 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion: 77 Resp: 263239
Ion Ratio Lower Upper
77 100
105 100.8 77.9 116.9
106 97.9 74.6 111.8



#8
Phenol
Concen: 27.794 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion: 94 Resp: 505052
Ion Ratio Lower Upper
94 100
65 26.6 21.0 31.6
66 36.2 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 27.588 ng/ul
RT: 7.246 min Scan# 707
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion: 93 Resp: 411709
Ion Ratio Lower Upper
93 100
63 70.4 58.3 87.5
95 31.5 25.9 38.9



#12
2-Chlorophenol
Concen: 27.644 ng/ul
RT: 7.375 min Scan# 729
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:128 Resp: 404516
Ion Ratio Lower Upper
128 100
64 45.0 37.0 55.4
130 32.2 26.8 40.2









#13
2-Methylphenol
Concen: 26.338 ng/ul
RT: 8.263 min Scan# 81
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:108 Resp: 366394
Ion Ratio Lower Upper
108 100
107 90.4 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.275 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. -0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion: 45 Resp: 478588
Ion Ratio Lower Upper
45 100
77 16.5 12.7 19.1
79 11.8 10.5 15.7









#16
Acetophenone
Concen: 26.101 ng/ul
RT: 8.640 min Scan# 941
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:105 Resp: 583319
Ion Ratio Lower Upper
105 100
77 74.9 59.4 89.0
51 24.2 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 22.230 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. 0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion: 70 Resp: 273297
Ion Ratio Lower Upper
70 100
42 49.5 41.0 61.4
101 10.7 7.8 11.8
130 21.9 16.9 25.3









#18
4-Methylphenol
Concen: 25.641 ng/ul
RT: 8.593 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:108 Resp: 389511
Ion Ratio Lower Upper
108 100
107 114.0 90.6 136.0



#19
Hexachloroethane
Concen: 28.916 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:117 Resp: 173467
Ion Ratio Lower Upper
117 100
201 117.6 83.7 125.5
199 73.2 52.3 78.5









#22
Nitrobenzene
Concen: 29.870 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion: 77 Resp: 439226
Ion Ratio Lower Upper
77 100
123 45.9 35.8 53.8
65 13.0 10.5 15.7



#23
Isophorone
Concen: 25.556 ng/ul
RT: 9.546 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion: 82 Resp: 766210
Ion Ratio Lower Upper
82 100
95 7.3 5.5 8.3
138 18.7 13.5 20.3









#25
2-Nitrophenol
Concen: 31.855 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:139 Resp: 213863
Ion Ratio Lower Upper
139 100
65 45.3 39.1 58.7
109 31.3 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 28.552 ng/ul
RT: 9.793 min Scan# 1140
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:107 Resp: 374903
Ion Ratio Lower Upper
107 100
121 51.5 40.0 60.0
122 84.4 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 26.862 ng/ul
 RT: 10.028 min Scan# 1180
 Delta R.T. -0.000 min
 Lab File: BP019101.D
 Acq: 27 Dec 2023 01:42




Instrument :
 BNA_P
 ClientSampleId :
 SLCS031


Tgt Ion: 93 Resp: 508455
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.1 39.1
 123 11.0 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 30.664 ng/ul
 RT: 10.257 min Scan# 1219
 Delta R.T. -0.000 min
 Lab File: BP019101.D
 Acq: 27 Dec 2023 01:42

Tgt Ion: 162 Resp: 378694
 Ion Ratio Lower Upper
 162 100
 164 65.9 51.7 77.5
 98 34.5 29.2 43.8









#30
Naphthalene
Concen: 29.449 ng/ul
RT: 10.651 min Scan# 1286
Delta R.T. -0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:128 Resp: 1213955
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 25.404 ng/ul
RT: 10.775 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:127 Resp: 405710
Ion Ratio Lower Upper
127 100
129 32.0 25.6 38.4





#33

Hexachlorobutadiene

Concen: 35.413 ng/ul

RT: 10.934 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Instrument :

BNA_P

ClientSampleId :

SLCS031





Tgt Ion:225 Resp: 292903

Ion Ratio Lower Upper

225 100

223 63.7 49.5 74.3

227 63.1 50.9 76.3



#34

Caprolactam

Concen: 26.771 ng/ul

RT: 11.563 min Scan# 1441

Delta R.T. 0.006 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42




Tgt Ion:113 Resp: 101377


Ion Ratio Lower Upper

113 100

55 153.7 133.7 200.5

56 116.1 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 25.996 ng/ul
RT: 11.904 min Scan# 1499
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:107 Resp: 359657
Ion Ratio Lower Upper
107 100
144 25.6 21.0 31.6
142 79.0 63.6 95.4



#36
2-Methylnaphthalene
Concen: 27.209 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:142 Resp: 794614
Ion Ratio Lower Upper
142 100
141 90.4 71.4 107.0









#37
1-Methylnaphthalene
Concen: 26.466 ng/ul
RT: 12.487 min Scan# 1598
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:142 Resp: 780534
Ion Ratio Lower Upper
142 100
141 93.6 73.5 110.3
116 3.8 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.771 ng/ul
RT: 12.634 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:216 Resp: 498435
Ion Ratio Lower Upper
216 100
214 78.7 63.9 95.9
179 20.1 16.3 24.5
108 14.5 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 32.050 ng/ul

RT: 12.610 min Scan# 1619

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Instrument :

BNA_P

ClientSampleId :

SLCS031





Tgt Ion:237 Resp: 272310

Ion Ratio Lower Upper

237 100

235 62.5 49.4 74.0

272 13.5 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 32.332 ng/ul

RT: 12.881 min Scan# 1665

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42




Tgt Ion:196 Resp: 298147


Ion Ratio Lower Upper

196 100

198 98.2 76.9 115.3

200 30.5 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 31.794 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:196 Resp: 317853
Ion Ratio Lower Upper
196 100
198 95.1 77.9 116.9
200 31.2 25.2 37.8



#43
1,1'-Biphenyl
Concen: 30.410 ng/ul
RT: 13.281 min Scan# 1733
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:154 Resp: 1039114
Ion Ratio Lower Upper
154 100
153 41.3 32.8 49.2
76 12.2 10.7 16.1









#44
 2-Chloronaphthalene
 Concen: 31.241 ng/ul
 RT: 13.322 min Scan# 1740
 Delta R.T. -0.000 min
 Lab File: BP019101.D
 Acq: 27 Dec 2023 01:42




Instrument :
 BNA_P
 ClientSampleId :
 SLCS031


Tgt Ion:	162	Resp:	843125
Ion Ratio	Lower	Upper	
162	100		
164	32.7	25.7	38.5
127	36.0	30.3	45.5



#45
 2-Nitroaniline
 Concen: 28.959 ng/ul
 RT: 13.534 min Scan# 1776
 Delta R.T. -0.000 min
 Lab File: BP019101.D
 Acq: 27 Dec 2023 01:42

Tgt Ion:	65	Resp:	211870
Ion Ratio	Lower	Upper	
65	100		
92	72.4	56.9	85.3
138	111.0	88.2	132.2









#47
Dimethylphthalate
Concen: 29.193 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:163 Resp: 1013318
Ion Ratio Lower Upper
163 100
194 4.3 3.2 4.8
164 10.3 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 30.241 ng/ul
RT: 14.034 min Scan# 1861
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:165 Resp: 198320
Ion Ratio Lower Upper
165 100
89 51.9 46.6 70.0
121 20.1 17.3 25.9









#50
Acenaphthylene
Concen: 30.254 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:152 Resp: 1320771
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 12.8 10.6 16.0



#51
3-Nitroaniline
Concen: 28.401 ng/ul
RT: 14.363 min Scan# 1917
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:138 Resp: 185922
Ion Ratio Lower Upper
138 100
108 11.5 9.5 14.3
92 115.1 93.1 139.7









#52
Acenaphthene
Concen: 30.132 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:153 Resp: 869744
Ion Ratio Lower Upper
153 100
152 47.7 38.1 57.1
154 88.2 71.7 107.5



#53
2,4-Dinitrophenol
Concen: 32.510 ng/ul
RT: 14.569 min Scan# 1952
Delta R.T. -0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:184 Resp: 112993
Ion Ratio Lower Upper
184 100
63 57.1 48.2 72.4
154 62.3 49.1 73.7









#55
4-Nitrophenol
Concen: 32.360 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:109 Resp: 141510
Ion Ratio Lower Upper
109 100
139 128.6 109.0 163.4
65 121.6 106.2 159.4



#56
Dibenzofuran
Concen: 30.388 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:168 Resp: 1219077
Ion Ratio Lower Upper
168 100
139 35.5 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 32.343 ng/ul
RT: 14.822 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:165 Resp: 298357
Ion Ratio Lower Upper
165 100
63 37.8 32.7 49.1
182 2.9 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 33.039 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:232 Resp: 277054
Ion Ratio Lower Upper
232 100
131 36.0 33.8 50.8
130 1.7 1.8 2.8#
166 27.8 24.7 37.1









#59
Diethylphthalate
Concen: 28.774 ng/ul
RT: 15.275 min Scan# 2072
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:149 Resp: 977349
Ion Ratio Lower Upper
149 100
177 23.3 18.2 27.4
150 12.4 9.9 14.9



#61
Fluorene
Concen: 30.822 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:166 Resp: 978402
Ion Ratio Lower Upper
166 100
165 98.1 78.0 117.0
167 13.3 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 31.554 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:204 Resp: 525336
Ion Ratio Lower Upper
204 100
206 33.0 27.1 40.7
141 51.8 45.8 68.8



#63
4-Nitroaniline
Concen: 33.012 ng/ul
RT: 15.528 min Scan# 2115
Delta R.T. 0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:138 Resp: 205878
Ion Ratio Lower Upper
138 100
92 49.5 42.3 63.5
108 80.1 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 33.797 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:198 Resp: 189729
Ion Ratio Lower Upper
198 100
182 4.2 3.4 5.2
77 22.5 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 28.628 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:169 Resp: 830736
Ion Ratio Lower Upper
169 100
168 65.6 53.1 79.7
167 36.2 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 30.260 ng/ul
RT: 16.386 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:248 Resp: 334539
Ion Ratio Lower Upper
248 100
250 97.4 79.3 118.9
141 62.4 57.2 85.8



#69
Hexachlorobenzene
Concen: 32.602 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:284 Resp: 402423
Ion Ratio Lower Upper
284 100
142 27.6 25.8 38.6
249 30.3 24.9 37.3





#70

Atrazine

Concen: 34.917 ng/ul

RT: 16.669 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Instrument :

BNA_P

ClientSampleId :

SLCS031





Tgt Ion:200 Resp: 301237

Ion Ratio Lower Upper

200 100

173 25.0 21.0 31.4

215 48.8 38.2 57.2



#71

Pentachlorophenol

Concen: 33.195 ng/ul

RT: 16.845 min Scan# 2339

Delta R.T. -0.006 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42




Tgt Ion:266 Resp: 243688


Ion Ratio Lower Upper

266 100

264 63.0 50.2 75.2

268 65.0 50.5 75.7





#72

Phenanthrene

Concen: 30.298 ng/ul

RT: 17.239 min Scan# 2406

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Instrument :

BNA_P

ClientSampleId :

SLCS031





Tgt Ion:178 Resp: 1602455

Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 18.6 15.4 23.2



#74

Anthracene

Concen: 30.533 ng/ul

RT: 17.333 min Scan# 2422

Delta R.T. 0.006 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42




Tgt Ion:178 Resp: 1618683


Ion Ratio Lower Upper

178 100

179 15.2 12.5 18.7

176 18.5 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.239 ng/uL
RT: 13.245 min Scan# 1727
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:216 Resp: 488159
Ion Ratio Lower Upper
216 100
214 78.2 61.4 92.2
218 49.0 39.2 58.8



#76
Pentachlorobenzene
Concen: 31.375 ng/uL
RT: 14.763 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:250 Resp: 466660
Ion Ratio Lower Upper
250 100
248 64.1 50.6 75.8
252 64.2 51.3 76.9









#77
 Carbazole
 Concen: 34.563 ng/ul
 RT: 17.610 min Scan# 2469
 Delta R.T. -0.000 min
 Lab File: BP019101.D
 Acq: 27 Dec 2023 01:42




Instrument :
 BNA_P
 ClientSampleId :
 SLCS031


Tgt Ion:167 Resp: 1460380
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.1 25.7
 139 12.0 10.2 15.4



#78
 Di-n-butylphthalate
 Concen: 28.609 ng/ul
 RT: 18.163 min Scan# 2563
 Delta R.T. -0.000 min
 Lab File: BP019101.D
 Acq: 27 Dec 2023 01:42

Tgt Ion:149 Resp: 1609557
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.2
 104 4.8 3.9 5.9









#80
Fluoranthene
Concen: 23.179 ng/ul
RT: 19.210 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:202 Resp: 2037708
Ion Ratio Lower Upper
202 100
101 9.8 9.1 13.7
100 7.4 7.1 10.7



#82
Pyrene
Concen: 23.992 ng/ul
RT: 19.563 min Scan# 2801
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:202 Resp: 2127803
Ion Ratio Lower Upper
202 100
101 11.4 11.0 16.4
100 9.3 9.1 13.7









#83
Butylbenzylphthalate
Concen: 23.469 ng/ul
RT: 20.445 min Scan# 2951
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:	149	Resp:	761962
Ion Ratio	Lower	Upper	
149	100		
91	63.5	56.6	85.0
206	23.8	17.4	26.0



#84
3,3'-Dichlorobenzidine
Concen: 31.882 ng/ul
RT: 21.210 min Scan# 3081
Delta R.T. -0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:	252	Resp:	740232
Ion Ratio	Lower	Upper	
252	100		
254	63.9	51.0	76.4
126	9.8	7.8	11.6





#85

Benzo(a)anthracene

Concen: 29.772 ng/ul

RT: 21.274 min Scan# 3092

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Instrument :

BNA_P

ClientSampleId :

SLCS031





Tgt Ion:228 Resp: 2308846

Ion Ratio Lower Upper

228 100

229 19.7 15.7 23.5

226 27.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 24.400 ng/ul

RT: 21.204 min Scan# 3080

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42




Tgt Ion:149 Resp: 1088842


Ion Ratio Lower Upper

149 100

167 28.0 22.2 33.2

279 5.0 3.3 4.9#





#87

Chrysene

Concen: 30.592 ng/ul

RT: 21.327 min Scan# 3101

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Instrument :

BNA_P

ClientSampleId :

SLCS031





Tgt Ion:228 Resp: 2184705

Ion Ratio Lower Upper

228 100

226 29.8 23.9 35.9

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 23.208 ng/ul




RT: 22.110 min Scan# 3234


Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Tgt Ion:149 Resp: 1924960





#90

Benzo(b)fluoranthene

Concen: 30.304 ng/ul

RT: 22.927 min Scan# 3372

Delta R.T. 0.006 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Instrument :

BNA_P

ClientSampleId :

SLCS031





Tgt Ion:252 Resp: 2538203

Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 8.0 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 29.417 ng/ul

RT: 22.974 min Scan# 3381

Delta R.T. -0.000 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42




Tgt Ion:252 Resp: 2383986


Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 8.1 8.0 12.0









#93
Benzo(a)pyrene
Concen: 30.821 ng/ul
RT: 23.539 min Scan# 3477
Delta R.T. -0.000 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42




Instrument :
BNA_P
ClientSampleId :
SLCS031


Tgt Ion:252 Resp: 2358828
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 9.2 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 33.354 ng/ul
RT: 26.092 min Scan# 3911
Delta R.T. 0.006 min
Lab File: BP019101.D
Acq: 27 Dec 2023 01:42

Tgt Ion:276 Resp: 3049896
Ion Ratio Lower Upper
276 100
138 19.7 19.4 29.2
277 24.7 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 33.186 ng/ul

RT: 26.109 min Scan# 3913

Delta R.T. 0.006 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Instrument :

BNA_P

ClientSampleId :

SLCS031





Tgt Ion:278 Resp: 2524378

Ion Ratio Lower Upper

278 100

139 13.2 13.0 19.6

279 23.7 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 33.715 ng/ul

RT: 26.845 min Scan# 4039

Delta R.T. 0.012 min

Lab File: BP019101.D

Acq: 27 Dec 2023 01:42

Tgt Ion:276 Resp: 2481076

Ion Ratio Lower Upper

276 100

138 18.4 18.1 27.1

277 23.9 19.4 29.0

