

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122723\
 Data File : BP019122.D
 Acq On : 27 Dec 2023 19:32
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726

Quant Time: Dec 28 03:45:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 28 03:15:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	156874	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	592774	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	363321	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	813399	20.000	ng/u1	0.00
79) Chrysene-d12	21.316	240	889569	20.000	ng/u1	0.00
88) Perylene-d12	23.680	264	1053538	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	34203	7.446	ng/uL	0.00
4) Pyridine-d5	3.652	84	224957	19.512	ng/u1	0.00
7) Phenol-d5	6.981	99	271920	19.329	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.146	67	163239	20.057	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	215949	19.041	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	212283	19.382	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	99997	19.515	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	106824	19.290	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	221042	19.766	ng/u1	0.00
31) 4-Chloroaniline-d4	10.752	131	266513	19.376	ng/u1	0.00
46) Dimethylphthalate-d6	13.869	166	615261	19.172	ng/u1	0.00
49) Acenaphthylene-d8	14.145	160	673304	19.336	ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	101555	18.740	ng/u1	0.00
60) Fluorene-d10	15.445	176	521744	18.819	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	96865	18.165	ng/u1	0.00
73) Anthracene-d10	17.310	188	815951	19.209	ng/u1	0.00
81) Pyrene-d10	19.551	212	1023503	18.879	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.527	264	1156357	19.095	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.270	88	37871	7.499	ng/uL	96
5) Pyridine	3.675	79	228568	19.532	ng/u1	99
6) Benzaldehyde	6.952	77	129199	19.780	ng/u1	96
8) Phenol	7.011	94	270785	19.488	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.240	93	226453	20.142	ng/u1	97
12) 2-Chlorophenol	7.369	128	217929	18.774	ng/u1	100
13) 2-Methylphenol	8.258	108	200499	19.505	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.334	45	267625	20.179	ng/u1	99
16) Acetophenone	8.634	105	325030	19.814	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.622	70	153899	18.815	ng/u1	98
18) 4-Methylphenol	8.587	108	213067	19.257	ng/u1	99
19) Hexachloroethane	8.881	117	94900	19.491	ng/u1	97
22) Nitrobenzene	9.010	77	240659	19.224	ng/u1	97
23) Isophorone	9.540	82	430512	19.070	ng/u1	99
25) 2-Nitrophenol	9.722	139	115473	19.529	ng/u1	97
26) 2,4-Dimethylphenol	9.793	107	210217	20.194	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.028	93	293118	19.645	ng/u1	100
29) 2,4-Dichlorophenol	10.257	162	213867	19.571	ng/u1	100
30) Naphthalene	10.657	128	685235	19.173	ng/u1	100
32) 4-Chloroaniline	10.775	127	261060	19.779	ng/u1	98
33) Hexachlorobutadiene	10.934	225	166209	19.763	ng/u1	98
34) Caprolactam	11.569	113	49365	17.102	ng/u1	97
35) 4-Chloro-3-methylphenol	11.904	107	204925	19.103	ng/u1	98
36) 2-Methylnaphthalene	12.269	142	467664	19.408	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122723\
 Data File : BP019122.D
 Acq On : 27 Dec 2023 19:32
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726


Quant Time: Dec 28 03:45:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 28 03:15:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	465311	19.149	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.640	216	297810	19.902	ng/ul	100
40) Hexachlorocyclopentadiene	12.616	237	163976	19.743	ng/ul	97
41) 2,4,6-Trichlorophenol	12.887	196	169159	18.943	ng/ul	99
42) 2,4,5-Trichlorophenol	12.957	196	181729	19.171	ng/ul	97
43) 1,1'-Biphenyl	13.287	154	608143	19.188	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	486626	19.135	ng/ul	99
45) 2-Nitroaniline	13.540	65	117711	19.281	ng/ul	91
47) Dimethylphthalate	13.916	163	608520	19.210	ng/ul	100
48) 2,6-Dinitrotoluene	14.040	165	110990	19.067	ng/ul	96
50) Acenaphthylene	14.175	152	760793	19.419	ng/ul	100
51) 3-Nitroaniline	14.369	138	107784	18.940	ng/ul	98
52) Acenaphthene	14.516	153	500879	18.864	ng/ul	99
53) 2,4-Dinitrophenol	14.581	184	59832	15.955	ng/ul	96
55) 4-Nitrophenol	14.681	109	80719	18.638	ng/ul	96
56) Dibenzofuran	14.851	168	722756	19.061	ng/ul	97
57) 2,4-Dinitrotoluene	14.828	165	163260	19.071	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.081	232	157213	18.982	ng/ul	99
59) Diethylphthalate	15.287	149	584455	19.029	ng/ul	99
61) Fluorene	15.504	166	578141	19.049	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.504	204	316365	19.088	ng/ul	98
63) 4-Nitroaniline	15.534	138	107641	18.800	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.592	198	105509	18.688	ng/ul	99
67) N-Nitrosodiphenylamine	15.716	169	489132	19.703	ng/ul	100
68) 4-Bromophenyl-phenylether	16.398	248	201251	19.906	ng/ul	98
69) Hexachlorobenzene	16.510	284	244565	19.795	ng/ul	98
70) Atrazine	16.681	200	111167	15.868	ng/ul	99
71) Pentachlorophenol	16.857	266	136131	18.767	ng/ul	97
72) Phenanthrene	17.251	178	932538	18.918	ng/ul	99
74) Anthracene	17.345	178	951841	19.396	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.245	216	285811	20.209	ng/uL	97
76) Pentachlorobenzene	14.769	250	284049	20.212	ng/uL	99
77) Carbazole	17.622	167	826023	20.236	ng/ul	99
78) Di-n-butylphthalate	18.175	149	988130	19.519	ng/ul	100
80) Fluoranthene	19.228	202	1168280	18.779	ng/ul	99
82) Pyrene	19.580	202	1221835	18.746	ng/ul	100
83) Butylbenzylphthalate	20.463	149	451636	19.398	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.239	252	430915	19.364	ng/ul	99
85) Benzo(a)anthracene	21.298	228	1325519	18.859	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.227	149	686225	20.016	ng/ul	99
87) Chrysene	21.351	228	1242363	18.956	ng/ul	99
89) Di-n-octyl phthalate	22.145	149	1190866	19.766	ng/ul	100
90) Benzo(b)fluoranthene	22.957	252	1391053	18.798	ng/ul	100
91) Benzo(k)fluoranthene	23.004	252	1396726	19.208	ng/ul	99
93) Benzo(a)pyrene	23.574	252	1321303	18.970	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.139	276	1688019	18.906	ng/ul	98
95) Dibenzo(a,h)anthracene	26.157	278	1412460	19.016	ng/ul	99
96) Benzo(g,h,i)perylene	26.892	276	1365195	19.080	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 28 03:45:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 28 03:15:29 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.499 ng/uL
RT: 3.270 min Scan# 31
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 88 Resp: 37871
Ion Ratio Lower Upper
88 100
43 24.5 21.0 31.4
58 67.2 56.6 84.8



#5
Pyridine
Concen: 19.532 ng/uL
RT: 3.675 min Scan# 100
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion: 79 Resp: 228568
Ion Ratio Lower Upper
79 100
52 66.1 52.4 78.6
51 31.7 25.9 38.9









#6
Benzaldehyde
Concen: 19.780 ng/ul
RT: 6.952 min Scan# 612
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 77 Resp: 129199
Ion Ratio Lower Upper
77 100
105 102.8 78.6 117.8
106 96.8 74.8 112.2



#8
Phenol
Concen: 19.488 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion: 94 Resp: 270785
Ion Ratio Lower Upper
94 100
65 26.3 21.1 31.7
66 35.9 29.3 43.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.142 ng/ul
 RT: 7.240 min Scan# 706
 Delta R.T. -0.000 min
 Lab File: BP019122.D
 Acq: 27 Dec 2023 19:32

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726

Tgt Ion: 93 Resp: 226453




Ion	Ratio	Lower	Upper
93	100		
63	71.3	59.4	89.0
95	32.2	26.6	40.0




#12
 2-Chlorophenol
 Concen: 18.774 ng/ul
 RT: 7.369 min Scan# 728
 Delta R.T. -0.006 min
 Lab File: BP019122.D
 Acq: 27 Dec 2023 19:32

Tgt Ion: 128 Resp: 217929

Ion	Ratio	Lower	Upper
128	100		
64	45.2	35.9	53.9
130	31.4	25.3	37.9





#13

2-Methylphenol

Concen: 19.505 ng/ul

RT: 8.258 min Scan# 81

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :





SSTD020726

Tgt Ion:108 Resp: 200499

Ion Ratio Lower Upper

108 100

107 90.6 72.6 109.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.179 ng/ul

RT: 8.334 min Scan# 892

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion: 45 Resp: 267625


Ion Ratio Lower Upper

45 100

77 16.9 13.6 20.4

79 13.1 9.7 14.5









#16
Acetophenone
Concen: 19.814 ng/ul
RT: 8.634 min Scan# 944
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:105 Resp: 325030
Ion Ratio Lower Upper
105 100
77 75.8 59.8 89.6
51 25.3 19.3 28.9



#17
N-Nitroso-di-n-propylamine
Concen: 18.815 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion: 70 Resp: 153899
Ion Ratio Lower Upper
70 100
42 51.9 40.6 61.0
101 9.8 7.3 10.9
130 21.5 16.1 24.1









#18
4-Methylphenol
Concen: 19.257 ng/ul
RT: 8.587 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:108 Resp: 213067
Ion Ratio Lower Upper
108 100
107 111.1 88.1 132.1



#19
Hexachloroethane
Concen: 19.491 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:117 Resp: 94900
Ion Ratio Lower Upper
117 100
201 114.3 92.5 138.7
199 75.1 56.9 85.3









#22
Nitrobenzene
Concen: 19.224 ng/ul
RT: 9.010 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 77 Resp: 240659
Ion Ratio Lower Upper
77 100
123 46.5 35.5 53.3
65 13.0 10.3 15.5



#23
Isophorone
Concen: 19.070 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion: 82 Resp: 430512
Ion Ratio Lower Upper
82 100
95 7.7 5.9 8.9
138 17.8 14.7 22.1









#25
2-Nitrophenol
Concen: 19.529 ng/ul
RT: 9.722 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:139 Resp: 115473
Ion Ratio Lower Upper
139 100
65 50.3 38.8 58.2
109 30.3 26.2 39.2



#26
2,4-Dimethylphenol
Concen: 20.194 ng/ul
RT: 9.793 min Scan# 1140
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:107 Resp: 210217
Ion Ratio Lower Upper
107 100
121 50.9 39.8 59.6
122 86.1 66.4 99.6









#27
Bis(2-Chloroethoxy)methane
Concen: 19.645 ng/ul
RT: 10.028 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 93 Resp: 293118
Ion Ratio Lower Upper
93 100
95 33.5 26.7 40.1
123 10.8 8.5 12.7



#29
2,4-Dichlorophenol
Concen: 19.571 ng/ul
RT: 10.257 min Scan# 1219
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion: 162 Resp: 213867
Ion Ratio Lower Upper
162 100
164 65.8 52.6 78.8
98 34.1 26.7 40.1





#30

Naphthalene

Concen: 19.173 ng/ul

RT: 10.657 min Scan# 1287

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:128 Resp: 685235

Ion Ratio Lower Upper

128 100

129 10.7 8.6 13.0

127 13.0 10.6 15.8



#32

4-Chloroaniline

Concen: 19.779 ng/ul

RT: 10.775 min Scan# 1307

Delta R.T. -0.006 min

Lab File: BP019122.D




Acq: 27 Dec 2023 19:32


Tgt Ion:127 Resp: 261060

Ion Ratio Lower Upper

127 100

129 32.9 25.6 38.4





#33

Hexachlorobutadiene

Concen: 19.763 ng/ul

RT: 10.934 min Scan# 1334

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:225 Resp: 166209

Ion Ratio Lower Upper

225 100

223 64.7 50.5 75.7

227 64.4 53.0 79.4



#34

Caprolactam

Concen: 17.102 ng/ul

RT: 11.569 min Scan# 1442

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:113 Resp: 49365


Ion Ratio Lower Upper

113 100

55 159.6 122.6 184.0

56 117.9 93.0 139.4









#35
4-Chloro-3-methylphenol
Concen: 19.103 ng/ul
RT: 11.904 min Scan# 1499
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:107 Resp: 204925
Ion Ratio Lower Upper
107 100
144 25.3 21.0 31.6
142 79.3 64.8 97.2



#36
2-Methylnaphthalene
Concen: 19.408 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:142 Resp: 467664
Ion Ratio Lower Upper
142 100
141 88.9 71.7 107.5





#37

1-Methylnaphthalene

Concen: 19.149 ng/ul

RT: 12.487 min Scan# 1599

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:142 Resp: 465311

Ion Ratio Lower Upper

142 100

141 92.4 74.4 111.6

116 3.7 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.902 ng/ul

RT: 12.640 min Scan# 1624

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Tgt Ion:216 Resp: 297810




Ion Ratio Lower Upper


216 100

214 78.4 62.6 94.0

179 19.3 15.8 23.8

108 14.3 11.9 17.9





#40

Hexachlorocyclopentadiene

Concen: 19.743 ng/ul

RT: 12.616 min Scan# 1620

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:237 Resp: 163976

Ion Ratio Lower Upper

237 100

235 61.7 51.2 76.8

272 13.2 10.9 16.3



#41

2,4,6-Trichlorophenol

Concen: 18.943 ng/ul

RT: 12.887 min Scan# 1666

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:196 Resp: 169159


Ion Ratio Lower Upper

196 100

198 97.2 77.4 116.0

200 30.3 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 19.171 ng/ul

RT: 12.957 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:196 Resp: 181729

Ion Ratio Lower Upper

196 100

198 99.7 77.3 115.9

200 31.6 25.7 38.5



#43

1,1'-Biphenyl

Concen: 19.188 ng/ul

RT: 13.287 min Scan# 1734

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:154 Resp: 608143


Ion Ratio Lower Upper

154 100

153 39.6 31.9 47.9

76 11.6 9.5 14.3





#44

2-Chloronaphthalene

Concen: 19.135 ng/ul

RT: 13.322 min Scan# 1740

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:162 Resp: 486626

Ion Ratio Lower Upper

162 100

164 32.1 26.2 39.4

127 36.6 28.8 43.2



#45

2-Nitroaniline

Concen: 19.281 ng/ul

RT: 13.540 min Scan# 1777

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion: 65 Resp: 117711


Ion Ratio Lower Upper

65 100

92 71.4 60.2 90.4

138 103.4 92.9 139.3









#47
Dimethylphthalate
Concen: 19.210 ng/ul
RT: 13.916 min Scan# 1841
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:163 Resp: 608520
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 10.2 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 19.067 ng/ul
RT: 14.040 min Scan# 1862
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:165 Resp: 110990
Ion Ratio Lower Upper
165 100
89 52.8 45.1 67.7
121 19.9 16.7 25.1









#50
Acenaphthylene
Concen: 19.419 ng/ul
RT: 14.175 min Scan# 1885
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:152 Resp: 760793
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 13.1 10.6 15.8



#51
3-Nitroaniline
Concen: 18.940 ng/ul
RT: 14.369 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:138 Resp: 107784
Ion Ratio Lower Upper
138 100
108 12.0 9.8 14.6
92 113.0 91.9 137.9









#52
Acenaphthene
Concen: 18.864 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:153 Resp: 500879
Ion Ratio Lower Upper
153 100
152 46.5 38.1 57.1
154 87.2 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 15.955 ng/ul
RT: 14.581 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:184 Resp: 59832
Ion Ratio Lower Upper
184 100
63 54.7 48.5 72.7
154 62.1 49.3 73.9









#55
4-Nitrophenol
Concen: 18.638 ng/ul
RT: 14.681 min Scan# 1972
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:109 Resp: 80719
Ion Ratio Lower Upper
109 100
139 127.1 102.6 154.0
65 123.9 92.9 139.3



#56
Dibenzofuran
Concen: 19.061 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:168 Resp: 722756
Ion Ratio Lower Upper
168 100
139 35.6 27.4 41.0









#57
2,4-Dinitrotoluene
Concen: 19.071 ng/ul
RT: 14.828 min Scan# 1996
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:165 Resp: 163260
Ion Ratio Lower Upper
165 100
63 38.6 31.5 47.3
182 2.7 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.982 ng/ul
RT: 15.081 min Scan# 2039
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:232 Resp: 157213
Ion Ratio Lower Upper
232 100
131 37.0 29.3 43.9
130 1.8 1.4 2.2
166 28.0 22.6 33.8









#59
Diethylphthalate
Concen: 19.029 ng/ul
RT: 15.287 min Scan# 2074
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:149 Resp: 584455
Ion Ratio Lower Upper
149 100
177 23.2 18.7 28.1
150 12.6 9.8 14.8



#61
Fluorene
Concen: 19.049 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:166 Resp: 578141
Ion Ratio Lower Upper
166 100
165 98.3 79.3 118.9
167 13.5 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 19.088 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:204 Resp: 316365
Ion Ratio Lower Upper
204 100
206 33.2 26.3 39.5
141 51.1 42.2 63.2



#63
4-Nitroaniline
Concen: 18.800 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:138 Resp: 107641
Ion Ratio Lower Upper
138 100
92 53.6 42.2 63.4
108 80.8 66.2 99.4









#66
4,6-Dinitro-2-methylphenol
Concen: 18.688 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:198 Resp: 105509
Ion Ratio Lower Upper
198 100
182 4.6 4.0 6.0
77 21.2 17.4 26.2



#67
N-Nitrosodiphenylamine
Concen: 19.703 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:169 Resp: 489132
Ion Ratio Lower Upper
169 100
168 67.2 53.8 80.8
167 36.6 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 19.906 ng/ul
RT: 16.398 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:248 Resp: 201251
Ion Ratio Lower Upper
248 100
250 98.1 78.3 117.5
141 64.0 48.6 72.8



#69
Hexachlorobenzene
Concen: 19.795 ng/ul
RT: 16.510 min Scan# 2282
Delta R.T. -0.000 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:284 Resp: 244565
Ion Ratio Lower Upper
284 100
142 26.6 22.9 34.3
249 30.1 23.7 35.5





#70

Atrazine

Concen: 15.868 ng/ul

RT: 16.681 min Scan# 211

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:200 Resp: 111167

Ion Ratio Lower Upper

200 100

173 24.1 20.2 30.2

215 49.2 39.9 59.9



#71

Pentachlorophenol

Concen: 18.767 ng/ul

RT: 16.857 min Scan# 2341

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:266 Resp: 136131


Ion Ratio Lower Upper

266 100

264 63.5 50.0 75.0

268 62.9 52.8 79.2





#72

Phenanthrene

Concen: 18.918 ng/ul

RT: 17.251 min Scan# 2408

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:178 Resp: 932538

Ion Ratio Lower Upper

178 100

179 15.4 11.9 17.9

176 19.4 15.0 22.6



#74

Anthracene

Concen: 19.396 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:178 Resp: 951841


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 18.7 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.209 ng/uL

RT: 13.245 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:216 Resp: 285811

Ion Ratio Lower Upper

216 100

214 80.2 61.6 92.4

218 49.2 38.9 58.3



#76

Pentachlorobenzene

Concen: 20.212 ng/uL

RT: 14.769 min Scan# 1986

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:250 Resp: 284049


Ion Ratio Lower Upper

250 100

248 63.3 50.4 75.6

252 64.1 52.1 78.1





#77

Carbazole

Concen: 20.236 ng/ul

RT: 17.622 min Scan# 2471

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:167 Resp: 826023

Ion Ratio Lower Upper

167 100

166 21.1 17.4 26.0

139 12.2 9.8 14.6



#78

Di-n-butylphthalate

Concen: 19.519 ng/ul

RT: 18.175 min Scan# 2565

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:149 Resp: 988130


Ion Ratio Lower Upper

149 100

150 9.1 7.3 10.9

104 5.0 3.9 5.9





#80

Fluoranthene

Concen: 18.779 ng/ul

RT: 19.228 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:202 Resp: 1168280

Ion Ratio Lower Upper

202 100

101 9.9 7.8 11.6

100 7.8 6.0 9.0



#82

Pyrene

Concen: 18.746 ng/ul

RT: 19.580 min Scan# 2804

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:202 Resp: 1221835


Ion Ratio Lower Upper

202 100

101 11.6 9.2 13.8

100 9.7 7.8 11.6





#83

Butylbenzylphthalate

Concen: 19.398 ng/ul

RT: 20.463 min Scan# 2955

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:149 Resp: 451636

Ion Ratio Lower Upper

149 100

91 71.2 55.6 83.4

206 22.0 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 19.364 ng/ul

RT: 21.239 min Scan# 3086

Delta R.T. -0.000 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:252 Resp: 430915


Ion Ratio Lower Upper

252 100

254 65.3 51.5 77.3

126 8.8 7.6 11.4









#85
Benzo(a)anthracene
Concen: 18.859 ng/ul
RT: 21.298 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:228 Resp: 1325519
Ion Ratio Lower Upper
228 100
229 19.4 15.5 23.3
226 26.7 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.016 ng/ul
RT: 21.227 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BP019122.D
Acq: 27 Dec 2023 19:32

Tgt Ion:149 Resp: 686225
Ion Ratio Lower Upper
149 100
167 27.2 22.5 33.7
279 4.8 4.0 6.0





#87

Chrysene

Concen: 18.956 ng/ul

RT: 21.351 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:228 Resp: 1242363

Ion Ratio Lower Upper

228 100

226 30.1 23.4 35.2

229 19.3 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 19.766 ng/ul




RT: 22.145 min Scan# 3240


Delta R.T. -0.006 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Tgt Ion:149 Resp: 1190866





#90

Benzo(b)fluoranthene

Concen: 18.798 ng/ul

RT: 22.957 min Scan# 3378

Delta R.T. -0.012 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:252 Resp: 1391053

Ion Ratio Lower Upper

252 100

253 22.4 17.8 26.8

125 8.4 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 19.208 ng/ul

RT: 23.004 min Scan# 3386

Delta R.T. -0.012 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:252 Resp: 1396726


Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 8.6 6.7 10.1





#93

Benzo(a)pyrene

Concen: 18.970 ng/ul

RT: 23.574 min Scan# 34

Delta R.T. -0.012 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:252 Resp: 1321303

Ion Ratio Lower Upper

252 100

253 22.2 17.5 26.3

125 9.2 7.4 11.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.906 ng/ul

RT: 26.139 min Scan# 3919

Delta R.T. -0.012 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32




Tgt Ion:276 Resp: 1688019


Ion Ratio Lower Upper

276 100

138 20.1 16.6 25.0

277 25.4 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 19.016 ng/ul

RT: 26.157 min Scan# 3922

Delta R.T. -0.012 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:278 Resp: 1412460

Ion Ratio Lower Upper

278 100

139 12.6 11.0 16.4

279 23.7 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 19.080 ng/ul

RT: 26.892 min Scan# 4047

Delta R.T. -0.012 min

Lab File: BP019122.D

Acq: 27 Dec 2023 19:32

Tgt Ion:276 Resp: 1365195

Ion Ratio Lower Upper

276 100

138 17.2 15.2 22.8

277 23.6 19.4 29.0

