

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122923\
 Data File : BM043679.D
 Acq On : 29 Dec 2023 19:40
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020175

Quant Time: Dec 29 22:36:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 22:31:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	206541	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	884584	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.598	164	501493	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	1028004	20.000	ng/u1	0.00
79) Chrysene-d12	21.521	240	921346	20.000	ng/u1	0.00
88) Perylene-d12	23.938	264	1027309	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	44740	7.040	ng/uL	0.00
4) Pyridine-d5	3.793	84	332632	18.186	ng/u1	0.00
7) Phenol-d5	7.128	99	433313	18.442	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.298	67	261736	17.589	ng/u1	0.00
11) 2-Chlorophenol-d4	7.486	132	332004	18.507	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	326915	17.930	ng/u1	0.00
21) Nitrobenzene-d5	9.133	128	159600	18.648	ng/u1	0.00
24) 2-Nitrophenol-d4	9.851	143	170645	19.297	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.386	165	301875	19.338	ng/u1	0.00
31) 4-Chloroaniline-d4	10.916	131	426824	18.585	ng/u1	0.00
46) Dimethylphthalate-d6	14.015	166	790581	17.512	ng/u1	0.00
49) Acenaphthylene-d8	14.292	160	976750	18.239	ng/u1	0.00
54) 4-Nitrophenol-d4	14.804	143	177175	20.736	ng/u1	0.01
60) Fluorene-d10	15.592	176	694739	18.610	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.715	200	107781	15.853	ng/u1	0.00
73) Anthracene-d10	17.445	188	1059881	18.384	ng/u1	0.00
81) Pyrene-d10	19.733	212	1140583	15.680	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.780	264	1136495	17.927	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.404	88	51503	7.335	ng/uL	95
5) Pyridine	3.810	79	358420	18.949	ng/u1	95
6) Benzaldehyde	7.104	77	204593	18.365	ng/u1	99
8) Phenol	7.151	94	424842	18.379	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.392	93	329553	17.561	ng/u1	99
12) 2-Chlorophenol	7.522	128	333927	18.273	ng/u1	98
13) 2-Methylphenol	8.410	108	313181	17.843	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.486	45	559706	17.761	ng/u1	97
16) Acetophenone	8.798	105	486888	17.442	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.781	70	271005	16.266	ng/u1	98
18) 4-Methylphenol	8.739	108	337625	17.749	ng/u1	100
19) Hexachloroethane	9.033	117	128366	16.659	ng/u1	93
22) Nitrobenzene	9.175	77	391510	18.484	ng/u1	99
23) Isophorone	9.704	82	719410	16.744	ng/u1	100
25) 2-Nitrophenol	9.886	139	181201	19.280	ng/u1	97
26) 2,4-Dimethylphenol	9.945	107	314250	18.770	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.192	93	434750	17.137	ng/u1	100
29) 2,4-Dichlorophenol	10.416	162	291059	19.213	ng/u1	98
30) Naphthalene	10.822	128	1037540	18.404	ng/u1	99
32) 4-Chloroaniline	10.939	127	415893	18.832	ng/u1	99
33) Hexachlorobutadiene	11.086	225	148804	19.265	ng/u1	99
34) Caprolactam	11.727	113	90566	18.707	ng/u1	96
35) 4-Chloro-3-methylphenol	12.057	107	321739	18.081	ng/u1	100
36) 2-Methylnaphthalene	12.427	142	677887	17.664	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122923\
 Data File : BM043679.D
 Acq On : 29 Dec 2023 19:40
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020175


Quant Time: Dec 29 22:36:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 22:31:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.645	142	690118	17.829	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.786	216	283786	18.572	ng/ul	97
40) Hexachlorocyclopentadiene	12.757	237	96606	10.076	ng/ul	98
41) 2,4,6-Trichlorophenol	13.033	196	203686	18.753	ng/ul	98
42) 2,4,5-Trichlorophenol	13.104	196	222182	19.266	ng/ul	99
43) 1,1'-Biphenyl	13.433	154	849912	17.779	ng/ul	100
44) 2-Chloronaphthalene	13.474	162	672459	18.348	ng/ul	99
45) 2-Nitroaniline	13.692	65	231656	18.849	ng/ul	99
47) Dimethylphthalate	14.063	163	791552	17.754	ng/ul	100
48) 2,6-Dinitrotoluene	14.186	165	162357	18.836	ng/ul	95
50) Acenaphthylene	14.321	152	1128166	18.345	ng/ul	99
51) 3-Nitroaniline	14.515	138	199128	20.124	ng/ul	98
52) Acenaphthene	14.663	153	737156	18.187	ng/ul	99
53) 2,4-Dinitrophenol	14.721	184	102146	22.262	ng/ul	97
55) 4-Nitrophenol	14.821	109	141494	21.326	ng/ul	98
56) Dibenzofuran	14.998	168	964561	18.426	ng/ul	98
57) 2,4-Dinitrotoluene	14.974	165	236035	19.382	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.221	232	171676	19.430	ng/ul	96
59) Diethylphthalate	15.421	149	832688	18.225	ng/ul	99
61) Fluorene	15.645	166	829860	18.557	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.645	204	356915	18.144	ng/ul	97
63) 4-Nitroaniline	15.674	138	216138	23.864	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.727	198	114219	15.735	ng/ul	98
67) N-Nitrosodiphenylamine	15.857	169	671385	17.003	ng/ul	100
68) 4-Bromophenyl-phenylether	16.539	248	205472	16.871	ng/ul	99
69) Hexachlorobenzene	16.639	284	243981	17.486	ng/ul	99
70) Atrazine	16.815	200	170991	19.756	ng/ul	100
71) Pentachlorophenol	16.992	266	185856	22.033	ng/ul	98
72) Phenanthrene	17.386	178	1250075	17.942	ng/ul	99
74) Anthracene	17.480	178	1290910	18.359	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.392	216	283284	16.780	ng/uL	99
76) Pentachlorobenzene	14.910	250	272087	16.551	ng/uL	96
77) Carbazole	17.751	167	1233515	20.554	ng/ul	100
78) Di-n-butylphthalate	18.321	149	1518304	19.352	ng/ul	100
80) Fluoranthene	19.397	202	1384427	15.369	ng/ul	98
82) Pyrene	19.762	202	1463724	15.806	ng/ul	97
83) Butylbenzylphthalate	20.668	149	742667	18.549	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.444	252	428941	18.903	ng/ul	98
85) Benzo(a)anthracene	21.509	228	1415543	18.373	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.444	149	1133456	19.885	ng/ul	100
87) Chrysene	21.562	228	1287012	18.569	ng/ul	100
89) Di-n-octyl phthalate	22.386	149	1943482	19.776	ng/ul	100
90) Benzo(b)fluoranthene	23.197	252	1403023	17.822	ng/ul	99
91) Benzo(k)fluoranthene	23.250	252	1331225	17.338	ng/ul	100
93) Benzo(a)pyrene	23.832	252	1294210	18.019	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.450	276	1631181	19.020	ng/ul	100
95) Dibenzo(a,h)anthracene	26.474	278	1359844	19.148	ng/ul	99
96) Benzo(g,h,i)perylene	27.215	276	1280128	19.042	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 29 22:36:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 22:31:57 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.335 ng/uL
RT: 3.404 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion: 88 Resp: 51503
Ion Ratio Lower Upper
88 100
43 36.9 26.3 39.5
58 82.3 63.1 94.7



#5
Pyridine
Concen: 18.949 ng/uL
RT: 3.810 min Scan# 106
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion: 79 Resp: 358420
Ion Ratio Lower Upper
79 100
52 84.0 62.4 93.6
51 39.3 30.6 45.8









#6
Benzaldehyde
Concen: 18.365 ng/ul
RT: 7.104 min Scan# 666
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion: 77 Resp: 204593
Ion Ratio Lower Upper
77 100
105 93.7 73.4 110.0
106 90.1 71.4 107.0



#8
Phenol
Concen: 18.379 ng/ul
RT: 7.151 min Scan# 674
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion: 94 Resp: 424842
Ion Ratio Lower Upper
94 100
65 29.4 23.7 35.5
66 37.0 30.2 45.4









#10
Bis(2-Chloroethyl)ether
Concen: 17.561 ng/ul
RT: 7.392 min Scan# 715
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion: 93 Resp: 329553
Ion Ratio Lower Upper
93 100
63 82.8 65.8 98.8
95 32.0 26.2 39.2



#12
2-Chlorophenol
Concen: 18.273 ng/ul
RT: 7.522 min Scan# 737
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:128 Resp: 333927
Ion Ratio Lower Upper
128 100
64 48.8 40.5 60.7
130 33.1 25.6 38.4









#13
2-Methylphenol
Concen: 17.843 ng/ul
RT: 8.410 min Scan# 887
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:108 Resp: 313181
Ion Ratio Lower Upper
108 100
107 87.5 69.0 103.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.761 ng/ul
RT: 8.486 min Scan# 901
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion: 45 Resp: 559706
Ion Ratio Lower Upper
45 100
77 11.7 10.6 15.8
79 9.2 7.6 11.4









#16
Acetophenone
Concen: 17.442 ng/ul
RT: 8.798 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:105 Resp: 486888
Ion Ratio Lower Upper
105 100
77 80.1 65.1 97.7
51 31.0 24.3 36.5



#17
N-Nitroso-di-n-propylamine
Concen: 16.266 ng/ul
RT: 8.781 min Scan# 951
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion: 70 Resp: 271005
Ion Ratio Lower Upper
70 100
42 64.4 50.0 75.0
101 8.9 7.4 11.0
130 19.8 15.5 23.3









#18
4-Methylphenol
Concen: 17.749 ng/ul
RT: 8.739 min Scan# 943
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:108 Resp: 337625
Ion Ratio Lower Upper
108 100
107 108.9 87.1 130.7



#19
Hexachloroethane
Concen: 16.659 ng/ul
RT: 9.033 min Scan# 994
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:117 Resp: 128366
Ion Ratio Lower Upper
117 100
201 81.7 59.8 89.6
199 52.2 39.0 58.6





#22

Nitrobenzene

Concen: 18.484 ng/ul

RT: 9.175 min Scan# 1018

Delta R.T. -0.000 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion: 77 Resp: 391510

Ion Ratio Lower Upper

77 100

123 43.6 35.6 53.4

65 13.2 10.6 15.8



#23

Isophorone

Concen: 16.744 ng/ul

RT: 9.704 min Scan# 1108

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion: 82 Resp: 719410


Ion Ratio Lower Upper

82 100

95 6.6 5.5 8.3

138 17.2 13.6 20.4









#25
2-Nitrophenol
Concen: 19.280 ng/ul
RT: 9.886 min Scan# 1138
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:139 Resp: 181201
Ion Ratio Lower Upper
139 100
65 53.3 44.6 66.8
109 29.3 24.6 36.8



#26
2,4-Dimethylphenol
Concen: 18.770 ng/ul
RT: 9.945 min Scan# 1149
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:107 Resp: 314250
Ion Ratio Lower Upper
107 100
121 50.9 40.3 60.5
122 89.1 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 17.137 ng/ul

RT: 10.192 min Scan# 1191

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion: 93 Resp: 434750

Ion Ratio Lower Upper

93 100

95 32.4 25.9 38.9

123 11.6 9.0 13.6



#29

2,4-Dichlorophenol

Concen: 19.213 ng/ul

RT: 10.416 min Scan# 1229

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:162 Resp: 291059


Ion Ratio Lower Upper

162 100

164 65.4 51.0 76.4

98 38.5 32.1 48.1





#30

Naphthalene

Concen: 18.404 ng/ul

RT: 10.822 min Scan# 1298

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:128 Resp: 1037540

Ion Ratio Lower Upper

128 100

129 10.9 8.8 13.2

127 12.5 10.2 15.4



#32

4-Chloroaniline

Concen: 18.832 ng/ul

RT: 10.939 min Scan# 1318

Delta R.T. 0.006 min

Lab File: BM043679.D




Acq: 29 Dec 2023 19:40


Tgt Ion:127 Resp: 415893

Ion Ratio Lower Upper

127 100

129 32.5 25.4 38.2









#33
Hexachlorobutadiene
Concen: 19.265 ng/ul
RT: 11.086 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:225 Resp: 148804
Ion Ratio Lower Upper
225 100
223 62.3 49.4 74.0
227 64.1 50.4 75.6



#34
Caprolactam
Concen: 18.707 ng/ul
RT: 11.727 min Scan# 1452
Delta R.T. 0.011 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:113 Resp: 90566
Ion Ratio Lower Upper
113 100
55 193.3 148.9 223.3
56 142.4 111.4 167.2









#35
4-Chloro-3-methylphenol
Concen: 18.081 ng/ul
RT: 12.057 min Scan# 1507
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:107 Resp: 321739
Ion Ratio Lower Upper
107 100
144 25.7 20.3 30.5
142 80.6 64.3 96.5



#36
2-Methylnaphthalene
Concen: 17.664 ng/ul
RT: 12.427 min Scan# 1571
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:142 Resp: 677887
Ion Ratio Lower Upper
142 100
141 87.7 69.9 104.9









#37
1-Methylnaphthalene
Concen: 17.829 ng/ul
RT: 12.645 min Scan# 1608
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:142 Resp: 690118
Ion Ratio Lower Upper
142 100
141 90.2 73.0 109.4
116 3.8 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.572 ng/ul
RT: 12.786 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:216 Resp: 283786
Ion Ratio Lower Upper
216 100
214 78.9 64.6 97.0
179 23.2 20.0 30.0
108 21.8 19.4 29.0









#40
Hexachlorocyclopentadiene
Concen: 10.076 ng/ul
RT: 12.757 min Scan# 1627
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:237 Resp: 96606
Ion Ratio Lower Upper
237 100
235 62.3 48.2 72.4
272 12.8 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 18.753 ng/ul
RT: 13.033 min Scan# 1674
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:196 Resp: 203686
Ion Ratio Lower Upper
196 100
198 95.1 78.3 117.5
200 31.0 25.1 37.7





#42

2,4,5-Trichlorophenol

Concen: 19.266 ng/ul

RT: 13.104 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:196 Resp: 222182

Ion Ratio Lower Upper

196 100

198 94.5 75.1 112.7

200 31.4 24.6 37.0



#43

1,1'-Biphenyl

Concen: 17.779 ng/ul

RT: 13.433 min Scan# 1742

Delta R.T. -0.000 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:154 Resp: 849912


Ion Ratio Lower Upper

154 100

153 39.8 31.8 47.6

76 14.1 11.7 17.5





#44

2-Chloronaphthalene

Concen: 18.348 ng/ul

RT: 13.474 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:162 Resp: 672459

Ion Ratio Lower Upper

162 100

164 32.7 26.4 39.6

127 38.9 32.2 48.4



#45

2-Nitroaniline

Concen: 18.849 ng/ul

RT: 13.692 min Scan# 1786

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion: 65 Resp: 231656


Ion Ratio Lower Upper

65 100

92 63.9 51.4 77.2

138 99.7 78.4 117.6









#47
Dimethylphthalate
Concen: 17.754 ng/ul
RT: 14.063 min Scan# 1849
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:163 Resp: 791552
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.4
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 18.836 ng/ul
RT: 14.186 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:165 Resp: 162357
Ion Ratio Lower Upper
165 100
89 68.1 50.7 76.1
121 24.4 18.6 27.8





#50

Acenaphthylene

Concen: 18.345 ng/ul

RT: 14.321 min Scan# 1893

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:152 Resp: 1128166

Ion Ratio Lower Upper

152 100

151 19.9 15.6 23.4

153 12.8 10.2 15.4



#51

3-Nitroaniline

Concen: 20.124 ng/ul

RT: 14.515 min Scan# 1926

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:138 Resp: 199128


Ion Ratio Lower Upper

138 100

108 10.9 8.4 12.6

92 115.4 94.6 142.0





#52

Acenaphthene

Concen: 18.187 ng/ul

RT: 14.663 min Scan# 1951

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:153 Resp: 737156

Ion Ratio Lower Upper

153 100

152 47.7 38.1 57.1

154 89.3 70.4 105.6



#53

2,4-Dinitrophenol

Concen: 22.262 ng/ul

RT: 14.721 min Scan# 1961

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:184 Resp: 102146


Ion Ratio Lower Upper

184 100

63 85.2 64.9 97.3

154 67.5 53.1 79.7









#55
4-Nitrophenol
Concen: 21.326 ng/ul
RT: 14.821 min Scan# 1976
Delta R.T. 0.011 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:109 Resp: 141494
Ion Ratio Lower Upper
109 100
139 134.1 107.0 160.6
65 142.4 110.4 165.6



#56
Dibenzofuran
Concen: 18.426 ng/ul
RT: 14.998 min Scan# 2008
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:168 Resp: 964561
Ion Ratio Lower Upper
168 100
139 38.3 31.8 47.8





#57

2,4-Dinitrotoluene

Concen: 19.382 ng/ul

RT: 14.974 min Scan# 2004

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:165 Resp: 236035

Ion Ratio Lower Upper

165 100

63 51.7 43.2 64.8

182 2.6 2.3 3.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.430 ng/ul

RT: 15.221 min Scan# 2046

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Tgt Ion:232 Resp: 171676




Ion Ratio Lower Upper


232 100

131 55.3 47.4 71.0

130 2.8 2.4 3.6

166 36.5 30.1 45.1









#59
Diethylphthalate
Concen: 18.225 ng/ul
RT: 15.421 min Scan# 2080
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:149 Resp: 832688
Ion Ratio Lower Upper
149 100
177 19.9 16.6 24.8
150 12.1 9.7 14.5



#61
Fluorene
Concen: 18.557 ng/ul
RT: 15.645 min Scan# 2118
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:166 Resp: 829860
Ion Ratio Lower Upper
166 100
165 97.7 76.9 115.3
167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.144 ng/ul
RT: 15.645 min Scan# 2118
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:204 Resp: 356915
Ion Ratio Lower Upper
204 100
206 32.2 25.3 37.9
141 65.0 54.2 81.2



#63
4-Nitroaniline
Concen: 23.864 ng/ul
RT: 15.674 min Scan# 2123
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:138 Resp: 216138
Ion Ratio Lower Upper
138 100
92 53.3 43.9 65.9
108 80.5 67.0 100.6









#66
4,6-Dinitro-2-methylphenol
Concen: 15.735 ng/ul
RT: 15.727 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:198 Resp: 114219
Ion Ratio Lower Upper
198 100
182 5.1 3.8 5.6
77 35.4 27.6 41.4



#67
N-Nitrosodiphenylamine
Concen: 17.003 ng/ul
RT: 15.857 min Scan# 2154
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:169 Resp: 671385
Ion Ratio Lower Upper
169 100
168 67.9 54.1 81.1
167 37.0 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 16.871 ng/ul
RT: 16.539 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:248 Resp: 205472
Ion Ratio Lower Upper
248 100
250 98.1 78.6 118.0
141 86.2 70.1 105.1



#69
Hexachlorobenzene
Concen: 17.486 ng/ul
RT: 16.639 min Scan# 2287
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:284 Resp: 243981
Ion Ratio Lower Upper
284 100
142 41.8 34.5 51.7
249 29.8 23.8 35.6





#70

Atrazine

Concen: 19.756 ng/ul

RT: 16.815 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:200 Resp: 170991

Ion Ratio Lower Upper

200 100

173 29.3 23.4 35.2

215 47.0 37.8 56.8



#71

Pentachlorophenol

Concen: 22.033 ng/ul

RT: 16.992 min Scan# 2347

Delta R.T. 0.011 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:266 Resp: 185856


Ion Ratio Lower Upper

266 100

264 60.8 49.9 74.9

268 62.9 51.0 76.6





#72

Phenanthrene

Concen: 17.942 ng/ul

RT: 17.386 min Scan# 2413

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:178 Resp: 1250075

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 19.9 15.5 23.3



#74

Anthracene

Concen: 18.359 ng/ul

RT: 17.480 min Scan# 2430

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:178 Resp: 1290910


Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 18.8 15.2 22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 16.780 ng/uL

RT: 13.392 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:216 Resp: 283284

Ion Ratio Lower Upper

216 100

214 79.9 63.0 94.6

218 48.1 38.6 57.8



#76

Pentachlorobenzene

Concen: 16.551 ng/uL

RT: 14.910 min Scan# 1993

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:250 Resp: 272087


Ion Ratio Lower Upper

250 100

248 63.8 49.0 73.4

252 65.0 49.6 74.4









#77
 Carbazole
 Concen: 20.554 ng/ul
 RT: 17.751 min Scan# 2475
 Delta R.T. 0.006 min
 Lab File: BM043679.D
 Acq: 29 Dec 2023 19:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020175


Tgt Ion:167 Resp: 1233515
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.3 25.9
 139 13.4 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 19.352 ng/ul
 RT: 18.321 min Scan# 2573
 Delta R.T. 0.006 min
 Lab File: BM043679.D
 Acq: 29 Dec 2023 19:40

Tgt Ion:149 Resp: 1518304
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 5.1 4.2 6.2









#80
Fluoranthene
Concen: 15.369 ng/ul
RT: 19.397 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:202 Resp: 1384427
Ion Ratio Lower Upper
202 100
101 13.1 11.4 17.0
100 9.6 8.2 12.4



#82
Pyrene
Concen: 15.806 ng/ul
RT: 19.762 min Scan# 2818
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:202 Resp: 1463724
Ion Ratio Lower Upper
202 100
101 14.7 12.7 19.1
100 11.8 10.1 15.1





#83

Butylbenzylphthalate

Concen: 18.549 ng/ul

RT: 20.668 min Scan# 2972

Delta R.T. -0.000 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:149 Resp: 742667

Ion Ratio Lower Upper

149 100

91 73.4 60.5 90.7

206 16.8 13.9 20.9



#84

3,3'-Dichlorobenzidine

Concen: 18.903 ng/ul

RT: 21.444 min Scan# 3104

Delta R.T. -0.000 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:252 Resp: 428941


Ion Ratio Lower Upper

252 100

254 64.5 50.7 76.1

126 14.3 12.3 18.5









#85
Benzo(a)anthracene
Concen: 18.373 ng/ul
RT: 21.509 min Scan# 3114
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:228 Resp: 1415543
Ion Ratio Lower Upper
228 100
229 19.3 15.6 23.4
226 27.5 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.885 ng/ul
RT: 21.444 min Scan# 3104
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:149 Resp: 1133456
Ion Ratio Lower Upper
149 100
167 27.0 21.6 32.4
279 3.5 2.8 4.2





#87

Chrysene

Concen: 18.569 ng/ul

RT: 21.562 min Scan# 31

Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:228 Resp: 1287012

Ion Ratio Lower Upper

228 100

226 30.3 24.2 36.2

229 19.6 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 19.776 ng/ul




RT: 22.386 min Scan# 3264


Delta R.T. 0.006 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Tgt Ion:149 Resp: 1943482









#90
Benzo(b)fluoranthene
Concen: 17.822 ng/ul
RT: 23.197 min Scan# 3402
Delta R.T. -0.000 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40




Instrument :
BNA_P
ClientSampleId :
SSTD020175


Tgt Ion:252 Resp: 1403023
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.2
125 12.9 10.4 15.6



#91
Benzo(k)fluoranthene
Concen: 17.338 ng/ul
RT: 23.250 min Scan# 3411
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:252 Resp: 1331225
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 12.5 10.2 15.4





#93

Benzo(a)pyrene

Concen: 18.019 ng/ul

RT: 23.832 min Scan# 3

Delta R.T. 0.011 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40

Instrument :

BNA_P

ClientSampleId :

SSTD020175





Tgt Ion:252 Resp: 1294210

Ion Ratio Lower Upper

252 100

253 22.0 17.9 26.9

125 13.4 11.4 17.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.020 ng/ul

RT: 26.450 min Scan# 3955

Delta R.T. 0.012 min

Lab File: BM043679.D

Acq: 29 Dec 2023 19:40




Tgt Ion:276 Resp: 1631181


Ion Ratio Lower Upper

276 100

138 28.3 22.6 34.0

277 24.7 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 19.148 ng/ul
RT: 26.474 min Scan# 3957
Delta R.T. 0.012 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Instrument :
BNA_P
ClientSampleId :
SSTD020175

Tgt Ion:278 Resp: 1359844
Ion Ratio Lower Upper
278 100
139 20.2 16.0 24.0
279 23.8 19.4 29.0



#96
Benzo(g,h,i)perylene
Concen: 19.042 ng/ul
RT: 27.215 min Scan# 4085
Delta R.T. 0.006 min
Lab File: BM043679.D
Acq: 29 Dec 2023 19:40

Tgt Ion:276 Resp: 1280128
Ion Ratio Lower Upper
276 100
138 26.5 22.4 33.6
277 23.0 19.1 28.7

