

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
 Data File : BP019171.D
 Acq On : 29 Dec 2023 18:06
 Operator : MA/JU
 Sample : 06020-09
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH7X0

Quant Time: Dec 29 21:54:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	191197	20.000	ng/ul	0.00
20) Naphthalene-d8	10.598	136	733426	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.445	164	442364	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.204	188	976433	20.000	ng/ul	0.00
79) Chrysene-d12	21.304	240	1113104	20.000	ng/ul	0.00
88) Perylene-d12	23.668	264	1305465	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	24982	4.462	ng/uL	0.00
4) Pyridine-d5	3.652	84	149005	10.604	ng/ul	0.00
7) Phenol-d5	6.975	99	154023	8.983	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	374243	37.729	ng/ul	0.00
11) 2-Chlorophenol-d4	7.334	132	397334	28.745	ng/ul	0.00
15) 4-Methylphenol-d8	8.516	113	270612	20.272	ng/ul	0.00
21) Nitrobenzene-d5	8.963	128	235639	37.167	ng/ul	0.00
24) 2-Nitrophenol-d4	9.687	143	240831	35.148	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	454560	32.853	ng/ul	0.00
31) 4-Chloroaniline-d4	10.746	131	553385	32.516	ng/ul	0.00
46) Dimethylphthalate-d6	13.863	166	1525072	39.031	ng/ul	0.00
49) Acenaphthylene-d8	14.139	160	1626457	38.363	ng/ul	0.00
54) 4-Nitrophenol-d4	14.669	143	45201	6.851	ng/ul	0.00
60) Fluorene-d10	15.439	176	1305169	38.665	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	208939	32.641	ng/ul	0.00
73) Anthracene-d10	17.304	188	2090560	40.998	ng/ul	0.00
81) Pyrene-d10	19.545	212	2729255	40.232	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.515	264	3075665	40.988	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
Data File : BP019171.D
Acq On : 29 Dec 2023 18:06
Operator : MA/JU
Sample : 06020-09
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH7X0

Quant Time: Dec 29 21:54:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 03:48:11 2023
Response via : Initial Calibration

