

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
 Data File : BP019192.D
 Acq On : 30 Dec 2023 06:38
 Operator : MA/JU
 Sample : 06020-14
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH7X9

Quant Time: Jan 01 00:20:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	202316	20.000	ng/ul	0.00
20) Naphthalene-d8	10.593	136	762864	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.440	164	465986	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.198	188	1059158	20.000	ng/ul	0.00
79) Chrysene-d12	21.304	240	1208258	20.000	ng/ul	0.00
88) Perylene-d12	23.663	264	1431706	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	25809	4.357	ng/uL	0.00
4) Pyridine-d5	3.658	84	103097	6.934	ng/ul	0.00
7) Phenol-d5	6.975	99	123777	6.822	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	333111	31.737	ng/ul	0.00
11) 2-Chlorophenol-d4	7.328	132	343049	23.454	ng/ul	0.00
15) 4-Methylphenol-d8	8.516	113	217937	15.429	ng/ul	0.00
21) Nitrobenzene-d5	8.963	128	199763	30.292	ng/ul	0.00
24) 2-Nitrophenol-d4	9.681	143	203747	28.588	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	392438	27.269	ng/ul	0.00
31) 4-Chloroaniline-d4	10.740	131	426225	24.078	ng/ul	0.00
46) Dimethylphthalate-d6	13.863	166	1462537	35.533	ng/ul	0.00
49) Acenaphthylene-d8	14.134	160	1440411	32.253	ng/ul	0.00
54) 4-Nitrophenol-d4	14.663	143	43063	6.196	ng/ul	0.00
60) Fluorene-d10	15.439	176	1234815	34.727	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	208471	30.024	ng/ul	-0.01
73) Anthracene-d10	17.298	188	1885700	34.092	ng/ul	0.00
81) Pyrene-d10	19.545	212	2756681	37.436	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.510	264	3081173	37.441	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
Data File : BP019192.D
Acq On : 30 Dec 2023 06:38
Operator : MA/JU
Sample : 06020-14
Misc :
ALS Vial : 71 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH7X9

Quant Time: Jan 01 00:20:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 03:48:11 2023
Response via : Initial Calibration

