

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
 Data File : BP019167.D
 Acq On : 29 Dec 2023 15:49
 Operator : MA/JU
 Sample : 06021-10
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH7Z2

Quant Time: Dec 29 18:18:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	176426	20.000	ng/ul	0.00
20) Naphthalene-d8	10.598	136	647804	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.445	164	380994	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.198	188	865770	20.000	ng/ul	0.00
79) Chrysene-d12	21.304	240	950534	20.000	ng/ul	0.00
88) Perylene-d12	23.668	264	1155177	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	19876	3.848	ng/uL	0.00
4) Pyridine-d5	3.658	84	76127	5.871	ng/ul	0.00
7) Phenol-d5	6.981	99	85272	5.390	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	244873	26.753	ng/ul	0.00
11) 2-Chlorophenol-d4	7.334	132	249487	19.560	ng/ul	0.00
15) 4-Methylphenol-d8	8.522	113	152548	12.384	ng/ul	0.00
21) Nitrobenzene-d5	8.969	128	141911	25.342	ng/ul	0.00
24) 2-Nitrophenol-d4	9.687	143	137124	22.657	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	276244	22.604	ng/ul	0.00
31) 4-Chloroaniline-d4	10.745	131	343072	22.823	ng/ul	0.00
46) Dimethylphthalate-d6	13.863	166	957118	28.441	ng/ul	0.00
49) Acenaphthylene-d8	14.139	160	1034948	28.343	ng/ul	0.00
54) 4-Nitrophenol-d4	14.669	143	20831	3.666	ng/ul	0.00
60) Fluorene-d10	15.439	176	864563	29.738	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	117105	20.633	ng/ul	0.00
73) Anthracene-d10	17.298	188	1510193	33.402	ng/ul	0.00
81) Pyrene-d10	19.545	212	1919033	33.127	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.515	264	2215417	33.365	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
 Data File : BP019167.D
 Acq On : 29 Dec 2023 15:49
 Operator : MA/JU
 Sample : 06021-10
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH7Z2

Quant Time: Dec 29 18:18:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

