

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
 Data File : BP019170.D
 Acq On : 29 Dec 2023 17:32
 Operator : MA/JU
 Sample : 06020-08
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH7W9

Quant Time: Dec 29 18:21:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	175075	20.000	ng/u1	0.00
20) Naphthalene-d8	10.599	136	680246	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	407054	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.198	188	889174	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	924346	20.000	ng/u1	0.00
88) Perylene-d12	23.669	264	1102595	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	23022	4.491	ng/uL	0.00
4) Pyridine-d5	3.658	84	73064	5.679	ng/u1	0.00
7) Phenol-d5	6.975	99	129225	8.231	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	295953	32.584	ng/u1	0.00
11) 2-Chlorophenol-d4	7.334	132	318854	25.191	ng/u1	0.00
15) 4-Methylphenol-d8	8.516	113	214941	17.584	ng/u1	0.00
21) Nitrobenzene-d5	8.963	128	179402	30.509	ng/u1	0.00
24) 2-Nitrophenol-d4	9.687	143	181969	28.633	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	352725	27.486	ng/u1	0.00
31) 4-Chloroaniline-d4	10.746	131	409726	25.957	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	1192989	33.181	ng/u1	0.00
49) Acenaphthylene-d8	14.134	160	1262818	32.370	ng/u1	0.00
54) 4-Nitrophenol-d4	14.663	143	37914	6.245	ng/u1	0.00
60) Fluorene-d10	15.440	176	1031687	33.215	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	157829	27.076	ng/u1	0.00
73) Anthracene-d10	17.298	188	1771980	38.160	ng/u1	0.00
81) Pyrene-d10	19.545	212	2146727	38.107	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.516	264	2449147	38.644	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
Data File : BP019170.D
Acq On : 29 Dec 2023 17:32
Operator : MA/JU
Sample : 06020-08
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH7W9

Quant Time: Dec 29 18:21:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 03:48:11 2023
Response via : Initial Calibration

