

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
 Data File : BP019180.D
 Acq On : 29 Dec 2023 23:47
 Operator : MA/JU
 Sample : PB158075BL
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK075

Quant Time: Dec 30 00:38:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	219285	20.000	ng/ul	0.00
20) Naphthalene-d8	10.593	136	797475	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.445	164	452059	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.198	188	1004071	20.000	ng/ul	0.00
79) Chrysene-d12	21.298	240	1114297	20.000	ng/ul	-0.01
88) Perylene-d12	23.663	264	1348370	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	36725	5.720	ng/uL	0.00
4) Pyridine-d5	3.652	84	446008	27.676	ng/ul	0.00
7) Phenol-d5	6.975	99	508540	25.860	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	304481	26.764	ng/ul	0.00
11) 2-Chlorophenol-d4	7.328	132	408673	25.778	ng/ul	0.00
15) 4-Methylphenol-d8	8.516	113	380205	24.834	ng/ul	0.00
21) Nitrobenzene-d5	8.963	128	181403	26.314	ng/ul	0.00
24) 2-Nitrophenol-d4	9.681	143	189989	25.501	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	377418	25.087	ng/ul	0.00
31) 4-Chloroaniline-d4	10.740	131	498611	26.945	ng/ul	0.00
46) Dimethylphthalate-d6	13.863	166	1000366	25.053	ng/ul	0.00
49) Acenaphthylene-d8	14.134	160	1142778	26.377	ng/ul	0.00
54) 4-Nitrophenol-d4	14.657	143	145606	21.595	ng/ul	0.00
60) Fluorene-d10	15.440	176	872699	25.299	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	134042	20.364	ng/ul	0.00
73) Anthracene-d10	17.298	188	1390765	26.523	ng/ul	0.00
81) Pyrene-d10	19.539	212	1764948	25.990	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.510	264	2069524	26.702	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122823\
Data File : BP019180.D
Acq On : 29 Dec 2023 23:47
Operator : MA/JU
Sample : PB158075BL
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK075

Quant Time: Dec 30 00:38:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 03:48:11 2023
Response via : Initial Calibration

