

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123022\
 Data File : BP013370.D
 Acq On : 30 Dec 2022 13:58
 Operator : CG/JU
 Sample : PB149946BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK946

Quant Time: Dec 30 23:37:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	352897	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	1548406	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.569	164	1101244	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.327	188	2607772	20.000	ng/ul	0.00
79) Chrysene-d12	21.415	240	2449476	20.000	ng/ul	0.00
88) Perylene-d12	23.880	264	2245114	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.316	96	53590	6.493	ng/uL	0.00
4) Pyridine-d5	3.746	84	750209	31.997	ng/ul	0.00
7) Phenol-d5	7.087	99	898633	31.263	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.251	67	604516	34.863	ng/ul	0.00
11) 2-Chlorophenol-d4	7.457	132	719578	31.801	ng/ul	0.00
15) 4-Methylphenol-d8	8.640	113	731125	31.043	ng/ul	0.00
21) Nitrobenzene-d5	9.098	128	360631	31.700	ng/ul	0.00
24) 2-Nitrophenol-d4	9.822	143	404408	31.575	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	746408	30.003	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	1079029	30.724	ng/ul	0.00
46) Dimethylphthalate-d6	13.981	166	2753594	32.535	ng/ul	0.00
49) Acenaphthylene-d8	14.263	160	2885133	31.023	ng/ul	0.00
54) 4-Nitrophenol-d4	14.775	143	364231	23.924	ng/ul	0.00
60) Fluorene-d10	15.569	176	2401199	33.535	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	385057	22.945	ng/ul	0.00
73) Anthracene-d10	17.427	188	3873152	32.921	ng/ul	0.00
81) Pyrene-d10	19.663	212	4700707	33.433	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.721	264	3698638	33.051	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.369	88	184	0.021	ng/ul#	21
5) Pyridine	3.769	79	271	0.012	ng/ul#	1
6) Benzaldehyde	7.057	77	154	0.014	ng/ul#	32
8) Phenol	7.128	94	327	0.011	ng/ul#	1
10) Bis(2-Chloroethyl)ether	7.346	93	571	0.024	ng/ul	94
12) 2-Chlorophenol	7.487	128	128	0.006	ng/ul#	1
13) 2-Methylphenol	8.375	108	159	0.007	ng/ul	91
14) 2,2'-oxybis(1-Chloropr...	8.446	45	188	0.006	ng/ul#	1
16) Acetophenone	8.757	105	843	0.023	ng/ul#	1
17) N-Nitroso-di-n-propyla...	8.734	70	189	0.010	ng/ul#	27
18) 4-Methylphenol	8.698	108	298	0.012	ng/ul	94
19) Hexachloroethane	9.016	117	242	0.026	ng/ul#	7
22) Nitrobenzene	9.128	77	738	0.028	ng/ul#	66
23) Isophorone	9.669	82	217	0.004	ng/ul#	85
25) 2-Nitrophenol	9.845	139	191	0.014	ng/ul#	1
26) 2,4-Dimethylphenol	10.075	107	61	0.002	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.151	93	498	0.014	ng/ul#	85
29) 2,4-Dichlorophenol	10.381	162	103	0.004	ng/ul#	1
30) Naphthalene	10.792	128	1124	0.014	ng/ul#	61
32) 4-Chloroaniline	10.904	127	303	0.009	ng/ul#	1
33) Hexachlorobutadiene	11.069	225	192	0.011	ng/ul#	1
34) Caprolactam	11.751	113	198	0.024	ng/ul	79
35) 4-Chloro-3-methylphenol	12.028	107	169	0.007	ng/ul#	82
36) 2-Methylnaphthalene	12.398	142	590	0.011	ng/ul	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123022\
 Data File : BP013370.D
 Acq On : 30 Dec 2022 13:58
 Operator : CG/JU
 Sample : PB149946BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK946

Quant Time: Dec 30 23:37:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.622	142	516	0.009	ng/ul#	93
39) 1,2,4,5-Tetrachloroben...	12.769	216	237	0.007	ng/ul#	63
40) Hexachlorocyclopentadiene	12.745	237	17	0.001	ng/ul#	50
41) 2,4,6-Trichlorophenol	13.010	196	93	0.004	ng/ul#	1
42) 2,4,5-Trichlorophenol	13.081	196	13	0.001	ng/ul#	1
43) 1,1'-Biphenyl	13.410	154	810	0.010	ng/ul#	76
44) 2-Chloronaphthalene	13.463	162	601	0.009	ng/ul#	74
45) 2-Nitroaniline	13.657	65	141	0.009	ng/ul#	35
47) Dimethylphthalate	14.028	163	900	0.011	ng/ul#	52
48) 2,6-Dinitrotoluene	14.151	165	86	0.006	ng/ul#	54
50) Acenaphthylene	14.292	152	1333	0.013	ng/ul#	86
51) 3-Nitroaniline	14.486	138	73	0.005	ng/ul#	1
52) Acenaphthene	14.633	153	624	0.009	ng/ul#	66
53) 2,4-Dinitrophenol	14.698	184	22	0.002	ng/ul#	1
55) 4-Nitrophenol	14.786	109	451	0.043	ng/ul#	1
56) Dibenzofuran	14.910	168	56	0.001	ng/ul	92
57) 2,4-Dinitrotoluene	14.933	165	318	0.014	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.192	232	262	0.012	ng/ul#	69
59) Diethylphthalate	15.386	149	1389	0.017	ng/ul#	82
61) Fluorene	15.616	166	753	0.010	ng/ul#	83
62) 4-Chlorophenyl-phenyle...	15.610	204	540	0.013	ng/ul#	87
63) 4-Nitroaniline	15.645	138	59	0.005	ng/ul#	68
66) 4,6-Dinitro-2-methylph...	15.686	198	1631	0.099	ng/ul#	1
67) N-Nitrosodiphenylamine	15.822	169	460	0.006	ng/ul	97
68) 4-Bromophenyl-phenylether	16.510	248	244	0.009	ng/ul#	69
69) Hexachlorobenzene	16.627	284	318	0.009	ng/ul#	74
70) Atrazine	16.786	200	231	0.008	ng/ul#	76
71) Pentachlorophenol	16.986	266	316	0.016	ng/ul#	72
72) Phenanthrene	17.374	178	1715	0.013	ng/ul#	68
74) Anthracene	17.457	178	1345	0.010	ng/ul#	69
75) 1,2,3,4-Tetrachloroben...	13.375	216	249	0.007	ng/ul#	46
76) Pentachlorobenzene	14.892	250	277	0.007	ng/ul#	81
77) Carbazole	17.745	167	833	0.007	ng/ul#	93
78) Di-n-butylphthalate	18.274	149	3284	0.024	ng/ul#	90
80) Fluoranthene	19.339	202	1732	0.011	ng/ul#	90
82) Pyrene	19.692	202	1523	0.009	ng/ul#	73
83) Butylbenzylphthalate	20.557	149	272	0.004	ng/ul#	71
84) 3,3'-Dichlorobenzidine	21.327	252	517	0.009	ng/ul#	36
85) Benzo(a)anthracene	21.415	228	8853	0.054	ng/ul#	74
86) Bis(2-ethylhexyl)phtha...	21.310	149	2940	0.033	ng/ul#	71
87) Chrysene	21.451	228	3986	0.026	ng/ul#	75
89) Di-n-octyl phthalate	22.251	149	1102	0.009	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	5996	0.042	ng/ul#	22
91) Benzo(k)fluoranthene	23.174	252	3976	0.028	ng/ul#	28
93) Benzo(a)pyrene	23.768	252	2816	0.023	ng/ul#	1
94) Indeno(1,2,3-cd)pyrene	26.462	276	1354	0.009	ng/ul	96
95) Dibenzo(a,h)anthracene	26.480	278	519	0.004	ng/ul#	22
96) Benzo(g,h,i)perylene	27.245	276	342	0.003	ng/ul#	61


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123022\
 Data File : BP013370.D
 Acq On : 30 Dec 2022 13:58
 Operator : CG/JU
 Sample : PB149946BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK946

Quant Time: Dec 30 23:37:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 0.021 ng/uL
RT: 3.369 min Scan# 31
Delta R.T. 0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion: 88 Resp: 184
Ion Ratio Lower Upper
88 100
43 69.3 22.4 33.6#
58 6.4 58.2 87.4#



#5
Pyridine
Concen: 0.012 ng/uL
RT: 3.769 min Scan# 99
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion: 79 Resp: 271
Ion Ratio Lower Upper
79 100
52 2881.5 60.7 91.1#
51 344.8 28.9 43.3#









#6
Benzaldehyde
Concen: 0.014 ng/ul
RT: 7.057 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion: 77 Resp: 154
Ion Ratio Lower Upper
77 100
105 45.1 81.5 122.3#
106 18.3 76.9 115.3#



#8
Phenol
Concen: 0.011 ng/ul
RT: 7.128 min Scan# 670
Delta R.T. 0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion: 94 Resp: 327
Ion Ratio Lower Upper
94 100
65 85.5 21.1 31.7#
66 152.9 27.0 40.6#









#10
Bis(2-Chloroethyl)ether
Concen: 0.024 ng/ul
RT: 7.346 min Scan# 707
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion: 93	Resp: 571
Ion Ratio	Lower Upper
93 100	
63 68.4	59.8 89.8
95 31.7	26.4 39.6



#12
2-Chlorophenol
Concen: 0.006 ng/ul
RT: 7.487 min Scan# 731
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:128	Resp:	128	
Ion	Ratio	Lower	Upper
128	100		
64	211.7	34.9	52.3#
130	28.9	25.8	38.6





#13

2-Methylphenol

Concen: 0.007 ng/ul

RT: 8.375 min Scan# 882

Delta R.T. -0.006 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :





SBLK946

Tgt Ion:108 Resp: 159

Ion Ratio Lower Upper

108 100

107 94.7 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.006 ng/ul

RT: 8.446 min Scan# 894

Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58




Tgt Ion: 45 Resp: 188


Ion Ratio Lower Upper

45 100

77 99.6 11.0 16.6#

79 55.5 9.3 13.9#





#16

Acetophenone

Concen: 0.023 ng/ul

RT: 8.757 min Scan# 947

Delta R.T. -0.006 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :

SBLK946





Tgt Ion:105 Resp: 843

Ion Ratio Lower Upper

105 100

77 207.1 59.3 88.9#

51 173.5 19.6 29.4#



#17

N-Nitroso-di-n-propylamine

Concen: 0.010 ng/ul

RT: 8.734 min Scan# 943

Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Tgt Ion: 70 Resp: 189




Ion Ratio Lower Upper


70 100

42 132.4 45.5 68.3#

101 22.5 7.9 11.9#

130 14.5 17.4 26.2#





#18

4-Methylphenol

Concen: 0.012 ng/ul

RT: 8.698 min Scan# 91

Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :





SBLK946

Tgt Ion:108 Resp: 298

Ion Ratio Lower Upper

108 100

107 103.4 87.4 131.2



#19

Hexachloroethane

Concen: 0.026 ng/ul

RT: 9.016 min Scan# 991

Delta R.T. -0.000 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58




Tgt Ion:117 Resp: 242


Ion Ratio Lower Upper

117 100

201 4.7 93.7 140.5#

199 8.5 58.6 88.0#









#22
Nitrobenzene
Concen: 0.028 ng/ul
RT: 9.128 min Scan# 1010
Delta R.T. -0.018 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion: 77 Resp: 738
Ion Ratio Lower Upper
77 100
123 23.7 38.3 57.5#
65 23.4 10.3 15.5#



#23
Isophorone
Concen: 0.004 ng/ul
RT: 9.669 min Scan# 1102
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion: 82 Resp: 217
Ion Ratio Lower Upper
82 100
95 9.7 5.4 8.2#
138 11.3 15.3 22.9#









#25
2-Nitrophenol
Concen: 0.014 ng/ul
RT: 9.845 min Scan# 1133
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:139 Resp: 191
Ion Ratio Lower Upper
139 100
65 636.5 36.6 54.8#
109 92.9 21.6 32.4#



#26
2,4-Dimethylphenol
Concen: 0.002 ng/ul
RT: 10.075 min Scan# 1171
Delta R.T. 0.159 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:107 Resp: 61
Ion Ratio Lower Upper
107 100
121 57.6 41.8 62.6
122 92.4 72.6 108.8





#27

Bis(2-Chloroethoxy)methane

Concen: 0.014 ng/ul

RT: 10.151 min Scan# 1183

Delta R.T. -0.000 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :

SBLK946





Tgt Ion: 93 Resp: 498

Ion Ratio Lower Upper

93 100

95 23.9 26.0 39.0#

123 6.3 9.2 13.8#



#29

2,4-Dichlorophenol

Concen: 0.004 ng/ul

RT: 10.381 min Scan# 1223

Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58




Tgt Ion:162 Resp: 103


Ion Ratio Lower Upper

162 100

164 5780.4 52.5 78.7#

98 388.9 26.6 39.8#









#30
Naphthalene
Concen: 0.014 ng/ul
RT: 10.792 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:128 Resp: 1124
Ion Ratio Lower Upper
128 100
129 31.2 8.7 13.1#
127 2.6 10.4 15.6#



#32
4-Chloroaniline
Concen: 0.009 ng/ul
RT: 10.904 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:127 Resp: 303
Ion Ratio Lower Upper
127 100
129 148.0 25.6 38.4#









#33
Hexachlorobutadiene
Concen: 0.011 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:225 Resp: 192
Ion Ratio Lower Upper
225 100
223 238.7 49.2 73.8#
227 3.8 50.2 75.2#



#34
Caprolactam
Concen: 0.024 ng/ul
RT: 11.751 min Scan# 1456
Delta R.T. 0.059 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:113 Resp: 198
Ion Ratio Lower Upper
113 100
55 147.4 141.9 212.9
56 100.0 99.6 149.4





#35

4-Chloro-3-methylphenol

Concen: 0.007 ng/ul

RT: 12.028 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :

SBLK946





Tgt Ion:107 Resp: 169

Ion Ratio Lower Upper

107 100

144 5.4 21.4 32.2#

142 91.7 65.8 98.6



#36

2-Methylnaphthalene

Concen: 0.011 ng/ul

RT: 12.398 min Scan# 1566

Delta R.T. -0.006 min

Lab File: BP013370.D




Acq: 30 Dec 2022 13:58


Tgt Ion:142 Resp: 590

Ion Ratio Lower Upper

142 100

141 97.0 70.8 106.2









#37
1-Methylnaphthalene
Concen: 0.009 ng/ul
RT: 12.622 min Scan# 1604
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:142 Resp: 516
Ion Ratio Lower Upper
142 100
141 86.5 72.7 109.1
116 23.8 2.7 4.1#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 0.007 ng/ul
RT: 12.769 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:216 Resp: 237
Ion Ratio Lower Upper
216 100
214 39.7 61.9 92.9#
179 13.7 15.0 22.6#
108 28.1 11.0 16.4#









#40
Hexachlorocyclopentadiene
Concen: 0.001 ng/ul
RT: 12.745 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:237	Resp:	17
Ion Ratio	Lower	Upper
237	100	
235	20.8	49.7 74.5#
272	0.0	10.8 16.2#



#41
2,4,6-Trichlorophenol
Concen: 0.004 ng/ul
RT: 13.010 min Scan# 1670
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:196	Resp:	93
Ion Ratio	Lower	Upper
196	100	
198	175.7	77.0 115.6#
200	191.0	25.3 37.9#









#42
2,4,5-Trichlorophenol
Concen: 0.001 ng/ul
RT: 13.081 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:196 Resp: 13
Ion Ratio Lower Upper
196 100
198 795.0 77.1 115.7#
200 115.0 25.4 38.2#



#43
1,1'-Biphenyl
Concen: 0.010 ng/ul
RT: 13.410 min Scan# 1738
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:154 Resp: 810
Ion Ratio Lower Upper
154 100
153 20.6 31.5 47.3#
76 12.7 9.6 14.4









#44
2-Chloronaphthalene
Concen: 0.009 ng/ul
RT: 13.463 min Scan# 1747
Delta R.T. 0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:162 Resp: 601
Ion Ratio Lower Upper
162 100
164 1.8 26.3 39.5#
127 36.6 29.3 43.9



#45
2-Nitroaniline
Concen: 0.009 ng/ul
RT: 13.657 min Scan# 1780
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion: 65 Resp: 141
Ion Ratio Lower Upper
65 100
92 87.8 56.4 84.6#
138 13.8 90.3 135.5#





#47

Dimethylphthalate

Concen: 0.011 ng/ul

RT: 14.028 min Scan# 1843

Delta R.T. -0.000 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :

SBLK946





Tgt Ion:163 Resp: 900

Ion Ratio Lower Upper

163 100

194 25.5 3.4 5.0#

164 25.7 8.3 12.5#



#48

2,6-Dinitrotoluene

Concen: 0.006 ng/ul

RT: 14.151 min Scan# 1864

Delta R.T. -0.006 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58




Tgt Ion:165 Resp: 86


Ion Ratio Lower Upper

165 100

89 92.5 43.0 64.4#

121 30.5 16.1 24.1#









#50
Acenaphthylene
Concen: 0.013 ng/ul
RT: 14.292 min Scan# 1888
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:152 Resp: 1333
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.6
153 26.6 10.6 16.0#



#51
3-Nitroaniline
Concen: 0.005 ng/ul
RT: 14.486 min Scan# 1921
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:138 Resp: 73
Ion Ratio Lower Upper
138 100
108 66.3 8.6 13.0#
92 427.5 88.4 132.6#









#52
Acenaphthene
Concen: 0.009 ng/ul
RT: 14.633 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:153 Resp: 624
Ion Ratio Lower Upper
153 100
152 66.7 38.2 57.4#
154 123.3 70.8 106.2#



#53
2,4-Dinitrophenol
Concen: 0.002 ng/ul
RT: 14.698 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:184 Resp: 22
Ion Ratio Lower Upper
184 100
63 394.7 48.4 72.6#
154 92.1 45.5 68.3#









#55
4-Nitrophenol
Concen: 0.043 ng/ul
RT: 14.786 min Scan# 1972
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:109 Resp: 451
Ion Ratio Lower Upper
109 100
139 109.1 120.8 181.2#
65 1450.5 107.9 161.9#



#56
Dibenzofuran
Concen: 0.001 ng/ul
RT: 14.910 min Scan# 1993
Delta R.T. -0.065 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:168 Resp: 56
Ion Ratio Lower Upper
168 100
139 30.4 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 0.014 ng/ul
RT: 14.933 min Scan# 1998
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:165 Resp: 318
Ion Ratio Lower Upper
165 100
63 48.5 31.8 47.6#
182 44.1 2.1 3.1#



#58
2,3,4,6-Tetrachlorophenol
Concen: 0.012 ng/ul
RT: 15.192 min Scan# 2041
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:232 Resp: 262
Ion Ratio Lower Upper
232 100
131 8.3 31.8 47.6#
130 16.2 1.8 2.6#
166 27.2 22.9 34.3









#59
Diethylphthalate
Concen: 0.017 ng/ul
RT: 15.386 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:149	Resp:	1389
Ion Ratio	Lower	Upper
149	100	
177	33.1	18.4 27.6#
150	7.7	9.9 14.9#



#61
Fluorene
Concen: 0.010 ng/ul
RT: 15.616 min Scan# 2113
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:166	Resp:	753
Ion Ratio	Lower	Upper
166	100	
165	114.5	77.4 116.2
167	8.1	10.6 16.0#









#62
4-Chlorophenyl-phenylether
Concen: 0.013 ng/ul
RT: 15.610 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:204 Resp: 540
Ion Ratio Lower Upper
204 100
206 38.0 26.3 39.5
141 39.9 40.4 60.6#



#63
4-Nitroaniline
Concen: 0.005 ng/ul
RT: 15.645 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:138 Resp: 59
Ion Ratio Lower Upper
138 100
92 85.9 39.0 58.4#
108 55.3 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 0.099 ng/ul
RT: 15.686 min Scan# 2125
Delta R.T. -0.024 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:198 Resp: 1631
Ion Ratio Lower Upper
198 100
182 418.0 3.3 4.9#
77 1197.2 16.6 25.0#



#67
N-Nitrosodiphenylamine
Concen: 0.006 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:169 Resp: 460
Ion Ratio Lower Upper
169 100
168 62.8 53.0 79.6
167 37.0 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 0.009 ng/ul
RT: 16.510 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:248 Resp: 244
Ion Ratio Lower Upper
248 100
250 50.0 77.4 116.0#
141 63.2 48.3 72.5



#69
Hexachlorobenzene
Concen: 0.009 ng/ul
RT: 16.627 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:284 Resp: 318
Ion Ratio Lower Upper
284 100
142 15.0 22.6 34.0#
249 45.0 24.6 37.0#





#70

Atrazine

Concen: 0.008 ng/ul

RT: 16.786 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :

SBLK946





Tgt Ion:200 Resp: 231

Ion Ratio Lower Upper

200 100

173 20.2 20.5 30.7#

215 28.6 39.9 59.9#



#71

Pentachlorophenol

Concen: 0.016 ng/ul

RT: 16.986 min Scan# 2346

Delta R.T. 0.006 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58




Tgt Ion:266 Resp: 316


Ion Ratio Lower Upper

266 100

264 25.0 50.9 76.3#

268 58.6 51.4 77.2









#72
Phenanthrene
Concen: 0.013 ng/ul
RT: 17.374 min Scan# 2412
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:178 Resp: 1715
Ion Ratio Lower Upper
178 100
179 31.7 12.5 18.7#
176 7.3 15.4 23.0#



#74
Anthracene
Concen: 0.010 ng/ul
RT: 17.457 min Scan# 2426
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:178 Resp: 1345
Ion Ratio Lower Upper
178 100
179 39.1 12.5 18.7#
176 22.7 14.8 22.2#





#75

1,2,3,4-Tetrachlorobenzene

Concen: 0.007 ng/uL

RT: 13.375 min Scan# 1732

Delta R.T. -0.000 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :

SBLK946





Tgt Ion:216 Resp: 249

Ion Ratio Lower Upper

216 100

214 3.0 62.2 93.2#

218 50.3 38.9 58.3



#76

Pentachlorobenzene

Concen: 0.007 ng/uL

RT: 14.892 min Scan# 1990

Delta R.T. -0.000 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58




Tgt Ion:250 Resp: 277


Ion Ratio Lower Upper

250 100

248 44.7 50.6 76.0#

252 74.7 51.1 76.7









#77
Carbazole
Concen: 0.007 ng/ul
RT: 17.745 min Scan# 2473
Delta R.T. 0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:167 Resp: 833
Ion Ratio Lower Upper
167 100
166 20.7 17.2 25.8
139 5.3 9.4 14.2#



#78
Di-n-butylphthalate
Concen: 0.024 ng/ul
RT: 18.274 min Scan# 2565
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:149 Resp: 3284
Ion Ratio Lower Upper
149 100
150 8.2 7.5 11.3
104 12.0 3.8 5.6#









#80
Fluoranthene
Concen: 0.011 ng/ul
RT: 19.339 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:202 Resp: 1732
Ion Ratio Lower Upper
202 100
101 16.5 8.2 12.2#
100 7.4 6.1 9.1



#82
Pyrene
Concen: 0.009 ng/ul
RT: 19.692 min Scan# 2806
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:202 Resp: 1523
Ion Ratio Lower Upper
202 100
101 17.5 9.4 14.2#
100 25.0 7.8 11.8#





#83

Butylbenzylphthalate

Concen: 0.004 ng/ul

RT: 20.557 min Scan# 2953

Delta R.T. -0.006 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :

SBLK946





Tgt Ion:149 Resp: 272

Ion Ratio Lower Upper

149 100

91 48.8 53.4 80.0#

206 46.6 17.9 26.9#



#84

3,3'-Dichlorobenzidine

Concen: 0.009 ng/ul

RT: 21.327 min Scan# 3084

Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58




Tgt Ion:252 Resp: 517


Ion Ratio Lower Upper

252 100

254 18.4 51.8 77.8#

126 44.0 7.4 11.2#









#85
Benzo(a)anthracene
Concen: 0.054 ng/ul
RT: 21.415 min Scan# 3099
Delta R.T. 0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:228 Resp: 8853
Ion Ratio Lower Upper
228 100
229 29.0 16.1 24.1#
226 11.1 22.0 33.0#



#86
Bis(2-ethylhexyl)phthalate
Concen: 0.033 ng/ul
RT: 21.310 min Scan# 3081
Delta R.T. -0.006 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:149 Resp: 2940
Ion Ratio Lower Upper
149 100
167 13.4 22.5 33.7#
279 16.8 3.8 5.8#





#87

Chrysene

Concen: 0.026 ng/ul

RT: 21.451 min Scan# 3105

Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Instrument :

BNA_P

ClientSampleId :

SBLK946





Tgt Ion:228 Resp: 3986

Ion Ratio Lower Upper

228 100

226 51.1 24.5 36.7#

229 17.6 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 0.009 ng/ul




RT: 22.251 min Scan# 3241


Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Tgt Ion:149 Resp: 1102









#90
Benzo(b)fluoranthene
Concen: 0.042 ng/ul
RT: 23.121 min Scan# 3389
Delta R.T. -0.018 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58




Instrument :
BNA_P
ClientSampleId :
SBLK946


Tgt Ion:252 Resp: 5996
Ion Ratio Lower Upper
252 100
253 55.5 17.7 26.5#
125 43.9 6.8 10.2#



#91
Benzo(k)fluoranthene
Concen: 0.028 ng/ul
RT: 23.174 min Scan# 3398
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:252 Resp: 3976
Ion Ratio Lower Upper
252 100
253 44.3 17.7 26.5#
125 57.5 6.8 10.2#





#93

Benzo(a)pyrene

Concen: 0.023 ng/ul

RT: 23.768 min Scan# 3499

Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58





Instrument :

BNA_P

ClientSampleId :

SBLK946

Tgt Ion:252	Resp:	2816
Ion Ratio	Lower	Upper
252	100	
253	80.9	17.5 26.3#
125	69.4	7.5 11.3#



#94

Indeno(1,2,3-cd)pyrene

Concen: 0.009 ng/ul




RT: 26.462 min Scan# 3957


Delta R.T. -0.012 min

Lab File: BP013370.D

Acq: 30 Dec 2022 13:58

Tgt Ion:276	Resp:	1354
Ion Ratio	Lower	Upper
276	100	
138	21.2	17.1 25.7
277	21.8	20.0 30.0









#95
Dibenzo(a,h)anthracene
Concen: 0.004 ng/ul
RT: 26.480 min Scan# 3958
Delta R.T. -0.012 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Instrument :
BNA_P
ClientSampleId :
SBLK946

Tgt Ion:278 Resp: 519
Ion Ratio Lower Upper
278 100
139 29.0 11.1 16.7#
279 73.7 19.0 28.6#



#96
Benzo(g,h,i)perylene
Concen: 0.003 ng/ul
RT: 27.245 min Scan# 4090
Delta R.T. -0.018 min
Lab File: BP013370.D
Acq: 30 Dec 2022 13:58

Tgt Ion:276 Resp: 342
Ion Ratio Lower Upper
276 100
138 40.1 15.4 23.2#
277 40.2 19.1 28.7#

